

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/10/2025 02/10/2025

Client Sample No.: CCAL01 Date Analyzed: 02/10/2025

Lab Sample No.: AR1660CCC500 Data File : PR070257.D Time Analyzed: 17:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.844	4.745	4.945	496.040	500.000	-0.8
Aroclor-1016-2	4.865	4.766	4.966	497.710	500.000	-0.5
Aroclor-1016-3	4.929	4.830	5.030	492.900	500.000	-1.4
Aroclor-1016-4	5.032	4.932	5.132	484.240	500.000	-3.2
Aroclor-1016-5	5.346	5.247	5.447	491.450	500.000	-1.7
Aroclor-1260-1	6.558	6.459	6.659	486.740	500.000	-2.7
Aroclor-1260-2	6.841	6.742	6.942	502.590	500.000	0.5
Aroclor-1260-3	7.231	7.133	7.333	502.780	500.000	0.6
Aroclor-1260-4	7.476	7.376	7.576	505.990	500.000	1.2
Aroclor-1260-5	7.817	7.716	7.916	528.480	500.000	5.7
Decachlorobiphenyl	9.510	9.412	9.612	52.470	50.000	4.9
Tetrachloro-m-xylene	3.624	3.525	3.725	49.720	50.000	-0.6

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/10/2025 02/10/2025

Client Sample No.: CCAL01 Date Analyzed: 02/10/2025

Lab Sample No.: AR1660CCC500 Data File : PR070257.D Time Analyzed: 17:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.074	3.974	4.174	500.530	500.000	0.1
Aroclor-1016-2	4.091	3.992	4.192	497.290	500.000	-0.5
Aroclor-1016-3	4.271	4.172	4.372	501.580	500.000	0.3
Aroclor-1016-4	4.323	4.223	4.423	487.850	500.000	-2.4
Aroclor-1016-5	4.540	4.441	4.641	483.830	500.000	-3.2
Aroclor-1260-1	5.635	5.536	5.736	503.350	500.000	0.7
Aroclor-1260-2	5.841	5.743	5.943	510.750	500.000	2.2
Aroclor-1260-3	6.000	5.901	6.101	508.940	500.000	1.8
Aroclor-1260-4	6.507	6.407	6.607	507.580	500.000	1.5
Aroclor-1260-5	6.773	6.674	6.874	531.850	500.000	6.4
Decachlorobiphenyl	8.225	8.126	8.326	52.600	50.000	5.2
Tetrachloro-m-xylene	2.956	2.856	3.056	50.190	50.000	0.4