

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/10/2025 02/10/2025

Client Sample No.: CCAL02 Date Analyzed: 02/10/2025

Lab Sample No.: AR1660CCC500 Data File : PR070268.D Time Analyzed: 20:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.844	4.745	4.945	511.770	500.000	2.4
Aroclor-1016-2	4.865	4.766	4.966	505.410	500.000	1.1
Aroclor-1016-3	4.930	4.830	5.030	499.260	500.000	-0.1
Aroclor-1016-4	5.032	4.932	5.132	494.670	500.000	-1.1
Aroclor-1016-5	5.346	5.247	5.447	495.660	500.000	-0.9
Aroclor-1260-1	6.559	6.459	6.659	497.030	500.000	-0.6
Aroclor-1260-2	6.841	6.742	6.942	510.980	500.000	2.2
Aroclor-1260-3	7.233	7.133	7.333	505.970	500.000	1.2
Aroclor-1260-4	7.475	7.376	7.576	522.730	500.000	4.5
Aroclor-1260-5	7.816	7.716	7.916	532.580	500.000	6.5
Decachlorobiphenyl	9.509	9.412	9.612	52.800	50.000	5.6
Tetrachloro-m-xylene	3.624	3.525	3.725	50.370	50.000	0.7

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/10/2025 02/10/2025

Client Sample No.: CCAL02 Date Analyzed: 02/10/2025

Lab Sample No.: AR1660CCC500 Data File : PR070268.D Time Analyzed: 20:04

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.073	3.974	4.174	508.970	500.000	1.8
Aroclor-1016-2	4.091	3.992	4.192	508.290	500.000	1.7
Aroclor-1016-3	4.271	4.172	4.372	513.270	500.000	2.7
Aroclor-1016-4	4.322	4.223	4.423	499.740	500.000	-0.1
Aroclor-1016-5	4.540	4.441	4.641	496.860	500.000	-0.6
Aroclor-1260-1	5.635	5.536	5.736	511.560	500.000	2.3
Aroclor-1260-2	5.842	5.743	5.943	514.720	500.000	2.9
Aroclor-1260-3	5.999	5.901	6.101	515.340	500.000	3.1
Aroclor-1260-4	6.507	6.407	6.607	516.390	500.000	3.3
Aroclor-1260-5	6.774	6.674	6.874	541.920	500.000	8.4
Decachlorobiphenyl	8.224	8.126	8.326	53.340	50.000	6.7
Tetrachloro-m-xylene	2.956	2.856	3.056	50.770	50.000	1.5