

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR021025

Review By	yogesh	Review On	2/11/2025 11:14:49 AM
Supervise By	Ankita	Supervise On	2/11/2025 12:39:14 PM
SubDirectory	PR021025	HP Acquire Method	HP Processing Method PR021025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PR070231.D	10 Feb 2025 11:01		AJMA	Ok
2	I.BLK	I.BLK	PR070232.D	10 Feb 2025 11:15		AJMA	Ok
3	AR1660ICC1000	AR1660ICC1000	PR070233.D	10 Feb 2025 11:30		AJMA	Ok
4	AR1660ICC750	AR1660ICC750	PR070234.D	10 Feb 2025 11:45		AJMA	Ok
5	AR1660ICC500	AR1660ICC500	PR070235.D	10 Feb 2025 12:00		AJMA	Ok
6	AR1660ICC250	AR1660ICC250	PR070236.D	10 Feb 2025 12:14		AJMA	Ok
7	AR1660ICC050	AR1660ICC050	PR070237.D	10 Feb 2025 12:29		AJMA	Ok
8	AR1221ICC500	AR1221ICC500	PR070238.D	10 Feb 2025 12:44		AJMA	Ok
9	AR1232ICC500	AR1232ICC500	PR070239.D	10 Feb 2025 12:58		AJMA	Ok
10	AR1242ICC500	AR1242ICC500	PR070240.D	10 Feb 2025 13:13		AJMA	Ok
11	AR1248ICC500	AR1248ICC500	PR070241.D	10 Feb 2025 13:28		AJMA	Ok
12	AR1254ICC500	AR1254ICC500	PR070242.D	10 Feb 2025 13:42		AJMA	Ok
13	AR1262ICC500	AR1262ICC500	PR070243.D	10 Feb 2025 13:57		AJMA	Ok
14	AR1268ICC500	AR1268ICC500	PR070244.D	10 Feb 2025 14:12		AJMA	Ok,M
15	PR021025ICV500	ICVPR021025	PR070245.D	10 Feb 2025 14:26		AJMA	Ok
16	AR1660CCC500	AR1660CCC500	PR070246.D	10 Feb 2025 14:41		AJMA	Ok
17	I.BLK	I.BLK	PR070247.D	10 Feb 2025 14:56		AJMA	Ok
18	PB166204BL	PB166204BL	PR070248.D	10 Feb 2025 15:10		AJMA	Ok

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR021025

Review By	yogesh	Review On	2/11/2025 11:14:49 AM
Supervise By	Ankita	Supervise On	2/11/2025 12:39:14 PM
SubDirectory	PR021025	HP Acquire Method	HP Processing Method PR021025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1168-01	LOD-MDL-SOIL-01-QT	PR070249.D	10 Feb 2025 15:25	AR1660 LOD 25 PPB	AJMA	Ok,M
20	Q1168-01	LOD-MDL-SOIL-01-QT	PR070250.D	10 Feb 2025 15:40	AR1221 LOD 25 PPB	AJMA	Ok
21	Q1168-01	LOD-MDL-SOIL-01-QT	PR070251.D	10 Feb 2025 15:54	AR1232 LOD 25 PPB	AJMA	Ok,M
22	Q1168-01	LOD-MDL-SOIL-01-QT	PR070252.D	10 Feb 2025 16:09	AR1242 LOD 25 PPB	AJMA	Ok,M
23	Q1168-01	LOD-MDL-SOIL-01-QT	PR070253.D	10 Feb 2025 16:24	AR1248 LOD 25 PPB	AJMA	Ok,M
24	Q1168-01	LOD-MDL-SOIL-01-QT	PR070254.D	10 Feb 2025 16:38	AR1254 LOD 25 PPB	AJMA	Ok,M
25	Q1168-01	LOD-MDL-SOIL-01-QT	PR070255.D	10 Feb 2025 16:53	AR1262 LOD 25 PPB	AJMA	Ok,M
26	Q1168-01	LOD-MDL-SOIL-01-QT	PR070256.D	10 Feb 2025 17:08	AR1268 LOD 25 PPB	AJMA	Ok,M
27	AR1660CCC500	AR1660CCC500	PR070257.D	10 Feb 2025 17:22		AJMA	Ok
28	I.BLK	I.BLK	PR070258.D	10 Feb 2025 17:37		AJMA	Ok
29	PB166205BL	PB166205BL	PR070259.D	10 Feb 2025 17:52		AJMA	Ok
30	Q1168-07	LOD-MDL-WATER-01-Q	PR070260.D	10 Feb 2025 18:06	AR1660 LOD 25 PPB	AJMA	Ok,M
31	Q1168-07	LOD-MDL-WATER-01-Q	PR070261.D	10 Feb 2025 18:21	AR1221 LOD 25 PPB	AJMA	Ok,M
32	Q1168-07	LOD-MDL-WATER-01-Q	PR070262.D	10 Feb 2025 18:36	AR1232 LOD 25 PPB	AJMA	Ok,M
33	Q1168-07	LOD-MDL-WATER-01-Q	PR070263.D	10 Feb 2025 18:50	AR1242 LOD 25 PPB	AJMA	Ok,M
34	Q1168-07	LOD-MDL-WATER-01-Q	PR070264.D	10 Feb 2025 19:05	AR1248 LOD 25 PPB	AJMA	Ok
35	Q1168-07	LOD-MDL-WATER-01-Q	PR070265.D	10 Feb 2025 19:20	AR1254 LOD 25 PPB	AJMA	Ok,M
36	Q1168-07	LOD-MDL-WATER-01-Q	PR070266.D	10 Feb 2025 19:34	AR1262 LOD 25 PPB	AJMA	Ok,M
37	Q1168-07	LOD-MDL-WATER-01-Q	PR070267.D	10 Feb 2025 19:49	AR1268 LOD 25 PPB	AJMA	Ok,M

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR021025

Review By	yogesh	Review On	2/11/2025 11:14:49 AM
Supervise By	Ankita	Supervise On	2/11/2025 12:39:14 PM
SubDirectory	PR021025	HP Acquire Method	HP Processing Method PR021025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	AR1660CCC500	AR1660CCC500	PR070268.D	10 Feb 2025 20:04		AJMA	Ok
39	I.BLK	I.BLK	PR070269.D	10 Feb 2025 20:18		AJMA	Ok
40	Q1168-01	LOD-MDL-SOIL-01-QT	PR070270.D	10 Feb 2025 20:33	AR1660 LOD 40 PPB	AJMA	Ok,M
41	Q1168-01	LOD-MDL-SOIL-01-QT	PR070271.D	10 Feb 2025 20:48	AR1221 LOD 40 PPB	AJMA	Ok,M
42	Q1168-01	LOD-MDL-SOIL-01-QT	PR070272.D	10 Feb 2025 21:02	AR1232 LOD 40 PPB	AJMA	Ok,M
43	Q1168-01	LOD-MDL-SOIL-01-QT	PR070273.D	10 Feb 2025 21:17	AR1242 LOD 40 PPB	AJMA	Ok
44	Q1168-01	LOD-MDL-SOIL-01-QT	PR070274.D	10 Feb 2025 21:32	AR1248 LOD 40 PPB	AJMA	Ok
45	Q1168-01	LOD-MDL-SOIL-01-QT	PR070275.D	10 Feb 2025 21:47	AR1254 LOD 40 PPB	AJMA	Ok
46	Q1168-01	LOD-MDL-SOIL-01-QT	PR070276.D	10 Feb 2025 22:01	AR1262 LOD 40 PPB	AJMA	Ok,M
47	Q1168-01	LOD-MDL-SOIL-01-QT	PR070277.D	10 Feb 2025 22:16	AR1268 LOD 40 PPB	AJMA	Ok,M
48	AR1660CCC500	AR1660CCC500	PR070278.D	10 Feb 2025 22:30		AJMA	Ok
49	I.BLK	I.BLK	PR070279.D	10 Feb 2025 22:45		AJMA	Ok
50	Q1168-07	LOD-MDL-WATER-01-Q	PR070280.D	10 Feb 2025 23:00	AR1660 LOD 40 PPB	AJMA	Ok
51	Q1168-07	LOD-MDL-WATER-01-Q	PR070281.D	10 Feb 2025 23:15	AR1221 LOD 40 PPB	AJMA	Ok,M
52	Q1168-07	LOD-MDL-WATER-01-Q	PR070282.D	10 Feb 2025 23:29	AR1232 LOD 40 PPB	AJMA	Ok,M
53	Q1168-07	LOD-MDL-WATER-01-Q	PR070283.D	10 Feb 2025 23:44	AR1242 LOD 40 PPB	AJMA	Ok,M
54	Q1168-07	LOD-MDL-WATER-01-Q	PR070284.D	10 Feb 2025 23:59	AR1248 LOD 40 PPB	AJMA	Ok
55	Q1168-07	LOD-MDL-WATER-01-Q	PR070285.D	11 Feb 2025 00:13	AR1254 LOD 40 PPB	AJMA	Ok,M
56	Q1168-07	LOD-MDL-WATER-01-Q	PR070286.D	11 Feb 2025 00:28	AR1262 LOD 40 PPB	AJMA	Ok,M

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR021025

Review By	yogesh	Review On	2/11/2025 11:14:49 AM
Supervise By	Ankita	Supervise On	2/11/2025 12:39:14 PM
SubDirectory	PR021025	HP Acquire Method	HP Processing Method PR021025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

57	Q1168-07	LOD-MDL-WATER-01-0	PR070287.D	11 Feb 2025 00:43	AR1268 LOD 40 PPB	AJMA	Ok,M
58	AR1660CCC500	AR1660CCC500	PR070288.D	11 Feb 2025 00:57		AJMA	Ok
59	I.BLK	I.BLK	PR070289.D	11 Feb 2025 01:12		AJMA	Ok
60	PB166227BL	PB166227BL	PR070290.D	11 Feb 2025 01:27		AJMA	Ok
61	Q1168-02	LOQ-SOIL-02-QT1-202	PR070291.D	11 Feb 2025 01:41	AR1660 LOQ 50 PPB	AJMA	Ok
62	Q1168-02	LOQ-SOIL-02-QT1-202	PR070292.D	11 Feb 2025 01:56	AR1221 LOQ 50 PPB	AJMA	Ok,M
63	Q1168-02	LOQ-SOIL-02-QT1-202	PR070293.D	11 Feb 2025 02:11	AR1232 LOQ 50 PPB	AJMA	Ok,M
64	Q1168-02	LOQ-SOIL-02-QT1-202	PR070294.D	11 Feb 2025 02:25	AR1242 LOQ 50 PPB	AJMA	Ok
65	Q1168-02	LOQ-SOIL-02-QT1-202	PR070295.D	11 Feb 2025 02:40	AR1248 LOQ 50 PPB	AJMA	Ok
66	Q1168-02	LOQ-SOIL-02-QT1-202	PR070296.D	11 Feb 2025 02:55	AR1254 LOQ 50 PPB	AJMA	Ok
67	Q1168-02	LOQ-SOIL-02-QT1-202	PR070297.D	11 Feb 2025 03:09	AR1262 LOQ 50 PPB	AJMA	Ok,M
68	Q1168-02	LOQ-SOIL-02-QT1-202	PR070298.D	11 Feb 2025 03:24	AR1268 LOQ 50 PPB	AJMA	Not Ok
69	AR1660CCC500	AR1660CCC500	PR070299.D	11 Feb 2025 03:39		AJMA	Ok
70	I.BLK	I.BLK	PR070300.D	11 Feb 2025 03:53		AJMA	Ok
71	PB166228BL	PB166228BL	PR070301.D	11 Feb 2025 04:08		AJMA	Ok,M
72	Q1168-08	LOQ-WATER-02-QT1-2	PR070302.D	11 Feb 2025 04:23	AR1660 LOQ 50 PPB	AJMA	Ok
73	Q1168-08	LOQ-WATER-02-QT1-2	PR070303.D	11 Feb 2025 04:37	AR1221 LOQ 50 PPB	AJMA	Ok,M
74	Q1168-08	LOQ-WATER-02-QT1-2	PR070304.D	11 Feb 2025 04:52	AR1232 LOQ 50 PPB	AJMA	Ok,M
75	Q1168-08	LOQ-WATER-02-QT1-2	PR070305.D	11 Feb 2025 05:07	AR1242 LOQ 50 PPB	AJMA	Ok,M

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR021025

Review By	yogesh	Review On	2/11/2025 11:14:49 AM
Supervise By	Ankita	Supervise On	2/11/2025 12:39:14 PM
SubDirectory	PR021025	HP Acquire Method	HP Processing Method PR021025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

76	Q1168-08	LOQ-WATER-02-QT1-2	PR070306.D	11 Feb 2025 05:22	AR1248 LOQ 50 PPB	AJMA	Ok,M
77	Q1168-08	LOQ-WATER-02-QT1-2	PR070307.D	11 Feb 2025 05:36	AR1254 LOQ 50 PPB	AJMA	Ok,M
78	Q1168-08	LOQ-WATER-02-QT1-2	PR070308.D	11 Feb 2025 05:51	AR1262 LOQ 50 PPB	AJMA	Ok,M
79	Q1168-08	LOQ-WATER-02-QT1-2	PR070309.D	11 Feb 2025 06:06	AR1268 LOQ 50 PPB	AJMA	Not Ok
80	AR1660CCC500	AR1660CCC500	PR070310.D	11 Feb 2025 06:20		AJMA	Ok
81	I.BLK	I.BLK	PR070311.D	11 Feb 2025 06:35		AJMA	Ok
82	Q1168-03	MDL-SOIL-03-QT1-202	PR070312.D	11 Feb 2025 06:50	AR1660 MDL 25 PPB	AJMA	Ok,M
83	Q1168-03	MDL-SOIL-03-QT1-202	PR070313.D	11 Feb 2025 07:04	AR1221 MDL 25 PPB	AJMA	Ok,M
84	Q1168-03	MDL-SOIL-03-QT1-202	PR070314.D	11 Feb 2025 07:19	AR1232 MDL 25 PPB	AJMA	Ok,M
85	Q1168-03	MDL-SOIL-03-QT1-202	PR070315.D	11 Feb 2025 07:34	AR1242 MDL 25 PPB	AJMA	Ok,M
86	Q1168-03	MDL-SOIL-03-QT1-202	PR070316.D	11 Feb 2025 07:48	AR1248 MDL 25 PPB	AJMA	Ok,M
87	Q1168-03	MDL-SOIL-03-QT1-202	PR070317.D	11 Feb 2025 08:03	AR1254 MDL 25 PPB	AJMA	Ok,M
88	Q1168-03	MDL-SOIL-03-QT1-202	PR070318.D	11 Feb 2025 08:18	AR1262 MDL 25 PPB	AJMA	Ok,M
89	Q1168-03	MDL-SOIL-03-QT1-202	PR070319.D	11 Feb 2025 08:32	AR1268 MDL 25 PPB	AJMA	Ok,M
90	AR1660CCC500	AR1660CCC500	PR070320.D	11 Feb 2025 08:47		AJMA	Ok,M
91	I.BLK	I.BLK	PR070321.D	11 Feb 2025 09:02		AJMA	Ok
92	Q1168-09	MDL-WATER-03-QT1-2	PR070322.D	11 Feb 2025 09:16	AR1660 MDL 25 PPB	AJMA	Ok
93	Q1168-09	MDL-WATER-03-QT1-2	PR070323.D	11 Feb 2025 09:31	AR1221 MDL 25 PPB	AJMA	Ok,M
94	Q1168-09	MDL-WATER-03-QT1-2	PR070324.D	11 Feb 2025 09:46	AR1232 MDL 25 PPB	AJMA	Ok,M

Instrument ID: ECD_R

Daily Analysis Runlog For Sequence/QC Batch ID # PR021025

Review By	yogesh	Review On	2/11/2025 11:14:49 AM
Supervise By	Ankita	Supervise On	2/11/2025 12:39:14 PM
SubDirectory	PR021025	HP Acquire Method	HP Processing Method PR021025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

95	Q1168-09	MDL-WATER-03-QT1-2	PR070325.D	11 Feb 2025 10:00	AR1242 MDL 25 PPB	AJMA	Ok,M
96	Q1168-09	MDL-WATER-03-QT1-2	PR070326.D	11 Feb 2025 10:15	AR1248 MDL 25 PPB	AJMA	Ok,M
97	Q1168-09	MDL-WATER-03-QT1-2	PR070327.D	11 Feb 2025 10:30	AR1254 MDL 25 PPB	AJMA	Ok,M
98	Q1168-09	MDL-WATER-03-QT1-2	PR070328.D	11 Feb 2025 10:45	AR1262 MDL 25 PPB	AJMA	Ok,M
99	Q1168-09	MDL-WATER-03-QT1-2	PR070329.D	11 Feb 2025 10:59	AR1268 MDL 25 PPB	AJMA	Ok,M
100	Q1168-02	LOQ-SOIL-02-QT1-202	PR070330.D	11 Feb 2025 12:21	AR1268 LOQ 50 PPB, DCB high in both column,	AJMA	Ok,M
101	Q1168-08	LOQ-WATER-02-QT1-2	PR070331.D	11 Feb 2025 12:35	AR1268 LOQ 50 PPB, DCB high in both column,	AJMA	Ok,M
102	AR1660CCC500	AR1660CCC500	PR070332.D	11 Feb 2025 12:50		AJMA	Ok
103	I.BLK	I.BLK	PR070333.D	11 Feb 2025 13:05		AJMA	Ok

M : Manual Integration