

Surrogate Summary

SDG No.: Q1168

Client: Chemtech Consulting Group

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PR070258.D	PIBLK-PR070258.D	Tetrachloro-m-xylene	1	20	20.2	101		60	140
		Decachlorobiphenyl	1	20	22.2	111		60	140
		Tetrachloro-m-xylene	2	20	20.3	102		60	140
		Decachlorobiphenyl	2	20	22.1	110		60	140
PB166205BL	PB166205BL	Tetrachloro-m-xylene	1	20	18.6	93		10	157
		Decachlorobiphenyl	1	20	22.0	110		10	173
		Tetrachloro-m-xylene	2	20	18.3	92		10	157
		Decachlorobiphenyl	2	20	21.4	107		10	173
Q1168-07	LOD-MDL-WATER-01-QT1-2025	Tetrachloro-m-xylene	1	20	16.0	80		10	157
		Decachlorobiphenyl	1	20	19.0	95		10	173
		Tetrachloro-m-xylene	2	20	16.2	81		10	157
		Decachlorobiphenyl	2	20	18.6	93		10	173
I.BLK-PR070269.D	PIBLK-PR070269.D	Tetrachloro-m-xylene	1	20	20.3	101		60	140
		Decachlorobiphenyl	1	20	22.4	112		60	140
		Tetrachloro-m-xylene	2	20	20.6	103		60	140
		Decachlorobiphenyl	2	20	22.1	111		60	140
I.BLK-PR070333.D	PIBLK-PR070333.D	Tetrachloro-m-xylene	1	20	21.0	105		60	140
		Decachlorobiphenyl	1	20	24.1	121		60	140
		Tetrachloro-m-xylene	2	20	20.4	102		60	140
		Decachlorobiphenyl	2	20	23.8	119		60	140