

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM **Case No.:** Q1168 **SAS No.:** Q1168 **SDG NO.:** Q1168

Instrument ID: ECD_L **Calibration Date(s):** 01/21/2025 01/21/2025
Calibration Times: 10:57 11:51

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PL093728.D</u>		CF 075 = <u>PL093729.D</u>			
CF 050 = <u>PL093730.D</u>		CF 025 = <u>PL093731.D</u>		CF 005 = <u>PL093732.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	1660930000	1679650000	1932670000	1802720000	2426830000	1900560000	17
4,4'-DDE	2179870000	2169930000	2489080000	2321590000	3012520000	2434600000	14
4,4'-DDT	1755570000	1766710000	2016720000	1907120000	2414170000	1972060000	14
Aldrin	2924220000	2896750000	3292630000	3099660000	4146570000	3271970000	16
alpha-BHC	3537700000	3490280000	3918110000	3562830000	4660310000	3833850000	13
alpha-Chlordane	2458070000	2458490000	2788200000	2666580000	3570690000	2788400000	16
beta-BHC	1393460000	1394440000	1618290000	1508890000	2121530000	1607320000	19
Decachlorobiphenyl	1768480000	1816480000	2098320000	2018470000	2757820000	2091910000	19
delta-BHC	3233860000	3194550000	3605880000	3303370000	4188780000	3505290000	12
Dieldrin	2456580000	2440810000	2788190000	2639340000	3554340000	2775850000	17
Endosulfan I	2304400000	2298550000	2637060000	2528610000	3445650000	2642860000	18
Endosulfan II	2084130000	2100600000	2413950000	2287820000	3160260000	2409350000	18
Endosulfan sulfate	1923100000	1945070000	2248580000	2190510000	3011450000	2263740000	20
Endrin	2079430000	2060990000	2363220000	2218560000	3001890000	2344820000	17
Endrin aldehyde	1673120000	1696040000	1958970000	1896570000	2495580000	1944060000	17
Endrin ketone	2196850000	2205550000	2539700000	2413910000	3257130000	2522630000	17
gamma-BHC (Lindane)	3375960000	3339350000	3767250000	3460830000	4470850000	3682850000	13
gamma-Chlordane	2455830000	2471830000	2815630000	2678390000	3515170000	2787370000	16
Heptachlor	2922500000	2901690000	3325290000	3144100000	4093120000	3277340000	15
Heptachlor epoxide	2568680000	2575960000	2953630000	2835830000	3935020000	2973820000	19
Methoxychlor	907284000	922109000	1080370000	1020090000	1287130000	1043400000	15
Tetrachloro-m-xylene	2397870000	2402980000	2740040000	2595500000	3327420000	2692760000	14

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM **Case No.:** Q1168 **SAS No.:** Q1168 **SDG NO.:** Q1168

Instrument ID: ECD_L **Calibration Date(s):** 01/21/2025 01/21/2025
Calibration Times: 10:57 11:51

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PL093728.D</u>		CF 075 = <u>PL093729.D</u>			
CF 050 = <u>PL093730.D</u>		CF 025 = <u>PL093731.D</u>		CF 005 = <u>PL093732.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	3134000000	3054730000	3379160000	2910470000	3304320000	3156540000	6
4,4'-DDE	3891920000	3807640000	4253650000	3749010000	4345130000	4009470000	7
4,4'-DDT	3270010000	3177800000	3542860000	3046890000	3232670000	3254050000	6
Aldrin	4482990000	4370810000	4856520000	4222470000	4876190000	4561800000	6
alpha-BHC	4914190000	4768640000	5271080000	4480730000	5010260000	4888980000	6
alpha-Chlordane	4056970000	3962110000	4424110000	3914810000	4574820000	4186560000	7
beta-BHC	1863440000	1842720000	2072180000	1889740000	2319100000	1997440000	10
Decachlorobiphenyl	3226690000	3193800000	3627020000	3320620000	4152210000	3504070000	11
delta-BHC	4741230000	4607910000	5098810000	4368820000	4939430000	4751240000	6
Dieldrin	4189300000	4076770000	4553570000	3958830000	4699760000	4295650000	7
Endosulfan I	3734100000	3661580000	4099030000	3635320000	4254550000	3876920000	7
Endosulfan II	3553260000	3487640000	3912960000	3484510000	4080760000	3703830000	7
Endosulfan sulfate	3408630000	3353240000	3757030000	3348270000	3963240000	3566080000	8
Endrin	3607760000	3481170000	3870730000	3406140000	4097610000	3692680000	8
Endrin aldehyde	2861460000	2820180000	3183430000	2892290000	3465840000	3044640000	9
Endrin ketone	3965120000	3881890000	4400080000	3907370000	4821740000	4195240000	10
gamma-BHC (Lindane)	4713370000	4597010000	5084610000	4384810000	4926270000	4741210000	6
gamma-Chlordane	4137240000	4016860000	4483010000	3935490000	4615500000	4237620000	7
Heptachlor	4505180000	4413750000	4924840000	4345980000	5084220000	4654790000	7
Heptachlor epoxide	4026840000	3946880000	4424170000	3927960000	4575440000	4180260000	7
Methoxychlor	1651870000	1634200000	1870410000	1643810000	2140390000	1788140000	12
Tetrachloro-m-xylene	3101220000	3058550000	3437230000	3066200000	3657590000	3264160000	8