

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM **Case No.:** Q1168 **SAS No.:** Q1168 **SDG NO.:** Q1168

Instrument ID: ECD_D **Date(s) Analyzed:** 01/22/2025 01/22/2025

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	6.25	6.15	6.35	24342100
		2	6.45	6.35	6.55	34232400
		3	7.15	7.05	7.25	66136500
		4	7.57	7.47	7.67	84346400
		5	7.93	7.83	8.03	49170000

INITIAL CALIBRATION OF MULTICOMPONENT ANALYTES

Contract: CHEM02

Lab Code: CHEM **Case No.:** Q1168 **SAS No.:** Q1168 **SDG NO.:** Q1168

Instrument ID: ECD_D **Date(s) Analyzed:** 01/22/2025 01/22/2025

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	AMOUNT (ng)	PEAK	RT	RT WINDOW		CALIBRATION FACTOR
				FROM	TO	
Toxaphene	500	1	5.48	5.38	5.58	114468000
		2	5.65	5.55	5.75	76949100
		3	6.76	6.66	6.86	360443000
		4	7.21	7.11	7.31	262909000
		5	7.34	7.24	7.44	180637000