

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

Continuing Calib Date: 02/07/2025 Initial Calibration Date(s): 01/22/2025 01/22/2025

Continuing Calib Time: 09:34 Initial Calibration Time(s): 11:51 12:47

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.09	9.08	8.98	9.18	-0.01
Tetrachloro-m-xylene	3.56	3.56	3.46	3.66	0.00
alpha-BHC	4.01	4.00	3.90	4.10	-0.01
beta-BHC	4.52	4.52	4.42	4.62	0.00
delta-BHC	4.77	4.77	4.67	4.87	0.00
gamma-BHC (Lindane)	4.34	4.34	4.24	4.44	0.00
Heptachlor	4.94	4.94	4.84	5.04	0.00
Aldrin	5.28	5.28	5.18	5.38	0.00
Heptachlor epoxide	5.70	5.70	5.60	5.80	0.00
Endosulfan I	6.08	6.08	5.98	6.18	0.00
Dieldrin	6.36	6.35	6.25	6.45	-0.01
4,4'-DDE	6.20	6.20	6.10	6.30	0.00
Endrin	6.58	6.58	6.48	6.68	0.00
Endosulfan II	6.79	6.79	6.69	6.89	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.15	7.05	7.25	-0.01
4,4'-DDT	7.03	7.03	6.93	7.13	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.64	7.63	7.53	7.73	-0.01
Endrin aldehyde	6.92	6.92	6.82	7.02	0.00
alpha-Chlordane	6.04	6.03	5.93	6.13	0.00
gamma-Chlordane	5.95	5.95	5.85	6.05	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG NO.: Q1168

Continuing Calib Date: 02/07/2025 Initial Calibration Date(s): 01/22/2025 01/22/2025

Continuing Calib Time: 09:34 Initial Calibration Time(s): 11:51 12:47

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.08	8.08	7.98	8.18	0.00
Tetrachloro-m-xylene	2.88	2.88	2.78	2.98	0.00
alpha-BHC	3.40	3.40	3.30	3.50	0.01
beta-BHC	4.03	4.03	3.93	4.13	0.00
delta-BHC	4.27	4.27	4.17	4.37	0.01
gamma-BHC (Lindane)	3.73	3.73	3.63	3.83	0.00
Heptachlor	4.09	4.09	3.99	4.19	0.00
Aldrin	4.37	4.37	4.27	4.47	0.00
Heptachlor epoxide	4.88	4.88	4.78	4.98	0.00
Endosulfan I	5.25	5.25	5.15	5.35	0.00
Dieldrin	5.52	5.52	5.42	5.62	0.00
4,4'-DDE	5.38	5.38	5.28	5.48	0.00
Endrin	5.80	5.79	5.69	5.89	-0.01
Endosulfan II	6.09	6.09	5.99	6.19	0.00
4,4'-DDD	5.94	5.94	5.84	6.04	0.00
Endosulfan sulfate	6.49	6.49	6.39	6.59	0.00
4,4'-DDT	6.19	6.19	6.09	6.29	0.00
Methoxychlor	6.76	6.76	6.66	6.86	0.00
Endrin ketone	7.00	7.00	6.90	7.10	0.00
Endrin aldehyde	6.27	6.26	6.16	6.36	0.00
alpha-Chlordane	5.20	5.20	5.10	5.30	0.00
gamma-Chlordane	5.13	5.13	5.03	5.23	0.00