

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
 ProjectID: LOD-LOQ Study
 Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG No.: Q1168

Calibration Sequence : FE012325		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	159126227	93604	FE052027.D
850	78434502	92276	FE052028.D
340	32838892	96585	FE052029.D
170	18083163	106372	FE052030.D
85	8992058	105789	FE052031.D
AVG RF : 98925		% RSD : 6.792	AVG RT : 15.2554