

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CHEM02
 ProjectID: LOD-LOQ Study
 Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG No.: Q1168

Calibration Sequence : FF011425		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	222185351	130697	FF015220.D
850	98300301	115647	FF015221.D
340	41298428	121466	FF015222.D
170	22822884	134252	FF015223.D
85	10986372	129251	FF015224.D
AVG RF : 126263		% RSD : 5.981	AVG RT : 15.0138