

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG No.: Q1168
DataFile: FE052057.D Analyst Name: YP\AJ Analyst Date: 01-24-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	50195783	100392	106182	5.453