

TPH GC CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: CHEM02
ProjectID: LOD-LOQ Study
Lab Code: CHEM Case No.: Q1168 SAS No.: Q1168 SDG No.: Q1168
DataFile: FG015135.D Analyst Name: YP\AJ Analyst Date: 01-23-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
850	130883693	153981	131220	17.346