

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** Q1168 **SAS No.:** Q1168 **SDG NO.:** Q1168

Continuing Calib Date: 01/30/2025 **Initial Calibration Date(s):** 01/30/2025 01/30/2025

Continuing Calib Time: 12:01 **Initial Calibration Time(s):** 09:39 10:27

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	5.97	5.97	5.87	6.07	0.00
EDB	2.62	2.62	2.52	2.72	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** Q1168 **SAS No.:** Q1168 **SDG NO.:** Q1168

Continuing Calib Date: 01/30/2025 **Initial Calibration Date(s):** 01/30/2025 01/30/2025

Continuing Calib Time: 12:01 **Initial Calibration Time(s):** 09:39 10:27

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	5.00	5.00	4.90	5.10	0.01
EDB	1.98	1.98	1.88	2.08	0.00