



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Chemtech Consulting Group

Project Name: NJ Drinking Water PT

Project # N/A

Chemtech Project # Q1172

Test Name: Metals Group7

A. Number of Samples and Date of Receipt:

11 Water samples were received on 01/15/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Hardness, Calcium, Hardness, Total, Mercury, Metals Group3, Metals Group6, Metals Group7, Silica, Turbidity, VOCGC Group 1, VOCMS Group1, VOCMS Group2, VOCMS Group3, VOCMS Group4 and VOCMS Group5. This data package contains results for Metals Group7.

C. Analytical Techniques:

The analysis and digestion of Metals Group7 was based on method 200.8.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

E. Additional Comments:

Collision cell is being used to remove potential interferences. The analytes Na, Mg, Al, K, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As are being analyzed with collision cell and analytes Be, B, Ca, Ti, Se, Sr, Zr, Mo, Ag, Cd, Sn, Sb, Ba, Tl, Pb, U are being analyzed with Non-Collision Cell. Helium gas is used for the Collision Cell analysis.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____