



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Chemtech Consulting Group

SDG No.: Q1172

Project: NJ Drinking Water PT

RunNo.: LB134679

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Turbidity	ICV	NTU	9.474	10	95	90-110	02/13/2025
Sample ID: Turbidity	CCV1	NTU	9.508	10	95	90-110	02/13/2025
Sample ID: Turbidity	CCV2	NTU	9.282	10	93	90-110	02/13/2025

Initial and Continuing Calibration Verification

Client: Chemtech Consulting Group

SDG No.: Q1172

Project: NJ Drinking Water PT

RunNo.: LB134680

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Turbidity	ICV	NTU	9.474	10	95	90-110	02/13/2025
Sample ID: Turbidity	CCV1	NTU	9.508	10	95	90-110	02/13/2025
Sample ID: Turbidity	CCV2	NTU	9.282	10	93	90-110	02/13/2025