

Initial and Continuing Calibration Blank Summary

Client: Chemtech Consulting Group

SDG No.: Q1172

Project: NJ Drinking Water PT

RunNo.: LB134679

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Turbidity	NTU	< 0.5000	0.5000	U	0.075	1.0	02/13/2025
Sample ID: CCB1 Turbidity	NTU	0.084	0.5000	J	0.075	1	02/13/2025
Sample ID: CCB2 Turbidity	NTU	< 0.5000	0.5000	U	0.075	1	02/13/2025

Initial and Continuing Calibration Blank Summary

Client: Chemtech Consulting Group

SDG No.: Q1172

Project: NJ Drinking Water PT

RunNo.: LB134680

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Turbidity	NTU	< 0.5000	0.5000	U	0.075	1.0	02/13/2025
Sample ID: CCB1 Turbidity	NTU	0.084	0.5000	J	0.075	1	02/13/2025
Sample ID: CCB2 Turbidity	NTU	< 0.5000	0.5000	U	0.075	1	02/13/2025