

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134671

Review By	jaswal	Review On	2/11/2025 10:37:57 PM
Supervise By	Mohan	Supervise On	2/12/2025 11:24:41 AM

STD. NAME	STD REF.#
ICAL Standard	MP84450,MP84451,MP84452,MP84453,MP84454,MP84456,MP84457
ICV Standard	MP84458
CCV Standard	MP84460
ICSA Standard	
CRI Standard	MP84462
LCS Standard	
Chk Standard	MP84459,MP84461,MP84463,MP84447

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	02/11/25 15:00		Mohan	OK
2	S0.05	S0.05	CAL2	02/11/25 15:02		Mohan	OK
3	S0.2	S0.2	CAL3	02/11/25 15:04		Mohan	OK
4	S2.5	S2.5	CAL4	02/11/25 15:06		Mohan	OK
5	S5	S5	CAL5	02/11/25 15:09		Mohan	OK
6	S7.5	S7.5	CAL6	02/11/25 15:11		Mohan	OK
7	S10	S10	CAL7	02/11/25 15:13		Mohan	OK
8	ICV96	ICV96	ICV	02/11/25 15:17		Mohan	OK
9	ICB96	ICB96	ICB	02/11/25 15:19		Mohan	OK
10	CCV56	CCV56	CCV	02/11/25 15:21		Mohan	OK
11	CCB56	CCB56	CCB	02/11/25 15:24		Mohan	OK
12	CRA	CRA	CRDL	02/11/25 15:28		Mohan	OK
13	HighStd	HighStd	HIGH STD	02/11/25 15:31		Mohan	OK
14	ChkStd	ChkStd	SAM	02/11/25 15:33		Mohan	OK
15	PB166684BL	PB166684BL	MB	02/11/25 15:35		Mohan	OK
16	Q1300-02	COMP	SAM	02/11/25 15:47		Mohan	OK
17	Q1322-01	MANHOLE	SAM	02/11/25 15:50		Mohan	OK
18	Q1322-01DUP	MANHOLEDUP	DUP	02/11/25 15:52		Mohan	OK

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB134671

Review By	jaswal	Review On	2/11/2025 10:37:57 PM
Supervise By	Mohan	Supervise On	2/12/2025 11:24:41 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84450,MP84451,MP84452,MP84453,MP84454,MP84456,MP84457		
ICV Standard	MP84458		
CCV Standard	MP84460		
ICSA Standard			
CRI Standard	MP84462		
LCS Standard			
Chk Standard	MP84459,MP84461,MP84463,MP84447		

19	Q1322-01MS	MANHOLEMS	MS	02/11/25 15:57		Mohan	OK
20	CCV57	CCV57	CCV	02/11/25 15:59		Mohan	OK
21	CCB57	CCB57	CCB	02/11/25 16:02		Mohan	OK
22	Q1322-01MSD	MANHOLEMSD	MSD	02/11/25 16:04		Mohan	OK
23	Q1172-05	PT-HG-WS	SAM	02/11/25 16:06		Mohan	OK
24	Q1322-01L	MANHOLEL	SD	02/11/25 16:13		Mohan	OK
25	Q1322-01A	MANHOLEA	PS	02/11/25 16:15		Mohan	OK
26	PB166684BS	PB166684BS	LCS	02/11/25 16:18		Mohan	OK
27	CCV58	CCV58	CCV	02/11/25 16:20		Mohan	OK
28	CCB58	CCB58	CCB	02/11/25 16:22		Mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134704

Review By	kareem	Review On	2/13/2025 5:06:02 PM
Supervise By	JANVI	Supervise On	2/13/2025 5:07:41 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84381,MP84383,MP84384,MP84385,MP84386,MP84388		
ICV Standard	MP84446		
CCV Standard	MP84392		
ICSA Standard	MP84230,MP84231		
CRI Standard			
LCS Standard	MP84387		
Chk Standard	MP84218,MP84219		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	02/12/25 11:47		Kareem	OK
2	S1	S1	CAL2	02/12/25 11:52		Kareem	OK
3	S2	S2	CAL3	02/12/25 11:56		Kareem	OK
4	S3	S3	CAL4	02/12/25 12:00		Kareem	OK
5	S4	S4	CAL5	02/12/25 12:04		Kareem	OK
6	S5	S5	CAL6	02/12/25 12:09		Kareem	OK
7	ICV01	ICV01	ICV	02/12/25 12:59		Kareem	OK
8	LLICV01	LLICV01	LLICV	02/12/25 13:08		Kareem	OK
9	ICB01	ICB01	ICB	02/12/25 13:12		Kareem	OK
10	CRI01	CRI01	CRDL	02/12/25 13:17		Kareem	OK
11	ICSA01	ICSA01	ICSA	02/12/25 13:21		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	02/12/25 13:25		Kareem	OK
13	ICSADL	ICSADL	ICSA	02/12/25 13:31		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	02/12/25 13:35		Kareem	OK
15	CCV01	CCV01	CCV	02/12/25 13:39		Kareem	OK
16	CCB01	CCB01	CCB	02/12/25 13:43		Kareem	OK
17	Q1341-01	HAR-ROLLOFFS	SAM	02/12/25 13:48		Kareem	OK
18	Q1343-01	WC-9	SAM	02/12/25 13:52		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134704

Review By	kareem	Review On	2/13/2025 5:06:02 PM
Supervise By	JANVI	Supervise On	2/13/2025 5:07:41 PM

STD. NAME	STD REF.#
ICAL Standard	MP84204,MP84381,MP84383,MP84384,MP84385,MP84386,MP84388
ICV Standard	MP84446
CCV Standard	MP84392
ICSA Standard	MP84230,MP84231
CRI Standard	
LCS Standard	MP84387
Chk Standard	MP84218,MP84219

19	Q1343-05	WC-10	SAM	02/12/25 13:56		Kareem	OK
20	Q1343-09	WC-7	SAM	02/12/25 14:00		Kareem	OK
21	Q1343-09DUP	WC-7DUP	DUP	02/12/25 14:04		Kareem	OK
22	Q1343-09L	WC-7L	SD	02/12/25 14:08		Kareem	OK
23	Q1343-09MS	WC-7MS	MS	02/12/25 14:13	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
24	Q1343-09MSD	WC-7MSD	MSD	02/12/25 14:17	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
25	Q1343-09A	WC-7A	PS	02/12/25 14:21	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
26	Q1343-13	WC-13	SAM	02/12/25 14:25		Kareem	OK
27	CCV02	CCV02	CCV	02/12/25 14:29		Kareem	OK
28	CCB02	CCB02	CCB	02/12/25 14:33		Kareem	OK
29	Q1343-17	WC-14	SAM	02/12/25 14:37		Kareem	OK
30	PB166634BL	PB166634BL	MB	02/12/25 14:41		Kareem	OK
31	PB166634BS	PB166634BS	LCS	02/12/25 14:46	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
32	Q1324-01	WATER TREATMENT	SAM	02/12/25 14:50		Kareem	OK
33	Q1324-01DUP	WATER TREATMENT	DUP	02/12/25 14:55		Kareem	OK
34	Q1324-01L	WATER TREATMENT	SD	02/12/25 14:59		Kareem	OK
35	Q1324-01MS	WATER TREATMENT	MS	02/12/25 15:03	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
36	Q1324-01MSD	WATER TREATMENT	MSD	02/12/25 15:07	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134704

Review By	kareem	Review On	2/13/2025 5:06:02 PM
Supervise By	JANVI	Supervise On	2/13/2025 5:07:41 PM

STD. NAME	STD REF.#
ICAL Standard	MP84204,MP84381,MP84383,MP84384,MP84385,MP84386,MP84388
ICV Standard	MP84446
CCV Standard	MP84392
ICSA Standard	MP84230,MP84231
CRI Standard	
LCS Standard	MP84387
Chk Standard	MP84218,MP84219

37	Q1324-01A	WATER TREATMENT	PS	02/12/25 15:12	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
38	PB166635BL	PB166635BL	MB	02/12/25 15:16		Kareem	OK
39	CCV03	CCV03	CCV	02/12/25 15:20		Kareem	OK
40	CCB03	CCB03	CCB	02/12/25 15:24		Kareem	OK
41	PB166635BS	PB166635BS	LCS	02/12/25 15:29	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
42	PB166651TB	PB166651TB	MB	02/12/25 15:33		Kareem	OK
43	Q1343-04	WC-9	SAM	02/12/25 15:37		Kareem	OK
44	Q1343-08	WC-10	SAM	02/12/25 15:42		Kareem	OK
45	Q1343-12	WC-7	SAM	02/12/25 15:46		Kareem	OK
46	Q1343-16	WC-13	SAM	02/12/25 15:51		Kareem	OK
47	Q1343-20	WC-14	SAM	02/12/25 15:55		Kareem	OK
48	Q1344-02	SOIL-1	SAM	02/12/25 15:59		Kareem	OK
49	Q1344-04	SOIL-2	SAM	02/12/25 16:04		Kareem	OK
50	Q1346-02	SOIL-COMP	SAM	02/12/25 16:08		Kareem	OK
51	CCV04	CCV04	CCV	02/12/25 16:13		Kareem	OK
52	CCB04	CCB04	CCB	02/12/25 16:17		Kareem	OK
53	Q1346-02DUP	SOIL-COMPDUP	DUP	02/12/25 16:21		Kareem	OK
54	Q1346-02L	SOIL-COMPL	SD	02/12/25 16:26		Kareem	OK
55	Q1346-02MS	SOIL-COMPMS	MS	02/12/25 16:30	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134704

Review By	kareem	Review On	2/13/2025 5:06:02 PM
Supervise By	JANVI	Supervise On	2/13/2025 5:07:41 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84381,MP84383,MP84384,MP84385,MP84386,MP84388		
ICV Standard	MP84446		
CCV Standard	MP84392		
ICSA Standard	MP84230,MP84231		
CRI Standard			
LCS Standard	MP84387		
Chk Standard	MP84218,MP84219		

56	Q1346-02MSD	SOIL-COMPMSD	MSD	02/12/25 16:34	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
57	Q1346-02A	SOIL-COMPA	PS	02/12/25 16:39	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
58	PB166680BL	PB166680BL	MB	02/12/25 16:43		Kareem	OK
59	PB166680BS	PB166680BS	LCS	02/12/25 16:47	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
60	Q1172-04	PT-TM-WS	SAM	02/12/25 16:55		Kareem	OK
61	Q1346-01	SOIL-COMP	SAM	02/12/25 17:00		Kareem	OK
62	CCV05	CCV05	CCV	02/12/25 17:04		Kareem	OK
63	CCB05	CCB05	CCB	02/12/25 17:08		Kareem	OK
64	PB166704BL	PB166704BL	MB	02/12/25 17:12		Kareem	OK
65	PB166704BS	PB166704BS	LCS	02/12/25 17:17	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
66	Q1346-03	CONCRETE	SAM	02/12/25 17:21		Kareem	OK
67	Q1353-01	346	SAM	02/12/25 17:25		Kareem	OK
68	Q1353-01DUP	346DUP	DUP	02/12/25 17:29		Kareem	OK
69	Q1353-01L	346L	SD	02/12/25 17:34		Kareem	OK
70	Q1353-01MS	346MS	MS	02/12/25 17:38	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
71	Q1353-01MSD	346MSD	MSD	02/12/25 17:42	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
72	Q1353-01A	346A	PS	02/12/25 17:46	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
73	Q1354-01	NB-08-021125	SAM	02/12/25 17:50		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134704

Review By	kareem	Review On	2/13/2025 5:06:02 PM
Supervise By	JANVI	Supervise On	2/13/2025 5:07:41 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84381,MP84383,MP84384,MP84385,MP84386,MP84388		
ICV Standard	MP84446		
CCV Standard	MP84392		
ICSA Standard	MP84230,MP84231		
CRI Standard			
LCS Standard	MP84387		
Chk Standard	MP84218,MP84219		

74	CCV06	CCV06	CCV	02/12/25 17:55		Kareem	OK
75	CCB06	CCB06	CCB	02/12/25 17:59		Kareem	OK
76	PB166414BL	PB166414BL	MB	02/12/25 18:08		Kareem	OK
77	PB166414BS	PB166414BS	LCS	02/12/25 18:12	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
78	PB166688BL	PB166688BL	MB	02/12/25 18:16		Kareem	OK
79	PB166688BS	PB166688BS	LCS	02/12/25 18:20	0.1 ML M6010 AND M6001 WERE ADDED	Kareem	OK
80	LR1	LR1	HIGH STD	02/12/25 18:24		Kareem	OK
81	LR2	LR2	HIGH STD	02/12/25 18:29		Kareem	OK
82	CCV07	CCV07	CCV	02/12/25 18:34		Kareem	OK
83	CCB07	CCB07	CCB	02/12/25 18:38		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134743

Review By	kareem	Review On	2/19/2025 10:19:34 AM
Supervise By	Janvi	Supervise On	2/21/2025 4:44:58 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84223,MP84224,MP84225,MP84226,MP84228		
ICV Standard	MP84446		
CCV Standard	MP84232		
ICSA Standard	MP84230,MP84231		
CRI Standard	MP84228		
LCS Standard			
Chk Standard	MP84218,MP84219		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	02/18/25 11:19		Kareem	OK
2	S1	S1	CAL2	02/18/25 11:24		Kareem	OK
3	S2	S2	CAL3	02/18/25 11:28		Kareem	OK
4	S3	S3	CAL4	02/18/25 11:32		Kareem	OK
5	S4	S4	CAL5	02/18/25 11:37		Kareem	OK
6	S5	S5	CAL6	02/18/25 11:41		Kareem	OK
7	ICV01	ICV01	ICV	02/18/25 11:58		Kareem	OK
8	LLICV01	LLICV01	LLICV	02/18/25 12:10		Kareem	OK
9	ICB01	ICB01	ICB	02/18/25 12:14		Kareem	OK
10	CRI01	CRI01	CRDL	02/18/25 12:18		Kareem	OK
11	ICSA01	ICSA01	ICSA	02/18/25 12:23		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	02/18/25 12:29		Kareem	OK
13	ICSADL	ICSADL	ICSA	02/18/25 12:33		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	02/18/25 12:37		Kareem	OK
15	CCV01	CCV01	CCV	02/18/25 12:41		Kareem	OK
16	CCB01	CCB01	CCB	02/18/25 12:45		Kareem	OK
17	Q1322-01	MANHOLE	SAM	02/18/25 12:50		Kareem	OK
18	Q1322-01DUP	MANHOLEDUP	DUP	02/18/25 12:54		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134743

Review By	kareem	Review On	2/19/2025 10:19:34 AM
Supervise By	Janvi	Supervise On	2/21/2025 4:44:58 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84223,MP84224,MP84225,MP84226,MP84228		
ICV Standard	MP84446		
CCV Standard	MP84232		
ICSA Standard	MP84230,MP84231		
CRI Standard	MP84228		
LCS Standard			
Chk Standard	MP84218,MP84219		

19	Q1322-01L	MANHOLEL	SD	02/18/25 12:59		Kareem	OK
20	Q1322-01MS	MANHOLEMS	MS	02/18/25 13:03	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
21	Q1322-01MSD	MANHOLEMSD	MSD	02/18/25 13:07	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
22	Q1322-01A	MANHOLEA	PS	02/18/25 13:11	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
23	PB166737BL	PB166737BL	MB	02/18/25 13:15		Kareem	OK
24	PB166737BS	PB166737BS	LCS	02/18/25 13:20	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
25	Q1366-01	HD-01-2132025	SAM	02/18/25 13:24		Kareem	OK
26	Q1366-01DUP	HD-01-2132025DUP	DUP	02/18/25 13:28		Kareem	OK
27	CCV02	CCV02	CCV	02/18/25 13:32		Kareem	OK
28	CCB02	CCB02	CCB	02/18/25 13:36		Kareem	OK
29	Q1366-01L	HD-01-2132025L	SD	02/18/25 13:41		Kareem	OK
30	Q1366-01MS	HD-01-2132025MS	MS	02/18/25 13:45	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
31	Q1366-01MSD	HD-01-2132025MSD	MSD	02/18/25 13:49	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
32	Q1366-01A	HD-01-2132025A	PS	02/18/25 13:53	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134743

Review By	kareem	Review On	2/19/2025 10:19:34 AM
Supervise By	Janvi	Supervise On	2/21/2025 4:44:58 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84223,MP84224,MP84225,MP84226,MP84228		
ICV Standard	MP84446		
CCV Standard	MP84232		
ICSA Standard	MP84230,MP84231		
CRI Standard	MP84228		
LCS Standard			
Chk Standard	MP84218,MP84219		

33	PB166731BL	PB166731BL	MB	02/18/25 13:58		Kareem	OK
34	PB166731BS	PB166731BS	LCS	02/18/25 14:02	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
35	Q1172-03	PT-MIN-WS	SAM	02/18/25 14:06		Kareem	OK
36	Q1172-06	PT-SIO2-WS	SAM	02/18/25 14:10		Kareem	OK
37	Q1372-01	AU-05-021425	SAM	02/18/25 14:15		Kareem	OK
38	Q1374-01	TR-05-02142025	SAM	02/18/25 14:19		Kareem	OK
39	CCV03	CCV03	CCV	02/18/25 14:23		Kareem	OK
40	CCB03	CCB03	CCB	02/18/25 14:27		Kareem	OK
41	Q1374-01DUP	TR-05-02142025DUP	DUP	02/18/25 14:32		Kareem	OK
42	Q1374-01L	TR-05-02142025L	SD	02/18/25 14:36		Kareem	OK
43	Q1374-01MS	TR-05-02142025MS	MS	02/18/25 14:40	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
44	Q1374-01MSD	TR-05-02142025MSD	MSD	02/18/25 14:44	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
45	Q1374-01A	TR-05-02142025A	PS	02/18/25 14:48	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
46	PB166747BL	PB166747BL	MB	02/18/25 14:52		Kareem	OK
47	PB166747BS	PB166747BS	LCS	02/18/25 15:01	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
48	PB166735TB	PB166735TB	MB	02/18/25 15:05		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134743

Review By	kareem	Review On	2/19/2025 10:19:34 AM
Supervise By	Janvi	Supervise On	2/21/2025 4:44:58 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84223,MP84224,MP84225,MP84226,MP84228		
ICV Standard	MP84446		
CCV Standard	MP84232		
ICSA Standard	MP84230,MP84231		
CRI Standard	MP84228		
LCS Standard			
Chk Standard	MP84218,MP84219		

49	Q1373-01	VB27161	SAM	02/18/25 15:10		Kareem	OK
50	Q1373-02	VDR25992	SAM	02/18/25 15:14		Kareem	OK
51	CCV04	CCV04	CCV	02/18/25 15:19		Kareem	OK
52	CCB04	CCB04	CCB	02/18/25 15:23		Kareem	OK
53	Q1373-02DUP	VDR25992DUP	DUP	02/18/25 15:27		Kareem	OK
54	Q1373-02L	VDR25992L	SD	02/18/25 15:32		Kareem	OK
55	Q1373-02MS	VDR25992MS	MS	02/18/25 15:36	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
56	Q1373-02MSD	VDR25992MSD	MSD	02/18/25 15:40	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
57	Q1373-02A	VDR25992A	PS	02/18/25 15:44	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
58	PB166757BL	PB166757BL	MB	02/18/25 15:48		Kareem	OK
59	PB166757BS	PB166757BS	LCS	02/18/25 15:53	0.1 ML M6010 AND M6001 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
60	LR1	LR1	HIGH STD	02/18/25 15:57		Kareem	OK
61	LR2	LR2	HIGH STD	02/18/25 16:02		Kareem	OK
62	CCV05	CCV05	CCV	02/18/25 16:06		Kareem	OK
63	CCB05	CCB05	CCB	02/18/25 16:11		Kareem	OK