

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Chemtech Consulting Group

SDG No.: Q1172

Instrument: CV1

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166684BL		WATER		Batch Number:	PB166684		Prep Date:	02/11/2025	
	Mercury	0.20	<0.20	U	0.20	CV	02/11/2025	15:35	LB134671

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Chemtech Consulting Group

SDG No.: Q1172

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB166688BL	WATER			Batch Number:	PB166688		Prep Date:	02/12/2025	
	Aluminum	24.4	<100	J	100	P	02/12/2025	18:16	LB134704
	Antimony	50.0	<50.0	U	50.0	P	02/12/2025	18:16	LB134704
	Arsenic	20.0	<20.0	U	20.0	P	02/12/2025	18:16	LB134704
	Barium	100	<100	U	100	P	02/12/2025	18:16	LB134704
	Beryllium	6.00	<6.00	U	6.00	P	02/12/2025	18:16	LB134704
	Boron	100	<100	U	100	P	02/12/2025	18:16	LB134704
	Cadmium	6.00	<6.00	U	6.00	P	02/12/2025	18:16	LB134704
	Calcium	2000	<2000	U	2000	P	02/12/2025	18:16	LB134704
	Chromium	10.0	<10.0	U	10.0	P	02/12/2025	18:16	LB134704
	Copper	20.0	<20.0	U	20.0	P	02/12/2025	18:16	LB134704
	Iron	35.2	<100	J	100	P	02/12/2025	18:16	LB134704
	Lead	12.0	<12.0	U	12.0	P	02/12/2025	18:16	LB134704
	Magnesium	2000	<2000	U	2000	P	02/12/2025	18:16	LB134704
	Manganese	3.12	<20.0	J	20.0	P	02/12/2025	18:16	LB134704
	Molybdenum	200	<200	U	200	P	02/12/2025	18:16	LB134704
	Nickel	40.0	<40.0	U	40.0	P	02/12/2025	18:16	LB134704
	Potassium	2000	<2000	U	2000	P	02/12/2025	18:16	LB134704
	Selenium	20.0	<20.0	U	20.0	P	02/12/2025	18:16	LB134704
	Silver	10.0	<10.0	U	10.0	P	02/12/2025	18:16	LB134704
	Sodium	2000	<2000	U	2000	P	02/12/2025	18:16	LB134704
	Thallium	40.0	<40.0	U	40.0	P	02/12/2025	18:16	LB134704
	Vanadium	40.0	<40.0	U	40.0	P	02/12/2025	18:16	LB134704
	Zinc	3.99	<40.0	J	40.0	P	02/12/2025	18:16	LB134704