

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Chemtech Consulting Group **SDG No.:** Q1172
Contract: CHEM02 **Lab Code:** CHEM **Case No.:** Q1172 **SAS No.:** Q1172

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166684BS Mercury	ug/L	4.0	3.75		94	85 - 115	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Chemtech Consulting Group **SDG No.:** Q1172
Contract: CHEM02 **Lab Code:** CHEM **Case No.:** Q1172 **SAS No.:** Q1172

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166688BS							
Aluminum	ug/L	2000	1920		96	85 - 115	P
Antimony	ug/L	800	754		94	85 - 115	P
Arsenic	ug/L	800	731		91	85 - 115	P
Barium	ug/L	200	196		98	85 - 115	P
Beryllium	ug/L	200	186		93	85 - 115	P
Boron	ug/L	300	262		87	85 - 115	P
Cadmium	ug/L	200	188		94	85 - 115	P
Calcium	ug/L	1000	1060	J	106	85 - 115	P
Chromium	ug/L	400	414		104	85 - 115	P
Copper	ug/L	300	300		100	85 - 115	P
Iron	ug/L	3000	3090		103	85 - 115	P
Lead	ug/L	1000	936		94	85 - 115	P
Magnesium	ug/L	2000	1960	J	98	85 - 115	P
Manganese	ug/L	200	203		102	85 - 115	P
Molybdenum	ug/L	400	407		102	85 - 115	P
Nickel	ug/L	500	490		98	85 - 115	P
Potassium	ug/L	10000	9590		96	85 - 115	P
Selenium	ug/L	2000	1760		88	85 - 115	P
Silver	ug/L	75.0	72.9		97	85 - 115	P
Sodium	ug/L	3000	3180		106	85 - 115	P
Thallium	ug/L	2000	1780		89	85 - 115	P
Vanadium	ug/L	300	292		97	85 - 115	P
Zinc	ug/L	200	196		98	85 - 115	P