



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Chemtech Consulting Group SDG No.: Q1172
Contract: CHEM02 Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Aluminum	19.6	20.0	98	50 - 150	P	02/14/2025	11:48	LB134731
	Antimony	1.98	2.0	99	70 - 130	P	02/14/2025	11:48	LB134731
	Arsenic	1.01	1.0	101	70 - 130	P	02/14/2025	11:48	LB134731
	Barium	9.68	10.0	97	70 - 130	P	02/14/2025	11:48	LB134731
	Beryllium	1.08	1.0	108	70 - 130	P	02/14/2025	11:48	LB134731
	Cadmium	1.07	1.0	107	70 - 130	P	02/14/2025	11:48	LB134731
	Chromium	2.10	2.0	105	70 - 130	P	02/14/2025	11:48	LB134731
	Copper	1.46	2.0	73	70 - 130	P	02/14/2025	11:48	LB134731
	Iron	46.4	50.0	93	50 - 150	P	02/14/2025	11:48	LB134731
	Lead	1.10	1.0	110	70 - 130	P	02/14/2025	11:48	LB134731
	Manganese	1.28	1.0	128	50 - 150	P	02/14/2025	11:48	LB134731
	Molybdenum	5.05	5.0	101	70 - 130	P	02/14/2025	11:48	LB134731
	Nickel	0.93	1.0	93	70 - 130	P	02/14/2025	11:48	LB134731
	Selenium	5.32	5.0	106	70 - 130	P	02/14/2025	11:48	LB134731
	Silver	0.97	1.0	97	70 - 130	P	02/14/2025	11:48	LB134731
	Thallium	1.02	1.0	102	70 - 130	P	02/14/2025	11:48	LB134731
	Vanadium	5.14	5.0	103	70 - 130	P	02/14/2025	11:48	LB134731
	Zinc	5.38	5.0	108	50 - 150	P	02/14/2025	11:48	LB134731