



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Chemtech Consulting Group SDG No.: Q1172
Contract: CHEM02 Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	20.0	+/-20.0	U	20.0	P	02/14/2025	11:03	LB134731
	Antimony	2.00	+/-2.00	U	2.00	P	02/14/2025	11:03	LB134731
	Arsenic	1.00	+/-1.00	U	1.00	P	02/14/2025	11:03	LB134731
	Barium	10.0	+/-10.0	U	10.0	P	02/14/2025	11:03	LB134731
	Beryllium	1.00	+/-1.00	U	1.00	P	02/14/2025	11:03	LB134731
	Cadmium	1.00	+/-1.00	U	1.00	P	02/14/2025	11:03	LB134731
	Chromium	2.00	+/-2.00	U	2.00	P	02/14/2025	11:03	LB134731
	Copper	2.00	+/-2.00	U	2.00	P	02/14/2025	11:03	LB134731
	Iron	50.0	+/-50.0	U	50.0	P	02/14/2025	11:03	LB134731
	Lead	0.30	+/-1.00	J	1.00	P	02/14/2025	11:03	LB134731
	Manganese	0.44	+/-1.00	J	1.00	P	02/14/2025	11:03	LB134731
	Nickel	1.00	+/-1.00	U	1.00	P	02/14/2025	11:03	LB134731
	Molybdenum	5.00	+/-5.00	U	5.00	P	02/14/2025	11:03	LB134731
	Selenium	5.00	+/-5.00	U	5.00	P	02/14/2025	11:03	LB134731
	Silver	1.00	+/-1.00	U	1.00	P	02/14/2025	11:03	LB134731
	Thallium	1.00	+/-1.00	U	1.00	P	02/14/2025	11:03	LB134731
	Vanadium	5.00	+/-5.00	U	5.00	P	02/14/2025	11:03	LB134731
	Zinc	5.00	+/-5.00	U	5.00	P	02/14/2025	11:03	LB134731

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Chemtech Consulting Group		SDG No.: Q1172							
Contract: CHEM02	Lab Code: CHEM	Case No.: Q1172	SAS No.: Q1172						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	20.0	+/-20.0	U	20.0	P	02/14/2025	11:35	LB134731
	Antimony	2.00	+/-2.00	U	2.00	P	02/14/2025	11:35	LB134731
	Arsenic	1.00	+/-1.00	U	1.00	P	02/14/2025	11:35	LB134731
	Barium	10.0	+/-10.0	U	10.0	P	02/14/2025	11:35	LB134731
	Beryllium	1.00	+/-1.00	U	1.00	P	02/14/2025	11:35	LB134731
	Cadmium	1.00	+/-1.00	U	1.00	P	02/14/2025	11:35	LB134731
	Chromium	2.00	+/-2.00	U	2.00	P	02/14/2025	11:35	LB134731
	Copper	2.00	+/-2.00	U	2.00	P	02/14/2025	11:35	LB134731
	Iron	50.0	+/-50.0	U	50.0	P	02/14/2025	11:35	LB134731
	Lead	0.30	+/-1.00	J	1.00	P	02/14/2025	11:35	LB134731
	Manganese	0.49	+/-1.00	J	1.00	P	02/14/2025	11:35	LB134731
	Nickel	1.00	+/-1.00	U	1.00	P	02/14/2025	11:35	LB134731
	Molybdenum	5.00	+/-5.00	U	5.00	P	02/14/2025	11:35	LB134731
	Selenium	5.00	+/-5.00	U	5.00	P	02/14/2025	11:35	LB134731
	Silver	1.00	+/-1.00	U	1.00	P	02/14/2025	11:35	LB134731
	Thallium	1.00	+/-1.00	U	1.00	P	02/14/2025	11:35	LB134731
	Vanadium	5.00	+/-5.00	U	5.00	P	02/14/2025	11:35	LB134731
	Zinc	5.00	+/-5.00	U	5.00	P	02/14/2025	11:35	LB134731
CCB02	Aluminum	20.0	+/-20.0	U	20.0	P	02/14/2025	12:32	LB134731
	Antimony	2.00	+/-2.00	U	2.00	P	02/14/2025	12:32	LB134731
	Arsenic	1.00	+/-1.00	U	1.00	P	02/14/2025	12:32	LB134731
	Barium	10.0	+/-10.0	U	10.0	P	02/14/2025	12:32	LB134731
	Beryllium	1.00	+/-1.00	U	1.00	P	02/14/2025	12:32	LB134731
	Cadmium	1.00	+/-1.00	U	1.00	P	02/14/2025	12:32	LB134731
	Chromium	2.00	+/-2.00	U	2.00	P	02/14/2025	12:32	LB134731
	Copper	2.00	+/-2.00	U	2.00	P	02/14/2025	12:32	LB134731
	Iron	50.0	+/-50.0	U	50.0	P	02/14/2025	12:32	LB134731
	Lead	1.00	+/-1.00	U	1.00	P	02/14/2025	12:32	LB134731
	Manganese	0.44	+/-1.00	J	1.00	P	02/14/2025	12:32	LB134731
	Nickel	1.00	+/-1.00	U	1.00	P	02/14/2025	12:32	LB134731
	Molybdenum	5.00	+/-5.00	U	5.00	P	02/14/2025	12:32	LB134731
	Selenium	5.00	+/-5.00	U	5.00	P	02/14/2025	12:32	LB134731
	Silver	1.00	+/-1.00	U	1.00	P	02/14/2025	12:32	LB134731
	Thallium	1.00	+/-1.00	U	1.00	P	02/14/2025	12:32	LB134731
	Vanadium	5.00	+/-5.00	U	5.00	P	02/14/2025	12:32	LB134731
	Zinc	5.00	+/-5.00	U	5.00	P	02/14/2025	12:32	LB134731