

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Chemtech Consulting Group **SDG No.:** Q1172
Contract: CHEM02 **Lab Code:** CHEM **Case No.:** Q1172 **SAS No.:** Q1172

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166661BS							
Aluminum	ug/L	10000	10100		101	85 - 115	P
Antimony	ug/L	500	485		97	85 - 115	P
Arsenic	ug/L	500	486		97	85 - 115	P
Barium	ug/L	2500	2390		96	85 - 115	P
Beryllium	ug/L	500	501		100	85 - 115	P
Cadmium	ug/L	500	474		95	85 - 115	P
Chromium	ug/L	500	497		99	85 - 115	P
Copper	ug/L	5000	4860		97	85 - 115	P
Iron	ug/L	25000	25500		102	85 - 115	P
Lead	ug/L	2500	2490		100	85 - 115	P
Manganese	ug/L	5000	4970		99	85 - 115	P
Nickel	ug/L	500	477		95	85 - 115	P
Selenium	ug/L	500	505		101	85 - 115	P
Molybdenum	ug/L	5000	4980		100	85 - 115	P
Silver	ug/L	500	486		97	85 - 115	P
Thallium	ug/L	500	497		99	85 - 115	P
Vanadium	ug/L	500	499		100	85 - 115	P
Zinc	ug/L	5000	5030		101	85 - 115	P