

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** Q1172 **SAS No.:** Q1172 **SDG NO.:** Q1172

GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 01/30/2025 01/30/2025

Client Sample No.: CCAL02 **Date Analyzed:** 01/30/2025

Lab Sample No.: M8011.504.1 0.1 PF **Data File :** PQ069904.D **Time Analyzed:** 13:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	5.971	5.870	6.070	0.100	0.100	0.0
EDB	2.618	2.516	2.716	0.100	0.100	0.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** Q1172 **SAS No.:** Q1172 **SDG NO.:** Q1172

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 01/30/2025 01/30/2025

Client Sample No.: CCAL02 **Date Analyzed:** 01/30/2025

Lab Sample No.: M8011.504.1 0.1 PF **Data File :** PQ069904.D **Time Analyzed:** 13:23

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	4.995	4.895	5.095	0.100	0.100	0.0
EDB	1.980	1.878	2.078	0.100	0.100	0.0