284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Cover Page

Order ID : Q1172

Project ID : NJ Drinking Water PT

Client : Chemtech Consulting Group

Lab Sample Number	Client Sample Number
Q1172-01	PT-TURB-WS
Q1172-02	PT-TURB-WS
Q1172-03	PT-MIN-WS
Q1172-04	PT-TM-WS
Q1172-05	PT-HG-WS
Q1172-06	PT-SIO2-WS
Q1172-07	PT-RVOA-WS
Q1172-08	PT-UNRVOA-WS
Q1172-09	PT-THM-WS
Q1172-10	PT-ADD-WS
Q1172-11	PT-EDBCP-WS

I certify that the data package is in compliance with the terms and conditions of the contract,both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 3/1/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

284 Sheffield Street, Mountainside NJ 07092 Tel. 908-789-8900 Fax: 908-789-8922

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following " Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value**.** This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This is flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: **Q1172**

	Completed
For thorough review, the report must have the following:	
GENERAL:	
Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)	✓
Check chain-of-custody for proper relinquish/return of samples	✓
Is the chain of custody signed and complete	✓
Check internal chain-of-custody for proper relinquish/return of samples /sample extracts	✓
Collect information for each project id from server. Were all requirements followed	✓
COVER PAGE:	
Do numbers of samples correspond to the number of samples in the Chain of Custody on login page	✓
Do lab numbers and client Ids on cover page agree with the Chain of Custody	✓
CHAIN OF CUSTODY:	
Do requested analyses on Chain of Custody agree with form I results	✓
Do requested analyses on Chain of Custody agree with the log-in page	✓
Were the correct method log-in for analysis according to the Analytical Request and Chain of Castody	✓
Were the samples received within hold time	✓
Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle	✓
ANALYTICAL:	
Was method requirement followed?	✓
Was client requirement followed?	✓
Does the case narrative summarize all QC failure?	✓
All runlogs and manual integration are reviewed for requirements	✓
All manual calculations and /or hand notations verified	✓

QA Review Signature: **KETAN PATEL** **Date:** **03/01/2025**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

LAB CHRONICLE

OrderID:	Q1172	**OrderDate:**	1/23/2025 2:53:00 PM
Client:	Chemtech Consulting Group	**Project:**	NJ Drinking Water PT
Contact:	QA Officer	**Location:**	QA Office,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1172-07	**PT-RVOA-WS**	**Water**			**01/13/25**			**01/15/25**
			VOCMS Group1	524.2			02/11/25	
Q1172-07DL	**PT-RVOA-WSDL**	**Water**			**01/13/25**			**01/15/25**
			VOCMS Group1	524.2			02/11/25	
Q1172-08	**PT-UNRVOA-WS**	**Water**			**01/13/25**			**01/15/25**
			VOCMS Group2	524.2			02/11/25	
Q1172-08DL	**PT-UNRVOA-WSDL**	**Water**			**01/13/25**			**01/15/25**
			VOCMS Group2	524.2			02/13/25	
Q1172-09	**PT-THM-WS**	**Water**			**01/13/25**			**01/15/25**
			VOCMS Group3				02/11/25	
			VOCMS Group4	524.2			02/11/25	
Q1172-09DL	**PT-THM-WSDL**	**Water**			**01/13/25**			**01/15/25**
			VOCMS Group3				02/11/25	
			VOCMS Group4	524.2			02/11/25	
Q1172-10	**PT-ADD-WS**	**Water**			**01/13/25**			**01/15/25**
			VOCMS Group3	524.2			02/12/25	
Q1172-10DL	**PT-ADD-WSDL**	**Water**			**01/13/25**			**01/15/25**
			VOCMS Group3	524.2			02/12/25	
Q1172-11	**PT-EDBCP-WS**	**Water**			**01/13/25**			**01/15/25**
			VOCMS Group5	524.2			02/11/25	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1172

Client: Chemtech Consulting Group

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	**PT-ADD-WS**							
Q1172-10	PT-ADD-WS	Water	1,1,2-Trichloro-1,2,2-trifluoroe	19.8	E	0.14	0.50	ug/L
Q1172-10	PT-ADD-WS	Water	tert-Butyl Alcohol	32.7		8.60	10.0	ug/L
Q1172-10	PT-ADD-WS	Water	Methyl tert-Butyl Ether	7.60		0.12	0.50	ug/L
Q1172-10	PT-ADD-WS	Water	n-propylbenzene	11.6		0.16	0.50	ug/L
Q1172-10	PT-ADD-WS	Water	Isopropyl Ether	10.7		0.12	0.50	ug/L
			Total Voc :	82.4				
			Total Concentration:	82.4				
Client ID:	**PT-ADD-WSDL**							
Q1172-10DL	PT-ADD-WSDL	Water	1,1,2-Trichloro-1,2,2-trifluoroe	20.4	D	0.70	2.50	ug/L
Q1172-10DL	PT-ADD-WSDL	Water	Methyl tert-Butyl Ether	7.30	D	0.60	2.50	ug/L
Q1172-10DL	PT-ADD-WSDL	Water	n-propylbenzene	9.60	D	0.80	2.50	ug/L
Q1172-10DL	PT-ADD-WSDL	Water	Isopropyl Ether	10.2	D	0.60	2.50	ug/L
			Total Voc :	47.5				
			Total Concentration:	47.5				

QC SUMMARY

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Surrogate Summary

SDG No.: Q1172

Client: Chemtech Consulting Group

Analytical Method: SW524.2

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits Low	Limits High
Q1172-10	PT-ADD-WS	1,2-Dichlorobenzene-d4	1	0.77	77	70	130
		4-Bromofluorobenzene	1	0.73	73	70	130
Q1172-10DL	PT-ADD-WSDL	1,2-Dichlorobenzene-d4	1	0.82	82	70	130
		4-Bromofluorobenzene	1	0.77	77	70	130
VU0212WBL01	VU0212WBL01	1,2-Dichlorobenzene-d4	1	0.78	78	70	130
		4-Bromofluorobenzene	1	0.82	82	70	130
VU0212WBS01	VU0212WBS01	1,2-Dichlorobenzene-d4	1	0.98	98	70	130
		4-Bromofluorobenzene	1	1.03	103	70	130
VU0212WBSD0	VU0212WBSD01	1,2-Dichlorobenzene-d4	1	0.96	96	70	130
		4-Bromofluorobenzene	1	1.02	103	70	130

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1172

Client: Chemtech Consulting Group

Analytical Method: SW524.2 **Datafile :** VU063251.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits Low	Limits High	RPD
VU0212WBS01	Dichlorodifluoromethane	2	2.00	ug/L	100			70	130	
	Chloromethane	2	2.00	ug/L	100			70	130	
	Vinyl Chloride	2	2.00	ug/L	100			70	130	
	Bromomethane	2	2.20	ug/L	110			70	130	
	Chloroethane	2	2.00	ug/L	100			70	130	
	Tetrahydrofuran	4	3.80	ug/L	95			70	130	
	Trichlorofluoromethane	2	2.00	ug/L	100			70	130	
	1,1,2-Trichloro-1,2,2-trifluoroethane	2	2.00	ug/L	100			70	130	
	tert-Butyl Alcohol	20	21.2	ug/L	106			70	130	
	Diethyl Ether	2	1.90	ug/L	95			70	130	
	1,1-Dichloroethene	2	2.00	ug/L	100			70	130	
	Acrylonitrile	4	4.10	ug/L	103			70	130	
	Acetone	10	10.3	ug/L	103			70	130	
	Carbon disulfide	2	1.90	ug/L	95			70	130	
	Methyl tert-butyl Ether	2	1.90	ug/L	95			70	130	
	Methyl acrylate	2	2.00	ug/L	100			70	130	
	Methylene Chloride	2	2.00	ug/L	100			70	130	
	trans-1,2-Dichloroethene	2	2.00	ug/L	100			70	130	
	1,1-Dichloroethane	2	2.00	ug/L	100			70	130	
	Cyclohexane	2	1.90	ug/L	95			70	130	
	2-Butanone	10	8.90	ug/L	89			70	130	
	Carbon Tetrachloride	2	1.90	ug/L	95			70	130	
	2,2-Dichloropropane	2	2.00	ug/L	100			70	130	
	cis-1,2-Dichloroethene	2	2.00	ug/L	100			70	130	
	Bromochloromethane	2	2.00	ug/L	100			70	130	
	Chloroform	2	2.00	ug/L	100			70	130	
	1,1,1-Trichloroethane	2	2.00	ug/L	100			70	130	
	Methylcyclohexane	2	2.00	ug/L	100			70	130	
	1,1-Dichloropropene	2	2.00	ug/L	100			70	130	
	Propionitrile	10	10.4	ug/L	104			70	130	
	Benzene	2	2.00	ug/L	100			70	130	
	1,2-Dichloroethane	2	2.00	ug/L	100			70	130	
	Trichloroethene	2	2.00	ug/L	100			70	130	
	1,2-Dichloropropane	2	2.00	ug/L	100			70	130	
	1-Chlorobutane	2	1.90	ug/L	95			70	130	
	Dibromomethane	2	2.00	ug/L	100			70	130	
	Bromodichloromethane	2	2.00	ug/L	100			70	130	
	4-Methyl-2-Pentanone	10	9.50	ug/L	95			70	130	
	Toluene	2	2.00	ug/L	100			70	130	
	t-1,3-Dichloropropene	2	2.00	ug/L	100			70	130	
	cis-1,3-Dichloropropene	2	2.00	ug/L	100			70	130	
	1,1,2-Trichloroethane	2	2.00	ug/L	100			70	130	
	1,3-Dichloropropane	2	2.00	ug/L	100			70	130	
	2-Hexanone	10	9.00	ug/L	90			70	130	
	Dibromochloromethane	2	2.00	ug/L	100			70	130	
	1,2-Dibromoethane	2	1.90	ug/L	95			70	130	
	Tetrachloroethene	2	2.10	ug/L	105			70	130	
	Chlorobenzene	2	2.00	ug/L	100			70	130	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1172

Client: Chemtech Consulting Group

Analytical Method: SW524.2 **Datafile :** VU063251.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits Low	Limits High	RPD
VU0212WBS01	1,1,1,2-Tetrachloroethane	2	2.00	ug/L	100			70	130	
	Hexachloroethane	2	1.90	ug/L	95			70	130	
	Ethyl Benzene	2	1.90	ug/L	95			70	130	
	m/p-Xylenes	4	3.90	ug/L	98			70	130	
	o-Xylene	2	2.00	ug/L	100			70	130	
	Styrene	2	1.90	ug/L	95			70	130	
	Bromoform	2	1.90	ug/L	95			70	130	
	Isopropylbenzene	2	2.00	ug/L	100			70	130	
	1,1,2,2-Tetrachloroethane	2	2.00	ug/L	100			70	130	
	1,2,3-Trichloropropane	2	2.00	ug/L	100			70	130	
	Bromobenzene	2	2.00	ug/L	100			70	130	
	N-propylbenzene	2	1.90	ug/L	95			70	130	
	2-Chlorotoluene	2	2.00	ug/L	100			70	130	
	1,3,5-Trimethylbenzene	2	2.00	ug/L	100			70	130	
	4-Chlorotoluene	2	2.00	ug/L	100			70	130	
	tert-Butylbenzene	2	2.00	ug/L	100			70	130	
	1,2,4-Trimethylbenzene	2	1.90	ug/L	95			70	130	
	Sec-butylbenzene	2	2.00	ug/L	100			70	130	
	p-Isopropyltoluene	2	2.00	ug/L	100			70	130	
	1,3-Dichlorobenzene	2	2.00	ug/L	100			70	130	
	1,4-Dichlorobenzene	2	2.00	ug/L	100			70	130	
	n-Butylbenzene	2	1.90	ug/L	95			70	130	
	1,2-Dichlorobenzene	2	2.00	ug/L	100			70	130	
	1,2-Dibromo-3-Chloropropane	2	1.70	ug/L	85			70	130	
	1,2,4-Trichlorobenzene	2	1.80	ug/L	90			70	130	
	Hexachlorobutadiene	2	2.10	ug/L	105			70	130	
	Naphthalene	2	1.70	ug/L	85			70	130	
	1,2,3-Trichlorobenzene	2	1.60	ug/L	80			70	130	
	Iodomethane	2	2.00	ug/L	100			70	130	
	Allyl Chloride	2	2.00	ug/L	100			70	130	
	Methacrylonitrile	2	2.00	ug/L	100			70	130	
	Isopropyl Ether	2	1.90	ug/L	95			70	130	
	Methyl methacrylate	4	3.90	ug/L	98			70	130	

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1172

Client: Chemtech Consulting Group

Analytical Method: SW524.2 **Datafile :** VU063252.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits Low	Limits High	RPD
VU0212WBSD01	Dichlorodifluoromethane	2	2.10	ug/L	105	5		70	130	20
	Chloromethane	2	2.10	ug/L	105	5		70	130	20
	Vinyl Chloride	2	2.10	ug/L	105	5		70	130	20
	Bromomethane	2	2.30	ug/L	115	4		70	130	20
	Chloroethane	2	2.10	ug/L	105	5		70	130	20
	Tetrahydrofuran	4	3.80	ug/L	95	0		70	130	20
	Trichlorofluoromethane	2	2.10	ug/L	105	5		70	130	20
	1,1,2-Trichloro-1,2,2-trifluoroethane	2	2.10	ug/L	105	5		70	130	20
	tert-Butyl Alcohol	20	15.6	ug/L	78	30	*	70	130	20
	Diethyl Ether	2	2.10	ug/L	105	10		70	130	20
	1,1-Dichloroethene	2	2.10	ug/L	105	5		70	130	20
	Acrylonitrile	4	4.30	ug/L	108	5		70	130	20
	Acetone	10	10.0	ug/L	100	3		70	130	20
	Carbon disulfide	2	2.00	ug/L	100	5		70	130	20
	Methyl tert-butyl Ether	2	2.00	ug/L	100	5		70	130	20
	Methyl acrylate	2	2.00	ug/L	100	0		70	130	20
	Methylene Chloride	2	2.10	ug/L	105	5		70	130	20
	trans-1,2-Dichloroethene	2	2.00	ug/L	100	0		70	130	20
	1,1-Dichloroethane	2	2.10	ug/L	105	5		70	130	20
	Cyclohexane	2	2.00	ug/L	100	5		70	130	20
	2-Butanone	10	9.20	ug/L	92	3		70	130	20
	Carbon Tetrachloride	2	2.10	ug/L	105	10		70	130	20
	2,2-Dichloropropane	2	2.10	ug/L	105	5		70	130	20
	cis-1,2-Dichloroethene	2	2.10	ug/L	105	5		70	130	20
	Bromochloromethane	2	2.10	ug/L	105	5		70	130	20
	Chloroform	2	2.10	ug/L	105	5		70	130	20
	1,1,1-Trichloroethane	2	2.10	ug/L	105	5		70	130	20
	Methylcyclohexane	2	2.00	ug/L	100	0		70	130	20
	1,1-Dichloropropene	2	2.10	ug/L	105	5		70	130	20
	Propionitrile	10	9.80	ug/L	98	6		70	130	20
	Benzene	2	2.10	ug/L	105	5		70	130	20
	1,2-Dichloroethane	2	2.10	ug/L	105	5		70	130	20
	Trichloroethene	2	2.10	ug/L	105	5		70	130	20
	1,2-Dichloropropane	2	2.10	ug/L	105	5		70	130	20
	1-Chlorobutane	2	2.00	ug/L	100	5		70	130	20
	Dibromomethane	2	2.10	ug/L	105	5		70	130	20
	Bromodichloromethane	2	2.10	ug/L	105	5		70	130	20
	4-Methyl-2-Pentanone	10	9.70	ug/L	97	2		70	130	20
	Toluene	2	2.10	ug/L	105	5		70	130	20
	t-1,3-Dichloropropene	2	2.10	ug/L	105	5		70	130	20
	cis-1,3-Dichloropropene	2	2.00	ug/L	100	0		70	130	20
	1,1,2-Trichloroethane	2	2.10	ug/L	105	5		70	130	20
	1,3-Dichloropropane	2	2.10	ug/L	105	5		70	130	20
	2-Hexanone	10	9.10	ug/L	91	1		70	130	20
	Dibromochloromethane	2	2.10	ug/L	105	5		70	130	20
	1,2-Dibromoethane	2	2.10	ug/L	105	10		70	130	20
	Tetrachloroethene	2	2.10	ug/L	105	0		70	130	20
	Chlorobenzene	2	2.10	ug/L	105	5		70	130	20

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1172

Client: Chemtech Consulting Group

Analytical Method: SW524.2 **Datafile :** VU063252.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Limits Low	Limits High	RPD
VU0212WBSD01	1,1,1,2-Tetrachloroethane	2	2.10	ug/L	105	5		70	130	20
	Hexachloroethane	2	2.00	ug/L	100	5		70	130	20
	Ethyl Benzene	2	2.00	ug/L	100	5		70	130	20
	m/p-Xylenes	4	4.10	ug/L	103	5		70	130	20
	o-Xylene	2	2.10	ug/L	105	5		70	130	20
	Styrene	2	2.00	ug/L	100	5		70	130	20
	Bromoform	2	2.00	ug/L	100	5		70	130	20
	Isopropylbenzene	2	2.00	ug/L	100	0		70	130	20
	1,1,2,2-Tetrachloroethane	2	2.00	ug/L	100	0		70	130	20
	1,2,3-Trichloropropane	2	1.80	ug/L	90	11		70	130	20
	Bromobenzene	2	2.10	ug/L	105	5		70	130	20
	N-propylbenzene	2	1.90	ug/L	95	0		70	130	20
	2-Chlorotoluene	2	2.00	ug/L	100	0		70	130	20
	1,3,5-Trimethylbenzene	2	2.00	ug/L	100	0		70	130	20
	4-Chlorotoluene	2	2.10	ug/L	105	5		70	130	20
	tert-Butylbenzene	2	2.00	ug/L	100	0		70	130	20
	1,2,4-Trimethylbenzene	2	2.00	ug/L	100	5		70	130	20
	Sec-butylbenzene	2	2.10	ug/L	105	5		70	130	20
	p-Isopropyltoluene	2	2.00	ug/L	100	0		70	130	20
	1,3-Dichlorobenzene	2	2.10	ug/L	105	5		70	130	20
	1,4-Dichlorobenzene	2	2.10	ug/L	105	5		70	130	20
	n-Butylbenzene	2	1.90	ug/L	95	0		70	130	20
	1,2-Dichlorobenzene	2	2.00	ug/L	100	0		70	130	20
	1,2-Dibromo-3-Chloropropane	2	2.00	ug/L	100	16		70	130	20
	1,2,4-Trichlorobenzene	2	1.90	ug/L	95	5		70	130	20
	Hexachlorobutadiene	2	2.10	ug/L	105	0		70	130	20
	Naphthalene	2	2.00	ug/L	100	16		70	130	20
	1,2,3-Trichlorobenzene	2	1.90	ug/L	95	17		70	130	20
	Iodomethane	2	2.20	ug/L	110	10		70	130	20
	Allyl Chloride	2	2.00	ug/L	100	0		70	130	20
	Methacrylonitrile	2	1.90	ug/L	95	5		70	130	20
	Isopropyl Ether	2	2.00	ug/L	100	5		70	130	20
	Methyl methacrylate	4	4.10	ug/L	103	5		70	130	20

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VU0212WBL01

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172 SDG NO.: Q1172

Lab File ID: VU063250.D Lab Sample ID: VU0212WBL01

Date Analyzed: 02/12/2025 Time Analyzed: 12:17

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

Instrument ID: MSVOA_U

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VU0212WBS01	VU0212WBS01	VU063251.D	02/12/2025
VU0212WBSD01	VU0212WBSD01	VU063252.D	02/12/2025
PT-ADD-WS	Q1172-10	VU063253.D	02/12/2025
PT-ADD-WSDL	Q1172-10DL	VU063254.D	02/12/2025

COMMENTS:

Raw Data: **VU063218.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172 SDG NO.: Q1172

Lab File ID: VU063218.D BFB Injection Date: 02/10/2025

Instrument ID: MSVOA_U BFB Injection Time: 09:06

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.5
75	30.0 - 60.0% of mass 95	52.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	0.4 (0.5) 1
174	50.0 - 100.0% of mass 95	79.1
175	5.0 - 9.0% of mass 174	5.8 (7.3) 1
176	95.0 - 101.0% of mass 174	77.5 (98.1) 1
177	5.0 - 9.0% of mass 176	4.8 (6.2) 2

1-Value is % mass 174 2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDICC0.5	VSTDICC0.5	VU063219.D	02/10/2025	12:59
VSTDICC001	VSTDICC001	VU063220.D	02/10/2025	13:23
VSTDICC002	VSTDICC002	VU063221.D	02/10/2025	13:58
VSTDICC005	VSTDICC005	VU063222.D	02/10/2025	14:23
VSTDICCC010	VSTDICCC010	VU063223.D	02/10/2025	15:06
VSTDICC015	VSTDICC015	VU063224.D	02/10/2025	15:33

Raw Data: **VU063248.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172 SDG NO.: Q1172

Lab File ID: VU063248.D BFB Injection Date: 02/12/2025

Instrument ID: MSVOA_U BFB Injection Time: 10:13

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.3
75	30.0 - 60.0% of mass 95	51.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	1 (1.2) 1
174	50.0 - 100.0% of mass 95	78.7
175	5.0 - 9.0% of mass 174	5.9 (7.5) 1
176	95.0 - 101.0% of mass 174	75.7 (96.2) 1
177	5.0 - 9.0% of mass 176	5 (6.6) 2

1-Value is % mass 174 2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC010	VSTDCCC010	VU063249.D	02/12/2025	10:48
VU0212WBL01	VU0212WBL01	VU063250.D	02/12/2025	12:17
VU0212WBS01	VU0212WBS01	VU063251.D	02/12/2025	12:56
VU0212WBSD01	VU0212WBSD01	VU063252.D	02/12/2025	13:20
PT-ADD-WS	Q1172-10	VU063253.D	02/12/2025	14:20
PT-ADD-WSDL	Q1172-10DL	VU063254.D	02/12/2025	14:58

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172 SDG NO.: Q1172

Lab File ID: VU063249.D Date Analyzed: 02/12/2025

Instrument ID: MSVOA_U Time Analyzed: 10:48

GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	54296	6.11	0	0.00	0	0.00
UPPER LIMIT	70584.8	6.607	0		0	
LOWER LIMIT	38007.2	5.607	0		0	
EPA SAMPLE NO.						
PT-ADD-WS	52826	6.10	0	0.00	0	0.00
PT-ADD-WSDL	48656	6.11	0	0.00	0	0.00
VU0212WBL01	47954	6.11	0	0.00	0	0.00
VU0212WBS01	51131	6.10	0	0.00	0	0.00
VU0212WBSD01	51180	6.11	0	0.00	0	0.00

IS1 = Fluorobenzene
IS2 =
IS3 =

AREA UPPER LIMIT = +30% of internal standard area
AREA LOWER LIMIT = -30% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

SAMPLE DATA

Raw Data: **VU063253.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	01/13/25
Project:	NJ Drinking Water PT	Date Received:	01/15/25
Client Sample ID:	PT-ADD-WS	SDG No.:	Q1172
Lab Sample ID:	Q1172-10	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063253.D	1		02/12/25 14:20	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.14	U	0.14	0.50	ug/L
74-87-3	Chloromethane	0.13	U	0.13	0.50	ug/L
75-01-4	Vinyl Chloride	0.13	U	0.13	0.50	ug/L
74-83-9	Bromomethane	0.18	U	0.18	0.50	ug/L
75-00-3	Chloroethane	0.14	U	0.14	0.50	ug/L
109-99-9	Tetrahydrofuran	0.44	U	0.44	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.21	U	0.21	0.50	ug/L
76-13-1	1,1,2-Trichloro-1,2,2-trifluoroethane	19.8	E	0.14	0.50	ug/L
75-65-0	tert-Butyl Alcohol	32.7		8.60	10.0	ug/L
60-29-7	Diethyl Ether	0.13	U	0.13	0.50	ug/L
75-35-4	1,1-Dichloroethene	0.12	U	0.12	0.50	ug/L
107-13-1	Acrylonitrile	0.44	U	0.44	1.00	ug/L
67-64-1	Acetone	1.10	U	1.10	2.50	ug/L
75-15-0	Carbon Disulfide	0.13	U	0.13	0.50	ug/L
1634-04-4	Methyl tert-Butyl Ether	7.60		0.12	0.50	ug/L
96-33-3	Methyl acrylate	0.28	U	0.28	0.50	ug/L
75-09-2	Methylene Chloride	0.47	U	0.47	0.50	ug/L
156-60-5	trans-1,2-Dichloroethene	0.14	U	0.14	0.50	ug/L
75-34-3	1,1-Dichloroethane	0.13	U	0.13	0.50	ug/L
110-82-7	Cyclohexane	0.14	U	0.14	0.50	ug/L
78-93-3	2-Butanone	0.68	U	0.68	2.50	ug/L
56-23-5	Carbon Tetrachloride	0.14	U	0.14	0.50	ug/L
594-20-7	2,2-Dichloropropane	0.14	U	0.14	0.50	ug/L
156-59-2	cis-1,2-Dichloroethene	0.13	U	0.13	0.50	ug/L
74-97-5	Bromochloromethane	0.16	U	0.16	0.50	ug/L
67-66-3	Chloroform	0.13	U	0.13	0.50	ug/L
71-55-6	1,1,1-Trichloroethane	0.12	U	0.12	0.50	ug/L
108-87-2	Methylcyclohexane	0.12	U	0.12	0.50	ug/L
563-58-6	1,1-Dichloropropene	0.11	U	0.11	0.50	ug/L
107-12-0	Propionitrile	1.00	U	1.00	2.50	ug/L

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	01/13/25
Project:	NJ Drinking Water PT	Date Received:	01/15/25
Client Sample ID:	PT-ADD-WS	SDG No.:	Q1172
Lab Sample ID:	Q1172-10	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063253.D	1		02/12/25 14:20	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
71-43-2	Benzene	0.11	U	0.11	0.50	ug/L
107-06-2	1,2-Dichloroethane	0.16	U	0.16	0.50	ug/L
79-01-6	Trichloroethene	0.13	U	0.13	0.50	ug/L
78-87-5	1,2-Dichloropropane	0.13	U	0.13	0.50	ug/L
109-69-3	1-Chlorobutane	0.12	U	0.12	0.50	ug/L
74-95-3	Dibromomethane	0.14	U	0.14	0.50	ug/L
75-27-4	Bromodichloromethane	0.12	U	0.12	0.50	ug/L
108-10-1	4-Methyl-2-Pentanone	0.60	U	0.60	2.50	ug/L
108-88-3	Toluene	0.11	U	0.11	0.50	ug/L
10061-02-6	t-1,3-Dichloropropene	0.11	U	0.11	0.50	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.11	U	0.11	0.50	ug/L
79-00-5	1,1,2-Trichloroethane	0.13	U	0.13	0.50	ug/L
142-28-9	1,3-Dichloropropane	0.13	U	0.13	0.50	ug/L
591-78-6	2-Hexanone	0.57	U	0.57	2.50	ug/L
124-48-1	Dibromochloromethane	0.13	U	0.13	0.50	ug/L
106-93-4	1,2-Dibromoethane	0.13	U	0.13	0.50	ug/L
127-18-4	Tetrachloroethene	0.14	U	0.14	0.50	ug/L
108-90-7	Chlorobenzene	0.11	U	0.11	0.50	ug/L
630-20-6	1,1,1,2-Tetrachloroethane	0.13	U	0.13	0.50	ug/L
67-72-1	Hexachloroethane	0.12	U	0.12	0.50	ug/L
100-41-4	Ethyl Benzene	0.12	U	0.12	0.50	ug/L
179601-23-1	m/p-Xylenes	0.23	U	0.23	1.00	ug/L
1330-20-7	Total Xylenes	0.35	U	0.35	1.50	ug/L
95-47-6	o-Xylene	0.12	U	0.12	0.50	ug/L
100-42-5	Styrene	0.13	U	0.13	0.50	ug/L
75-25-2	Bromoform	0.14	U	0.14	0.50	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	0.50	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.12	U	0.12	0.50	ug/L
96-18-4	1,2,3-Trichloropropane	0.21	U	0.21	0.50	ug/L
108-86-1	Bromobenzene	0.13	U	0.13	0.50	ug/L
103-65-1	n-propylbenzene	11.6		0.16	0.50	ug/L

Raw Data: **VU063253.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	01/13/25
Project:	NJ Drinking Water PT	Date Received:	01/15/25
Client Sample ID:	PT-ADD-WS	SDG No.:	Q1172
Lab Sample ID:	Q1172-10	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063253.D	1		02/12/25 14:20	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
95-49-8	2-Chlorotoluene	0.14	U	0.14	0.50	ug/L
108-67-8	1,3,5-Trimethylbenzene	0.13	U	0.13	0.50	ug/L
106-43-4	4-Chlorotoluene	0.14	U	0.14	0.50	ug/L
98-06-6	tert-Butylbenzene	0.11	U	0.11	0.50	ug/L
95-63-6	1,2,4-Trimethylbenzene	0.13	U	0.13	0.50	ug/L
135-98-8	sec-Butylbenzene	0.13	U	0.13	0.50	ug/L
99-87-6	p-Isopropyltoluene	0.16	U	0.16	0.50	ug/L
541-73-1	1,3-Dichlorobenzene	0.13	U	0.13	0.50	ug/L
106-46-7	1,4-Dichlorobenzene	0.14	U	0.14	0.50	ug/L
104-51-8	n-Butylbenzene	0.28	U	0.28	0.50	ug/L
95-50-1	1,2-Dichlorobenzene	0.14	U	0.14	0.50	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.23	U	0.23	0.50	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.21	U	0.21	0.50	ug/L
87-68-3	Hexachlorobutadiene	0.14	U	0.14	0.50	ug/L
91-20-3	Naphthalene	0.31	U	0.31	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.25	U	0.25	0.50	ug/L
74-88-4	Iodomethane	0.16	U	0.16	1.00	ug/L
107-05-1	Allyl Chloride	0.11	U	0.11	0.50	ug/L
126-98-7	Methacrylonitrile	0.19	U	0.19	0.50	ug/L
108-20-3	Isopropyl Ether	10.7		0.12	0.50	ug/L
80-62-6	Methyl methacrylate	0.24	U	0.24	1.00	ug/L
SURROGATES						
2199-69-1	1,2-Dichlorobenzene-d4	0.77		70 - 130	77%	SPK: 1
460-00-4	4-Bromofluorobenzene	0.73		70 - 130	73%	SPK: 1
INTERNAL STANDARDS						
462-06-6	Fluorobenzene	52800	6.103			
	Tert-Amyl methyl ether					

Raw Data: **VU063253.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	01/13/25
Project:	NJ Drinking Water PT	Date Received:	01/15/25
Client Sample ID:	PT-ADD-WS	SDG No.:	Q1172
Lab Sample ID:	Q1172-10	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063253.D	1		02/12/25 14:20	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063253.D
Acq On : 12 Feb 2025 14:20
Operator : MD/SY
Sample : Q1172-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Quant Time: Feb 13 03:12:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Fluorobenzene	6.103	96	52826	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.627	95	12765	0.733	ug/l	0.00
Spiked Amount 1.000			Recovery = 73.000%			
68) 1,2-Dichlorobenzene-d4	12.190	152	13885	0.766	ug/l	0.00
Spiked Amount 1.000			Recovery = 77.000%			
Target Compounds						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.570	101	259834	19.753	ug/l	100
24) tert-Butyl Alcohol	3.161	59	43566	32.713	ug/l	98
25) Methyl tert-Butyl Ether	3.348	73	253519	7.574	ug/l	99
31) Isopropyl Ether	3.975	45	442152	10.742	ug/l	100
32) Ethyl-t-butyl ether	4.483	59	1993904	53.267	ug/l	99
33) Tert-Amyl methyl ether	5.926	73	1646111	50.332	ug/l	100
73) n-propylbenzene	10.897	120	194126	11.610	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063253.D
Acq On : 12 Feb 2025 14:20
Operator : MD/SY
Sample : Q1172-10
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Quant Time: Feb 13 03:12:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.103 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063253.D
Acq: 12 Feb 2025 14:20

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Tgt Ion: 96 Resp: 52826

Ion	Ratio	Lower	Upper
96	100		
70	20.3	15.6	23.4
95	9.0	7.3	10.9
97	6.1	0.0	0.0#





#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 19.753 ug/l
RT: 2.570 min Scan# 398
Delta R.T. -0.000 min
Lab File: VU063253.D
Acq: 12 Feb 2025 14:20

Tgt Ion:101 Resp: 259834

Ion	Ratio	Lower	Upper
101	100		
85	44.5	35.4	53.0
151	73.4	58.5	87.7





#24
tert-Butyl Alcohol
Concen: 32.713 ug/l
RT: 3.161 min Scan# 5
Delta R.T. 0.003 min
Lab File: VU063253.D
Acq: 12 Feb 2025 14:20

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Tgt Ion: 59 Resp: 43566

Ion	Ratio	Lower	Upper
59	100		
57	8.8	7.5	11.3





#25
Methyl tert-Butyl Ether
Concen: 7.574 ug/l
RT: 3.348 min Scan# 640
Delta R.T. -0.000 min
Lab File: VU063253.D
Acq: 12 Feb 2025 14:20

Tgt Ion: 73 Resp: 253519

Ion	Ratio	Lower	Upper
73	100		
43	19.1	14.9	22.3





#31
Isopropyl Ether
Concen: 10.742 ug/l
RT: 3.975 min Scan# 8
Delta R.T. -0.000 min
Lab File: VU063253.D
Acq: 12 Feb 2025 14:20

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Tgt Ion: 45 Resp: 442152

Ion	Ratio	Lower	Upper
45	100		
43	51.7	25.7	77.1





#32
Ethyl-t-butyl ether
Concen: 53.267 ug/l
RT: 4.483 min Scan# 993
Delta R.T. -0.000 min
Lab File: VU063253.D
Acq: 12 Feb 2025 14:20

Tgt Ion: 59 Resp: 1993904

Ion	Ratio	Lower	Upper
59	100		
87	41.3	32.6	49.0





#33
Tert-Amyl methyl ether
Concen: 50.332 ug/l
RT: 5.926 min Scan# 14
Delta R.T. -0.000 min
Lab File: VU063253.D
Acq: 12 Feb 2025 14:20

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Tgt Ion: 73 Resp: 1646111

Ion	Ratio	Lower	Upper
73	100		
87	23.5	18.9	28.3
71	11.2	9.0	13.4





#57
4-Bromofluorobenzene
Concen: 0.733 ug/l
RT: 10.627 min Scan# 2904
Delta R.T. 0.003 min
Lab File: VU063253.D
Acq: 12 Feb 2025 14:20

Tgt Ion: 95 Resp: 12765

Ion	Ratio	Lower	Upper
95	100		
174	82.6	58.6	88.0
176	80.3	58.2	87.4





#68
1,2-Dichlorobenzene-d4
Concen: 0.766 ug/l
RT: 12.190 min Scan# 3
Delta R.T. 0.006 min
Lab File: VU063253.D
Acq: 12 Feb 2025 14:20

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WS

Tgt Ion:152 Resp: 13885

Ion	Ratio	Lower	Upper
152	100		
115	69.7	0.0	275.2
150	159.0	0.0	658.4





#73
n-propylbenzene
Concen: 11.610 ug/l
RT: 10.897 min Scan# 2988
Delta R.T. 0.003 min
Lab File: VU063253.D
Acq: 12 Feb 2025 14:20

Tgt Ion:120 Resp: 194126

Ion	Ratio	Lower	Upper
120	100		
91	469.0	369.8	554.6



Raw Data: **VU063254.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	01/13/25
Project:	NJ Drinking Water PT	Date Received:	01/15/25
Client Sample ID:	PT-ADD-WSDL	SDG No.:	Q1172
Lab Sample ID:	Q1172-10DL	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063254.D	5		02/12/25 14:58	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.70	UD	0.70	2.50	ug/L
74-87-3	Chloromethane	0.65	UD	0.65	2.50	ug/L
75-01-4	Vinyl Chloride	0.65	UD	0.65	2.50	ug/L
74-83-9	Bromomethane	0.90	UD	0.90	2.50	ug/L
75-00-3	Chloroethane	0.70	UD	0.70	2.50	ug/L
109-99-9	Tetrahydrofuran	2.20	UD	2.20	5.00	ug/L
75-69-4	Trichlorofluoromethane	1.10	UD	1.10	2.50	ug/L
76-13-1	1,1,2-Trichloro-1,2,2-trifluoroethane	20.4	D	0.70	2.50	ug/L
75-65-0	tert-Butyl Alcohol	43.2	UD	43.2	50.0	ug/L
60-29-7	Diethyl Ether	0.65	UD	0.65	2.50	ug/L
75-35-4	1,1-Dichloroethene	0.60	UD	0.60	2.50	ug/L
107-13-1	Acrylonitrile	2.20	UD	2.20	5.00	ug/L
67-64-1	Acetone	5.50	UD	5.50	12.5	ug/L
75-15-0	Carbon Disulfide	0.65	UD	0.65	2.50	ug/L
1634-04-4	Methyl tert-Butyl Ether	7.30	D	0.60	2.50	ug/L
96-33-3	Methyl acrylate	1.40	UD	1.40	2.50	ug/L
75-09-2	Methylene Chloride	2.40	UD	2.40	2.50	ug/L
156-60-5	trans-1,2-Dichloroethene	0.70	UD	0.70	2.50	ug/L
75-34-3	1,1-Dichloroethane	0.65	UD	0.65	2.50	ug/L
110-82-7	Cyclohexane	0.70	UD	0.70	2.50	ug/L
78-93-3	2-Butanone	3.40	UD	3.40	12.5	ug/L
56-23-5	Carbon Tetrachloride	0.70	UD	0.70	2.50	ug/L
594-20-7	2,2-Dichloropropane	0.70	UD	0.70	2.50	ug/L
156-59-2	cis-1,2-Dichloroethene	0.65	UD	0.65	2.50	ug/L
74-97-5	Bromochloromethane	0.80	UD	0.80	2.50	ug/L
67-66-3	Chloroform	0.65	UD	0.65	2.50	ug/L
71-55-6	1,1,1-Trichloroethane	0.60	UD	0.60	2.50	ug/L
108-87-2	Methylcyclohexane	0.60	UD	0.60	2.50	ug/L
563-58-6	1,1-Dichloropropene	0.55	UD	0.55	2.50	ug/L
107-12-0	Propionitrile	5.20	UD	5.20	12.5	ug/L

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	01/13/25
Project:	NJ Drinking Water PT	Date Received:	01/15/25
Client Sample ID:	PT-ADD-WSDL	SDG No.:	Q1172
Lab Sample ID:	Q1172-10DL	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063254.D	5		02/12/25 14:58	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
71-43-2	Benzene	0.55	UD	0.55	2.50	ug/L
107-06-2	1,2-Dichloroethane	0.80	UD	0.80	2.50	ug/L
79-01-6	Trichloroethene	0.65	UD	0.65	2.50	ug/L
78-87-5	1,2-Dichloropropane	0.65	UD	0.65	2.50	ug/L
109-69-3	1-Chlorobutane	0.60	UD	0.60	2.50	ug/L
74-95-3	Dibromomethane	0.70	UD	0.70	2.50	ug/L
75-27-4	Bromodichloromethane	0.60	UD	0.60	2.50	ug/L
108-10-1	4-Methyl-2-Pentanone	3.00	UD	3.00	12.5	ug/L
108-88-3	Toluene	0.55	UD	0.55	2.50	ug/L
10061-02-6	t-1,3-Dichloropropene	0.55	UD	0.55	2.50	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.55	UD	0.55	2.50	ug/L
79-00-5	1,1,2-Trichloroethane	0.65	UD	0.65	2.50	ug/L
142-28-9	1,3-Dichloropropane	0.65	UD	0.65	2.50	ug/L
591-78-6	2-Hexanone	2.90	UD	2.90	12.5	ug/L
124-48-1	Dibromochloromethane	0.65	UD	0.65	2.50	ug/L
106-93-4	1,2-Dibromoethane	0.65	UD	0.65	2.50	ug/L
127-18-4	Tetrachloroethene	0.70	UD	0.70	2.50	ug/L
108-90-7	Chlorobenzene	0.55	UD	0.55	2.50	ug/L
630-20-6	1,1,1,2-Tetrachloroethane	0.65	UD	0.65	2.50	ug/L
67-72-1	Hexachloroethane	0.60	UD	0.60	2.50	ug/L
100-41-4	Ethyl Benzene	0.60	UD	0.60	2.50	ug/L
179601-23-1	m/p-Xylenes	1.20	UD	1.20	5.00	ug/L
1330-20-7	Total Xylenes	1.80	UD	1.80	7.50	ug/L
95-47-6	o-Xylene	0.60	UD	0.60	2.50	ug/L
100-42-5	Styrene	0.65	UD	0.65	2.50	ug/L
75-25-2	Bromoform	0.70	UD	0.70	2.50	ug/L
98-82-8	Isopropylbenzene	0.65	UD	0.65	2.50	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.60	UD	0.60	2.50	ug/L
96-18-4	1,2,3-Trichloropropane	1.10	UD	1.10	2.50	ug/L
108-86-1	Bromobenzene	0.65	UD	0.65	2.50	ug/L
103-65-1	n-propylbenzene	9.60	D	0.80	2.50	ug/L

Raw Data: **VU063254.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	01/13/25
Project:	NJ Drinking Water PT	Date Received:	01/15/25
Client Sample ID:	PT-ADD-WSDL	SDG No.:	Q1172
Lab Sample ID:	Q1172-10DL	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063254.D	5		02/12/25 14:58	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
95-49-8	2-Chlorotoluene	0.70	UD	0.70	2.50	ug/L
108-67-8	1,3,5-Trimethylbenzene	0.65	UD	0.65	2.50	ug/L
106-43-4	4-Chlorotoluene	0.70	UD	0.70	2.50	ug/L
98-06-6	tert-Butylbenzene	0.55	UD	0.55	2.50	ug/L
95-63-6	1,2,4-Trimethylbenzene	0.65	UD	0.65	2.50	ug/L
135-98-8	sec-Butylbenzene	0.65	UD	0.65	2.50	ug/L
99-87-6	p-Isopropyltoluene	0.80	UD	0.80	2.50	ug/L
541-73-1	1,3-Dichlorobenzene	0.65	UD	0.65	2.50	ug/L
106-46-7	1,4-Dichlorobenzene	0.70	UD	0.70	2.50	ug/L
104-51-8	n-Butylbenzene	1.40	UD	1.40	2.50	ug/L
95-50-1	1,2-Dichlorobenzene	0.70	UD	0.70	2.50	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	1.20	UD	1.20	2.50	ug/L
120-82-1	1,2,4-Trichlorobenzene	1.10	UD	1.10	2.50	ug/L
87-68-3	Hexachlorobutadiene	0.70	UD	0.70	2.50	ug/L
91-20-3	Naphthalene	1.60	UD	1.60	5.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	1.30	UD	1.30	2.50	ug/L
74-88-4	Iodomethane	0.80	UD	0.80	5.00	ug/L
107-05-1	Allyl Chloride	0.55	UD	0.55	2.50	ug/L
126-98-7	Methacrylonitrile	0.95	UD	0.95	2.50	ug/L
108-20-3	Isopropyl Ether	10.2	D	0.60	2.50	ug/L
80-62-6	Methyl methacrylate	1.20	UD	1.20	5.00	ug/L
SURROGATES						
2199-69-1	1,2-Dichlorobenzene-d4	0.82		70 - 130	82%	SPK: 1
460-00-4	4-Bromofluorobenzene	0.77		70 - 130	77%	SPK: 1
INTERNAL STANDARDS						
462-06-6	Fluorobenzene	48700	6.107			
	Tert-Amyl methyl ether					

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	01/13/25
Project:	NJ Drinking Water PT	Date Received:	01/15/25
Client Sample ID:	PT-ADD-WSDL	SDG No.:	Q1172
Lab Sample ID:	Q1172-10DL	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063254.D	5		02/12/25 14:58	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Sample Results: **VU063254.D**

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063254.D
Acq On : 12 Feb 2025 14:58
Operator : MD/SY
Sample : Q1172-10DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WSDL

Quant Time: Feb 13 03:12:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Fluorobenzene	6.107	96	48656	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.627	95	12366	0.771	ug/l	0.00
Spiked Amount 1.000			Recovery = 77.000%			
68) 1,2-Dichlorobenzene-d4	12.190	152	13603	0.815	ug/l	0.00
Spiked Amount 1.000			Recovery = 82.000%			
Target Compounds						Qvalue
8) 1,1,2-Trichloro-1,2,2-...	2.570	101	49347	4.073	ug/l	99
24) tert-Butyl Alcohol	3.171	59	11052	3.957	ug/l #	90
25) Methyl tert-Butyl Ether	3.351	73	44947	1.458	ug/l	97
31) Isopropyl Ether	3.978	45	76959	2.030	ug/l	96
32) Ethyl-t-butyl ether	4.486	59	334596	9.705	ug/l	99
33) Tert-Amyl methyl ether	5.926	73	269177	8.936	ug/l	100
73) n-propylbenzene	10.897	120	29507	1.916	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063254.D
Acq On : 12 Feb 2025 14:58
Operator : MD/SY
Sample : Q1172-10DL 5X
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WSDL

Quant Time: Feb 13 03:12:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.107 min Scan# 14
Delta R.T. 0.004 min
Lab File: VU063254.D
Acq: 12 Feb 2025 14:58

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WSDL

Tgt Ion: 96 Resp: 48656

Ion	Ratio	Lower	Upper
96	100		
70	20.5	15.6	23.4
95	8.2	7.3	10.9
97	6.2	0.0	0.0#





#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 4.073 ug/l
RT: 2.570 min Scan# 398
Delta R.T. -0.000 min
Lab File: VU063254.D
Acq: 12 Feb 2025 14:58

Tgt Ion:101 Resp: 49347

Ion	Ratio	Lower	Upper
101	100		
85	45.0	35.4	53.0
151	74.5	58.5	87.7





#24
tert-Butyl Alcohol
Concen: 3.957 ug/l
RT: 3.171 min Scan# 5
Delta R.T. 0.013 min
Lab File: VU063254.D
Acq: 12 Feb 2025 14:58

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WSDL

Tgt Ion: 59 Resp: 11052

Ion	Ratio	Lower	Upper
59	100		
57	5.8	7.5	11.3#





#25
Methyl tert-Butyl Ether
Concen: 1.458 ug/l
RT: 3.351 min Scan# 641
Delta R.T. 0.003 min
Lab File: VU063254.D
Acq: 12 Feb 2025 14:58

Tgt Ion: 73 Resp: 44947

Ion	Ratio	Lower	Upper
73	100		
43	20.0	14.9	22.3



Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WSDL



#31
Isopropyl Ether
Concen: 2.030 ug/l
RT: 3.978 min Scan# 8
Delta R.T. 0.003 min
Lab File: VU063254.D
Acq: 12 Feb 2025 14:58

Tgt Ion: 45 Resp: 76959

Ion	Ratio	Lower	Upper
45	100		
43	54.1	25.7	77.1





#32
Ethyl-t-butyl ether
Concen: 9.705 ug/l
RT: 4.486 min Scan# 994
Delta R.T. 0.003 min
Lab File: VU063254.D
Acq: 12 Feb 2025 14:58

Tgt Ion: 59 Resp: 334596

Ion	Ratio	Lower	Upper
59	100		
87	41.7	32.6	49.0





#33
Tert-Amyl methyl ether
Concen: 8.936 ug/l
RT: 5.926 min Scan# 14
Delta R.T. -0.000 min
Lab File: VU063254.D
Acq: 12 Feb 2025 14:58

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WSDL

Tgt Ion: 73 Resp: 269177

Ion	Ratio	Lower	Upper
73	100		
87	23.6	18.9	28.3
71	11.7	9.0	13.4





#57
4-Bromofluorobenzene
Concen: 0.771 ug/l
RT: 10.627 min Scan# 2904
Delta R.T. 0.003 min
Lab File: VU063254.D
Acq: 12 Feb 2025 14:58

Tgt Ion: 95 Resp: 12366

Ion	Ratio	Lower	Upper
95	100		
174	76.7	58.6	88.0
176	79.4	58.2	87.4





#68
1,2-Dichlorobenzene-d4
Concen: 0.815 ug/l
RT: 12.190 min Scan# 3
Delta R.T. 0.006 min
Lab File: VU063254.D
Acq: 12 Feb 2025 14:58

Instrument :
MSVOA_U
ClientSampleId :
PT-ADD-WSDL

Tgt Ion:152 Resp: 13603

Ion	Ratio	Lower	Upper
152	100		
115	66.7	0.0	275.2
150	156.4	0.0	658.4





#73
n-propylbenzene
Concen: 1.916 ug/l
RT: 10.897 min Scan# 2988
Delta R.T. 0.003 min
Lab File: VU063254.D
Acq: 12 Feb 2025 14:58

Tgt Ion:120 Resp: 29507

Ion	Ratio	Lower	Upper
120	100		
91	481.3	369.8	554.6



CALIBRATION SUMMARY

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172 SDG No.: Q1172

Instrument ID: MSVOA_U Calibration Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Calibration Time(s): 12:59 15:33

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID: RRF0.5 = VU063219.D RRF001 = VU063220.D RRF002 = VU063221.D
RRF005 = VU063222.D RRF010 = VU063223.D RRF015 = VU063224.D

COMPOUND	RRF0.5	RRF001	RRF002	RRF005	RRF010	RRF015	$\overline{RRF}$	% RSD
Dichlorodifluoromethane	0.326	0.362	0.329	0.325	0.305	0.302	0.325	6.7
Chloromethane	0.390	0.400	0.389	0.368	0.352	0.344	0.374	6.1
Vinyl Chloride	0.364	0.390	0.367	0.376	0.366	0.357	0.370	3.1
Bromomethane	0.198	0.201	0.125	0.189	0.191		0.181	17.6
Chloroethane	0.245	0.259	0.233	0.223	0.222	0.216	0.233	7.1
Tetrahydrofuran	0.046	0.046	0.049	0.044	0.051	0.046	0.047	5.9
Trichlorofluoromethane	0.426	0.473	0.452	0.443	0.423	0.416	0.439	4.8
1,1,2-Trichloro-1,2,2-trifluoroeth	0.259	0.272	0.253	0.250	0.231	0.230	0.249	6.5
tert-Butyl Alcohol		0.039	0.027	0.022	0.024	0.021	0.026	27
Diethyl Ether	0.223	0.231	0.215	0.218	0.216	0.203	0.218	4.2
1,1-Dichloroethene	0.256	0.274	0.253	0.250	0.248	0.241	0.254	4.3
Acrylonitrile	0.051	0.056	0.059	0.060	0.065	0.059	0.059	7.6
Acetone	0.050	0.047	0.044	0.042	0.047	0.040	0.045	8.4
Carbon Disulfide	0.944	0.945	0.880	0.869	0.861	0.822	0.887	5.5
Methyl tert-Butyl Ether	0.582	0.652	0.618	0.642	0.673	0.635	0.634	4.9
Methyl acrylate	0.118	0.146	0.150	0.157	0.165	0.153	0.148	10.9
Methylene Chloride	0.332	0.333	0.313	0.309	0.304	0.289	0.313	5.4
trans-1,2-Dichloroethene	0.292	0.305	0.287	0.288	0.287	0.279	0.290	3
1,1-Dichloroethane	0.523	0.592	0.552	0.549	0.538	0.520	0.546	4.8
Cyclohexane	0.415	0.423	0.423	0.457	0.453	0.460	0.439	4.6
2-Butanone	0.071	0.072	0.067	0.071	0.079	0.070	0.072	5.3
Carbon Tetrachloride	0.371	0.417	0.379	0.383	0.378	0.369	0.383	4.6
2,2-Dichloropropane	0.434	0.463	0.413	0.421	0.418	0.407	0.426	4.7
cis-1,2-Dichloroethene	0.297	0.332	0.306	0.316	0.316	0.310	0.313	3.7
Bromochloromethane	0.133	0.150	0.132	0.139	0.137	0.130	0.137	5.3
Chloroform	0.546	0.588	0.543	0.556	0.546	0.526	0.551	3.7
1,1,1-Trichloroethane	0.433	0.466	0.448	0.459	0.442	0.429	0.446	3.3
Methylcyclohexane	0.377	0.448	0.418	0.461	0.448	0.458	0.435	7.4
1,1-Dichloropropene	0.368	0.416	0.404	0.404	0.408	0.399	0.400	4.1
Propionitrile	0.016	0.020	0.023	0.022	0.025	0.024	0.022	14.5

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Raw Data

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172 SDG No.: Q1172

Instrument ID: MSVOA_U Calibration Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Calibration Time(s): 12:59 15:33

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID: RRF0.5 = VU063219.D RRF001 = VU063220.D RRF002 = VU063221.D
RRF005 = VU063222.D RRF010 = VU063223.D RRF015 = VU063224.D

COMPOUND	RRF0.5	RRF001	RRF002	RRF005	RRF010	RRF015	$\overline{RRF}$	% RSD
Benzene	1.163	1.280	1.242	1.256	1.232	1.199	1.229	3.4
1,2-Dichloroethane	0.351	0.371	0.358	0.357	0.355	0.335	0.355	3.3
Trichloroethene	0.287	0.311	0.294	0.290	0.291	0.281	0.292	3.4
1,2-Dichloropropane	0.314	0.333	0.319	0.323	0.326	0.314	0.322	2.3
1-Chlorobutane	0.484	0.555	0.547	0.578	0.565	0.552	0.547	5.9
Dibromomethane	0.161	0.173	0.163	0.164	0.163	0.153	0.163	4
Bromodichloromethane	0.332	0.402	0.386	0.393	0.391	0.371	0.379	6.7
4-Methyl-2-Pentanone	0.147	0.155	0.166	0.176	0.197	0.182	0.171	10.6
Toluene	0.607	0.705	0.706	0.745	0.744	0.732	0.707	7.4
t-1,3-Dichloropropene	0.284	0.343	0.343	0.362	0.383	0.368	0.347	9.9
cis-1,3-Dichloropropene	0.378	0.427	0.430	0.435	0.459	0.444	0.429	6.4
1,1,2-Trichloroethane	0.211	0.225	0.219	0.221	0.228	0.214	0.220	3
1,3-Dichloropropane	0.354	0.408	0.393	0.400	0.405	0.379	0.390	5.2
2-Hexanone	0.094	0.109	0.115	0.118	0.136	0.128	0.117	12.6
Dibromochloromethane	0.238	0.257	0.246	0.256	0.267	0.252	0.253	4
1,2-Dibromoethane	0.194	0.206	0.202	0.209	0.219	0.205	0.206	4
Tetrachloroethene	0.242	0.258	0.236	0.250	0.234	0.225	0.241	4.9
Chlorobenzene	0.676	0.752	0.745	0.775	0.768	0.758	0.746	4.8
1,1,1,2-Tetrachloroethane	0.254	0.279	0.264	0.269	0.274	0.268	0.268	3.2
Hexachloroethane	0.179	0.215	0.214	0.210	0.222	0.232	0.212	8.5
Ethyl Benzene	1.075	1.229	1.246	1.363	1.403	1.399	1.286	10
m/p-Xylenes	0.374	0.444	0.475	0.526	0.536	0.528	0.480	13.2
o-Xylene	0.374	0.449	0.469	0.502	0.514	0.513	0.470	11.5
Styrene	0.562	0.675	0.721	0.815	0.856	0.862	0.748	15.8
Bromoform	0.127	0.139	0.142	0.145	0.156	0.151	0.143	7.1
Isopropylbenzene	0.889	1.035	1.086	1.184	1.212	1.226	1.106	11.7
1,1,2,2-Tetrachloroethane	0.260	0.304	0.306	0.292	0.310	0.303	0.296	6.2
1,2,3-Trichloropropane	0.220	0.222	0.246	0.213	0.204	0.230	0.222	6.5
Bromobenzene	0.253	0.303	0.295	0.308	0.314	0.314	0.298	7.8
n-propylbenzene	0.242	0.286	0.314	0.344	0.354	0.359	0.317	14.5

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Raw Data

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172 SDG No.: Q1172

Instrument ID: MSVOA_U Calibration Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Calibration Time(s): 12:59 15:33

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID: RRF0.5 = VU063219.D RRF001 = VU063220.D RRF002 = VU063221.D
RRF005 = VU063222.D RRF010 = VU063223.D RRF015 = VU063224.D

COMPOUND	RRF0.5	RRF001	RRF002	RRF005	RRF010	RRF015	$\overline{RRF}$	% RSD
2-Chlorotoluene	0.228	0.278	0.292	0.315	0.316	0.321	0.292	12.1
1,3,5-Trimethylbenzene	0.741	0.939	1.024	1.125	1.148	1.169	1.024	16
4-Chlorotoluene	0.237	0.292	0.297	0.320	0.321	0.327	0.299	11.2
tert-Butylbenzene	0.854	0.987	1.030	1.074	1.118	1.153	1.036	10.3
1,2,4-Trimethylbenzene	0.709	0.905	1.009	1.126	1.157	1.192	1.016	18.1
sec-Butylbenzene	1.016	1.226	1.318	1.415	1.442	1.498	1.319	13.4
p-Isopropyltoluene	0.745	0.942	1.033	1.137	1.168	1.222	1.041	17
1,3-Dichlorobenzene	0.531	0.551	0.589	0.589	0.596	0.611	0.578	5.2
1,4-Dichlorobenzene	0.485	0.514	0.581	0.593	0.601	0.617	0.565	9.4
n-Butylbenzene	0.693	0.822	0.907	0.992	1.044	1.142	0.933	17.3
1,2-Dichlorobenzene	0.481	0.538	0.571	0.563	0.574	0.604	0.555	7.6
1,2-Dibromo-3-Chloropropane	0.031	0.041	0.039	0.043	0.049	0.046	0.042	14.6
1,2,4-Trichlorobenzene	0.206	0.246	0.253	0.282	0.308	0.331	0.271	16.7
Hexachlorobutadiene	0.186	0.197	0.203	0.191	0.182	0.202	0.194	4.4
Naphthalene	0.315	0.335	0.379	0.441	0.556	0.592	0.436	26.5
1,2,3-Trichlorobenzene	0.219	0.237	0.239	0.275	0.303	0.318	0.265	15.1
1,2-Dichlorobenzene-d4	0.314	0.328	0.350	0.336	0.344	0.388	0.343	7.4
4-Bromofluorobenzene	0.302	0.311	0.330	0.321	0.358	0.356	0.330	7
Iodomethane		0.411	0.389	0.411	0.405	0.378	0.399	3.6
Allyl Chloride	0.340	0.390	0.369	0.364	0.365	0.357	0.364	4.5
Methacrylonitrile	0.062	0.067	0.079	0.087	0.098	0.091	0.080	17.3
Isopropyl Ether	0.718	0.786	0.766	0.785	0.814	0.808	0.779	4.5
Methyl methacrylate	0.112	0.126	0.128	0.148	0.161	0.147	0.137	13.3

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Response Factor Report MSVOA_U

Method Path : Z:\voasrv\HPCHEM1\MSVOA_U\Method\
Method File : 524U021025DW.M
Title : METHOD 524.2 VOLATILES DRINKING WATER
Last Update : Tue Feb 11 08:42:19 2025
Response Via : Initial Calibration

Calibration Files
0.5 =VU063219.D 1 =VU063220.D 2 =VU063221.D 5 =VU063222.D 10 =VU063223.D 15 =VU063224.D

		Compound	0.5	1	2	5	10	15	Avg	%RSD
1)	i	Fluorobenzene	----------	ISTD	----------					
2)	T	Dichlorodifluo...	0.326	0.362	0.329	0.325	0.305	0.302	0.325	6.65
3)	t	Chloromethane	0.390	0.400	0.389	0.368	0.352	0.344	0.374	6.06
4)	Rt	Vinyl Chloride	0.364	0.390	0.367	0.376	0.366	0.357	0.370	3.13
5)	T	Bromomethane	0.198	0.201	0.125	0.189	0.191		0.181	17.62
6)	T	Chloroethane	0.245	0.259	0.233	0.223	0.222	0.216	0.233	7.09
7)	T	Trichlorofluor...	0.426	0.473	0.452	0.443	0.423	0.416	0.439	4.85
8)		1,1,2-Trichlor...	0.259	0.272	0.253	0.250	0.231	0.230	0.249	6.54
9)	Rt	1,1-Dichloroet...	0.256	0.274	0.253	0.250	0.248	0.241	0.254	4.34
10)	t	Iodomethane		0.411	0.389	0.411	0.405	0.378	0.399	3.63
11)	t	Allyl Chloride	0.340	0.390	0.369	0.364	0.365	0.357	0.364	4.45
12)	t	Acrylonitrile	0.051	0.056	0.059	0.060	0.065	0.059	0.059	7.58
13)	T	Acetone	0.050	0.047	0.044	0.042	0.047	0.040	0.045	8.41
14)	T	Carbon Disulfide	0.944	0.945	0.880	0.869	0.861	0.822	0.887	5.49
15)	RT	Methylene Chlo...	0.332	0.333	0.313	0.309	0.304	0.289	0.313	5.41
16)	RT	trans-1,2-Dich...	0.292	0.305	0.287	0.288	0.287	0.279	0.290	3.01
17)	t	1,1-Dichloroet...	0.523	0.592	0.552	0.549	0.538	0.520	0.546	4.80
18)	T	2-Butanone	0.071	0.072	0.067	0.071	0.079	0.070	0.072	5.29
19)		Cyclohexane	0.415	0.423	0.423	0.457	0.453	0.460	0.439	4.65
20)		Methylcyclohexane	0.377	0.448	0.418	0.461	0.448	0.458	0.435	7.44
21)	T	2,2-Dichloropr...	0.434	0.463	0.413	0.421	0.418	0.407	0.426	4.71
22)	RT	cis-1,2-Dichlo...	0.297	0.332	0.306	0.316	0.316	0.310	0.313	3.74
23)	t	Diethyl Ether	0.223	0.231	0.215	0.218	0.216	0.203	0.218	4.22
24)	t	tert-Butyl Alc...		0.039	0.027	0.022	0.024	0.021	0.026	27.01
25)	t	Methyl tert-Bu...	0.582	0.652	0.618	0.642	0.673	0.635	0.634	4.89
26)	t	Bromochloromet...	0.133	0.150	0.132	0.139	0.137	0.130	0.137	5.31
27)	t	Chloroform	0.546	0.588	0.543	0.556	0.546	0.526	0.551	3.74
28)	RT	1,1,1-Trichlor...	0.433	0.466	0.448	0.459	0.442	0.429	0.446	3.26
29)	T	1,1-Dichloropr...	0.368	0.416	0.404	0.404	0.408	0.399	0.400	4.14
30)	RT	Carbon Tetrach...	0.371	0.417	0.379	0.383	0.378	0.369	0.383	4.62
31)	t	Isopropyl Ether	0.718	0.786	0.766	0.785	0.814	0.808	0.779	4.45
32)		Ethyl-t-butyl ...	0.648	0.686	0.691	0.726	0.755	0.746	0.709	5.72
33)		Tert-Amyl meth...	0.540	0.604	0.585	0.637	0.688	0.661	0.619	8.69
34)	t	Propionitrile	0.016	0.020	0.023	0.022	0.025	0.024	0.022	14.47
35)	RT	Benzene	1.163	1.280	1.242	1.256	1.232	1.199	1.229	3.40
36)	RT	1,2-Dichloroet...	0.351	0.371	0.358	0.357	0.355	0.335	0.355	3.35
37)	RT	Trichloroethene	0.287	0.311	0.294	0.290	0.291	0.281	0.292	3.43
38)	Rt	1,2-Dichloropr...	0.314	0.333	0.319	0.323	0.326	0.314	0.322	2.27
39)	t	Methacrylonitrile	0.062	0.067	0.079	0.087	0.098	0.091	0.080	17.33
40)	t	Methyl acrylate	0.118	0.146	0.150	0.157	0.165	0.153	0.148	10.87
41)	t	Tetrahydrofuran	0.046	0.046	0.049	0.044	0.051	0.046	0.047	5.91
42)	t	1-Chlorobutane	0.484	0.555	0.547	0.578	0.565	0.552	0.547	5.95
43)	T	Dibromomethane	0.161	0.173	0.163	0.164	0.163	0.153	0.163	3.96
44)	T	Bromodichlorom...	0.332	0.402	0.386	0.393	0.391	0.371	0.379	6.67
45)	T	4-Methyl-2-Pen...	0.147	0.155	0.166	0.176	0.197	0.182	0.171	10.64
46)	t	t-1,4-Dichloro...	0.071	0.084	0.088	0.072	0.084	0.082	0.080	8.79
47)	t	Methyl methacr...	0.112	0.126	0.128	0.148	0.161	0.147	0.137	13.27
48)	t	Ethyl methacry...	0.194	0.227	0.244	0.270	0.314	0.296	0.258	17.35
49)	Rt	Toluene	0.607	0.705	0.706	0.745	0.744	0.732	0.707	7.37
50)	T	t-1,3-Dichloro...	0.284	0.343	0.343	0.362	0.383	0.368	0.347	9.93
51)	T	cis-1,3-Dichlo...	0.378	0.427	0.430	0.435	0.459	0.444	0.429	6.40
52)	RT	1,1,2-Trichlor...	0.211	0.225	0.219	0.221	0.228	0.214	0.220	2.98
53)	t	1,3-Dichloropr...	0.354	0.408	0.393	0.400	0.405	0.379	0.390	5.20
54)	t	2-Hexanone	0.094	0.109	0.115	0.118	0.136	0.128	0.117	12.64
55)	t	Dibromochlorom...	0.238	0.257	0.246	0.256	0.267	0.252	0.253	3.96
56)	T	1,2-Dibromoethane	0.194	0.206	0.202	0.209	0.219	0.205	0.206	4.03
57)	S	4-Bromofluorob...	0.302	0.311	0.330	0.321	0.358	0.356	0.330	7.00

Method Path : Z:\voasrv\HPCHEM1\MSVOA_U\Method\
Method File : 524U021025DW.M

58)	RT	Tetrachloroethene	0.242	0.258	0.236	0.250	0.234	0.225	0.241	4.94
59)	Rt	Chlorobenzene	0.676	0.752	0.745	0.775	0.768	0.758	0.746	4.79
60)	T	1,1,1,2-Tetrac...	0.254	0.279	0.264	0.269	0.274	0.268	0.268	3.18
61)	t	Pentachloroethane	0.209	0.246	0.251	0.240	0.241	0.250	0.239	6.51
62)	t	Hexachloroethane	0.179	0.215	0.214	0.210	0.222	0.232	0.212	8.47
63)	Rt	Ethyl Benzene	1.075	1.229	1.246	1.363	1.403	1.399	1.286	9.96
64)	RT	m/p-Xylenes	0.374	0.444	0.475	0.526	0.536	0.528	0.480	13.21
65)	RT	o-Xylene	0.374	0.449	0.469	0.502	0.514	0.513	0.470	11.48
66)	RT	Styrene	0.562	0.675	0.721	0.815	0.856	0.862	0.748	15.82
67)	t	Bromoform	0.127	0.139	0.142	0.145	0.156	0.151	0.143	7.11
68)	S	1,2-Dichlorobe...	0.314	0.328	0.350	0.336	0.344	0.388	0.343	7.36
69)	T	Isopropylbenzene	0.889	1.035	1.086	1.184	1.212	1.226	1.106	11.74
70)	T	1,1,2,2-Tetrac...	0.260	0.304	0.306	0.292	0.310	0.303	0.296	6.25
71)	T	1,2,3-Trichlor...	0.220	0.222	0.246	0.213	0.204	0.230	0.222	6.49
72)	t	Bromobenzene	0.253	0.303	0.295	0.308	0.314	0.314	0.298	7.83
73)	t	n-propylbenzene	0.242	0.286	0.314	0.344	0.354	0.359	0.317	14.49
74)	t	2-Chlorotoluene	0.228	0.278	0.292	0.315	0.316	0.321	0.292	12.05
75)	t	1,3,5-Trimethy...	0.741	0.939	1.024	1.125	1.148	1.169	1.024	15.99
76)	t	4-Chlorotoluene	0.237	0.292	0.297	0.320	0.321	0.327	0.299	11.24
77)	t	tert-Butylbenzene	0.854	0.987	1.030	1.074	1.118	1.153	1.036	10.34
78)	t	1,2,4-Trimethy...	0.709	0.905	1.009	1.126	1.157	1.192	1.016	18.12
79)	t	sec-Butylbenzene	1.016	1.226	1.318	1.415	1.442	1.498	1.319	13.42
80)		Nitrobenzene		0.004	0.006	0.007	0.009	0.009	0.007	30.30
81)	t	p-Isopropyltol...	0.745	0.942	1.033	1.137	1.168	1.222	1.041	16.95
82)	t	1,3-Dichlorobe...	0.531	0.551	0.589	0.589	0.596	0.611	0.578	5.23
83)	Rt	1,4-Dichlorobe...	0.485	0.514	0.581	0.593	0.601	0.617	0.565	9.37
84)	t	n-Butylbenzene	0.693	0.822	0.907	0.992	1.044	1.142	0.933	17.26
85)	Rt	1,2-Dichlorobe...	0.481	0.538	0.571	0.563	0.574	0.604	0.555	7.60
86)	t	1,2-Dibromo-3-...	0.031	0.041	0.039	0.043	0.049	0.046	0.042	14.63
87)	Rt	1,2,4-Trichlor...	0.206	0.246	0.253	0.282	0.308	0.331	0.271	16.72
88)	t	Hexachlorobuta...	0.186	0.197	0.203	0.191	0.182	0.202	0.194	4.41
89)	t	Naphthalene	0.315	0.335	0.379	0.441	0.556	0.592	0.436	26.52
90)	t	1,2,3-Trichlor...	0.219	0.237	0.239	0.275	0.303	0.318	0.265	15.07

(#) = Out of Range



Response = 6.053e-001 * Amt - 3.976e-001
Coef of Det (r^2) = 0.994995 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA U\Method\524U021025DW.M
Calibration Table Last Updated: Tue Feb 11 08:42:19 2025



Response = 2.078e-002 * Amt + 1.449e-001
Coef of Det (r^2) = 0.994628 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA U\Method\524U021025DW.M
Calibration Table Last Updated: Tue Feb 11 08:42:19 2025



Response = 9.496e-003 * Amt - 3.646e-002
Coef of Det (r^2) = 0.998483 Curve Fit: Linear
Method Name: Z:\voasrv\HPCHEM1\MSVOA U\Method\524U021025DW.M
Calibration Table Last Updated: Tue Feb 11 08:42:19 2025

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063219.D
Acq On : 10 Feb 2025 12:59
Operator : MD/SY
Sample : VSTDICC0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 03:57:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Fluorobenzene	6.107	96	49900	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.627	95	15091	0.917	ug/l	0.00
Spiked Amount 1.000			Recovery =	92.000%		
68) 1,2-Dichlorobenzene-d4	12.187	152	15652	0.914	ug/l	0.00
Spiked Amount 1.000			Recovery =	91.000%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.380	85	8130	0.501	ug/l	99
3) Chloromethane	1.518	50	9740	0.522	ug/l	100
4) Vinyl Chloride	1.599	62	9090	0.492	ug/l	98
5) Bromomethane	1.849	94	4941	0.578	ug/l	82
6) Chloroethane	1.927	64	6121	0.526	ug/l	90
7) Trichlorofluoromethane	2.129	101	10617	0.485	ug/l	92
8) 1,1,2-Trichloro-1,2,2-...	2.570	101	6464	0.520	ug/l	95
9) 1,1-Dichloroethene	2.573	96	6378	0.504	ug/l	95
11) Allyl Chloride	2.917	41	8493	0.467	ug/l	91
12) Acrylonitrile	3.332	53	2569m	0.893	ug/l	
13) Acetone	2.618	43	6197	2.760	ug/l	90
14) Carbon Disulfide	2.782	76	23559	0.532	ug/l	100
15) Methylene Chloride	3.033	84	8284	0.530	ug/l	96
16) trans-1,2-Dichloroethene	3.348	96	7275	0.504	ug/l	99
17) 1,1-Dichloroethane	3.856	63	13044	0.479	ug/l	96
18) 2-Butanone	4.730	43	8910m	2.456	ug/l	
19) Cyclohexane	5.377	56	10342m	0.473	ug/l	
20) Methylcyclohexane	6.753	83	9399	0.433	ug/l	96
21) 2,2-Dichloropropane	4.653	77	10817	0.509	ug/l	95
22) cis-1,2-Dichloroethene	4.663	96	7417	0.475	ug/l	91
23) Diethyl Ether	2.367	59	5553	0.511	ug/l	95
25) Methyl tert-Butyl Ether	3.354	73	14531	0.460	ug/l	98
26) Bromochloromethane	4.968	128	3323	0.487	ug/l	95
27) Chloroform	5.081	83	13627	0.496	ug/l	93
28) 1,1,1-Trichloroethane	5.306	97	10801	0.485	ug/l	98
29) 1,1-Dichloropropene	5.518	75	9175	0.460	ug/l	99
30) Carbon Tetrachloride	5.515	117	9251	0.485	ug/l	93
31) Isopropyl Ether	3.978	45	17908	0.461	ug/l	98
32) Ethyl-t-butyl ether	4.492	59	16177	0.458	ug/l	99
33) Tert-Amyl methyl ether	5.930	73	13471	0.436	ug/l	99
34) Propionitrile	4.801	54	2034m	1.809	ug/l	
35) Benzene	5.769	78	29011	0.473	ug/l	98
36) 1,2-Dichloroethane	5.791	62	8750	0.494	ug/l #	95
37) Trichloroethene	6.537	130	7149	0.490	ug/l	94
38) 1,2-Dichloropropane	6.782	63	7842	0.489	ug/l	98
39) Methacrylonitrile	5.017	41	1535m	0.382	ug/l	
40) Methyl acrylate	4.920	55	2955m	0.403	ug/l	
41) Tetrahydrofuran	5.065	42	2289	0.970	ug/l #	58
42) 1-Chlorobutane	5.451	56	12081	0.443	ug/l	92
43) Dibromomethane	6.910	93	4024	0.495	ug/l	94
44) Bromodichloromethane	7.100	83	8280	0.438	ug/l #	99
45) 4-Methyl-2-Pentanone	7.785	43	18398	2.160	ug/l #	91
46) t-1,4-Dichloro-2-butene	10.823	75	3537m	0.839	ug/l	

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063219.D
Acq On : 10 Feb 2025 12:59
Operator : MD/SY
Sample : VSTDICC0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 03:57:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
47) Methyl methacrylate	6.965	69	5580m	0.815 ug/l	
48) Ethyl methacrylate	8.332	69	4839	0.377 ug/l	91
49) Toluene	7.962	92	15136	0.429 ug/l	97
50) t-1,3-Dichloropropene	8.209	75	7086	0.409 ug/l	99
51) cis-1,3-Dichloropropene	7.605	75	9424	0.441 ug/l	91
52) 1,1,2-Trichloroethane	8.393	97	5257	0.480 ug/l	94
53) 1,3-Dichloropropane	8.566	76	8834	0.454 ug/l	96
54) 2-Hexanone	8.685	43	11692m	2.011 ug/l	
55) Dibromochloromethane	8.801	129	5934	0.471 ug/l	97
56) 1,2-Dibromoethane	8.917	107	4834	0.470 ug/l	90
58) Tetrachloroethene	8.544	164	6044	0.503 ug/l	88
59) Chlorobenzene	9.441	112	16870	0.453 ug/l	96
60) 1,1,1,2-Tetrachloroethane	9.524	131	6349	0.475 ug/l	97
61) Pentachloroethane	11.418	117	5212	0.436 ug/l	98
62) Hexachloroethane	12.466	117	4467	0.423 ug/l	97
63) Ethyl Benzene	9.563	91	26810	0.418 ug/l	98
64) m/p-Xylenes	9.685	106	18658	0.778 ug/l	96
65) o-Xylene	10.094	106	9329	0.398 ug/l	99
66) Styrene	10.113	104	14012	0.375 ug/l	98
67) Bromoform	10.283	173	3160	0.442 ug/l	95
69) Isopropylbenzene	10.476	105	22191	0.402 ug/l	99
70) 1,1,2,2-Tetrachloroethane	10.772	83	6493	0.440 ug/l	97
71) 1,2,3-Trichloropropane	10.817	75	5487m	0.490 ug/l	
72) Bromobenzene	10.778	156	6302	0.424 ug/l	92
73) n-propylbenzene	10.901	120	6029	0.382 ug/l	97
74) 2-Chlorotoluene	10.981	126	5699	0.392 ug/l	97
75) 1,3,5-Trimethylbenzene	11.081	105	18487	0.362 ug/l	98
76) 4-Chlorotoluene	11.093	126	5912	0.396 ug/l	88
77) tert-Butylbenzene	11.412	119	21304	0.412 ug/l	95
78) 1,2,4-Trimethylbenzene	11.460	105	17694	0.349 ug/l	95
79) sec-Butylbenzene	11.637	105	25352	0.385 ug/l	100
81) p-Isopropyltoluene	11.785	119	18576	0.358 ug/l	98
82) 1,3-Dichlorobenzene	11.740	146	13255	0.460 ug/l	97
83) 1,4-Dichlorobenzene	11.833	146	12096	0.429 ug/l	99
84) n-Butylbenzene	12.206	91	17297	0.369 ug/l	95
85) 1,2-Dichlorobenzene	12.209	146	11996	0.433 ug/l	98
86) 1,2-Dibromo-3-Chloropr...	12.997	75	779	0.376 ug/l	90
87) 1,2,4-Trichlorobenzene	13.839	180	5130	0.380 ug/l	94
88) Hexachlorobutadiene	14.013	225	4651	0.482 ug/l	96
89) Naphthalene	14.097	128	7859	0.899 ug/l #	93
90) 1,2,3-Trichlorobenzene	14.328	180	5453	0.412 ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063219.D
Acq On : 10 Feb 2025 12:59
Operator : MD/SY
Sample : VSTDICC0.5
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Feb 11 03:57:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.107 min Scan# 14
Delta R.T. 0.004 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 96 Resp: 4990

Ion	Ratio	Lower	Upper
96	100		
70	20.5	15.6	23.4
95	8.6	7.3	10.9
97	5.9	0.0	0.0#



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#2
Dichlorodifluoromethane
Concen: 0.501 ug/l
RT: 1.380 min Scan# 28
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 85 Resp: 8130

Ion	Ratio	Lower	Upper
85	100		
87	31.2	16.0	48.0





#3
Chloromethane
Concen: 0.522 ug/l
RT: 1.518 min Scan# 7
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 50 Resp: 974

Ion	Ratio	Lower	Upper
50	100		
52	32.2	25.8	38.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#4
Vinyl Chloride
Concen: 0.492 ug/l
RT: 1.599 min Scan# 96
Delta R.T. -0.000 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 62 Resp: 9090

Ion	Ratio	Lower	Upper
62	100		
64	33.0	25.4	38.0





#5
Bromomethane
Concen: 0.578 ug/l
RT: 1.849 min Scan# 1
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 94 Resp: 494

Ion	Ratio	Lower	Upper
94	100		
96	111.7	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#6
Chloroethane
Concen: 0.526 ug/l
RT: 1.927 min Scan# 198
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 64 Resp: 6121

Ion	Ratio	Lower	Upper
64	100		
66	38.0	25.8	38.8





#7
Trichlorofluoromethane
Concen: 0.485 ug/l
RT: 2.129 min Scan# 2
Delta R.T. -0.000 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:101 Resp: 1061

Ion	Ratio	Lower	Upper
101	100		
103	70.6	51.4	77.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 0.520 ug/l
RT: 2.570 min Scan# 398
Delta R.T. -0.000 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:101 Resp: 6464

Ion	Ratio	Lower	Upper
101	100		
85	46.2	35.4	53.0
151	77.9	58.5	87.7





#9
1,1-Dichloroethene
Concen: 0.504 ug/l
RT: 2.573 min Scan# 3
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 96 Resp: 637

Ion	Ratio	Lower	Upper
96	100		
61	159.4	0.0	492.9
98	54.5	0.0	124.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#11
Allyl Chloride
Concen: 0.467 ug/l
RT: 2.917 min Scan# 506
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 41 Resp: 8493

Ion	Ratio	Lower	Upper
41	100		
39	80.2	57.9	86.9





#12
Acrylonitrile
Concen: 0.893 ug/l m
RT: 3.332 min Scan# 6
Delta R.T. 0.032 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 53 Resp: 256

Ion	Ratio	Lower	Upper
53	100		
52	57.8	64.2	96.2#
51	12.6	30.8	46.2#



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#13
Acetone
Concen: 2.760 ug/l
RT: 2.618 min Scan# 413
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 43 Resp: 6197

Ion	Ratio	Lower	Upper
43	100		
58	28.6	27.4	41.0





#14
Carbon Disulfide
Concen: 0.532 ug/l
RT: 2.782 min Scan# 4(
Delta R.T. -0.000 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 76 Resp: 2355(

Ion	Ratio	Lower	Upper
76	100		
78	9.1	7.2	10.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#15
Methylene Chloride
Concen: 0.530 ug/l
RT: 3.033 min Scan# 542
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 84 Resp: 8284

Ion	Ratio	Lower	Upper
84	100		
49	123.1	96.4	144.6
51	30.5	0.0	74.8
86	65.7	51.3	76.9





#16
trans-1,2-Dichloroethene
Concen: 0.504 ug/l
RT: 3.348 min Scan# 64
Delta R.T. 0.010 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 96 Resp: 727

Ion	Ratio	Lower	Upper
96	100		
61	140.2	113.4	170.2
98	63.1	51.2	76.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#17
1,1-Dichloroethane
Concen: 0.479 ug/l
RT: 3.856 min Scan# 798
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 63 Resp: 13044

Ion	Ratio	Lower	Upper
63	100		
98	4.7	3.3	9.9
100	4.2	2.1	6.2





#18
2-Butanone
Concen: 2.456 ug/l m
RT: 4.730 min Scan# 1
Delta R.T. 0.045 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 43 Resp: 891

Ion	Ratio	Lower	Upper
43	100		
57	10.9	0.0	17.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#19
Cyclohexane
Concen: 0.473 ug/l m
RT: 5.377 min Scan# 1271
Delta R.T. -0.000 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 56 Resp: 10342

Ion	Ratio	Lower	Upper
56	100		
69	29.7	24.5	36.7
84	76.6	65.2	97.8





#20
Methylcyclohexane
Concen: 0.433 ug/l
RT: 6.753 min Scan# 1
Delta R.T. -0.000 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 83 Resp: 939

Ion	Ratio	Lower	Upper
83	100		
55	84.6	63.1	94.7
98	44.2	35.2	52.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#21
2,2-Dichloropropane
Concen: 0.509 ug/l
RT: 4.653 min Scan# 1046
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 77 Resp: 10817

Ion	Ratio	Lower	Upper
77	100		
97	20.9	18.5	27.7





#22
cis-1,2-Dichloroethene
Concen: 0.475 ug/l
RT: 4.663 min Scan# 1
Delta R.T. 0.013 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 96 Resp: 741

Ion	Ratio	Lower	Upper
96	100		
61	163.8	0.0	373.3
98	60.6	31.9	95.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#23
Diethyl Ether
Concen: 0.511 ug/l
RT: 2.367 min Scan# 335
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 59 Resp: 5553

Ion	Ratio	Lower	Upper
59	100		
45	77.3	57.8	86.6
74	69.4	57.7	86.5





#25
Methyl tert-Butyl Ether
Concen: 0.460 ug/l
RT: 3.354 min Scan# 64
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 73 Resp: 1453

Ion	Ratio	Lower	Upper
73	100		
43	19.6	14.9	22.3

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#26
Bromochloromethane
Concen: 0.487 ug/l
RT: 4.968 min Scan# 1144
Delta R.T. 0.010 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:128 Resp: 3323

Ion	Ratio	Lower	Upper
128	100		
49	171.5	0.0	343.4
130	116.6	102.9	154.3





#27
Chloroform
Concen: 0.496 ug/l
RT: 5.081 min Scan# 1
Delta R.T. 0.010 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 83 Resp: 1362

Ion	Ratio	Lower	Upper
83	100		
85	59.2	0.0	129.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#28
1,1,1-Trichloroethane
Concen: 0.485 ug/l
RT: 5.306 min Scan# 1249
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 97 Resp: 10801

Ion	Ratio	Lower	Upper
97	100		
99	63.2	32.4	97.0
61	45.5	23.8	71.2





#29
1,1-Dichloropropene
Concen: 0.460 ug/l
RT: 5.518 min Scan# 1
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 75 Resp: 917

Ion	Ratio	Lower	Upper
75	100		
110	34.4	17.2	51.5
77	29.6	24.6	37.0

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#30
Carbon Tetrachloride
Concen: 0.485 ug/l
RT: 5.515 min Scan# 1314
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:117 Resp: 9251

Ion	Ratio	Lower	Upper
117	100		
119	88.9	76.7	115.1
121	27.7	25.5	38.3





#31
Isopropyl Ether
Concen: 0.461 ug/l
RT: 3.978 min Scan# 8
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 45 Resp: 1790

Ion	Ratio	Lower	Upper
45	100		
43	53.0	25.7	77.1



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#32
Ethyl-t-butyl ether
Concen: 0.458 ug/l
RT: 4.492 min Scan# 996
Delta R.T. 0.010 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 59 Resp: 16177

Ion	Ratio	Lower	Upper
59	100		
87	41.6	32.6	49.0





#33
Tert-Amyl methyl ether
Concen: 0.436 ug/l
RT: 5.930 min Scan# 14
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 73 Resp: 1347

Ion	Ratio	Lower	Upper
73	100		
87	24.1	18.9	28.3
71	12.0	9.0	13.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#34
Propionitrile
Concen: 1.809 ug/l m
RT: 4.801 min Scan# 1092
Delta R.T. 0.048 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 54 Resp: 2034

Ion	Ratio	Lower	Upper
54	100		
52	6.2	16.3	24.5#
55	3.5	11.8	17.6#
40	2.6	4.3	6.5#





#35
Benzene
Concen: 0.473 ug/l
RT: 5.769 min Scan# 1
Delta R.T. 0.010 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 78 Resp: 2901

Ion	Ratio	Lower	Upper
78	100		
77	22.6	19.0	28.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#36
1,2-Dichloroethane
Concen: 0.494 ug/l
RT: 5.791 min Scan# 1400
Delta R.T. 0.013 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 62 Resp: 8750

Ion	Ratio	Lower	Upper
62	100		
98	6.9	6.9	10.3#





#37
Trichloroethene
Concen: 0.490 ug/l
RT: 6.537 min Scan# 1
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:130 Resp: 714

Ion	Ratio	Lower	Upper
130	100		
95	98.3	83.2	124.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#38
1,2-Dichloropropane
Concen: 0.489 ug/l
RT: 6.782 min Scan# 1708
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 63 Resp: 7842

Ion	Ratio	Lower	Upper
63	100		
65	30.5	25.3	37.9





#39
Methacrylonitrile
Concen: 0.382 ug/l m
RT: 5.017 min Scan# 1
Delta R.T. 0.058 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 41 Resp: 153

Ion	Ratio	Lower	Upper
41	100		
67	14.5	64.5	96.7#
39	0.0	43.0	64.6#
52	0.0	25.3	37.9#



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#40
Methyl acrylate
Concen: 0.403 ug/l m
RT: 4.920 min Scan# 1129
Delta R.T. 0.087 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 55 Resp: 2955

Ion	Ratio	Lower	Upper
55	100		
85	0.0	13.3	19.9#
58	0.0	7.3	10.9#
42	0.0	6.9	10.3#





#41
Tetrahydrofuran
Concen: 0.970 ug/l
RT: 5.065 min Scan# 11
Delta R.T. 0.026 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 42 Resp: 228

Ion	Ratio	Lower	Upper
42	100		
72	18.8	41.5	62.3#
71	23.0	37.2	55.8#



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#42
1-Chlorobutane
Concen: 0.443 ug/l
RT: 5.451 min Scan# 1294
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 56 Resp: 12081

Ion	Ratio	Lower	Upper
56	100		
41	58.0	26.3	78.8





#43
Dibromomethane
Concen: 0.495 ug/l
RT: 6.910 min Scan# 1
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 93 Resp: 402

Ion	Ratio	Lower	Upper
93	100		
95	82.9	67.2	100.8
174	84.6	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#44
Bromodichloromethane
Concen: 0.438 ug/l
RT: 7.100 min Scan# 1807
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 83 Resp: 8280

Ion	Ratio	Lower	Upper
83	100		
85	65.2	51.7	77.5
127	10.9	6.7	10.1#





#45
4-Methyl-2-Pentanone
Concen: 2.160 ug/l
RT: 7.785 min Scan# 2
Delta R.T. 0.010 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 43 Resp: 1839

Ion	Ratio	Lower	Upper
43	100		
58	36.0	20.8	62.5
85	14.1	14.8	22.2#



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#46
t-1,4-Dichloro-2-butene
Concen: 0.839 ug/l m
RT: 10.823 min Scan# 2965
Delta R.T. -0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 75 Resp: 3537

Ion	Ratio	Lower	Upper
75	100		
53	59.0	64.5	96.7#
89	37.1	43.4	65.2#
88	23.4	31.2	46.8#





#47
Methyl methacrylate
Concen: 0.815 ug/l m
RT: 6.965 min Scan# 1
Delta R.T. 0.016 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 69 Resp: 558

Ion	Ratio	Lower	Upper
69	100		
41	97.2	0.0	217.0
39	25.7	47.7	71.5#
100	30.5	27.4	41.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#48
Ethyl methacrylate
Concen: 0.377 ug/l
RT: 8.332 min Scan# 2190
Delta R.T. 0.010 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 69 Resp: 4839

Ion	Ratio	Lower	Upper
69	100		
41	68.2	30.6	92.0





#49
Toluene
Concen: 0.429 ug/l
RT: 7.962 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 92 Resp: 1513

Ion	Ratio	Lower	Upper
92	100		
91	173.4	141.8	212.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#50
t-1,3-Dichloropropene
Concen: 0.409 ug/l
RT: 8.209 min Scan# 2152
Delta R.T. 0.010 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 75 Resp: 7086

Ion	Ratio	Lower	Upper
75	100		
77	32.1	25.9	38.9





#51
cis-1,3-Dichloropropene
Concen: 0.441 ug/l
RT: 7.605 min Scan# 1
Delta R.T. 0.010 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 75 Resp: 942

Ion	Ratio	Lower	Upper
75	100		
77	35.5	25.3	37.9
39	48.8	33.5	50.3



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#52
1,1,2-Trichloroethane
Concen: 0.480 ug/l
RT: 8.393 min Scan# 2209
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 97 Resp: 5257

Ion	Ratio	Lower	Upper
97	100		
83	83.1	73.0	109.4
85	60.5	46.3	69.5
99	63.2	48.5	72.7





#53
1,3-Dichloropropane
Concen: 0.454 ug/l
RT: 8.566 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 76 Resp: 883

Ion	Ratio	Lower	Upper
76	100		
78	30.9	26.3	39.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#54
2-Hexanone
Concen: 2.011 ug/l m
RT: 8.685 min Scan# 2300
Delta R.T. 0.013 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 43 Resp: 11692

Ion	Ratio	Lower	Upper
43	100		
58	44.7	38.0	78.0
57	10.5	0.0	39.1





#55
Dibromochloromethane
Concen: 0.471 ug/l
RT: 8.801 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:129 Resp: 593

Ion	Ratio	Lower	Upper
129	100		
127	73.0	60.4	90.6



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#56
1,2-Dibromoethane
Concen: 0.470 ug/l
RT: 8.917 min Scan# 2372
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:107 Resp: 4834

Ion	Ratio	Lower	Upper
107	100		
109	84.1	0.0	187.8





#57
4-Bromofluorobenzene
Concen: 0.917 ug/l
RT: 10.627 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 95 Resp: 1509

Ion	Ratio	Lower	Upper
95	100		
174	78.4	58.6	88.0
176	77.1	58.2	87.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#58
Tetrachloroethene
Concen: 0.503 ug/l
RT: 8.544 min Scan# 2256
Delta R.T. -0.000 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:164 Resp: 6044

Ion	Ratio	Lower	Upper
164	100		
166	147.2	101.4	152.0
129	105.0	77.0	115.4
131	103.2	76.3	114.5





#59
Chlorobenzene
Concen: 0.453 ug/l
RT: 9.441 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:112 Resp: 1687

Ion	Ratio	Lower	Upper
112	100		
114	30.0	25.7	38.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#60
1,1,1,2-Tetrachloroethane
Concen: 0.475 ug/l
RT: 9.524 min Scan# 2561
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:131 Resp: 6349

Ion	Ratio	Lower	Upper
131	100		
133	92.8	76.7	115.1
119	69.5	54.4	81.6





#61
Pentachloroethane
Concen: 0.436 ug/l
RT: 11.418 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:117 Resp: 521

Ion	Ratio	Lower	Upper
117	100		
167	82.8	65.0	97.6



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#62
Hexachloroethane
Concen: 0.423 ug/l
RT: 12.466 min Scan# 3476
Delta R.T. -0.000 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:117 Resp: 4467

Ion	Ratio	Lower	Upper
117	100		
201	79.0	61.3	91.9





#63
Ethyl Benzene
Concen: 0.418 ug/l
RT: 9.563 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 91 Resp: 2681

Ion	Ratio	Lower	Upper
91	100		
106	28.8	24.2	36.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#64
m/p-Xylenes
Concen: 0.778 ug/l
RT: 9.685 min Scan# 2611
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:106 Resp: 18658

Ion	Ratio	Lower	Upper
106	100		
91	201.9	166.9	250.3





#65
o-Xylene
Concen: 0.398 ug/l
RT: 10.094 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:106 Resp: 932

Ion	Ratio	Lower	Upper
106	100		
91	220.7	110.9	332.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#66
Styrene
Concen: 0.375 ug/l
RT: 10.113 min Scan# 2744
Delta R.T. 0.010 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:104 Resp: 14012

Ion	Ratio	Lower	Upper
104	100		
78	48.9	41.2	61.8
103	56.6	44.8	67.2





#67
Bromoform
Concen: 0.442 ug/l
RT: 10.283 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:173 Resp: 316

Ion	Ratio	Lower	Upper
173	100		
175	44.9	39.0	58.4
254	8.4	7.7	11.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#68
1,2-Dichlorobenzene-d4
Concen: 0.914 ug/l
RT: 12.187 min Scan# 3389
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:152 Resp: 15652

Ion	Ratio	Lower	Upper
152	100		
115	68.2	0.0	275.2
150	166.4	0.0	658.4





#69
Isopropylbenzene
Concen: 0.402 ug/l
RT: 10.476 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:105 Resp: 2219

Ion	Ratio	Lower	Upper
105	100		
120	25.3	12.8	38.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#70
1,1,2,2-Tetrachloroethane
Concen: 0.440 ug/l
RT: 10.772 min Scan# 2949
Delta R.T. -0.000 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 83 Resp: 6493

Ion	Ratio	Lower	Upper
83	100		
131	8.6	7.4	11.0
85	67.6	51.8	77.8





#71
1,2,3-Trichloropropane
Concen: 0.490 ug/l m
RT: 10.817 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 75 Resp: 548

Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0
110	27.4	0.0	77.0
97	18.0	0.0	42.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#72
Bromobenzene
Concen: 0.424 ug/l
RT: 10.778 min Scan# 2951
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:156 Resp: 6302

Ion	Ratio	Lower	Upper
156	100		
77	186.6	0.0	343.6
158	98.3	0.0	193.0





#73
n-propylbenzene
Concen: 0.382 ug/l
RT: 10.901 min Scan# 2
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:120 Resp: 602

Ion	Ratio	Lower	Upper
120	100		
91	469.0	369.8	554.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#74
2-Chlorotoluene
Concen: 0.392 ug/l
RT: 10.981 min Scan# 3014
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:126 Resp: 5699

Ion	Ratio	Lower	Upper
126	100		
91	306.6	0.0	623.8





#75
1,3,5-Trimethylbenzene
Concen: 0.362 ug/l
RT: 11.081 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:105 Resp: 1848

Ion	Ratio	Lower	Upper
105	100		
120	48.0	37.3	55.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#76
4-Chlorotoluene
Concen: 0.396 ug/l
RT: 11.093 min Scan# 3049
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:126 Resp: 5912

Ion	Ratio	Lower	Upper
126	100		
91	326.4	0.0	703.6





#77
tert-Butylbenzene
Concen: 0.412 ug/l
RT: 11.412 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:119 Resp: 2130

Ion	Ratio	Lower	Upper
119	100		
91	54.7	29.4	88.3
134	20.4	17.6	26.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#78
1,2,4-Trimethylbenzene
Concen: 0.349 ug/l
RT: 11.460 min Scan# 3163
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:105 Resp: 17694

Ion	Ratio	Lower	Upper
105	100		
120	47.2	21.9	65.7





#79
sec-Butylbenzene
Concen: 0.385 ug/l
RT: 11.637 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:105 Resp: 2535

Ion	Ratio	Lower	Upper
105	100		
134	18.9	15.1	22.7

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#81
p-Isopropyltoluene
Concen: 0.358 ug/l
RT: 11.785 min Scan# 3264
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:119 Resp: 18576

Ion	Ratio	Lower	Upper
119	100		
134	26.6	20.3	30.5
91	24.7	19.4	29.2





#82
1,3-Dichlorobenzene
Concen: 0.460 ug/l
RT: 11.740 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:146 Resp: 1325

Ion	Ratio	Lower	Upper
146	100		
111	37.7	32.8	49.2
148	62.2	51.1	76.7

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#83
1,4-Dichlorobenzene
Concen: 0.429 ug/l
RT: 11.833 min Scan# 3279
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:146 Resp: 12096

Ion	Ratio	Lower	Upper
146	100		
111	40.6	32.1	48.1
148	63.6	50.2	75.4





#84
n-Butylbenzene
Concen: 0.369 ug/l
RT: 12.206 min Scan# 3
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 91 Resp: 1729

Ion	Ratio	Lower	Upper
91	100		
92	47.1	41.8	62.8
134	23.1	18.6	28.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#85
1,2-Dichlorobenzene
Concen: 0.433 ug/l
RT: 12.209 min Scan# 3396
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:146 Resp: 11996

Ion	Ratio	Lower	Upper
146	100		
111	44.2	21.9	65.7
148	67.4	32.3	96.9





#86
1,2-Dibromo-3-Chloropropane
Concen: 0.376 ug/l
RT: 12.997 min Scan# 3
Delta R.T. 0.010 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion: 75 Resp: 77

Ion	Ratio	Lower	Upper
75	100		
155	72.8	63.5	95.3
157	91.0	81.8	122.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#87
1,2,4-Trichlorobenzene
Concen: 0.380 ug/l
RT: 13.839 min Scan# 3903
Delta R.T. 0.010 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:180 Resp: 5130

Ion	Ratio	Lower	Upper
180	100		
182	103.3	76.6	115.0
145	30.3	25.4	38.2





#88
Hexachlorobutadiene
Concen: 0.482 ug/l
RT: 14.013 min Scan# 3
Delta R.T. -0.000 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:225 Resp: 465

Ion	Ratio	Lower	Upper
225	100		
223	62.8	49.5	74.3
227	58.2	51.0	76.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#89
Naphthalene
Concen: 0.899 ug/l
RT: 14.097 min Scan# 3983
Delta R.T. 0.019 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:128 Resp: 7859

Ion	Ratio	Lower	Upper
128	100		
127	8.9	10.6	16.0#
129	11.5	8.6	13.0





#90
1,2,3-Trichlorobenzene
Concen: 0.412 ug/l
RT: 14.328 min Scan# 4
Delta R.T. 0.006 min
Lab File: VU063219.D
Acq: 10 Feb 2025 12:59

Tgt Ion:180 Resp: 545

Ion	Ratio	Lower	Upper
180	100		
182	96.5	78.2	117.2
145	30.3	26.1	39.1



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC0.5

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063220.D
Acq On : 10 Feb 2025 13:23
Operator : MD/SY
Sample : VSTDICC001
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 03:57:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Fluorobenzene	6.107	96	51567	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.627	95	16040	0.943	ug/l	0.00
Spiked Amount 1.000			Recovery =	94.000%		
68) 1,2-Dichlorobenzene-d4	12.187	152	16904	0.956	ug/l	0.00
Spiked Amount 1.000			Recovery =	96.000%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.380	85	18689	1.115	ug/l	97
3) Chloromethane	1.518	50	20645	1.070	ug/l	98
4) Vinyl Chloride	1.599	62	20135	1.055	ug/l	99
5) Bromomethane	1.846	94	10384	1.176	ug/l	96
6) Chloroethane	1.924	64	13381	1.113	ug/l	95
7) Trichlorofluoromethane	2.129	101	24378	1.078	ug/l	98
8) 1,1,2-Trichloro-1,2,2-...	2.567	101	14009	1.091	ug/l	98
9) 1,1-Dichloroethene	2.570	96	14118	1.079	ug/l	97
10) Iodomethane	2.711	142	21185	1.030	ug/l	98
11) Allyl Chloride	2.914	41	20123	1.071	ug/l	97
12) Acrylonitrile	3.322	53	5799	1.950	ug/l	99
13) Acetone	2.618	43	12193	5.255	ug/l	98
14) Carbon Disulfide	2.782	76	48738	1.065	ug/l	99
15) Methylene Chloride	3.033	84	17173	1.063	ug/l	95
16) trans-1,2-Dichloroethene	3.345	96	15731	1.054	ug/l	98
17) 1,1-Dichloroethane	3.856	63	30538	1.085	ug/l	98
18) 2-Butanone	4.711	43	18530m	4.944	ug/l	
19) Cyclohexane	5.374	56	21832m	0.965	ug/l	
20) Methylcyclohexane	6.753	83	23086	1.029	ug/l	99
21) 2,2-Dichloropropane	4.650	77	23852	1.086	ug/l	98
22) cis-1,2-Dichloroethene	4.657	96	17134	1.062	ug/l	98
23) Diethyl Ether	2.367	59	11921	1.062	ug/l	95
24) tert-Butyl Alcohol	3.165	59	19963	11.656	ug/l #	89
25) Methyl tert-Butyl Ether	3.351	73	33609	1.029	ug/l	98
26) Bromochloromethane	4.962	128	7742	1.098	ug/l	96
27) Chloroform	5.075	83	30302	1.067	ug/l	94
28) 1,1,1-Trichloroethane	5.300	97	24038	1.045	ug/l	96
29) 1,1-Dichloropropene	5.518	75	21427	1.040	ug/l	99
30) Carbon Tetrachloride	5.512	117	21505	1.090	ug/l	90
31) Isopropyl Ether	3.978	45	40509	1.008	ug/l	96
32) Ethyl-t-butyl ether	4.486	59	35372	0.968	ug/l	98
33) Tert-Amyl methyl ether	5.927	73	31129	0.975	ug/l	99
34) Propionitrile	4.782	54	5038m	4.335	ug/l	
35) Benzene	5.766	78	65980	1.041	ug/l	99
36) 1,2-Dichloroethane	5.788	62	19157	1.048	ug/l	97
37) Trichloroethene	6.534	130	16013	1.063	ug/l	97
38) 1,2-Dichloropropane	6.782	63	17158	1.035	ug/l	93
39) Methacrylonitrile	4.978	41	3478	0.838	ug/l #	84
40) Methyl acrylate	4.866	55	7528m	0.993	ug/l	
41) Tetrahydrofuran	5.049	42	4693	1.924	ug/l	91
42) 1-Chlorobutane	5.448	56	28598	1.014	ug/l	99
43) Dibromomethane	6.907	93	8944	1.065	ug/l	93
44) Bromodichloromethane	7.094	83	20720	1.060	ug/l	98

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063220.D
Acq On : 10 Feb 2025 13:23
Operator : MD/SY
Sample : VSTDICC001
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 03:57:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
45) 4-Methyl-2-Pentanone	7.782	43	39981	4.541 ug/l	96
46) t-1,4-Dichloro-2-butene	10.820	75	8630m	1.981 ug/l	
47) Methyl methacrylate	6.956	69	12987	1.836 ug/l	98
48) Ethyl methacrylate	8.328	69	11711	0.882 ug/l	100
49) Toluene	7.962	92	36357	0.998 ug/l	100
50) t-1,3-Dichloropropene	8.206	75	17681	0.988 ug/l	98
51) cis-1,3-Dichloropropene	7.599	75	22033	0.997 ug/l	99
52) 1,1,2-Trichloroethane	8.393	97	11610	1.025 ug/l	97
53) 1,3-Dichloropropane	8.566	76	21023	1.046 ug/l	99
54) 2-Hexanone	8.679	43	28078m	4.673 ug/l	
55) Dibromochloromethane	8.801	129	13259	1.018 ug/l	98
56) 1,2-Dibromoethane	8.917	107	10628	1.001 ug/l	94
58) Tetrachloroethene	8.544	164	13296	1.071 ug/l	95
59) Chlorobenzene	9.441	112	38774	1.008 ug/l	99
60) 1,1,1,2-Tetrachloroethane	9.525	131	14393	1.041 ug/l	99
61) Pentachloroethane	11.418	117	12686	1.027 ug/l	94
62) Hexachloroethane	12.463	117	11074	1.014 ug/l	93
63) Ethyl Benzene	9.560	91	63398	0.956 ug/l	99
64) m/p-Xylenes	9.685	106	45748	1.847 ug/l	99
65) o-Xylene	10.090	106	23138	0.954 ug/l	100
66) Styrene	10.110	104	34785	0.901 ug/l	99
67) Bromoform	10.280	173	7192	0.973 ug/l #	97
69) Isopropylbenzene	10.473	105	53372	0.936 ug/l	98
70) 1,1,2,2-Tetrachloroethane	10.772	83	15680	1.027 ug/l	97
71) 1,2,3-Trichloropropane	10.814	75	11462m	0.991 ug/l	
72) Bromobenzene	10.775	156	15611	1.016 ug/l	95
73) n-propylbenzene	10.898	120	14761	0.904 ug/l	95
74) 2-Chlorotoluene	10.975	126	14317	0.952 ug/l	95
75) 1,3,5-Trimethylbenzene	11.078	105	48403	0.916 ug/l	99
76) 4-Chlorotoluene	11.090	126	15039	0.975 ug/l	89
77) tert-Butylbenzene	11.409	119	50914	0.953 ug/l	98
78) 1,2,4-Trimethylbenzene	11.460	105	46689	0.891 ug/l	100
79) sec-Butylbenzene	11.634	105	63226	0.930 ug/l	99
80) Nitrobenzene	13.232	77	1014	5.911 ug/l #	57
81) p-Isopropyltoluene	11.785	119	48578	0.905 ug/l	99
82) 1,3-Dichlorobenzene	11.737	146	28406	0.953 ug/l	98
83) 1,4-Dichlorobenzene	11.830	146	26529	0.910 ug/l	94
84) n-Butylbenzene	12.200	91	42404	0.876 ug/l	97
85) 1,2-Dichlorobenzene	12.203	146	27742	0.969 ug/l	99
86) 1,2-Dibromo-3-Chloropr...	12.997	75	2103	0.982 ug/l	89
87) 1,2,4-Trichlorobenzene	13.836	180	12695	0.909 ug/l	99
88) Hexachlorobutadiene	14.010	225	10179	1.020 ug/l	97
89) Naphthalene	14.090	128	17271	1.192 ug/l	97
90) 1,2,3-Trichlorobenzene	14.325	180	12218	0.894 ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063220.D
Acq On : 10 Feb 2025 13:23
Operator : MD/SY
Sample : VSTDICC001
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Feb 11 03:57:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.107 min Scan# 14
Delta R.T. 0.004 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 96 Resp: 5156

Ion	Ratio	Lower	Upper
96	100		
70	18.8	15.6	23.4
95	9.2	7.3	10.9
97	6.3	0.0	0.0#



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#2
Dichlorodifluoromethane
Concen: 1.115 ug/l
RT: 1.380 min Scan# 28
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 85 Resp: 18689

Ion	Ratio	Lower	Upper
85	100		
87	33.8	16.0	48.0





#3
Chloromethane
Concen: 1.070 ug/l
RT: 1.518 min Scan# 7
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 50 Resp: 2064

Ion	Ratio	Lower	Upper
50	100		
52	31.5	25.8	38.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#4
Vinyl Chloride
Concen: 1.055 ug/l
RT: 1.599 min Scan# 96
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 62 Resp: 20135

Ion	Ratio	Lower	Upper
62	100		
64	32.5	25.4	38.0





#5
Bromomethane
Concen: 1.176 ug/l
RT: 1.846 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 94 Resp: 1038

Ion	Ratio	Lower	Upper
94	100		
96	90.4	75.7	113.5

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#6
Chloroethane
Concen: 1.113 ug/l
RT: 1.924 min Scan# 197
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 64 Resp: 13381

Ion	Ratio	Lower	Upper
64	100		
66	29.6	25.8	38.8





#7
Trichlorofluoromethane
Concen: 1.078 ug/l
RT: 2.129 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:101 Resp: 2437

Ion	Ratio	Lower	Upper
101	100		
103	66.0	51.4	77.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 1.091 ug/l
RT: 2.567 min Scan# 397
Delta R.T. -0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:101 Resp: 14009

Ion	Ratio	Lower	Upper
101	100		
85	44.4	35.4	53.0
151	75.2	58.5	87.7





#9
1,1-Dichloroethene
Concen: 1.079 ug/l
RT: 2.570 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 96 Resp: 1411

Ion	Ratio	Lower	Upper
96	100		
61	160.1	0.0	492.9
98	65.4	0.0	124.0

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#10
Iodomethane
Concen: 1.030 ug/l
RT: 2.711 min Scan# 442
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:142 Resp: 21185

Ion	Ratio	Lower	Upper
142	100		
127	44.1	34.5	51.7





#11
Allyl Chloride
Concen: 1.071 ug/l
RT: 2.914 min Scan# 5
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 41 Resp: 2012

Ion	Ratio	Lower	Upper
41	100		
39	70.1	57.9	86.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#12
Acrylonitrile
Concen: 1.950 ug/l
RT: 3.322 min Scan# 632
Delta R.T. 0.023 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 53 Resp: 5799

Ion	Ratio	Lower	Upper
53	100		
52	79.0	64.2	96.2
51	38.4	30.8	46.2





#13
Acetone
Concen: 5.255 ug/l
RT: 2.618 min Scan# 41
Delta R.T. 0.006 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 43 Resp: 1219

Ion	Ratio	Lower	Upper
43	100		
58	35.6	27.4	41.0

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#14
Carbon Disulfide
Concen: 1.065 ug/l
RT: 2.782 min Scan# 464
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 76 Resp: 48738

Ion	Ratio	Lower	Upper
76	100		
78	9.4	7.2	10.8





#15
Methylene Chloride
Concen: 1.063 ug/l
RT: 3.033 min Scan# 54
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 84 Resp: 1717

Ion	Ratio	Lower	Upper
84	100		
49	126.9	96.4	144.6
51	39.2	0.0	74.8
86	67.6	51.3	76.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#16
trans-1,2-Dichloroethene
Concen: 1.054 ug/l
RT: 3.345 min Scan# 639
Delta R.T. 0.006 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 96 Resp: 15731

Ion	Ratio	Lower	Upper
96	100		
61	140.9	113.4	170.2
98	61.5	51.2	76.8





#17
1,1-Dichloroethane
Concen: 1.085 ug/l
RT: 3.856 min Scan# 7
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 63 Resp: 3053

Ion	Ratio	Lower	Upper
63	100		
98	5.6	3.3	9.9
100	3.8	2.1	6.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#18
2-Butanone
Concen: 4.944 ug/l m
RT: 4.711 min Scan# 1064
Delta R.T. 0.026 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 43 Resp: 18530

Ion	Ratio	Lower	Upper
43	100		
57	8.7	0.0	17.0





#19
Cyclohexane
Concen: 0.965 ug/l m
RT: 5.374 min Scan# 1
Delta R.T. -0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 56 Resp: 2183

Ion	Ratio	Lower	Upper
56	100		
69	33.0	24.5	36.7
84	92.3	65.2	97.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#20
Methylcyclohexane
Concen: 1.029 ug/l
RT: 6.753 min Scan# 1699
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 83 Resp: 23086

Ion	Ratio	Lower	Upper
83	100		
55	80.3	63.1	94.7
98	44.1	35.2	52.8





#21
2,2-Dichloropropane
Concen: 1.086 ug/l
RT: 4.650 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 77 Resp: 2385

Ion	Ratio	Lower	Upper
77	100		
97	22.2	18.5	27.7



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#22
cis-1,2-Dichloroethene
Concen: 1.062 ug/l
RT: 4.657 min Scan# 1047
Delta R.T. 0.007 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 96 Resp: 17134

Ion	Ratio	Lower	Upper
96	100		
61	152.5	0.0	373.3
98	64.7	31.9	95.9





#23
Diethyl Ether
Concen: 1.062 ug/l
RT: 2.367 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 59 Resp: 1192

Ion	Ratio	Lower	Upper
59	100		
45	79.7	57.8	86.6
74	73.1	57.7	86.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#24
tert-Butyl Alcohol
Concen: 11.656 ug/l
RT: 3.165 min Scan# 583
Delta R.T. 0.007 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 59 Resp: 19963

Ion	Ratio	Lower	Upper
59	100		
57	5.3	7.5	11.3#





#25
Methyl tert-Butyl Ether
Concen: 1.029 ug/l
RT: 3.351 min Scan# 64
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 73 Resp: 3360

Ion	Ratio	Lower	Upper
73	100		
43	19.3	14.9	22.3

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#26
Bromochloromethane
Concen: 1.098 ug/l
RT: 4.962 min Scan# 1142
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:128 Resp: 7742

Ion	Ratio	Lower	Upper
128	100		
49	164.2	0.0	343.4
130	125.4	102.9	154.3





#27
Chloroform
Concen: 1.067 ug/l
RT: 5.075 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 83 Resp: 3030

Ion	Ratio	Lower	Upper
83	100		
85	69.4	0.0	129.8

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#28
1,1,1-Trichloroethane
Concen: 1.045 ug/l
RT: 5.300 min Scan# 1247
Delta R.T. -0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 97 Resp: 24038

Ion	Ratio	Lower	Upper
97	100		
99	63.6	32.4	97.0
61	51.8	23.8	71.2





#29
1,1-Dichloropropene
Concen: 1.040 ug/l
RT: 5.518 min Scan# 1
Delta R.T. 0.006 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 75 Resp: 2142

Ion	Ratio	Lower	Upper
75	100		
110	33.1	17.2	51.5
77	30.5	24.6	37.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#30
Carbon Tetrachloride
Concen: 1.090 ug/l
RT: 5.512 min Scan# 1313
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:117 Resp: 21505

Ion	Ratio	Lower	Upper
117	100		
119	86.4	76.7	115.1
121	26.7	25.5	38.3





#31
Isopropyl Ether
Concen: 1.008 ug/l
RT: 3.978 min Scan# 8
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 45 Resp: 4050

Ion	Ratio	Lower	Upper
45	100		
43	54.0	25.7	77.1



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#32
Ethyl-t-butyl ether
Concen: 0.968 ug/l
RT: 4.486 min Scan# 994
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 59 Resp: 35372

Ion	Ratio	Lower	Upper
59	100		
87	42.1	32.6	49.0





#33
Tert-Amyl methyl ether
Concen: 0.975 ug/l
RT: 5.927 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 73 Resp: 3112

Ion	Ratio	Lower	Upper
73	100		
87	23.3	18.9	28.3
71	10.9	9.0	13.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#34
Propionitrile
Concen: 4.335 ug/l m
RT: 4.782 min Scan# 1086
Delta R.T. 0.029 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 54 Resp: 5038

Ion	Ratio	Lower	Upper
54	100		
52	16.5	16.3	24.5
55	7.3	11.8	17.6#
40	2.7	4.3	6.5#





#35
Benzene
Concen: 1.041 ug/l
RT: 5.766 min Scan# 1
Delta R.T. 0.006 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 78 Resp: 6598

Ion	Ratio	Lower	Upper
78	100		
77	24.3	19.0	28.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#36
1,2-Dichloroethane
Concen: 1.048 ug/l
RT: 5.788 min Scan# 1399
Delta R.T. 0.010 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 62 Resp: 19157

Ion	Ratio	Lower	Upper
62	100		
98	7.4	6.9	10.3





#37
Trichloroethene
Concen: 1.063 ug/l
RT: 6.534 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:130 Resp: 1601

Ion	Ratio	Lower	Upper
130	100		
95	106.9	83.2	124.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#38
1,2-Dichloropropane
Concen: 1.035 ug/l
RT: 6.782 min Scan# 1708
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 63 Resp: 17158

Ion	Ratio	Lower	Upper
63	100		
65	28.0	25.3	37.9





#39
Methacrylonitrile
Concen: 0.838 ug/l
RT: 4.978 min Scan# 1
Delta R.T. 0.019 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 41 Resp: 347

Ion	Ratio	Lower	Upper
41	100		
67	101.4	64.5	96.7#
39	61.6	43.0	64.6
52	33.5	25.3	37.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#40
Methyl acrylate
Concen: 0.993 ug/l m
RT: 4.866 min Scan# 1112
Delta R.T. 0.032 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 55 Resp: 7528

Ion	Ratio	Lower	Upper
55	100		
85	0.0	13.3	19.9#
58	2.3	7.3	10.9#
42	3.9	6.9	10.3#





#41
Tetrahydrofuran
Concen: 1.924 ug/l
RT: 5.049 min Scan# 11
Delta R.T. 0.010 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 42 Resp: 469

Ion	Ratio	Lower	Upper
42	100		
72	47.9	41.5	62.3
71	37.6	37.2	55.8

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#42
1-Chlorobutane
Concen: 1.014 ug/l
RT: 5.448 min Scan# 1293
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 56 Resp: 28598

Ion	Ratio	Lower	Upper
56	100		
41	51.7	26.3	78.8





#43
Dibromomethane
Concen: 1.065 ug/l
RT: 6.907 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 93 Resp: 894

Ion	Ratio	Lower	Upper
93	100		
95	81.7	67.2	100.8
174	83.3	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#44
Bromodichloromethane
Concen: 1.060 ug/l
RT: 7.094 min Scan# 1805
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 83 Resp: 20720

Ion	Ratio	Lower	Upper
83	100		
85	66.3	51.7	77.5
127	9.2	6.7	10.1





#45
4-Methyl-2-Pentanone
Concen: 4.541 ug/l
RT: 7.782 min Scan# 2
Delta R.T. 0.006 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 43 Resp: 3998

Ion	Ratio	Lower	Upper
43	100		
58	38.7	20.8	62.5
85	17.7	14.8	22.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#46
t-1,4-Dichloro-2-butene
Concen: 1.981 ug/l m
RT: 10.820 min Scan# 2964
Delta R.T. -0.006 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 75 Resp: 8630

Ion	Ratio	Lower	Upper
75	100		
53	55.7	64.5	96.7#
89	36.6	43.4	65.2#
88	27.5	31.2	46.8#





#47
Methyl methacrylate
Concen: 1.836 ug/l
RT: 6.956 min Scan# 1
Delta R.T. 0.006 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 69 Resp: 1298

Ion	Ratio	Lower	Upper
69	100		
41	109.0	0.0	217.0
39	62.7	47.7	71.5
100	33.7	27.4	41.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#48
Ethyl methacrylate
Concen: 0.882 ug/l
RT: 8.328 min Scan# 2189
Delta R.T. 0.007 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 69 Resp: 11711

Ion	Ratio	Lower	Upper
69	100		
41	61.3	30.6	92.0





#49
Toluene
Concen: 0.998 ug/l
RT: 7.962 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 92 Resp: 3635

Ion	Ratio	Lower	Upper
92	100		
91	176.9	141.8	212.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#50
t-1,3-Dichloropropene
Concen: 0.988 ug/l
RT: 8.206 min Scan# 2151
Delta R.T. 0.007 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 75 Resp: 17681

Ion	Ratio	Lower	Upper
75	100		
77	31.4	25.9	38.9





#51
cis-1,3-Dichloropropene
Concen: 0.997 ug/l
RT: 7.599 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 75 Resp: 2203

Ion	Ratio	Lower	Upper
75	100		
77	32.0	25.3	37.9
39	41.5	33.5	50.3



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#52
1,1,2-Trichloroethane
Concen: 1.025 ug/l
RT: 8.393 min Scan# 2209
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 97 Resp: 11610

Ion	Ratio	Lower	Upper
97	100		
83	91.6	73.0	109.4
85	59.6	46.3	69.5
99	67.2	48.5	72.7





#53
1,3-Dichloropropane
Concen: 1.046 ug/l
RT: 8.566 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 76 Resp: 2102

Ion	Ratio	Lower	Upper
76	100		
78	32.5	26.3	39.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#54
2-Hexanone
Concen: 4.673 ug/l m
RT: 8.679 min Scan# 2298
Delta R.T. 0.006 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 43 Resp: 28078

Ion	Ratio	Lower	Upper
43	100		
58	46.0	38.0	78.0
57	16.4	0.0	39.1





#55
Dibromochloromethane
Concen: 1.018 ug/l
RT: 8.801 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:129 Resp: 1325

Ion	Ratio	Lower	Upper
129	100		
127	77.0	60.4	90.6



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#56
1,2-Dibromoethane
Concen: 1.001 ug/l
RT: 8.917 min Scan# 2372
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:107 Resp: 10628

Ion	Ratio	Lower	Upper
107	100		
109	99.6	0.0	187.8





#57
4-Bromofluorobenzene
Concen: 0.943 ug/l
RT: 10.627 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Tgt Ion: 95 Resp: 1604

Ion	Ratio	Lower	Upper
95	100		
174	84.4	58.6	88.0
176	78.3	58.2	87.4

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#58
Tetrachloroethene
Concen: 1.071 ug/l
RT: 8.544 min Scan# 2256
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:164 Resp: 13296

Ion	Ratio	Lower	Upper
164	100		
166	134.4	101.4	152.0
129	102.9	77.0	115.4
131	95.7	76.3	114.5





#59
Chlorobenzene
Concen: 1.008 ug/l
RT: 9.441 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:112 Resp: 3877

Ion	Ratio	Lower	Upper
112	100		
114	31.7	25.7	38.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#60
1,1,1,2-Tetrachloroethane
Concen: 1.041 ug/l
RT: 9.525 min Scan# 2561
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:131 Resp: 14393

Ion	Ratio	Lower	Upper
131	100		
133	95.6	76.7	115.1
119	69.6	54.4	81.6





#61
Pentachloroethane
Concen: 1.027 ug/l
RT: 11.418 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:117 Resp: 1268

Ion	Ratio	Lower	Upper
117	100		
167	76.1	65.0	97.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#62
Hexachloroethane
Concen: 1.014 ug/l
RT: 12.463 min Scan# 3475
Delta R.T. -0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:117 Resp: 11074

Ion	Ratio	Lower	Upper
117	100		
201	70.2	61.3	91.9





#63
Ethyl Benzene
Concen: 0.956 ug/l
RT: 9.560 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 91 Resp: 6339

Ion	Ratio	Lower	Upper
91	100		
106	30.7	24.2	36.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#64
m/p-Xylenes
Concen: 1.847 ug/l
RT: 9.685 min Scan# 2611
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:106 Resp: 45748

Ion	Ratio	Lower	Upper
106	100		
91	207.1	166.9	250.3





#65
o-Xylene
Concen: 0.954 ug/l
RT: 10.090 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:106 Resp: 2313

Ion	Ratio	Lower	Upper
106	100		
91	221.5	110.9	332.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#66
Styrene
Concen: 0.901 ug/l
RT: 10.110 min Scan# 2743
Delta R.T. 0.007 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:104 Resp: 34785

Ion	Ratio	Lower	Upper
104	100		
78	51.5	41.2	61.8
103	57.5	44.8	67.2





#67
Bromoform
Concen: 0.973 ug/l
RT: 10.280 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:173 Resp: 719

Ion	Ratio	Lower	Upper
173	100		
175	47.4	39.0	58.4
254	6.6	7.7	11.5#



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#68
1,2-Dichlorobenzene-d4
Concen: 0.956 ug/l
RT: 12.187 min Scan# 3389
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:152 Resp: 16904

Ion	Ratio	Lower	Upper
152	100		
115	72.2	0.0	275.2
150	168.3	0.0	658.4





#69
Isopropylbenzene
Concen: 0.936 ug/l
RT: 10.473 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:105 Resp: 5337

Ion	Ratio	Lower	Upper
105	100		
120	26.5	12.8	38.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#70
1,1,2,2-Tetrachloroethane
Concen: 1.027 ug/l
RT: 10.772 min Scan# 2949
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 83 Resp: 15680

Ion	Ratio	Lower	Upper
83	100		
131	8.2	7.4	11.0
85	62.2	51.8	77.8





#71
1,2,3-Trichloropropane
Concen: 0.991 ug/l m
RT: 10.814 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 75 Resp: 1146

Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0
110	32.3	0.0	77.0
97	17.4	0.0	42.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#72
Bromobenzene
Concen: 1.016 ug/l
RT: 10.775 min Scan# 2950
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:156 Resp: 15611

Ion	Ratio	Lower	Upper
156	100		
77	161.9	0.0	343.6
158	96.3	0.0	193.0





#73
n-propylbenzene
Concen: 0.904 ug/l
RT: 10.898 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:120 Resp: 1476

Ion	Ratio	Lower	Upper
120	100		
91	476.2	369.8	554.6



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#74
2-Chlorotoluene
Concen: 0.952 ug/l
RT: 10.975 min Scan# 3012
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:126 Resp: 14317

Ion	Ratio	Lower	Upper
126	100		
91	322.6	0.0	623.8





#75
1,3,5-Trimethylbenzene
Concen: 0.916 ug/l
RT: 11.078 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:105 Resp: 4840

Ion	Ratio	Lower	Upper
105	100		
120	47.2	37.3	55.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#76
4-Chlorotoluene
Concen: 0.975 ug/l
RT: 11.090 min Scan# 3048
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:126 Resp: 15039

Ion	Ratio	Lower	Upper
126	100		
91	327.4	0.0	703.6





#77
tert-Butylbenzene
Concen: 0.953 ug/l
RT: 11.409 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:119 Resp: 5091

Ion	Ratio	Lower	Upper
119	100		
91	57.4	29.4	88.3
134	23.4	17.6	26.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#78
1,2,4-Trimethylbenzene
Concen: 0.891 ug/l
RT: 11.460 min Scan# 3163
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:105 Resp: 46689

Ion	Ratio	Lower	Upper
105	100		
120	43.8	21.9	65.7





#79
sec-Butylbenzene
Concen: 0.930 ug/l
RT: 11.634 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:105 Resp: 6322

Ion	Ratio	Lower	Upper
105	100		
134	19.1	15.1	22.7

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#80
Nitrobenzene
Concen: 5.911 ug/l
RT: 13.232 min Scan# 3714
Delta R.T. 0.029 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 77 Resp: 1014

Ion	Ratio	Lower	Upper
77	100		
123	13.5	18.9	67.1#
65	0.0	11.9	15.1#





#81
p-Isopropyltoluene
Concen: 0.905 ug/l
RT: 11.785 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:119 Resp: 4857

Ion	Ratio	Lower	Upper
119	100		
134	25.0	20.3	30.5
91	24.1	19.4	29.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#82
1,3-Dichlorobenzene
Concen: 0.953 ug/l
RT: 11.737 min Scan# 3249
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:146 Resp: 28406

Ion	Ratio	Lower	Upper
146	100		
111	42.9	32.8	49.2
148	65.4	51.1	76.7





#83
1,4-Dichlorobenzene
Concen: 0.910 ug/l
RT: 11.830 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:146 Resp: 2652

Ion	Ratio	Lower	Upper
146	100		
111	41.5	32.1	48.1
148	69.0	50.2	75.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#84
n-Butylbenzene
Concen: 0.876 ug/l
RT: 12.200 min Scan# 3393
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 91 Resp: 42404

Ion	Ratio	Lower	Upper
91	100		
92	49.9	41.8	62.8
134	22.8	18.6	28.0





#85
1,2-Dichlorobenzene
Concen: 0.969 ug/l
RT: 12.203 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:146 Resp: 2774

Ion	Ratio	Lower	Upper
146	100		
111	43.0	21.9	65.7
148	65.4	32.3	96.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#86
1,2-Dibromo-3-Chloropropane
Concen: 0.982 ug/l
RT: 12.997 min Scan# 3641
Delta R.T. 0.010 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion: 75 Resp: 2103

Ion	Ratio	Lower	Upper
75	100		
155	71.2	63.5	95.3
157	89.2	81.8	122.6





#87
1,2,4-Trichlorobenzene
Concen: 0.909 ug/l
RT: 13.836 min Scan# 3
Delta R.T. 0.007 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:180 Resp: 1269

Ion	Ratio	Lower	Upper
180	100		
182	95.4	76.6	115.0
145	30.9	25.4	38.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#88
Hexachlorobutadiene
Concen: 1.020 ug/l
RT: 14.010 min Scan# 3956
Delta R.T. -0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:225 Resp: 10179

Ion	Ratio	Lower	Upper
225	100		
223	65.9	49.5	74.3
227	63.2	51.0	76.4





#89
Naphthalene
Concen: 1.192 ug/l
RT: 14.090 min Scan# 3
Delta R.T. 0.013 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:128 Resp: 1727

Ion	Ratio	Lower	Upper
128	100		
127	11.3	10.6	16.0
129	11.3	8.6	13.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC001

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#90
1,2,3-Trichlorobenzene
Concen: 0.894 ug/l
RT: 14.325 min Scan# 4054
Delta R.T. 0.003 min
Lab File: VU063220.D
Acq: 10 Feb 2025 13:23

Tgt Ion:180 Resp: 12218

Ion	Ratio	Lower	Upper
180	100		
182	93.2	78.2	117.2
145	31.9	26.1	39.1



ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063221.D
Acq On : 10 Feb 2025 13:58
Operator : MD/SY
Sample : VSTDICC002
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 03:58:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
Internal Standards							
1) Fluorobenzene	6.103	96	53430	1.000	ug/l		# 0.00
System Monitoring Compounds							
57) 4-Bromofluorobenzene	10.624	95	17627	1.000	ug/l		0.00
Spiked Amount 1.000			Recovery = 100.000%				
68) 1,2-Dichlorobenzene-d4	12.184	152	18679	1.019	ug/l		0.00
Spiked Amount 1.000			Recovery = 102.000%				
Target Compounds							Qvalue
2) Dichlorodifluoromethane	1.377	85	35165	2.026	ug/l		97
3) Chloromethane	1.515	50	41607	2.081	ug/l		97
4) Vinyl Chloride	1.596	62	39263	1.985	ug/l		100
5) Bromomethane	1.821	94	13311	1.455	ug/l		99
6) Chloroethane	1.904	64	24932	2.001	ug/l		97
7) Trichlorofluoromethane	2.116	101	48274	2.060	ug/l		99
8) 1,1,2-Trichloro-1,2,2-...	2.563	101	27021	2.031	ug/l		98
9) 1,1-Dichloroethene	2.563	96	27058	1.996	ug/l		97
10) Iodomethane	2.705	142	41566	1.951	ug/l		100
11) Allyl Chloride	2.907	41	39456	2.026	ug/l		99
12) Acrylonitrile	3.316	53	12654	4.107	ug/l	#	86
13) Acetone	2.618	43	23537	9.790	ug/l		93
14) Carbon Disulfide	2.776	76	94065	1.985	ug/l		98
15) Methylene Chloride	3.026	84	33474	1.999	ug/l		97
16) trans-1,2-Dichloroethene	3.338	96	30710	1.985	ug/l		99
17) 1,1-Dichloroethane	3.849	63	59018	2.024	ug/l		98
18) 2-Butanone	4.702	43	35904	9.245	ug/l		100
19) Cyclohexane	5.374	56	45197m	1.929	ug/l		
20) Methylcyclohexane	6.750	83	44636	1.921	ug/l		96
21) 2,2-Dichloropropane	4.647	77	44095	1.938	ug/l		99
22) cis-1,2-Dichloroethene	4.653	96	32738	1.958	ug/l		98
23) Diethyl Ether	2.364	59	22999	1.977	ug/l		98
24) tert-Butyl Alcohol	3.178	59	28594	18.780	ug/l	#	86
25) Methyl tert-Butyl Ether	3.348	73	66015	1.950	ug/l		98
26) Bromochloromethane	4.959	128	14111	1.931	ug/l		96
27) Chloroform	5.075	83	58030	1.972	ug/l		97
28) 1,1,1-Trichloroethane	5.300	97	47900	2.009	ug/l		99
29) 1,1-Dichloropropene	5.512	75	43166	2.022	ug/l		99
30) Carbon Tetrachloride	5.509	117	40459	1.979	ug/l		99
31) Isopropyl Ether	3.975	45	81829	1.966	ug/l		97
32) Ethyl-t-butyl ether	4.483	59	73881	1.951	ug/l		97
33) Tert-Amyl methyl ether	5.927	73	62542	1.891	ug/l		97
34) Propionitrile	4.776	54	12222	10.151	ug/l	#	92
35) Benzene	5.763	78	132758	2.022	ug/l		98
36) 1,2-Dichloroethane	5.782	62	38273	2.020	ug/l		100
37) Trichloroethene	6.531	130	31430	2.013	ug/l		98
38) 1,2-Dichloropropane	6.779	63	34066	1.983	ug/l		99
39) Methacrylonitrile	4.968	41	8407	1.955	ug/l		97
40) Methyl acrylate	4.856	55	16003m	2.038	ug/l		
41) Tetrahydrofuran	5.049	42	10378	4.106	ug/l	#	88
42) 1-Chlorobutane	5.444	56	58462	2.001	ug/l		98
43) Dibromomethane	6.910	93	17367	1.996	ug/l		95
44) Bromodichloromethane	7.094	83	41213	2.035	ug/l		99

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063221.D
Acq On : 10 Feb 2025 13:58
Operator : MD/SY
Sample : VSTDICC002
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 03:58:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
45) 4-Methyl-2-Pentanone	7.782	43	88945	9.751	ug/l	98
46) t-1,4-Dichloro-2-butene	10.820	75	18852m	4.176	ug/l	
47) Methyl methacrylate	6.955	69	27411	3.740	ug/l	95
48) Ethyl methacrylate	8.325	69	26103	1.897	ug/l	96
49) Toluene	7.959	92	75472	1.999	ug/l	100
50) t-1,3-Dichloropropene	8.203	75	36608	1.974	ug/l	98
51) cis-1,3-Dichloropropene	7.599	75	45949	2.006	ug/l	97
52) 1,1,2-Trichloroethane	8.390	97	23377	1.993	ug/l	98
53) 1,3-Dichloropropane	8.566	76	41970	2.015	ug/l	100
54) 2-Hexanone	8.679	43	61467	9.874	ug/l	97
55) Dibromochloromethane	8.798	129	26319	1.950	ug/l	98
56) 1,2-Dibromoethane	8.914	107	21639	1.967	ug/l	100
58) Tetrachloroethene	8.544	164	25204	1.959	ug/l	98
59) Chlorobenzene	9.438	112	79653	1.999	ug/l	99
60) 1,1,1,2-Tetrachloroethane	9.525	131	28159	1.966	ug/l	98
61) Pentachloroethane	11.415	117	26815	2.096	ug/l	95
62) Hexachloroethane	12.463	117	22824	2.017	ug/l	99
63) Ethyl Benzene	9.560	91	133142	1.938	ug/l	99
64) m/p-Xylenes	9.685	106	101491	3.954	ug/l	98
65) o-Xylene	10.090	106	50127	1.995	ug/l	97
66) Styrene	10.106	104	77021	1.926	ug/l	100
67) Bromoform	10.283	173	15177	1.981	ug/l	97
69) Isopropylbenzene	10.476	105	116075	1.965	ug/l	100
70) 1,1,2,2-Tetrachloroethane	10.772	83	32740	2.071	ug/l	99
71) 1,2,3-Trichloropropane	10.814	75	26243m	2.189	ug/l	
72) Bromobenzene	10.775	156	31558	1.983	ug/l	95
73) n-propylbenzene	10.897	120	33538	1.983	ug/l	96
74) 2-Chlorotoluene	10.978	126	31255	2.005	ug/l	97
75) 1,3,5-Trimethylbenzene	11.078	105	109465	2.000	ug/l	99
76) 4-Chlorotoluene	11.090	126	31781	1.988	ug/l	98
77) tert-Butylbenzene	11.409	119	110066	1.988	ug/l	100
78) 1,2,4-Trimethylbenzene	11.460	105	107777	1.985	ug/l	99
79) sec-Butylbenzene	11.634	105	140803	1.998	ug/l	100
80) Nitrobenzene	13.219	77	3216m	10.179	ug/l	
81) p-Isopropyltoluene	11.785	119	110370	1.984	ug/l	100
82) 1,3-Dichlorobenzene	11.737	146	62914	2.038	ug/l	99
83) 1,4-Dichlorobenzene	11.827	146	62036	2.055	ug/l	99
84) n-Butylbenzene	12.200	91	96950	1.933	ug/l	99
85) 1,2-Dichlorobenzene	12.203	146	61011	2.057	ug/l	98
86) 1,2-Dibromo-3-Chloropr...	12.994	75	4203	1.894	ug/l	98
87) 1,2,4-Trichlorobenzene	13.836	180	27020	1.867	ug/l	97
88) Hexachlorobutadiene	14.010	225	21652	2.094	ug/l	99
89) Naphthalene	14.084	128	40458	1.891	ug/l	98
90) 1,2,3-Trichlorobenzene	14.322	180	25501	1.801	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063221.D
Acq On : 10 Feb 2025 13:58
Operator : MD/SY
Sample : VSTDICC002
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Feb 11 03:58:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.103 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 96 Resp: 5343

Ion	Ratio	Lower	Upper
96	100		
70	19.4	15.6	23.4
95	8.3	7.3	10.9
97	6.3	0.0	0.0#



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#2
Dichlorodifluoromethane
Concen: 2.026 ug/l
RT: 1.377 min Scan# 27
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 85 Resp: 35165

Ion	Ratio	Lower	Upper
85	100		
87	30.5	16.0	48.0





#3
Chloromethane
Concen: 2.081 ug/l
RT: 1.515 min Scan# 7
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 50 Resp: 4160

Ion	Ratio	Lower	Upper
50	100		
52	33.8	25.8	38.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#4
Vinyl Chloride
Concen: 1.985 ug/l
RT: 1.596 min Scan# 95
Delta R.T. -0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 62 Resp: 39263

Ion	Ratio	Lower	Upper
62	100		
64	31.9	25.4	38.0





#5
Bromomethane
Concen: 1.455 ug/l
RT: 1.821 min Scan# 1
Delta R.T. -0.022 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 94 Resp: 1331

Ion	Ratio	Lower	Upper
94	100		
96	95.3	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#6
Chloroethane
Concen: 2.001 ug/l
RT: 1.904 min Scan# 191
Delta R.T. -0.019 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 64 Resp: 24932

Ion	Ratio	Lower	Upper
64	100		
66	34.0	25.8	38.8





#7
Trichlorofluoromethane
Concen: 2.060 ug/l
RT: 2.116 min Scan# 2
Delta R.T. -0.013 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:101 Resp: 4827

Ion	Ratio	Lower	Upper
101	100		
103	63.3	51.4	77.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 2.031 ug/l
RT: 2.563 min Scan# 396
Delta R.T. -0.006 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:101 Resp: 27021

Ion	Ratio	Lower	Upper
101	100		
85	42.9	35.4	53.0
151	71.2	58.5	87.7





#9
1,1-Dichloroethene
Concen: 1.996 ug/l
RT: 2.563 min Scan# 3
Delta R.T. -0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 96 Resp: 2705

Ion	Ratio	Lower	Upper
96	100		
61	160.9	0.0	492.9
98	59.4	0.0	124.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#10
Iodomethane
Concen: 1.951 ug/l
RT: 2.705 min Scan# 440
Delta R.T. -0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:142 Resp: 41566

Ion	Ratio	Lower	Upper
142	100		
127	43.1	34.5	51.7





#11
Allyl Chloride
Concen: 2.026 ug/l
RT: 2.907 min Scan# 5
Delta R.T. -0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 41 Resp: 3945

Ion	Ratio	Lower	Upper
41	100		
39	71.6	57.9	86.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#12
Acrylonitrile
Concen: 4.107 ug/l
RT: 3.316 min Scan# 630
Delta R.T. 0.016 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 53 Resp: 12654

Ion	Ratio	Lower	Upper
53	100		
52	73.3	64.2	96.2
51	22.6	30.8	46.2#





#13
Acetone
Concen: 9.790 ug/l
RT: 2.618 min Scan# 41
Delta R.T. 0.006 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 43 Resp: 2353

Ion	Ratio	Lower	Upper
43	100		
58	38.0	27.4	41.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#14
Carbon Disulfide
Concen: 1.985 ug/l
RT: 2.776 min Scan# 462
Delta R.T. -0.006 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 76 Resp: 94065

Ion	Ratio	Lower	Upper
76	100		
78	9.7	7.2	10.8





#15
Methylene Chloride
Concen: 1.999 ug/l
RT: 3.026 min Scan# 54
Delta R.T. -0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 84 Resp: 3347

Ion	Ratio	Lower	Upper
84	100		
49	114.9	96.4	144.6
51	37.6	0.0	74.8
86	62.1	51.3	76.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#16
trans-1,2-Dichloroethene
Concen: 1.985 ug/l
RT: 3.338 min Scan# 637
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 96 Resp: 30710

Ion	Ratio	Lower	Upper
96	100		
61	143.5	113.4	170.2
98	64.2	51.2	76.8





#17
1,1-Dichloroethane
Concen: 2.024 ug/l
RT: 3.849 min Scan# 7
Delta R.T. -0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 63 Resp: 5901

Ion	Ratio	Lower	Upper
63	100		
98	5.8	3.3	9.9
100	3.7	2.1	6.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#18
2-Butanone
Concen: 9.245 ug/l
RT: 4.702 min Scan# 1061
Delta R.T. 0.016 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 43 Resp: 35904

Ion	Ratio	Lower	Upper
43	100		
57	8.5	0.0	17.0





#19
Cyclohexane
Concen: 1.929 ug/l m
RT: 5.374 min Scan# 1
Delta R.T. -0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 56 Resp: 4519

Ion	Ratio	Lower	Upper
56	100		
69	30.4	24.5	36.7
84	86.2	65.2	97.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#20
Methylcyclohexane
Concen: 1.921 ug/l
RT: 6.750 min Scan# 1698
Delta R.T. -0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 83 Resp: 44636

Ion	Ratio	Lower	Upper
83	100		
55	82.9	63.1	94.7
98	45.2	35.2	52.8





#21
2,2-Dichloropropane
Concen: 1.938 ug/l
RT: 4.647 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 77 Resp: 4409

Ion	Ratio	Lower	Upper
77	100		
97	22.5	18.5	27.7



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#22
cis-1,2-Dichloroethene
Concen: 1.958 ug/l
RT: 4.653 min Scan# 1046
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 96 Resp: 32738

Ion	Ratio	Lower	Upper
96	100		
61	153.1	0.0	373.3
98	63.1	31.9	95.9





#23
Diethyl Ether
Concen: 1.977 ug/l
RT: 2.364 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 59 Resp: 2299

Ion	Ratio	Lower	Upper
59	100		
45	69.8	57.8	86.6
74	71.4	57.7	86.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#24
tert-Butyl Alcohol
Concen: 18.780 ug/l
RT: 3.178 min Scan# 587
Delta R.T. 0.019 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 59 Resp: 28594

Ion	Ratio	Lower	Upper
59	100		
57	4.4	7.5	11.3#





#25
Methyl tert-Butyl Ether
Concen: 1.950 ug/l
RT: 3.348 min Scan# 64
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 73 Resp: 6601

Ion	Ratio	Lower	Upper
73	100		
43	19.4	14.9	22.3

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#26
Bromochloromethane
Concen: 1.931 ug/l
RT: 4.959 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:128 Resp: 14111

Ion	Ratio	Lower	Upper
128	100		
49	179.3	0.0	343.4
130	130.6	102.9	154.3





#27
Chloroform
Concen: 1.972 ug/l
RT: 5.075 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 83 Resp: 5803

Ion	Ratio	Lower	Upper
83	100		
85	67.3	0.0	129.8

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#28
1,1,1-Trichloroethane
Concen: 2.009 ug/l
RT: 5.300 min Scan# 1247
Delta R.T. -0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 97 Resp: 47900

Ion	Ratio	Lower	Upper
97	100		
99	64.2	32.4	97.0
61	47.9	23.8	71.2





#29
1,1-Dichloropropene
Concen: 2.022 ug/l
RT: 5.512 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 75 Resp: 4316

Ion	Ratio	Lower	Upper
75	100		
110	32.7	17.2	51.5
77	30.8	24.6	37.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#30
Carbon Tetrachloride
Concen: 1.979 ug/l
RT: 5.509 min Scan# 1312
Delta R.T. -0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:117 Resp: 40459

Ion	Ratio	Lower	Upper
117	100		
119	96.0	76.7	115.1
121	30.5	25.5	38.3





#31
Isopropyl Ether
Concen: 1.966 ug/l
RT: 3.975 min Scan# 8
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 45 Resp: 8182

Ion	Ratio	Lower	Upper
45	100		
43	53.7	25.7	77.1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#32
Ethyl-t-butyl ether
Concen: 1.951 ug/l
RT: 4.483 min Scan# 993
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 59 Resp: 73881

Ion	Ratio	Lower	Upper
59	100		
87	39.2	32.6	49.0





#33
Tert-Amyl methyl ether
Concen: 1.891 ug/l
RT: 5.927 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 73 Resp: 6254

Ion	Ratio	Lower	Upper
73	100		
87	24.9	18.9	28.3
71	12.2	9.0	13.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#34
Propionitrile
Concen: 10.151 ug/l
RT: 4.776 min Scan# 1084
Delta R.T. 0.023 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 54 Resp: 12222

Ion	Ratio	Lower	Upper
54	100		
52	16.1	16.3	24.5#
55	11.4	11.8	17.6#
40	3.8	4.3	6.5#





#35
Benzene
Concen: 2.022 ug/l
RT: 5.763 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 78 Resp: 13275

Ion	Ratio	Lower	Upper
78	100		
77	24.6	19.0	28.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#36
1,2-Dichloroethane
Concen: 2.020 ug/l
RT: 5.782 min Scan# 1397
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 62 Resp: 38273

Ion	Ratio	Lower	Upper
62	100		
98	8.5	6.9	10.3





#37
Trichloroethene
Concen: 2.013 ug/l
RT: 6.531 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:130 Resp: 3143

Ion	Ratio	Lower	Upper
130	100		
95	105.6	83.2	124.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#38
1,2-Dichloropropane
Concen: 1.983 ug/l
RT: 6.779 min Scan# 1707
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 63 Resp: 34066

Ion	Ratio	Lower	Upper
63	100		
65	31.3	25.3	37.9





#39
Methacrylonitrile
Concen: 1.955 ug/l
RT: 4.968 min Scan# 1
Delta R.T. 0.010 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 41 Resp: 840

Ion	Ratio	Lower	Upper
41	100		
67	78.1	64.5	96.7
39	51.0	43.0	64.6
52	32.0	25.3	37.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#40
Methyl acrylate
Concen: 2.038 ug/l m
RT: 4.856 min Scan# 1109
Delta R.T. 0.023 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 55 Resp: 16003

Ion	Ratio	Lower	Upper
55	100		
85	8.6	13.3	19.9#
58	3.0	7.3	10.9#
42	5.5	6.9	10.3#





#41
Tetrahydrofuran
Concen: 4.106 ug/l
RT: 5.049 min Scan# 1
Delta R.T. 0.010 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 42 Resp: 1037

Ion	Ratio	Lower	Upper
42	100		
72	45.1	41.5	62.3
71	36.2	37.2	55.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#42
1-Chlorobutane
Concen: 2.001 ug/l
RT: 5.444 min Scan# 1292
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 56 Resp: 58462

Ion	Ratio	Lower	Upper
56	100		
41	51.0	26.3	78.8





#43
Dibromomethane
Concen: 1.996 ug/l
RT: 6.910 min Scan# 1
Delta R.T. 0.006 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 93 Resp: 1736

Ion	Ratio	Lower	Upper
93	100		
95	80.0	67.2	100.8
174	89.2	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#44
Bromodichloromethane
Concen: 2.035 ug/l
RT: 7.094 min Scan# 1805
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 83 Resp: 41213

Ion	Ratio	Lower	Upper
83	100		
85	64.0	51.7	77.5
127	7.4	6.7	10.1





#45
4-Methyl-2-Pentanone
Concen: 9.751 ug/l
RT: 7.782 min Scan# 2
Delta R.T. 0.006 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 43 Resp: 8894

Ion	Ratio	Lower	Upper
43	100		
58	40.2	20.8	62.5
85	17.7	14.8	22.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#46
t-1,4-Dichloro-2-butene
Concen: 4.176 ug/l m
RT: 10.820 min Scan# 2964
Delta R.T. -0.006 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 75 Resp: 18852

Ion	Ratio	Lower	Upper
75	100		
53	60.4	64.5	96.7#
89	39.0	43.4	65.2#
88	29.7	31.2	46.8#





#47
Methyl methacrylate
Concen: 3.740 ug/l
RT: 6.955 min Scan# 1
Delta R.T. 0.006 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 69 Resp: 2741

Ion	Ratio	Lower	Upper
69	100		
41	113.8	0.0	217.0
39	64.1	47.7	71.5
100	35.9	27.4	41.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#48
Ethyl methacrylate
Concen: 1.897 ug/l
RT: 8.325 min Scan# 2188
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 69 Resp: 26103

Ion	Ratio	Lower	Upper
69	100		
41	64.5	30.6	92.0





#49
Toluene
Concen: 1.999 ug/l
RT: 7.959 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 92 Resp: 7547

Ion	Ratio	Lower	Upper
92	100		
91	176.6	141.8	212.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#50
t-1,3-Dichloropropene
Concen: 1.974 ug/l
RT: 8.203 min Scan# 2150
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 75 Resp: 36608

Ion	Ratio	Lower	Upper
75	100		
77	31.1	25.9	38.9





#51
cis-1,3-Dichloropropene
Concen: 2.006 ug/l
RT: 7.599 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 75 Resp: 4594

Ion	Ratio	Lower	Upper
75	100		
77	29.7	25.3	37.9
39	44.0	33.5	50.3



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#52
1,1,2-Trichloroethane
Concen: 1.993 ug/l
RT: 8.390 min Scan# 2208
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 97 Resp: 23377

Ion	Ratio	Lower	Upper
97	100		
83	87.9	73.0	109.4
85	57.7	46.3	69.5
99	59.4	48.5	72.7





#53
1,3-Dichloropropane
Concen: 2.015 ug/l
RT: 8.566 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 76 Resp: 4197

Ion	Ratio	Lower	Upper
76	100		
78	33.0	26.3	39.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#54
2-Hexanone
Concen: 9.874 ug/l
RT: 8.679 min Scan# 2298
Delta R.T. 0.006 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 43 Resp: 61467

Ion	Ratio	Lower	Upper
43	100		
58	55.4	38.0	78.0
57	18.3	0.0	39.1





#55
Dibromochloromethane
Concen: 1.950 ug/l
RT: 8.798 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:129 Resp: 2631

Ion	Ratio	Lower	Upper
129	100		
127	77.0	60.4	90.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#56
1,2-Dibromoethane
Concen: 1.967 ug/l
RT: 8.914 min Scan# 2371
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:107 Resp: 21639

Ion	Ratio	Lower	Upper
107	100		
109	93.6	0.0	187.8





#57
4-Bromofluorobenzene
Concen: 1.000 ug/l
RT: 10.624 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 95 Resp: 1762

Ion	Ratio	Lower	Upper
95	100		
174	77.7	58.6	88.0
176	74.7	58.2	87.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#58
Tetrachloroethene
Concen: 1.959 ug/l
RT: 8.544 min Scan# 2256
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:164 Resp: 25204

Ion	Ratio	Lower	Upper
164	100		
166	128.7	101.4	152.0
129	100.2	77.0	115.4
131	93.7	76.3	114.5





#59
Chlorobenzene
Concen: 1.999 ug/l
RT: 9.438 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:112 Resp: 7965

Ion	Ratio	Lower	Upper
112	100		
114	31.4	25.7	38.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#60
1,1,1,2-Tetrachloroethane
Concen: 1.966 ug/l
RT: 9.525 min Scan# 2561
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:131 Resp: 28159

Ion	Ratio	Lower	Upper
131	100		
133	98.9	76.7	115.1
119	68.1	54.4	81.6





#61
Pentachloroethane
Concen: 2.096 ug/l
RT: 11.415 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:117 Resp: 2681

Ion	Ratio	Lower	Upper
117	100		
167	76.4	65.0	97.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#62
Hexachloroethane
Concen: 2.017 ug/l
RT: 12.463 min Scan# 3475
Delta R.T. -0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:117 Resp: 22824

Ion	Ratio	Lower	Upper
117	100		
201	75.4	61.3	91.9





#63
Ethyl Benzene
Concen: 1.938 ug/l
RT: 9.560 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 91 Resp: 13314

Ion	Ratio	Lower	Upper
91	100		
106	29.8	24.2	36.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#64
m/p-Xylenes
Concen: 3.954 ug/l
RT: 9.685 min Scan# 2611
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:106 Resp: 101491

Ion	Ratio	Lower	Upper
106	100		
91	205.7	166.9	250.3





#65
o-Xylene
Concen: 1.995 ug/l
RT: 10.090 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:106 Resp: 5012

Ion	Ratio	Lower	Upper
106	100		
91	217.6	110.9	332.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#66
Styrene
Concen: 1.926 ug/l
RT: 10.106 min Scan# 2742
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:104 Resp: 77021

Ion	Ratio	Lower	Upper
104	100		
78	52.1	41.2	61.8
103	56.0	44.8	67.2





#67
Bromoform
Concen: 1.981 ug/l
RT: 10.283 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:173 Resp: 1517

Ion	Ratio	Lower	Upper
173	100		
175	46.3	39.0	58.4
254	9.3	7.7	11.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#68
1,2-Dichlorobenzene-d4
Concen: 1.019 ug/l
RT: 12.184 min Scan# 3388
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:152 Resp: 18679

Ion	Ratio	Lower	Upper
152	100		
115	78.6	0.0	275.2
150	189.0	0.0	658.4





#69
Isopropylbenzene
Concen: 1.965 ug/l
RT: 10.476 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:105 Resp: 11607

Ion	Ratio	Lower	Upper
105	100		
120	25.7	12.8	38.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#70
1,1,2,2-Tetrachloroethane
Concen: 2.071 ug/l
RT: 10.772 min Scan# 2949
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 83 Resp: 32740

Ion	Ratio	Lower	Upper
83	100		
131	8.5	7.4	11.0
85	64.1	51.8	77.8





#71
1,2,3-Trichloropropane
Concen: 2.189 ug/l m
RT: 10.814 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 75 Resp: 2624

Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0
110	29.6	0.0	77.0
97	16.2	0.0	42.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#72
Bromobenzene
Concen: 1.983 ug/l
RT: 10.775 min Scan# 2950
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:156 Resp: 31558

Ion	Ratio	Lower	Upper
156	100		
77	179.9	0.0	343.6
158	99.7	0.0	193.0





#73
n-propylbenzene
Concen: 1.983 ug/l
RT: 10.897 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:120 Resp: 3353

Ion	Ratio	Lower	Upper
120	100		
91	452.1	369.8	554.6



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#74
2-Chlorotoluene
Concen: 2.005 ug/l
RT: 10.978 min Scan# 3013
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:126 Resp: 31255

Ion	Ratio	Lower	Upper
126	100		
91	317.2	0.0	623.8





#75
1,3,5-Trimethylbenzene
Concen: 2.000 ug/l
RT: 11.078 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:105 Resp: 10946

Ion	Ratio	Lower	Upper
105	100		
120	46.1	37.3	55.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#76
4-Chlorotoluene
Concen: 1.988 ug/l
RT: 11.090 min Scan# 3048
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:126 Resp: 31781

Ion	Ratio	Lower	Upper
126	100		
91	355.9	0.0	703.6





#77
tert-Butylbenzene
Concen: 1.988 ug/l
RT: 11.409 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:119 Resp: 11006

Ion	Ratio	Lower	Upper
119	100		
91	59.1	29.4	88.3
134	21.7	17.6	26.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#78
1,2,4-Trimethylbenzene
Concen: 1.985 ug/l
RT: 11.460 min Scan# 3163
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:105 Resp: 107777

Ion	Ratio	Lower	Upper
105	100		
120	44.7	21.9	65.7





#79
sec-Butylbenzene
Concen: 1.998 ug/l
RT: 11.634 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:105 Resp: 14080

Ion	Ratio	Lower	Upper
105	100		
134	18.7	15.1	22.7

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#80
Nitrobenzene
Concen: 10.179 ug/l m
RT: 13.219 min Scan# 3710
Delta R.T. 0.016 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 77 Resp: 3216

Ion	Ratio	Lower	Upper
77	100		
123	32.9	18.9	67.1
65	11.5	11.9	15.1#





#81
p-Isopropyltoluene
Concen: 1.984 ug/l
RT: 11.785 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:119 Resp: 11037

Ion	Ratio	Lower	Upper
119	100		
134	25.2	20.3	30.5
91	24.2	19.4	29.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#82
1,3-Dichlorobenzene
Concen: 2.038 ug/l
RT: 11.737 min Scan# 3249
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:146 Resp: 62914

Ion	Ratio	Lower	Upper
146	100		
111	41.4	32.8	49.2
148	63.0	51.1	76.7





#83
1,4-Dichlorobenzene
Concen: 2.055 ug/l
RT: 11.827 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:146 Resp: 6203

Ion	Ratio	Lower	Upper
146	100		
111	39.9	32.1	48.1
148	62.2	50.2	75.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#84
n-Butylbenzene
Concen: 1.933 ug/l
RT: 12.200 min Scan# 3393
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 91 Resp: 96950

Ion	Ratio	Lower	Upper
91	100		
92	51.1	41.8	62.8
134	23.6	18.6	28.0





#85
1,2-Dichlorobenzene
Concen: 2.057 ug/l
RT: 12.203 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:146 Resp: 6101

Ion	Ratio	Lower	Upper
146	100		
111	42.9	21.9	65.7
148	62.9	32.3	96.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#86
1,2-Dibromo-3-Chloropropane
Concen: 1.894 ug/l
RT: 12.994 min Scan# 3640
Delta R.T. 0.006 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion: 75 Resp: 4203

Ion	Ratio	Lower	Upper
75	100		
155	77.8	63.5	95.3
157	99.9	81.8	122.6





#87
1,2,4-Trichlorobenzene
Concen: 1.867 ug/l
RT: 13.836 min Scan# 3
Delta R.T. 0.006 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:180 Resp: 2702

Ion	Ratio	Lower	Upper
180	100		
182	93.1	76.6	115.0
145	30.9	25.4	38.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#88
Hexachlorobutadiene
Concen: 2.094 ug/l
RT: 14.010 min Scan# 3956
Delta R.T. -0.003 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:225 Resp: 21652

Ion	Ratio	Lower	Upper
225	100		
223	64.0	49.5	74.3
227	63.7	51.0	76.4





#89
Naphthalene
Concen: 1.891 ug/l
RT: 14.084 min Scan# 3
Delta R.T. 0.006 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:128 Resp: 4045

Ion	Ratio	Lower	Upper
128	100		
127	12.1	10.6	16.0
129	10.2	8.6	13.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC002

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#90
1,2,3-Trichlorobenzene
Concen: 1.801 ug/l
RT: 14.322 min Scan# 4053
Delta R.T. 0.000 min
Lab File: VU063221.D
Acq: 10 Feb 2025 13:58

Tgt Ion:180 Resp: 25501

Ion	Ratio	Lower	Upper
180	100		
182	96.0	78.2	117.2
145	32.5	26.1	39.1



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063222.D
Acq On : 10 Feb 2025 14:23
Operator : MD/SY
Sample : VSTDICC005
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 03:59:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Fluorobenzene	6.103	96	55729	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.624	95	17913	0.974	ug/l	0.00
Spiked Amount 1.000			Recovery = 97.000%			
68) 1,2-Dichlorobenzene-d4	12.184	152	18704	0.978	ug/l	0.00
Spiked Amount 1.000			Recovery = 98.000%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.380	85	90512	4.999	ug/l	99
3) Chloromethane	1.518	50	102545	4.918	ug/l	99
4) Vinyl Chloride	1.599	62	104693	5.075	ug/l	99
5) Bromomethane	1.843	94	52801	5.534	ug/l	99
6) Chloroethane	1.924	64	62097	4.779	ug/l	98
7) Trichlorofluoromethane	2.129	101	123320	5.044	ug/l	100
8) 1,1,2-Trichloro-1,2,2-...	2.570	101	69588	5.015	ug/l	99
9) 1,1-Dichloroethene	2.567	96	69679	4.928	ug/l	96
10) Iodomethane	2.708	142	114428	5.148	ug/l	100
11) Allyl Chloride	2.911	41	101481	4.996	ug/l	100
12) Acrylonitrile	3.306	53	33591	10.453	ug/l	99
13) Acetone	2.612	43	58550	23.349	ug/l	96
14) Carbon Disulfide	2.782	76	242244	4.900	ug/l	100
15) Methylene Chloride	3.033	84	86216	4.936	ug/l	99
16) trans-1,2-Dichloroethene	3.338	96	80123	4.965	ug/l	98
17) 1,1-Dichloroethane	3.853	63	152899	5.027	ug/l	100
18) 2-Butanone	4.689	43	98391	24.289	ug/l	97
19) Cyclohexane	5.374	56	127410m	5.213	ug/l	
20) Methylcyclohexane	6.750	83	128499	5.302	ug/l	100
21) 2,2-Dichloropropane	4.647	77	117171	4.938	ug/l	98
22) cis-1,2-Dichloroethene	4.653	96	87987	5.046	ug/l	98
23) Diethyl Ether	2.364	59	60657	5.000	ug/l	97
24) tert-Butyl Alcohol	3.161	59	61281	45.943	ug/l	98
25) Methyl tert-Butyl Ether	3.348	73	178876	5.066	ug/l	99
26) Bromochloromethane	4.959	128	38617	5.067	ug/l	99
27) Chloroform	5.071	83	155011	5.050	ug/l	99
28) 1,1,1-Trichloroethane	5.300	97	127899	5.144	ug/l	99
29) 1,1-Dichloropropene	5.512	75	112514	5.052	ug/l	99
30) Carbon Tetrachloride	5.509	117	106762	5.007	ug/l	97
31) Isopropyl Ether	3.975	45	218715	5.037	ug/l	97
32) Ethyl-t-butyl ether	4.486	59	202182	5.120	ug/l	100
33) Tert-Amyl methyl ether	5.927	73	177493	5.144	ug/l	98
34) Propionitrile	4.759	54	31294	24.919	ug/l	96
35) Benzene	5.759	78	349985	5.111	ug/l	100
36) 1,2-Dichloroethane	5.779	62	99553	5.037	ug/l	99
37) Trichloroethene	6.531	130	80810	4.962	ug/l	93
38) 1,2-Dichloropropane	6.779	63	90072	5.026	ug/l	98
39) Methacrylonitrile	4.962	41	24242	5.404	ug/l	97
40) Methyl acrylate	4.840	55	43877m	5.358	ug/l	
41) Tetrahydrofuran	5.042	42	24329	9.229	ug/l	97
42) 1-Chlorobutane	5.444	56	160975	5.283	ug/l	98
43) Dibromomethane	6.907	93	45659	5.031	ug/l	97
44) Bromodichloromethane	7.094	83	109612	5.189	ug/l	99

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063222.D
Acq On : 10 Feb 2025 14:23
Operator : MD/SY
Sample : VSTDICC005
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 03:59:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
45) 4-Methyl-2-Pentanone	7.779	43	245466	25.799	ug/l	99
46) t-1,4-Dichloro-2-butene	10.824	75	40225m	8.544	ug/l	
47) Methyl methacrylate	6.949	69	82725	10.822	ug/l	97
48) Ethyl methacrylate	8.322	69	75186	5.238	ug/l	99
49) Toluene	7.959	92	207542	5.270	ug/l	100
50) t-1,3-Dichloropropene	8.200	75	100975	5.221	ug/l	99
51) cis-1,3-Dichloropropene	7.595	75	121107	5.070	ug/l	96
52) 1,1,2-Trichloroethane	8.386	97	61701	5.042	ug/l	98
53) 1,3-Dichloropropane	8.563	76	111566	5.136	ug/l	99
54) 2-Hexanone	8.672	43	163849	25.235	ug/l	98
55) Dibromochloromethane	8.798	129	71240	5.061	ug/l	98
56) 1,2-Dibromoethane	8.914	107	58284	5.079	ug/l	99
58) Tetrachloroethene	8.544	164	69703	5.194	ug/l	97
59) Chlorobenzene	9.438	112	215989	5.198	ug/l	99
60) 1,1,1,2-Tetrachloroethane	9.521	131	74952	5.018	ug/l	98
61) Pentachloroethane	11.415	117	66930	5.016	ug/l	96
62) Hexachloroethane	12.463	117	58399	4.947	ug/l	99
63) Ethyl Benzene	9.560	91	379921	5.301	ug/l	99
64) m/p-Xylenes	9.682	106	293404	10.960	ug/l	99
65) o-Xylene	10.090	106	139940	5.340	ug/l	98
66) Styrene	10.103	104	227199	5.448	ug/l	99
67) Bromoform	10.280	173	40369	5.053	ug/l	99
69) Isopropylbenzene	10.473	105	329990	5.356	ug/l	99
70) 1,1,2,2-Tetrachloroethane	10.769	83	81338	4.932	ug/l	99
71) 1,2,3-Trichloropropane	10.814	75	59228m	4.736	ug/l	
72) Bromobenzene	10.772	156	85960	5.178	ug/l	98
73) n-propylbenzene	10.894	120	95937	5.439	ug/l	99
74) 2-Chlorotoluene	10.975	126	87769	5.399	ug/l	98
75) 1,3,5-Trimethylbenzene	11.078	105	313345	5.489	ug/l	100
76) 4-Chlorotoluene	11.087	126	89247	5.353	ug/l	99
77) tert-Butylbenzene	11.409	119	299259	5.183	ug/l	99
78) 1,2,4-Trimethylbenzene	11.457	105	313777	5.539	ug/l	99
79) sec-Butylbenzene	11.634	105	394157	5.362	ug/l	99
80) Nitrobenzene	13.209	77	10183m	23.083	ug/l	
81) p-Isopropyltoluene	11.782	119	316807	5.461	ug/l	100
82) 1,3-Dichlorobenzene	11.737	146	164073	5.096	ug/l	99
83) 1,4-Dichlorobenzene	11.827	146	165157	5.244	ug/l	99
84) n-Butylbenzene	12.200	91	276461	5.285	ug/l	98
85) 1,2-Dichlorobenzene	12.203	146	156956	5.072	ug/l	100
86) 1,2-Dibromo-3-Chloropr...	12.987	75	12084	5.220	ug/l	96
87) 1,2,4-Trichlorobenzene	13.830	180	78516	5.203	ug/l	98
88) Hexachlorobutadiene	14.010	225	53269	4.939	ug/l	99
89) Naphthalene	14.081	128	122840	4.286	ug/l	99
90) 1,2,3-Trichlorobenzene	14.319	180	76620	5.189	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063222.D
Acq On : 10 Feb 2025 14:23
Operator : MD/SY
Sample : VSTDICC005
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Feb 11 03:59:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.103 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 96 Resp: 5572

Ion	Ratio	Lower	Upper
96	100		
70	20.8	15.6	23.4
95	9.3	7.3	10.9
97	6.5	0.0	0.0#



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#2
Dichlorodifluoromethane
Concen: 4.999 ug/l
RT: 1.380 min Scan# 28
Delta R.T. 0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 85 Resp: 90512

Ion	Ratio	Lower	Upper
85	100		
87	32.3	16.0	48.0





#3
Chloromethane
Concen: 4.918 ug/l
RT: 1.518 min Scan# 7
Delta R.T. 0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 50 Resp: 10254

Ion	Ratio	Lower	Upper
50	100		
52	31.9	25.8	38.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#4
Vinyl Chloride
Concen: 5.075 ug/l
RT: 1.599 min Scan# 96
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 62 Resp: 104693

Ion	Ratio	Lower	Upper
62	100		
64	31.0	25.4	38.0





#5
Bromomethane
Concen: 5.534 ug/l
RT: 1.843 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 94 Resp: 5280

Ion	Ratio	Lower	Upper
94	100		
96	95.6	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#6
Chloroethane
Concen: 4.779 ug/l
RT: 1.924 min Scan# 197
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 64 Resp: 62097

Ion	Ratio	Lower	Upper
64	100		
66	33.3	25.8	38.8





#7
Trichlorofluoromethane
Concen: 5.044 ug/l
RT: 2.129 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:101 Resp: 12332

Ion	Ratio	Lower	Upper
101	100		
103	64.1	51.4	77.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 5.015 ug/l
RT: 2.570 min Scan# 398
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:101 Resp: 69588

Ion	Ratio	Lower	Upper
101	100		
85	46.1	35.4	53.0
151	72.8	58.5	87.7





#9
1,1-Dichloroethene
Concen: 4.928 ug/l
RT: 2.567 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 96 Resp: 6967

Ion	Ratio	Lower	Upper
96	100		
61	170.1	0.0	492.9
98	64.3	0.0	124.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#10
Iodomethane
Concen: 5.148 ug/l
RT: 2.708 min Scan# 441
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:142 Resp: 114428

Ion	Ratio	Lower	Upper
142	100		
127	43.1	34.5	51.7





#11
Allyl Chloride
Concen: 4.996 ug/l
RT: 2.911 min Scan# 5
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 41 Resp: 10148

Ion	Ratio	Lower	Upper
41	100		
39	72.1	57.9	86.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#12
Acrylonitrile
Concen: 10.453 ug/l
RT: 3.306 min Scan# 627
Delta R.T. 0.006 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 53 Resp: 33591

Ion	Ratio	Lower	Upper
53	100		
52	78.3	64.2	96.2
51	38.5	30.8	46.2





#13
Acetone
Concen: 23.349 ug/l
RT: 2.612 min Scan# 4
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 43 Resp: 5855

Ion	Ratio	Lower	Upper
43	100		
58	32.0	27.4	41.0

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#14
Carbon Disulfide
Concen: 4.900 ug/l
RT: 2.782 min Scan# 464
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 76 Resp: 242244

Ion	Ratio	Lower	Upper
76	100		
78	9.1	7.2	10.8





#15
Methylene Chloride
Concen: 4.936 ug/l
RT: 3.033 min Scan# 54
Delta R.T. 0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 84 Resp: 8621

Ion	Ratio	Lower	Upper
84	100		
49	118.4	96.4	144.6
51	37.9	0.0	74.8
86	64.8	51.3	76.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#16
trans-1,2-Dichloroethene
Concen: 4.965 ug/l
RT: 3.338 min Scan# 637
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 96 Resp: 80123

Ion	Ratio	Lower	Upper
96	100		
61	145.3	113.4	170.2
98	63.0	51.2	76.8





#17
1,1-Dichloroethane
Concen: 5.027 ug/l
RT: 3.853 min Scan# 7
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 63 Resp: 15289

Ion	Ratio	Lower	Upper
63	100		
98	6.6	3.3	9.9
100	4.3	2.1	6.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#18
2-Butanone
Concen: 24.289 ug/l
RT: 4.689 min Scan# 1057
Delta R.T. 0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 43 Resp: 98391

Ion	Ratio	Lower	Upper
43	100		
57	9.5	0.0	17.0





#19
Cyclohexane
Concen: 5.213 ug/l m
RT: 5.374 min Scan# 1
Delta R.T. -0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 56 Resp: 12741

Ion	Ratio	Lower	Upper
56	100		
69	30.9	24.5	36.7
84	83.4	65.2	97.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#20
Methylcyclohexane
Concen: 5.302 ug/l
RT: 6.750 min Scan# 1698
Delta R.T. -0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 83 Resp: 128499

Ion	Ratio	Lower	Upper
83	100		
55	78.9	63.1	94.7
98	43.6	35.2	52.8





#21
2,2-Dichloropropane
Concen: 4.938 ug/l
RT: 4.647 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 77 Resp: 11717

Ion	Ratio	Lower	Upper
77	100		
97	22.1	18.5	27.7



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#22
cis-1,2-Dichloroethene
Concen: 5.046 ug/l
RT: 4.653 min Scan# 1046
Delta R.T. 0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 96 Resp: 87987

Ion	Ratio	Lower	Upper
96	100		
61	152.1	0.0	373.3
98	63.4	31.9	95.9





#23
Diethyl Ether
Concen: 5.000 ug/l
RT: 2.364 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 59 Resp: 6065

Ion	Ratio	Lower	Upper
59	100		
45	69.8	57.8	86.6
74	69.7	57.7	86.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#24
tert-Butyl Alcohol
Concen: 45.943 ug/l
RT: 3.161 min Scan# 582
Delta R.T. 0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 59 Resp: 61281

Ion	Ratio	Lower	Upper
59	100		
57	8.6	7.5	11.3





#25
Methyl tert-Butyl Ether
Concen: 5.066 ug/l
RT: 3.348 min Scan# 64
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 73 Resp: 17887

Ion	Ratio	Lower	Upper
73	100		
43	19.0	14.9	22.3

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#26
Bromochloromethane
Concen: 5.067 ug/l
RT: 4.959 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:128 Resp: 38617

Ion	Ratio	Lower	Upper
128	100		
49	171.4	0.0	343.4
130	127.4	102.9	154.3





#27
Chloroform
Concen: 5.050 ug/l
RT: 5.071 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 83 Resp: 15501

Ion	Ratio	Lower	Upper
83	100		
85	64.0	0.0	129.8

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#28
1,1,1-Trichloroethane
Concen: 5.144 ug/l
RT: 5.300 min Scan# 1247
Delta R.T. -0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 97 Resp: 127899

Ion	Ratio	Lower	Upper
97	100		
99	64.1	32.4	97.0
61	46.2	23.8	71.2





#29
1,1-Dichloropropene
Concen: 5.052 ug/l
RT: 5.512 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 75 Resp: 11251

Ion	Ratio	Lower	Upper
75	100		
110	35.5	17.2	51.5
77	30.9	24.6	37.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#30
Carbon Tetrachloride
Concen: 5.007 ug/l
RT: 5.509 min Scan# 1312
Delta R.T. -0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:117 Resp: 106762

Ion	Ratio	Lower	Upper
117	100		
119	98.3	76.7	115.1
121	29.7	25.5	38.3





#31
Isopropyl Ether
Concen: 5.037 ug/l
RT: 3.975 min Scan# 8
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 45 Resp: 21871

Ion	Ratio	Lower	Upper
45	100		
43	53.4	25.7	77.1



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#32
Ethyl-t-butyl ether
Concen: 5.120 ug/l
RT: 4.486 min Scan# 994
Delta R.T. 0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 59 Resp: 202182

Ion	Ratio	Lower	Upper
59	100		
87	40.5	32.6	49.0





#33
Tert-Amyl methyl ether
Concen: 5.144 ug/l
RT: 5.927 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Tgt Ion: 73 Resp: 17749

Ion	Ratio	Lower	Upper
73	100		
87	24.4	18.9	28.3
71	12.2	9.0	13.4

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#34
Propionitrile
Concen: 24.919 ug/l
RT: 4.759 min Scan# 1079
Delta R.T. 0.006 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 54 Resp: 31294

Ion	Ratio	Lower	Upper
54	100		
52	21.5	16.3	24.5
55	11.9	11.8	17.6
40	5.4	4.3	6.5





#35
Benzene
Concen: 5.111 ug/l
RT: 5.759 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 78 Resp: 34998

Ion	Ratio	Lower	Upper
78	100		
77	23.6	19.0	28.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#36
1,2-Dichloroethane
Concen: 5.037 ug/l
RT: 5.779 min Scan# 1396
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 62 Resp: 99553

Ion	Ratio	Lower	Upper
62	100		
98	8.2	6.9	10.3





#37
Trichloroethene
Concen: 4.962 ug/l
RT: 6.531 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:130 Resp: 8081

Ion	Ratio	Lower	Upper
130	100		
95	111.3	83.2	124.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#38
1,2-Dichloropropane
Concen: 5.026 ug/l
RT: 6.779 min Scan# 1707
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 63 Resp: 90072

Ion	Ratio	Lower	Upper
63	100		
65	30.7	25.3	37.9





#39
Methacrylonitrile
Concen: 5.404 ug/l
RT: 4.962 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 41 Resp: 2424

Ion	Ratio	Lower	Upper
41	100		
67	84.7	64.5	96.7
39	53.0	43.0	64.6
52	31.2	25.3	37.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#40
Methyl acrylate
Concen: 5.358 ug/l m
RT: 4.840 min Scan# 1104
Delta R.T. 0.006 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 55 Resp: 43877

Ion	Ratio	Lower	Upper
55	100		
85	15.7	13.3	19.9
58	7.2	7.3	10.9#
42	8.0	6.9	10.3





#41
Tetrahydrofuran
Concen: 9.229 ug/l
RT: 5.042 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 42 Resp: 2432

Ion	Ratio	Lower	Upper
42	100		
72	53.6	41.5	62.3
71	48.7	37.2	55.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#42
1-Chlorobutane
Concen: 5.283 ug/l
RT: 5.444 min Scan# 1292
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 56 Resp: 160975

Ion	Ratio	Lower	Upper
56	100		
41	51.3	26.3	78.8





#43
Dibromomethane
Concen: 5.031 ug/l
RT: 6.907 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 93 Resp: 4565

Ion	Ratio	Lower	Upper
93	100		
95	81.3	67.2	100.8
174	92.0	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#44
Bromodichloromethane
Concen: 5.189 ug/l
RT: 7.094 min Scan# 1805
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 83 Resp: 109612

Ion	Ratio	Lower	Upper
83	100		
85	63.9	51.7	77.5
127	8.0	6.7	10.1





#45
4-Methyl-2-Pentanone
Concen: 25.799 ug/l
RT: 7.779 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 43 Resp: 24546

Ion	Ratio	Lower	Upper
43	100		
58	40.8	20.8	62.5
85	18.0	14.8	22.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#46
t-1,4-Dichloro-2-butene
Concen: 8.544 ug/l m
RT: 10.824 min Scan# 2965
Delta R.T. -0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 75 Resp: 40225

Ion	Ratio	Lower	Upper
75	100		
53	79.1	64.5	96.7
89	51.4	43.4	65.2
88	38.4	31.2	46.8





#47
Methyl methacrylate
Concen: 10.822 ug/l
RT: 6.949 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 69 Resp: 8272

Ion	Ratio	Lower	Upper
69	100		
41	103.9	0.0	217.0
39	58.1	47.7	71.5
100	33.0	27.4	41.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#48
Ethyl methacrylate
Concen: 5.238 ug/l
RT: 8.322 min Scan# 2187
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 69 Resp: 75186

Ion	Ratio	Lower	Upper
69	100		
41	62.0	30.6	92.0





#49
Toluene
Concen: 5.270 ug/l
RT: 7.959 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 92 Resp: 20754

Ion	Ratio	Lower	Upper
92	100		
91	177.1	141.8	212.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#50
t-1,3-Dichloropropene
Concen: 5.221 ug/l
RT: 8.200 min Scan# 2149
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 75 Resp: 100975

Ion	Ratio	Lower	Upper
75	100		
77	31.8	25.9	38.9





#51
cis-1,3-Dichloropropene
Concen: 5.070 ug/l
RT: 7.595 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 75 Resp: 12110

Ion	Ratio	Lower	Upper
75	100		
77	32.6	25.3	37.9
39	45.0	33.5	50.3



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#52
1,1,2-Trichloroethane
Concen: 5.042 ug/l
RT: 8.386 min Scan# 2207
Delta R.T. -0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 97 Resp: 61701

Ion	Ratio	Lower	Upper
97	100		
83	93.2	73.0	109.4
85	59.6	46.3	69.5
99	62.0	48.5	72.7





#53
1,3-Dichloropropane
Concen: 5.136 ug/l
RT: 8.563 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 76 Resp: 11156

Ion	Ratio	Lower	Upper
76	100		
78	32.3	26.3	39.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#54
2-Hexanone
Concen: 25.235 ug/l
RT: 8.672 min Scan# 2296
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 43 Resp: 163849

Ion	Ratio	Lower	Upper
43	100		
58	55.7	38.0	78.0
57	18.9	0.0	39.1





#55
Dibromochloromethane
Concen: 5.061 ug/l
RT: 8.798 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:129 Resp: 7124

Ion	Ratio	Lower	Upper
129	100		
127	77.1	60.4	90.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#56
1,2-Dibromoethane
Concen: 5.079 ug/l
RT: 8.914 min Scan# 2371
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:107 Resp: 58284

Ion	Ratio	Lower	Upper
107	100		
109	94.6	0.0	187.8





#57
4-Bromofluorobenzene
Concen: 0.974 ug/l
RT: 10.624 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Tgt Ion: 95 Resp: 1791

Ion	Ratio	Lower	Upper
95	100		
174	81.3	58.6	88.0
176	78.3	58.2	87.4

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#58
Tetrachloroethene
Concen: 5.194 ug/l
RT: 8.544 min Scan# 2256
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:164 Resp: 69703

Ion	Ratio	Lower	Upper
164	100		
166	125.2	101.4	152.0
129	99.7	77.0	115.4
131	98.7	76.3	114.5





#59
Chlorobenzene
Concen: 5.198 ug/l
RT: 9.438 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:112 Resp: 21598

Ion	Ratio	Lower	Upper
112	100		
114	31.7	25.7	38.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#60
1,1,1,2-Tetrachloroethane
Concen: 5.018 ug/l
RT: 9.521 min Scan# 2560
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:131 Resp: 74952

Ion	Ratio	Lower	Upper
131	100		
133	98.9	76.7	115.1
119	68.6	54.4	81.6





#61
Pentachloroethane
Concen: 5.016 ug/l
RT: 11.415 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:117 Resp: 6693

Ion	Ratio	Lower	Upper
117	100		
167	77.9	65.0	97.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#62
Hexachloroethane
Concen: 4.947 ug/l
RT: 12.463 min Scan# 3475
Delta R.T. -0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:117 Resp: 58399

Ion	Ratio	Lower	Upper
117	100		
201	75.3	61.3	91.9





#63
Ethyl Benzene
Concen: 5.301 ug/l
RT: 9.560 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 91 Resp: 37992

Ion	Ratio	Lower	Upper
91	100		
106	30.6	24.2	36.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#64
m/p-Xylenes
Concen: 10.960 ug/l
RT: 9.682 min Scan# 2610
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:106 Resp: 293404

Ion	Ratio	Lower	Upper
106	100		
91	209.6	166.9	250.3





#65
o-Xylene
Concen: 5.340 ug/l
RT: 10.090 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:106 Resp: 13994

Ion	Ratio	Lower	Upper
106	100		
91	219.4	110.9	332.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#66
Styrene
Concen: 5.448 ug/l
RT: 10.103 min Scan# 2741
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:104 Resp: 227199

Ion	Ratio	Lower	Upper
104	100		
78	52.1	41.2	61.8
103	56.8	44.8	67.2





#67
Bromoform
Concen: 5.053 ug/l
RT: 10.280 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:173 Resp: 4036

Ion	Ratio	Lower	Upper
173	100		
175	49.3	39.0	58.4
254	9.9	7.7	11.5

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#68
1,2-Dichlorobenzene-d4
Concen: 0.978 ug/l
RT: 12.184 min Scan# 3388
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:152 Resp: 18704

Ion	Ratio	Lower	Upper
152	100		
115	105.9	0.0	275.2
150	243.0	0.0	658.4





#69
Isopropylbenzene
Concen: 5.356 ug/l
RT: 10.473 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:105 Resp: 32999

Ion	Ratio	Lower	Upper
105	100		
120	25.3	12.8	38.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#70
1,1,2,2-Tetrachloroethane
Concen: 4.932 ug/l
RT: 10.769 min Scan# 2948
Delta R.T. -0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 83 Resp: 81338

Ion	Ratio	Lower	Upper
83	100		
131	9.0	7.4	11.0
85	63.7	51.8	77.8





#71
1,2,3-Trichloropropane
Concen: 4.736 ug/l m
RT: 10.814 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 75 Resp: 5922

Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0
110	35.6	0.0	77.0
97	18.1	0.0	42.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#72
Bromobenzene
Concen: 5.178 ug/l
RT: 10.772 min Scan# 2949
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:156 Resp: 85960

Ion	Ratio	Lower	Upper
156	100		
77	175.5	0.0	343.6
158	97.1	0.0	193.0





#73
n-propylbenzene
Concen: 5.439 ug/l
RT: 10.894 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:120 Resp: 9593

Ion	Ratio	Lower	Upper
120	100		
91	464.6	369.8	554.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#74
2-Chlorotoluene
Concen: 5.399 ug/l
RT: 10.975 min Scan# 3012
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:126 Resp: 87769

Ion	Ratio	Lower	Upper
126	100		
91	307.9	0.0	623.8





#75
1,3,5-Trimethylbenzene
Concen: 5.489 ug/l
RT: 11.078 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:105 Resp: 31334

Ion	Ratio	Lower	Upper
105	100		
120	46.6	37.3	55.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#76
4-Chlorotoluene
Concen: 5.353 ug/l
RT: 11.087 min Scan# 3047
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:126 Resp: 89247

Ion	Ratio	Lower	Upper
126	100		
91	349.6	0.0	703.6





#77
tert-Butylbenzene
Concen: 5.183 ug/l
RT: 11.409 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:119 Resp: 29925

Ion	Ratio	Lower	Upper
119	100		
91	59.8	29.4	88.3
134	22.2	17.6	26.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#78
1,2,4-Trimethylbenzene
Concen: 5.539 ug/l
RT: 11.457 min Scan# 3162
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:105 Resp: 313777

Ion	Ratio	Lower	Upper
105	100		
120	43.3	21.9	65.7





#79
sec-Butylbenzene
Concen: 5.362 ug/l
RT: 11.634 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:105 Resp: 39415

Ion	Ratio	Lower	Upper
105	100		
134	18.7	15.1	22.7



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#80
Nitrobenzene
Concen: 23.083 ug/l m
RT: 13.209 min Scan# 3707
Delta R.T. 0.006 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 77 Resp: 10183

Ion	Ratio	Lower	Upper
77	100		
123	33.5	18.9	67.1
65	12.1	11.9	15.1





#81
p-Isopropyltoluene
Concen: 5.461 ug/l
RT: 11.782 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:119 Resp: 31680

Ion	Ratio	Lower	Upper
119	100		
134	25.7	20.3	30.5
91	24.2	19.4	29.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#82
1,3-Dichlorobenzene
Concen: 5.096 ug/l
RT: 11.737 min Scan# 3249
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:146 Resp: 164073

Ion	Ratio	Lower	Upper
146	100		
111	41.7	32.8	49.2
148	64.2	51.1	76.7





#83
1,4-Dichlorobenzene
Concen: 5.244 ug/l
RT: 11.827 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:146 Resp: 16515

Ion	Ratio	Lower	Upper
146	100		
111	39.8	32.1	48.1
148	63.3	50.2	75.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#84
n-Butylbenzene
Concen: 5.285 ug/l
RT: 12.200 min Scan# 3393
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 91 Resp: 276461

Ion	Ratio	Lower	Upper
91	100		
92	50.7	41.8	62.8
134	23.0	18.6	28.0





#85
1,2-Dichlorobenzene
Concen: 5.072 ug/l
RT: 12.203 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:146 Resp: 15695

Ion	Ratio	Lower	Upper
146	100		
111	43.1	21.9	65.7
148	64.6	32.3	96.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#86
1,2-Dibromo-3-Chloropropane
Concen: 5.220 ug/l
RT: 12.987 min Scan# 3638
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion: 75 Resp: 12084

Ion	Ratio	Lower	Upper
75	100		
155	77.3	63.5	95.3
157	96.9	81.8	122.6





#87
1,2,4-Trichlorobenzene
Concen: 5.203 ug/l
RT: 13.830 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:180 Resp: 7851

Ion	Ratio	Lower	Upper
180	100		
182	97.4	76.6	115.0
145	30.4	25.4	38.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#88
Hexachlorobutadiene
Concen: 4.939 ug/l
RT: 14.010 min Scan# 3956
Delta R.T. -0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:225 Resp: 53269

Ion	Ratio	Lower	Upper
225	100		
223	63.8	49.5	74.3
227	63.6	51.0	76.4





#89
Naphthalene
Concen: 4.286 ug/l
RT: 14.081 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:128 Resp: 12284

Ion	Ratio	Lower	Upper
128	100		
127	13.3	10.6	16.0
129	9.8	8.6	13.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#90
1,2,3-Trichlorobenzene
Concen: 5.189 ug/l
RT: 14.319 min Scan# 4052
Delta R.T. -0.003 min
Lab File: VU063222.D
Acq: 10 Feb 2025 14:23

Tgt Ion:180 Resp: 76620

Ion	Ratio	Lower	Upper
180	100		
182	95.8	78.2	117.2
145	33.2	26.1	39.1



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063223.D
Acq On : 10 Feb 2025 15:06
Operator : MD/SY
Sample : VSTDICCC010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 04:00:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Fluorobenzene	6.103	96	57821	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.624	95	20703	1.085	ug/l	0.00
Spiked Amount 1.000			Recovery = 109.000%			
68) 1,2-Dichlorobenzene-d4	12.184	152	19878	1.002	ug/l	0.00
Spiked Amount 1.000			Recovery = 100.000%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.377	85	176288	9.384	ug/l	100
3) Chloromethane	1.515	50	203794	9.420	ug/l	100
4) Vinyl Chloride	1.599	62	211768	9.893	ug/l	100
5) Bromomethane	1.843	94	110441	11.156	ug/l	100
6) Chloroethane	1.924	64	128364	9.521	ug/l	100
7) Trichlorofluoromethane	2.129	101	244670	9.646	ug/l	100
8) 1,1,2-Trichloro-1,2,2-...	2.570	101	133515	9.273	ug/l	100
9) 1,1-Dichloroethene	2.567	96	143583	9.788	ug/l	100
10) Iodomethane	2.708	142	234351	10.162	ug/l	100
11) Allyl Chloride	2.911	41	211288	10.026	ug/l	100
12) Acrylonitrile	3.300	53	74970	22.485	ug/l	100
13) Acetone	2.612	43	136864	52.606	ug/l	100
14) Carbon Disulfide	2.782	76	497871	9.707	ug/l	100
15) Methylene Chloride	3.030	84	175770	9.699	ug/l	100
16) trans-1,2-Dichloroethene	3.338	96	166021	9.916	ug/l	100
17) 1,1-Dichloroethane	3.853	63	311213	9.862	ug/l	100
18) 2-Butanone	4.685	43	227157	54.048	ug/l	100
19) Cyclohexane	5.377	56	262184m	10.340	ug/l	
20) Methylcyclohexane	6.753	83	259112	10.304	ug/l	100
21) 2,2-Dichloropropane	4.647	77	241757	9.820	ug/l	100
22) cis-1,2-Dichloroethene	4.650	96	182447	10.086	ug/l	100
23) Diethyl Ether	2.364	59	125036	9.933	ug/l	100
24) tert-Butyl Alcohol	3.158	59	136525	106.650	ug/l	100
25) Methyl tert-Butyl Ether	3.348	73	388982	10.617	ug/l	100
26) Bromochloromethane	4.959	128	78959	9.986	ug/l	100
27) Chloroform	5.071	83	315662	9.912	ug/l	100
28) 1,1,1-Trichloroethane	5.303	97	255321	9.897	ug/l	100
29) 1,1-Dichloropropene	5.512	75	235647	10.198	ug/l	100
30) Carbon Tetrachloride	5.512	117	218368	9.870	ug/l	100
31) Isopropyl Ether	3.975	45	470393	10.441	ug/l	100
32) Ethyl-t-butyl ether	4.483	59	436320	10.649	ug/l	100
33) Tert-Amyl methyl ether	5.927	73	397764	11.111	ug/l	100
34) Propionitrile	4.753	54	71557	54.918	ug/l	100
35) Benzene	5.759	78	712587	10.030	ug/l	100
36) 1,2-Dichloroethane	5.779	62	205344	10.014	ug/l	100
37) Trichloroethene	6.531	130	168312	9.962	ug/l	100
38) 1,2-Dichloropropane	6.779	63	188716	10.148	ug/l	100
39) Methacrylonitrile	4.959	41	56424	12.123	ug/l	100
40) Methyl acrylate	4.833	55	95663	11.258	ug/l	100
41) Tetrahydrofuran	5.039	42	59516	21.760	ug/l	100
42) 1-Chlorobutane	5.444	56	326490	10.327	ug/l	100
43) Dibromomethane	6.904	93	94106	9.994	ug/l	100
44) Bromodichloromethane	7.094	83	225813	10.303	ug/l	100

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063223.D
Acq On : 10 Feb 2025 15:06
Operator : MD/SY
Sample : VSTDICCC010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 04:00:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
45) 4-Methyl-2-Pentanone	7.775	43	568881	57.627 ug/l	100
46) t-1,4-Dichloro-2-butene	10.827	75	97536m	19.967 ug/l	
47) Methyl methacrylate	6.949	69	186394	23.502 ug/l	100
48) Ethyl methacrylate	8.322	69	181630	12.196 ug/l	100
49) Toluene	7.959	92	430442	10.535 ug/l	100
50) t-1,3-Dichloropropene	8.200	75	221262	11.027 ug/l	100
51) cis-1,3-Dichloropropene	7.595	75	265202	10.700 ug/l	100
52) 1,1,2-Trichloroethane	8.389	97	131655	10.370 ug/l	100
53) 1,3-Dichloropropane	8.563	76	234008	10.382 ug/l	100
54) 2-Hexanone	8.672	43	392522	58.266 ug/l	100
55) Dibromochloromethane	8.798	129	154464	10.575 ug/l	100
56) 1,2-Dibromoethane	8.914	107	126673	10.638 ug/l	100
58) Tetrachloroethene	8.544	164	135324	9.719 ug/l	100
59) Chlorobenzene	9.438	112	443950	10.297 ug/l	100
60) 1,1,1,2-Tetrachloroethane	9.521	131	158424	10.222 ug/l	100
61) Pentachloroethane	11.415	117	139274	10.060 ug/l	100
62) Hexachloroethane	12.467	117	128179	10.465 ug/l	100
63) Ethyl Benzene	9.560	91	811424	10.912 ug/l	100
64) m/p-Xylenes	9.682	106	619391	22.300 ug/l	100
65) o-Xylene	10.090	106	297309	10.934 ug/l	100
66) Styrene	10.103	104	495181	11.443 ug/l	100
67) Bromoform	10.280	173	90220	10.883 ug/l	100
69) Isopropylbenzene	10.473	105	700805	10.963 ug/l	100
70) 1,1,2,2-Tetrachloroethane	10.772	83	179045	10.464 ug/l	100
71) 1,2,3-Trichloropropane	10.814	75	117848m	9.083 ug/l	
72) Bromobenzene	10.772	156	181776	10.554 ug/l	100
73) n-propylbenzene	10.894	120	204779	11.189 ug/l	100
74) 2-Chlorotoluene	10.975	126	182649	10.829 ug/l	100
75) 1,3,5-Trimethylbenzene	11.077	105	664048	11.211 ug/l	100
76) 4-Chlorotoluene	11.087	126	185864	10.745 ug/l	100
77) tert-Butylbenzene	11.409	119	646436	10.791 ug/l	100
78) 1,2,4-Trimethylbenzene	11.457	105	668930	11.382 ug/l	100
79) sec-Butylbenzene	11.634	105	833792	10.932 ug/l	100
80) Nitrobenzene	13.203	77	25795	50.821 ug/l	100
81) p-Isopropyltoluene	11.782	119	675304	11.219 ug/l	100
82) 1,3-Dichlorobenzene	11.737	146	344574	10.314 ug/l	100
83) 1,4-Dichlorobenzene	11.827	146	347616	10.638 ug/l	100
84) n-Butylbenzene	12.200	91	603657	11.122 ug/l	100
85) 1,2-Dichlorobenzene	12.203	146	332012	10.342 ug/l	100
86) 1,2-Dibromo-3-Chloropr...	12.987	75	28133	11.713 ug/l	100
87) 1,2,4-Trichlorobenzene	13.830	180	177895	11.361 ug/l	100
88) Hexachlorobutadiene	14.013	225	105064	9.389 ug/l	100
89) Naphthalene	14.077	128	321601	9.843 ug/l	100
90) 1,2,3-Trichlorobenzene	14.322	180	174986	11.421 ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063223.D
Acq On : 10 Feb 2025 15:06
Operator : MD/SY
Sample : VSTDICCC010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Feb 11 04:00:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.103 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 96 Resp: 5782

Ion	Ratio	Lower	Upper
96	100		
70	19.5	15.6	23.4
95	9.1	7.3	10.9
97	0.0	0.0	0.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#2
Dichlorodifluoromethane
Concen: 9.384 ug/l
RT: 1.377 min Scan# 27
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 85 Resp: 176288

Ion	Ratio	Lower	Upper
85	100		
87	32.0	16.0	48.0





#3
Chloromethane
Concen: 9.420 ug/l
RT: 1.515 min Scan# 7
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 50 Resp: 20379

Ion	Ratio	Lower	Upper
50	100		
52	32.3	25.8	38.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#4
Vinyl Chloride
Concen: 9.893 ug/l
RT: 1.599 min Scan# 96
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 62 Resp: 211768

Ion	Ratio	Lower	Upper
62	100		
64	31.7	25.4	38.0





#5
Bromomethane
Concen: 11.156 ug/l
RT: 1.843 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 94 Resp: 11044

Ion	Ratio	Lower	Upper
94	100		
96	94.6	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#6
Chloroethane
Concen: 9.521 ug/l
RT: 1.924 min Scan# 197
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 64 Resp: 128364

Ion	Ratio	Lower	Upper
64	100		
66	32.3	25.8	38.8





#7
Trichlorofluoromethane
Concen: 9.646 ug/l
RT: 2.129 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:101 Resp: 24467

Ion	Ratio	Lower	Upper
101	100		
103	64.3	51.4	77.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 9.273 ug/l
RT: 2.570 min Scan# 398
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:101 Resp: 133515

Ion	Ratio	Lower	Upper
101	100		
85	44.2	35.4	53.0
151	73.1	58.5	87.7





#9
1,1-Dichloroethene
Concen: 9.788 ug/l
RT: 2.567 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 96 Resp: 14358

Ion	Ratio	Lower	Upper
96	100		
61	164.3	0.0	492.9
98	62.0	0.0	124.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#10
Iodomethane
Concen: 10.162 ug/l
RT: 2.708 min Scan# 441
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:142 Resp: 234351

Ion	Ratio	Lower	Upper
142	100		
127	43.1	34.5	51.7





#11
Allyl Chloride
Concen: 10.026 ug/l
RT: 2.911 min Scan# 5
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 41 Resp: 21128

Ion	Ratio	Lower	Upper
41	100		
39	72.4	57.9	86.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#12
Acrylonitrile
Concen: 22.485 ug/l
RT: 3.300 min Scan# 625
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 53 Resp: 74970

Ion	Ratio	Lower	Upper
53	100		
52	80.2	64.2	96.2
51	38.5	30.8	46.2





#13
Acetone
Concen: 52.606 ug/l
RT: 2.612 min Scan# 41
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 43 Resp: 13686

Ion	Ratio	Lower	Upper
43	100		
58	34.2	27.4	41.0

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#14
Carbon Disulfide
Concen: 9.707 ug/l
RT: 2.782 min Scan# 464
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 76 Resp: 497871

Ion	Ratio	Lower	Upper
76	100		
78	9.0	7.2	10.8





#15
Methylene Chloride
Concen: 9.699 ug/l
RT: 3.030 min Scan# 54
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 84 Resp: 17577

Ion	Ratio	Lower	Upper
84	100		
49	120.5	96.4	144.6
51	37.4	0.0	74.8
86	64.1	51.3	76.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#16
trans-1,2-Dichloroethene
Concen: 9.916 ug/l
RT: 3.338 min Scan# 637
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 96 Resp: 166021

Ion	Ratio	Lower	Upper
96	100		
61	141.8	113.4	170.2
98	64.0	51.2	76.8





#17
1,1-Dichloroethane
Concen: 9.862 ug/l
RT: 3.853 min Scan# 7
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 63 Resp: 31121

Ion	Ratio	Lower	Upper
63	100		
98	6.6	3.3	9.9
100	4.1	2.1	6.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#18
2-Butanone
Concen: 54.048 ug/l
RT: 4.685 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 43 Resp: 227157

Ion	Ratio	Lower	Upper
43	100		
57	8.5	0.0	17.0





#19
Cyclohexane
Concen: 10.340 ug/l m
RT: 5.377 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 56 Resp: 26218

Ion	Ratio	Lower	Upper
56	100		
69	30.6	24.5	36.7
84	81.5	65.2	97.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#20
Methylcyclohexane
Concen: 10.304 ug/l
RT: 6.753 min Scan# 1699
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 83 Resp: 259112

Ion	Ratio	Lower	Upper
83	100		
55	78.9	63.1	94.7
98	44.0	35.2	52.8





#21
2,2-Dichloropropane
Concen: 9.820 ug/l
RT: 4.647 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 77 Resp: 24175

Ion	Ratio	Lower	Upper
77	100		
97	23.1	18.5	27.7



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#22
cis-1,2-Dichloroethene
Concen: 10.086 ug/l
RT: 4.650 min Scan# 1045
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 96 Resp: 182447

Ion	Ratio	Lower	Upper
96	100		
61	149.3	0.0	373.3
98	63.9	31.9	95.9





#23
Diethyl Ether
Concen: 9.933 ug/l
RT: 2.364 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 59 Resp: 12503

Ion	Ratio	Lower	Upper
59	100		
45	72.2	57.8	86.6
74	72.1	57.7	86.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#24
tert-Butyl Alcohol
Concen: 106.650 ug/l
RT: 3.158 min Scan# 581
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 59 Resp: 136525

Ion	Ratio	Lower	Upper
59	100		
57	9.4	7.5	11.3





#25
Methyl tert-Butyl Ether
Concen: 10.617 ug/l
RT: 3.348 min Scan# 64
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 73 Resp: 38898

Ion	Ratio	Lower	Upper
73	100		
43	18.6	14.9	22.3

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#26
Bromochloromethane
Concen: 9.986 ug/l
RT: 4.959 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:128 Resp: 78959

Ion	Ratio	Lower	Upper
128	100		
49	171.7	0.0	343.4
130	128.6	102.9	154.3





#27
Chloroform
Concen: 9.912 ug/l
RT: 5.071 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 83 Resp: 31566

Ion	Ratio	Lower	Upper
83	100		
85	64.9	0.0	129.8

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#28
1,1,1-Trichloroethane
Concen: 9.897 ug/l
RT: 5.303 min Scan# 1248
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 97 Resp: 255321

Ion	Ratio	Lower	Upper
97	100		
99	64.7	32.4	97.0
61	47.5	23.8	71.2





#29
1,1-Dichloropropene
Concen: 10.198 ug/l
RT: 5.512 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 75 Resp: 23564

Ion	Ratio	Lower	Upper
75	100		
110	34.3	17.2	51.5
77	30.8	24.6	37.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#30
Carbon Tetrachloride
Concen: 9.870 ug/l
RT: 5.512 min Scan# 1313
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:117 Resp: 218368

Ion	Ratio	Lower	Upper
117	100		
119	95.9	76.7	115.1
121	31.9	25.5	38.3





#31
Isopropyl Ether
Concen: 10.441 ug/l
RT: 3.975 min Scan# 8
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 45 Resp: 47039

Ion	Ratio	Lower	Upper
45	100		
43	51.4	25.7	77.1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#32
Ethyl-t-butyl ether
Concen: 10.649 ug/l
RT: 4.483 min Scan# 993
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 59 Resp: 436320

Ion	Ratio	Lower	Upper
59	100		
87	40.8	32.6	49.0





#33
Tert-Amyl methyl ether
Concen: 11.111 ug/l
RT: 5.927 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 73 Resp: 39776

Ion	Ratio	Lower	Upper
73	100		
87	23.6	18.9	28.3
71	11.2	9.0	13.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#34
Propionitrile
Concen: 54.918 ug/l
RT: 4.753 min Scan# 1077
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 54 Resp: 71557

Ion	Ratio	Lower	Upper
54	100		
52	20.4	16.3	24.5
55	14.7	11.8	17.6
40	5.4	4.3	6.5





#35
Benzene
Concen: 10.030 ug/l
RT: 5.759 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 78 Resp: 71258

Ion	Ratio	Lower	Upper
78	100		
77	23.7	19.0	28.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#36
1,2-Dichloroethane
Concen: 10.014 ug/l
RT: 5.779 min Scan# 1396
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 62 Resp: 205344

Ion	Ratio	Lower	Upper
62	100		
98	8.6	6.9	10.3





#37
Trichloroethene
Concen: 9.962 ug/l
RT: 6.531 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:130 Resp: 16831

Ion	Ratio	Lower	Upper
130	100		
95	104.0	83.2	124.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#38
1,2-Dichloropropane
Concen: 10.148 ug/l
RT: 6.779 min Scan# 1707
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 63 Resp: 188716

Ion	Ratio	Lower	Upper
63	100		
65	31.6	25.3	37.9





#39
Methacrylonitrile
Concen: 12.123 ug/l
RT: 4.959 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 41 Resp: 5642

Ion	Ratio	Lower	Upper
41	100		
67	80.6	64.5	96.7
39	53.8	43.0	64.6
52	31.6	25.3	37.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#40
Methyl acrylate
Concen: 11.258 ug/l
RT: 4.833 min Scan# 1102
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 55 Resp: 95663

Ion	Ratio	Lower	Upper
55	100		
85	16.6	13.3	19.9
58	9.1	7.3	10.9
42	8.6	6.9	10.3





#41
Tetrahydrofuran
Concen: 21.760 ug/l
RT: 5.039 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 42 Resp: 5951

Ion	Ratio	Lower	Upper
42	100		
72	51.9	41.5	62.3
71	46.5	37.2	55.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#42
1-Chlorobutane
Concen: 10.327 ug/l
RT: 5.444 min Scan# 1292
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 56 Resp: 326490

Ion	Ratio	Lower	Upper
56	100		
41	52.5	26.3	78.8





#43
Dibromomethane
Concen: 9.994 ug/l
RT: 6.904 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 93 Resp: 9410

Ion	Ratio	Lower	Upper
93	100		
95	84.0	67.2	100.8
174	94.6	75.7	113.5

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#44
Bromodichloromethane
Concen: 10.303 ug/l
RT: 7.094 min Scan# 1805
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 83 Resp: 225813

Ion	Ratio	Lower	Upper
83	100		
85	64.6	51.7	77.5
127	8.4	6.7	10.1





#45
4-Methyl-2-Pentanone
Concen: 57.627 ug/l
RT: 7.775 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 43 Resp: 56888

Ion	Ratio	Lower	Upper
43	100		
58	41.7	20.8	62.5
85	18.5	14.8	22.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#46
t-1,4-Dichloro-2-butene
Concen: 19.967 ug/l m
RT: 10.827 min Scan# 2966
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 75 Resp: 97536

Ion	Ratio	Lower	Upper
75	100		
53	80.6	64.5	96.7
89	54.3	43.4	65.2
88	39.0	31.2	46.8





#47
Methyl methacrylate
Concen: 23.502 ug/l
RT: 6.949 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 69 Resp: 18639

Ion	Ratio	Lower	Upper
69	100		
41	108.5	0.0	217.0
39	59.6	47.7	71.5
100	34.2	27.4	41.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#48
Ethyl methacrylate
Concen: 12.196 ug/l
RT: 8.322 min Scan# 2187
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 69 Resp: 181630

Ion	Ratio	Lower	Upper
69	100		
41	61.3	30.6	92.0





#49
Toluene
Concen: 10.535 ug/l
RT: 7.959 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 92 Resp: 43044

Ion	Ratio	Lower	Upper
92	100		
91	177.2	141.8	212.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#50
t-1,3-Dichloropropene
Concen: 11.027 ug/l
RT: 8.200 min Scan# 2149
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 75 Resp: 221262

Ion	Ratio	Lower	Upper
75	100		
77	32.4	25.9	38.9





#51
cis-1,3-Dichloropropene
Concen: 10.700 ug/l
RT: 7.595 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 75 Resp: 26520

Ion	Ratio	Lower	Upper
75	100		
77	31.6	25.3	37.9
39	41.9	33.5	50.3



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#52
1,1,2-Trichloroethane
Concen: 10.370 ug/l
RT: 8.389 min Scan# 2208
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 97 Resp: 131655

Ion	Ratio	Lower	Upper
97	100		
83	91.2	73.0	109.4
85	57.9	46.3	69.5
99	60.6	48.5	72.7





#53
1,3-Dichloropropane
Concen: 10.382 ug/l
RT: 8.563 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 76 Resp: 23400

Ion	Ratio	Lower	Upper
76	100		
78	32.9	26.3	39.5

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#54
2-Hexanone
Concen: 58.266 ug/l
RT: 8.672 min Scan# 2296
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 43 Resp: 392522

Ion	Ratio	Lower	Upper
43	100		
58	58.0	38.0	78.0
57	19.1	0.0	39.1





#55
Dibromochloromethane
Concen: 10.575 ug/l
RT: 8.798 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:129 Resp: 15446

Ion	Ratio	Lower	Upper
129	100		
127	75.5	60.4	90.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#56
1,2-Dibromoethane
Concen: 10.638 ug/l
RT: 8.914 min Scan# 2371
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:107 Resp: 126673

Ion	Ratio	Lower	Upper
107	100		
109	93.9	0.0	187.8





#57
4-Bromofluorobenzene
Concen: 1.085 ug/l
RT: 10.624 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 95 Resp: 2070

Ion	Ratio	Lower	Upper
95	100		
174	73.3	58.6	88.0
176	72.8	58.2	87.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#58
Tetrachloroethene
Concen: 9.719 ug/l
RT: 8.544 min Scan# 2256
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:164 Resp: 135324

Ion	Ratio	Lower	Upper
164	100		
166	126.7	101.4	152.0
129	96.2	77.0	115.4
131	95.4	76.3	114.5





#59
Chlorobenzene
Concen: 10.297 ug/l
RT: 9.438 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:112 Resp: 44395

Ion	Ratio	Lower	Upper
112	100		
114	32.1	25.7	38.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#60
1,1,1,2-Tetrachloroethane
Concen: 10.222 ug/l
RT: 9.521 min Scan# 2560
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:131 Resp: 158424

Ion	Ratio	Lower	Upper
131	100		
133	95.9	76.7	115.1
119	68.0	54.4	81.6





#61
Pentachloroethane
Concen: 10.060 ug/l
RT: 11.415 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:117 Resp: 13927

Ion	Ratio	Lower	Upper
117	100		
167	81.3	65.0	97.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#62
Hexachloroethane
Concen: 10.465 ug/l
RT: 12.467 min Scan# 3476
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:117 Resp: 128179

Ion	Ratio	Lower	Upper
117	100		
201	76.6	61.3	91.9





#63
Ethyl Benzene
Concen: 10.912 ug/l
RT: 9.560 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 91 Resp: 81142

Ion	Ratio	Lower	Upper
91	100		
106	30.2	24.2	36.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#64
m/p-Xylenes
Concen: 22.300 ug/l
RT: 9.682 min Scan# 2610
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:106 Resp: 619391

Ion	Ratio	Lower	Upper
106	100		
91	208.6	166.9	250.3





#65
o-Xylene
Concen: 10.934 ug/l
RT: 10.090 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:106 Resp: 29730

Ion	Ratio	Lower	Upper
106	100		
91	221.9	110.9	332.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#66
Styrene
Concen: 11.443 ug/l
RT: 10.103 min Scan# 2741
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:104 Resp: 495181

Ion	Ratio	Lower	Upper
104	100		
78	51.5	41.2	61.8
103	56.0	44.8	67.2





#67
Bromoform
Concen: 10.883 ug/l
RT: 10.280 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:173 Resp: 9022

Ion	Ratio	Lower	Upper
173	100		
175	48.7	39.0	58.4
254	9.6	7.7	11.5

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#68
1,2-Dichlorobenzene-d4
Concen: 1.002 ug/l
RT: 12.184 min Scan# 3388
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:152 Resp: 19878

Ion	Ratio	Lower	Upper
152	100		
115	137.6	0.0	275.2
150	329.2	0.0	658.4





#69
Isopropylbenzene
Concen: 10.963 ug/l
RT: 10.473 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:105 Resp: 70080

Ion	Ratio	Lower	Upper
105	100		
120	25.6	12.8	38.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#70
1,1,2,2-Tetrachloroethane
Concen: 10.464 ug/l
RT: 10.772 min Scan# 2949
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 83 Resp: 179045

Ion	Ratio	Lower	Upper
83	100		
131	9.2	7.4	11.0
85	64.8	51.8	77.8





#71
1,2,3-Trichloropropane
Concen: 9.083 ug/l m
RT: 10.814 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 75 Resp: 11784

Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0
110	38.5	0.0	77.0
97	21.1	0.0	42.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#72
Bromobenzene
Concen: 10.554 ug/l
RT: 10.772 min Scan# 2949
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:156 Resp: 181776

Ion	Ratio	Lower	Upper
156	100		
77	171.8	0.0	343.6
158	96.5	0.0	193.0





#73
n-propylbenzene
Concen: 11.189 ug/l
RT: 10.894 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:120 Resp: 20477

Ion	Ratio	Lower	Upper
120	100		
91	462.2	369.8	554.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#74
2-Chlorotoluene
Concen: 10.829 ug/l
RT: 10.975 min Scan# 3012
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:126 Resp: 182649

Ion	Ratio	Lower	Upper
126	100		
91	311.9	0.0	623.8





#75
1,3,5-Trimethylbenzene
Concen: 11.211 ug/l
RT: 11.077 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:105 Resp: 66404

Ion	Ratio	Lower	Upper
105	100		
120	46.6	37.3	55.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#76
4-Chlorotoluene
Concen: 10.745 ug/l
RT: 11.087 min Scan# 3047
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:126 Resp: 185864

Ion	Ratio	Lower	Upper
126	100		
91	351.8	0.0	703.6





#77
tert-Butylbenzene
Concen: 10.791 ug/l
RT: 11.409 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:119 Resp: 64643

Ion	Ratio	Lower	Upper
119	100		
91	58.9	29.4	88.3
134	22.0	17.6	26.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#78
1,2,4-Trimethylbenzene
Concen: 11.382 ug/l
RT: 11.457 min Scan# 3162
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:105 Resp: 668930

Ion	Ratio	Lower	Upper
105	100		
120	43.8	21.9	65.7





#79
sec-Butylbenzene
Concen: 10.932 ug/l
RT: 11.634 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:105 Resp: 83379

Ion	Ratio	Lower	Upper
105	100		
134	18.9	15.1	22.7

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#80
Nitrobenzene
Concen: 50.821 ug/l
RT: 13.203 min Scan# 3705
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 77 Resp: 25795

Ion	Ratio	Lower	Upper
77	100		
123	43.0	18.9	67.1
65	13.5	11.9	15.1





#81
p-Isopropyltoluene
Concen: 11.219 ug/l
RT: 11.782 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:119 Resp: 67530

Ion	Ratio	Lower	Upper
119	100		
134	25.4	20.3	30.5
91	24.3	19.4	29.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#82
1,3-Dichlorobenzene
Concen: 10.314 ug/l
RT: 11.737 min Scan# 3249
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:146 Resp: 344574

Ion	Ratio	Lower	Upper
146	100		
111	41.0	32.8	49.2
148	63.9	51.1	76.7





#83
1,4-Dichlorobenzene
Concen: 10.638 ug/l
RT: 11.827 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:146 Resp: 34761

Ion	Ratio	Lower	Upper
146	100		
111	40.1	32.1	48.1
148	62.8	50.2	75.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#84
n-Butylbenzene
Concen: 11.122 ug/l
RT: 12.200 min Scan# 3393
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 91 Resp: 603657

Ion	Ratio	Lower	Upper
91	100		
92	52.3	41.8	62.8
134	23.3	18.6	28.0





#85
1,2-Dichlorobenzene
Concen: 10.342 ug/l
RT: 12.203 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:146 Resp: 33201

Ion	Ratio	Lower	Upper
146	100		
111	43.8	21.9	65.7
148	64.6	32.3	96.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#86
1,2-Dibromo-3-Chloropropane
Concen: 11.713 ug/l
RT: 12.987 min Scan# 3638
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion: 75 Resp: 28133

Ion	Ratio	Lower	Upper
75	100		
155	79.4	63.5	95.3
157	102.2	81.8	122.6





#87
1,2,4-Trichlorobenzene
Concen: 11.361 ug/l
RT: 13.830 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:180 Resp: 17789

Ion	Ratio	Lower	Upper
180	100		
182	95.8	76.6	115.0
145	31.8	25.4	38.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#88
Hexachlorobutadiene
Concen: 9.389 ug/l
RT: 14.013 min Scan# 3957
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:225 Resp: 105064

Ion	Ratio	Lower	Upper
225	100		
223	61.9	49.5	74.3
227	63.7	51.0	76.4





#89
Naphthalene
Concen: 9.843 ug/l
RT: 14.077 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:128 Resp: 32160

Ion	Ratio	Lower	Upper
128	100		
127	13.3	10.6	16.0
129	10.8	8.6	13.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICCC010

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#90
1,2,3-Trichlorobenzene
Concen: 11.421 ug/l
RT: 14.322 min Scan# 4053
Delta R.T. 0.000 min
Lab File: VU063223.D
Acq: 10 Feb 2025 15:06

Tgt Ion:180 Resp: 174986

Ion	Ratio	Lower	Upper
180	100		
182	97.7	78.2	117.2
145	32.6	26.1	39.1



Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063224.D
Acq On : 10 Feb 2025 15:33
Operator : MD/SY
Sample : VSTDICC015
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 04:01:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Fluorobenzene	6.103	96	58075	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.624	95	20689	1.080	ug/l	0.00
Spiked Amount 1.000			Recovery = 108.000%			
68) 1,2-Dichlorobenzene-d4	12.184	152	22516	1.130	ug/l	0.00
Spiked Amount 1.000			Recovery = 113.000%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.377	85	263402	13.960	ug/l	100
3) Chloromethane	1.515	50	300062	13.809	ug/l	97
4) Vinyl Chloride	1.596	62	311009	14.466	ug/l	100
6) Chloroethane	1.904	64	188210	13.899	ug/l	99
7) Trichlorofluoromethane	2.116	101	362692	14.236	ug/l	100
8) 1,1,2-Trichloro-1,2,2-...	2.563	101	200170	13.842	ug/l	99
9) 1,1-Dichloroethene	2.560	96	210091	14.259	ug/l	98
10) Iodomethane	2.705	142	329697	14.234	ug/l	99
11) Allyl Chloride	2.904	41	311245	14.705	ug/l	100
12) Acrylonitrile	3.300	53	102914	30.731	ug/l	99
13) Acetone	2.612	43	172446	65.992	ug/l	100
14) Carbon Disulfide	2.776	76	716395	13.906	ug/l	100
15) Methylene Chloride	3.026	84	251659	13.825	ug/l	99
16) trans-1,2-Dichloroethene	3.335	96	242715	14.434	ug/l	98
17) 1,1-Dichloroethane	3.846	63	453378	14.304	ug/l	100
18) 2-Butanone	4.685	43	304814	72.208	ug/l	99
19) Cyclohexane	5.370	56	400586m	15.729	ug/l	
20) Methylcyclohexane	6.750	83	398925	15.795	ug/l	99
21) 2,2-Dichloropropane	4.644	77	354900	14.352	ug/l	100
22) cis-1,2-Dichloroethene	4.647	96	270006	14.861	ug/l	100
23) Diethyl Ether	2.364	59	177089	14.007	ug/l	100
24) tert-Butyl Alcohol	3.177	59	185785	146.971	ug/l	100
25) Methyl tert-Butyl Ether	3.348	73	553339	15.037	ug/l	99
26) Bromochloromethane	4.952	128	113229	14.257	ug/l	98
27) Chloroform	5.071	83	457911	14.316	ug/l	100
28) 1,1,1-Trichloroethane	5.300	97	373842	14.428	ug/l	99
29) 1,1-Dichloropropene	5.509	75	347711	14.983	ug/l	99
30) Carbon Tetrachloride	5.509	117	321059	14.448	ug/l	99
31) Isopropyl Ether	3.975	45	703540	15.547	ug/l	98
32) Ethyl-t-butyl ether	4.483	59	649553	15.784	ug/l	99
33) Tert-Amyl methyl ether	5.923	73	575729	16.013	ug/l	100
34) Propionitrile	4.753	54	103055	78.746	ug/l	99
35) Benzene	5.756	78	1044796	14.641	ug/l	99
36) 1,2-Dichloroethane	5.775	62	291840	14.170	ug/l	100
37) Trichloroethene	6.528	130	244774	14.424	ug/l	100
38) 1,2-Dichloropropane	6.775	63	273676	14.653	ug/l	99
39) Methacrylonitrile	4.956	41	79040	16.908	ug/l	99
40) Methyl acrylate	4.830	55	133276	15.616	ug/l	99
41) Tetrahydrofuran	5.036	42	79845	29.065	ug/l	99
42) 1-Chlorobutane	5.441	56	481169	15.153	ug/l	99
43) Dibromomethane	6.904	93	133485	14.114	ug/l	98
44) Bromodichloromethane	7.090	83	323143	14.680	ug/l	99
45) 4-Methyl-2-Pentanone	7.779	43	794494	80.129	ug/l	99

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063224.D
Acq On : 10 Feb 2025 15:33
Operator : MD/SY
Sample : VSTDICC015
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Quant Time: Feb 11 04:01:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

	Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
46)	t-1,4-Dichloro-2-butene	10.827	75	143327m	29.213 ug/l	
47)	Methyl methacrylate	6.946	69	256757	32.232 ug/l	100
48)	Ethyl methacrylate	8.322	69	257931	17.244 ug/l	100
49)	Toluene	7.955	92	638079	15.549 ug/l	100
50)	t-1,3-Dichloropropene	8.197	75	320246	15.891 ug/l	98
51)	cis-1,3-Dichloropropene	7.595	75	386542	15.527 ug/l	99
52)	1,1,2-Trichloroethane	8.386	97	186154	14.598 ug/l	98
53)	1,3-Dichloropropane	8.563	76	330381	14.594 ug/l	99
54)	2-Hexanone	8.672	43	557608	82.409 ug/l	99
55)	Dibromochloromethane	8.798	129	219151	14.939 ug/l	97
56)	1,2-Dibromoethane	8.910	107	178579	14.932 ug/l	99
58)	Tetrachloroethene	8.541	164	195759	13.998 ug/l	99
59)	Chlorobenzene	9.434	112	659996	15.241 ug/l	100
60)	1,1,1,2-Tetrachloroethane	9.521	131	233573	15.005 ug/l	99
61)	Pentachloroethane	11.415	117	217502	15.642 ug/l	100
62)	Hexachloroethane	12.463	117	202372	16.451 ug/l	100
63)	Ethyl Benzene	9.560	91	1218926	16.321 ug/l	99
64)	m/p-Xylenes	9.682	106	919492	32.960 ug/l	99
65)	o-Xylene	10.090	106	447318	16.379 ug/l	100
66)	Styrene	10.103	104	750530	17.269 ug/l	100
67)	Bromoform	10.280	173	131670	15.814 ug/l	99
69)	Isopropylbenzene	10.473	105	1068251	16.638 ug/l	100
70)	1,1,2,2-Tetrachloroethane	10.769	83	264258	15.376 ug/l	100
71)	1,2,3-Trichloropropane	10.814	75	200210m	15.363 ug/l	
72)	Bromobenzene	10.772	156	273315	15.799 ug/l	99
73)	n-propylbenzene	10.894	120	312734	17.012 ug/l	100
74)	2-Chlorotoluene	10.975	126	279514	16.499 ug/l	98
75)	1,3,5-Trimethylbenzene	11.077	105	1018594	17.122 ug/l	100
76)	4-Chlorotoluene	11.087	126	284996	16.404 ug/l	100
77)	tert-Butylbenzene	11.409	119	1004388	16.693 ug/l	100
78)	1,2,4-Trimethylbenzene	11.457	105	1038794	17.598 ug/l	99
79)	sec-Butylbenzene	11.630	105	1305019	17.036 ug/l	100
80)	Nitrobenzene	13.200	77	39246	75.007 ug/l	99
81)	p-Isopropyltoluene	11.782	119	1064335	17.605 ug/l	100
82)	1,3-Dichlorobenzene	11.733	146	532282	15.863 ug/l	99
83)	1,4-Dichlorobenzene	11.827	146	537516	16.378 ug/l	99
84)	n-Butylbenzene	12.196	91	994690	18.246 ug/l	100
85)	1,2-Dichlorobenzene	12.200	146	526346	16.323 ug/l	99
86)	1,2-Dibromo-3-Chloropr...	12.987	75	39960	16.565 ug/l	98
87)	1,2,4-Trichlorobenzene	13.830	180	288112	18.320 ug/l	99
88)	Hexachlorobutadiene	14.010	225	175875	15.648 ug/l	99
89)	Naphthalene	14.077	128	515515	15.328 ug/l	100
90)	1,2,3-Trichlorobenzene	14.318	180	277101	18.007 ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063224.D
Acq On : 10 Feb 2025 15:33
Operator : MD/SY
Sample : VSTDICC015
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Feb 11 04:01:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 03:56:39 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.103 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 96 Resp: 58075

Ion	Ratio	Lower	Upper
96	100		
70	21.2	15.6	23.4
95	8.9	7.3	10.9
97	0.0	0.0	0.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#2
Dichlorodifluoromethane
Concen: 13.960 ug/l
RT: 1.377 min Scan# 27
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 85 Resp: 263402

Ion	Ratio	Lower	Upper
85	100		
87	32.2	16.0	48.0





#3
Chloromethane
Concen: 13.809 ug/l
RT: 1.515 min Scan# 70
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 50 Resp: 300062

Ion	Ratio	Lower	Upper
50	100		
52	34.1	25.8	38.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#4
Vinyl Chloride
Concen: 14.466 ug/l
RT: 1.596 min Scan# 95
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 62 Resp: 311009

Ion	Ratio	Lower	Upper
62	100		
64	31.5	25.4	38.0





#6
Chloroethane
Concen: 13.899 ug/l
RT: 1.904 min Scan# 19
Delta R.T. -0.019 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 64 Resp: 188210

Ion	Ratio	Lower	Upper
64	100		
66	31.7	25.8	38.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#7
Trichlorofluoromethane
Concen: 14.236 ug/l
RT: 2.116 min Scan# 257
Delta R.T. -0.013 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:101 Resp: 362692

Ion	Ratio	Lower	Upper
101	100		
103	64.6	51.4	77.2





#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 13.842 ug/l
RT: 2.563 min Scan# 39
Delta R.T. -0.006 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Tgt Ion:101 Resp: 200170

Ion	Ratio	Lower	Upper
101	100		
85	44.3	35.4	53.0
151	72.0	58.5	87.7

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#9
1,1-Dichloroethene
Concen: 14.259 ug/l
RT: 2.560 min Scan# 395
Delta R.T. -0.006 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 96 Resp: 210091

Ion	Ratio	Lower	Upper
96	100		
61	166.6	0.0	492.9
98	63.7	0.0	124.0





#10
Iodomethane
Concen: 14.234 ug/l
RT: 2.705 min Scan# 44
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:142 Resp: 329697

Ion	Ratio	Lower	Upper
142	100		
127	43.7	34.5	51.7



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#11
Allyl Chloride
Concen: 14.705 ug/l
RT: 2.904 min Scan# 502
Delta R.T. -0.006 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 41 Resp: 311245

Ion	Ratio	Lower	Upper
41	100		
39	72.4	57.9	86.9





#12
Acrylonitrile
Concen: 30.731 ug/l
RT: 3.300 min Scan# 62
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 53 Resp: 102914

Ion	Ratio	Lower	Upper
53	100		
52	80.8	64.2	96.2
51	38.3	30.8	46.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#13
Acetone
Concen: 65.992 ug/l
RT: 2.612 min Scan# 411
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 43 Resp: 172446

Ion	Ratio	Lower	Upper
43	100		
58	34.3	27.4	41.0





#14
Carbon Disulfide
Concen: 13.906 ug/l
RT: 2.776 min Scan# 46
Delta R.T. -0.006 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 76 Resp: 716395

Ion	Ratio	Lower	Upper
76	100		
78	8.9	7.2	10.8

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#15
Methylene Chloride
Concen: 13.825 ug/l
RT: 3.026 min Scan# 540
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 84 Resp: 251659

Ion	Ratio	Lower	Upper
84	100		
49	119.7	96.4	144.6
51	37.9	0.0	74.8
86	65.1	51.3	76.9





#16
trans-1,2-Dichloroethene
Concen: 14.434 ug/l
RT: 3.335 min Scan# 63
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 96 Resp: 242715

Ion	Ratio	Lower	Upper
96	100		
61	143.8	113.4	170.2
98	62.9	51.2	76.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#17
1,1-Dichloroethane
Concen: 14.304 ug/l
RT: 3.846 min Scan# 795
Delta R.T. -0.006 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 63 Resp: 453378

Ion	Ratio	Lower	Upper
63	100		
98	6.5	3.3	9.9
100	4.3	2.1	6.2





#18
2-Butanone
Concen: 72.208 ug/l
RT: 4.685 min Scan# 10
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 43 Resp: 304814

Ion	Ratio	Lower	Upper
43	100		
57	8.7	0.0	17.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#19
Cyclohexane
Concen: 15.729 ug/l m
RT: 5.370 min Scan# 1269
Delta R.T. -0.006 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 56 Resp: 400586

Ion	Ratio	Lower	Upper
56	100		
69	30.3	24.5	36.7
84	81.7	65.2	97.8





#20
Methylcyclohexane
Concen: 15.795 ug/l
RT: 6.750 min Scan# 16
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 83 Resp: 398925

Ion	Ratio	Lower	Upper
83	100		
55	77.0	63.1	94.7
98	44.1	35.2	52.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#21
2,2-Dichloropropane
Concen: 14.352 ug/l
RT: 4.644 min Scan# 1043
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 77 Resp: 354900

Ion	Ratio	Lower	Upper
77	100		
97	22.9	18.5	27.7





#22
cis-1,2-Dichloroethene
Concen: 14.861 ug/l
RT: 4.647 min Scan# 10
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 96 Resp: 270006

Ion	Ratio	Lower	Upper
96	100		
61	149.6	0.0	373.3
98	63.7	31.9	95.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#23
Diethyl Ether
Concen: 14.007 ug/l
RT: 2.364 min Scan# 334
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 59 Resp: 177089

Ion	Ratio	Lower	Upper
59	100		
45	71.8	57.8	86.6
74	71.8	57.7	86.5





#24
tert-Butyl Alcohol
Concen: 146.971 ug/l
RT: 3.177 min Scan# 58
Delta R.T. 0.019 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 59 Resp: 185785

Ion	Ratio	Lower	Upper
59	100		
57	9.5	7.5	11.3



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#25
Methyl tert-Butyl Ether
Concen: 15.037 ug/l
RT: 3.348 min Scan# 640
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 73 Resp: 553339

Ion	Ratio	Lower	Upper
73	100		
43	18.9	14.9	22.3





#26
Bromochloromethane
Concen: 14.257 ug/l
RT: 4.952 min Scan# 11
Delta R.T. -0.006 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:128 Resp: 113229

Ion	Ratio	Lower	Upper
128	100		
49	170.1	0.0	343.4
130	126.0	102.9	154.3

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#27
Chloroform
Concen: 14.316 ug/l
RT: 5.071 min Scan# 1176
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 83 Resp: 457911

Ion	Ratio	Lower	Upper
83	100		
85	64.9	0.0	129.8





#28
1,1,1-Trichloroethane
Concen: 14.428 ug/l
RT: 5.300 min Scan# 12
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 97 Resp: 373842

Ion	Ratio	Lower	Upper
97	100		
99	64.4	32.4	97.0
61	46.5	23.8	71.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#29
1,1-Dichloropropene
Concen: 14.983 ug/l
RT: 5.509 min Scan# 1312
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 75 Resp: 347711

Ion	Ratio	Lower	Upper
75	100		
110	34.1	17.2	51.5
77	30.5	24.6	37.0





#30
Carbon Tetrachloride
Concen: 14.448 ug/l
RT: 5.509 min Scan# 13
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Tgt Ion:117 Resp: 321059

Ion	Ratio	Lower	Upper
117	100		
119	97.2	76.7	115.1
121	31.5	25.5	38.3

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#31
Isopropyl Ether
Concen: 15.547 ug/l
RT: 3.975 min Scan# 835
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 45 Resp: 703540

Ion	Ratio	Lower	Upper
45	100		
43	50.1	25.7	77.1





#32
Ethyl-t-butyl ether
Concen: 15.784 ug/l
RT: 4.483 min Scan# 99
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 59 Resp: 649553

Ion	Ratio	Lower	Upper
59	100		
87	41.1	32.6	49.0

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#33
Tert-Amyl methyl ether
Concen: 16.013 ug/l
RT: 5.923 min Scan# 1441
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 73 Resp: 575729

Ion	Ratio	Lower	Upper
73	100		
87	23.5	18.9	28.3
71	11.3	9.0	13.4





#34
Propionitrile
Concen: 78.746 ug/l
RT: 4.753 min Scan# 10
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 54 Resp: 103055

Ion	Ratio	Lower	Upper
54	100		
52	19.3	16.3	24.5
55	14.7	11.8	17.6
40	5.1	4.3	6.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#35
Benzene
Concen: 14.641 ug/l
RT: 5.756 min Scan# 1389
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 78 Resp: 1044796

Ion	Ratio	Lower	Upper
78	100		
77	24.2	19.0	28.4





#36
1,2-Dichloroethane
Concen: 14.170 ug/l
RT: 5.775 min Scan# 13
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 62 Resp: 291840

Ion	Ratio	Lower	Upper
62	100		
98	8.7	6.9	10.3



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#37
Trichloroethene
Concen: 14.424 ug/l
RT: 6.528 min Scan# 1629
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:130 Resp: 244774

Ion	Ratio	Lower	Upper
130	100		
95	103.9	83.2	124.8





#38
1,2-Dichloropropane
Concen: 14.653 ug/l
RT: 6.775 min Scan# 17
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 63 Resp: 273676

Ion	Ratio	Lower	Upper
63	100		
65	31.1	25.3	37.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#39
Methacrylonitrile
Concen: 16.908 ug/l
RT: 4.956 min Scan# 1140
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 41 Resp: 79040

Ion	Ratio	Lower	Upper
41	100		
67	80.5	64.5	96.7
39	53.2	43.0	64.6
52	32.7	25.3	37.9





#40
Methyl acrylate
Concen: 15.616 ug/l
RT: 4.830 min Scan# 11
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 55 Resp: 133276

Ion	Ratio	Lower	Upper
55	100		
85	17.1	13.3	19.9
58	8.6	7.3	10.9
42	8.6	6.9	10.3



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#41
Tetrahydrofuran
Concen: 29.065 ug/l
RT: 5.036 min Scan# 1165
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 42 Resp: 79845

Ion	Ratio	Lower	Upper
42	100		
72	51.6	41.5	62.3
71	45.6	37.2	55.8





#42
1-Chlorobutane
Concen: 15.153 ug/l
RT: 5.441 min Scan# 12
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 56 Resp: 481169

Ion	Ratio	Lower	Upper
56	100		
41	51.7	26.3	78.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#43
Dibromomethane
Concen: 14.114 ug/l
RT: 6.904 min Scan# 1746
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 93 Resp: 133485

Ion	Ratio	Lower	Upper
93	100		
95	82.7	67.2	100.8
174	91.6	75.7	113.5





#44
Bromodichloromethane
Concen: 14.680 ug/l
RT: 7.090 min Scan# 18
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 83 Resp: 323143

Ion	Ratio	Lower	Upper
83	100		
85	65.3	51.7	77.5
127	8.1	6.7	10.1



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#45
4-Methyl-2-Pentanone
Concen: 80.129 ug/l
RT: 7.779 min Scan# 2018
Delta R.T. 0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 43 Resp: 794494

Ion	Ratio	Lower	Upper
43	100		
58	42.1	20.8	62.5
85	19.0	14.8	22.2





#46
t-1,4-Dichloro-2-butene
Concen: 29.213 ug/l m
RT: 10.827 min Scan# 29
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 75 Resp: 143327

Ion	Ratio	Lower	Upper
75	100		
53	82.4	64.5	96.7
89	56.3	43.4	65.2
88	41.1	31.2	46.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#47
Methyl methacrylate
Concen: 32.232 ug/l
RT: 6.946 min Scan# 1759
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 69 Resp: 256757

Ion	Ratio	Lower	Upper
69	100		
41	108.6	0.0	217.0
39	59.9	47.7	71.5
100	34.7	27.4	41.0





#48
Ethyl methacrylate
Concen: 17.244 ug/l
RT: 8.322 min Scan# 21
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 69 Resp: 257931

Ion	Ratio	Lower	Upper
69	100		
41	61.3	30.6	92.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#49
Toluene
Concen: 15.549 ug/l
RT: 7.955 min Scan# 2073
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 92 Resp: 638079

Ion	Ratio	Lower	Upper
92	100		
91	177.1	141.8	212.6





#50
t-1,3-Dichloropropene
Concen: 15.891 ug/l
RT: 8.197 min Scan# 21
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 75 Resp: 320246

Ion	Ratio	Lower	Upper
75	100		
77	31.6	25.9	38.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#51
cis-1,3-Dichloropropene
Concen: 15.527 ug/l
RT: 7.595 min Scan# 1961
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 75 Resp: 386542

Ion	Ratio	Lower	Upper
75	100		
77	31.6	25.3	37.9
39	43.0	33.5	50.3



#52
1,1,2-Trichloroethane
Concen: 14.598 ug/l
RT: 8.386 min Scan# 22
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 97 Resp: 186154

Ion	Ratio	Lower	Upper
97	100		
83	89.6	73.0	109.4
85	57.5	46.3	69.5
99	63.2	48.5	72.7



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#53
1,3-Dichloropropane
Concen: 14.594 ug/l
RT: 8.563 min Scan# 2262
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 76 Resp: 330381

Ion	Ratio	Lower	Upper
76	100		
78	32.6	26.3	39.5





#54
2-Hexanone
Concen: 82.409 ug/l
RT: 8.672 min Scan# 22
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 43 Resp: 557608

Ion	Ratio	Lower	Upper
43	100		
58	58.4	38.0	78.0
57	19.8	0.0	39.1



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#55
Dibromochloromethane
Concen: 14.939 ug/l
RT: 8.798 min Scan# 2335
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:129 Resp: 219151

Ion	Ratio	Lower	Upper
129	100		
127	78.2	60.4	90.6





#56
1,2-Dibromoethane
Concen: 14.932 ug/l
RT: 8.910 min Scan# 23
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:107 Resp: 178579

Ion	Ratio	Lower	Upper
107	100		
109	92.5	0.0	187.8

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#57
4-Bromofluorobenzene
Concen: 1.080 ug/l
RT: 10.624 min Scan# 2903
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 95 Resp: 20689

Ion	Ratio	Lower	Upper
95	100		
174	81.3	58.6	88.0
176	77.0	58.2	87.4





#58
Tetrachloroethene
Concen: 13.998 ug/l
RT: 8.541 min Scan# 22
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:164 Resp: 195759

Ion	Ratio	Lower	Upper
164	100		
166	128.0	101.4	152.0
129	97.4	77.0	115.4
131	94.1	76.3	114.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#59
Chlorobenzene
Concen: 15.241 ug/l
RT: 9.434 min Scan# 2533
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:112 Resp: 659996

Ion	Ratio	Lower	Upper
112	100		
114	31.9	25.7	38.5





#60
1,1,1,2-Tetrachloroethane
Concen: 15.005 ug/l
RT: 9.521 min Scan# 25
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:131 Resp: 233573

Ion	Ratio	Lower	Upper
131	100		
133	95.3	76.7	115.1
119	68.6	54.4	81.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#61
Pentachloroethane
Concen: 15.642 ug/l
RT: 11.415 min Scan# 3149
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:117 Resp: 217502

Ion	Ratio	Lower	Upper
117	100		
167	80.9	65.0	97.6





#62
Hexachloroethane
Concen: 16.451 ug/l
RT: 12.463 min Scan# 34
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:117 Resp: 202372

Ion	Ratio	Lower	Upper
117	100		
201	76.5	61.3	91.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#63
Ethyl Benzene
Concen: 16.321 ug/l
RT: 9.560 min Scan# 2572
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 91 Resp: 1218926

Ion	Ratio	Lower	Upper
91	100		
106	29.8	24.2	36.2





#64
m/p-Xylenes
Concen: 32.960 ug/l
RT: 9.682 min Scan# 26
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:106 Resp: 919492

Ion	Ratio	Lower	Upper
106	100		
91	210.6	166.9	250.3

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#65
o-Xylene
Concen: 16.379 ug/l
RT: 10.090 min Scan# 2737
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:106 Resp: 447318

Ion	Ratio	Lower	Upper
106	100		
91	221.7	110.9	332.9





#66
Styrene
Concen: 17.269 ug/l
RT: 10.103 min Scan# 27
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:104 Resp: 750530

Ion	Ratio	Lower	Upper
104	100		
78	51.7	41.2	61.8
103	55.7	44.8	67.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#67
Bromoform
Concen: 15.814 ug/l
RT: 10.280 min Scan# 2796
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:173 Resp: 131670

Ion	Ratio	Lower	Upper
173	100		
175	47.8	39.0	58.4
254	9.7	7.7	11.5





#68
1,2-Dichlorobenzene-d4
Concen: 1.130 ug/l
RT: 12.184 min Scan# 33
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Tgt Ion:152 Resp: 22516

Ion	Ratio	Lower	Upper
152	100		
115	152.9	0.0	275.2
150	390.7	0.0	658.4

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#69
Isopropylbenzene
Concen: 16.638 ug/l
RT: 10.473 min Scan# 2856
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:105 Resp: 1068251

Ion	Ratio	Lower	Upper
105	100		
120	25.7	12.8	38.4





#70
1,1,2,2-Tetrachloroethane
Concen: 15.376 ug/l
RT: 10.769 min Scan# 29
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 83 Resp: 264258

Ion	Ratio	Lower	Upper
83	100		
131	9.2	7.4	11.0
85	64.5	51.8	77.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#71
1,2,3-Trichloropropane
Concen: 15.363 ug/l m
RT: 10.814 min Scan# 2962
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 75 Resp: 200210

Ion	Ratio	Lower	Upper
75	100		
77	47.8	0.0	0.0#
110	31.9	0.0	77.0
97	18.5	0.0	42.2





#72
Bromobenzene
Concen: 15.799 ug/l
RT: 10.772 min Scan# 29
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:156 Resp: 273315

Ion	Ratio	Lower	Upper
156	100		
77	174.0	0.0	343.6
158	96.2	0.0	193.0

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#73
n-propylbenzene
Concen: 17.012 ug/l
RT: 10.894 min Scan# 2987
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:120 Resp: 312734

Ion	Ratio	Lower	Upper
120	100		
91	462.7	369.8	554.6





#74
2-Chlorotoluene
Concen: 16.499 ug/l
RT: 10.975 min Scan# 30
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:126 Resp: 279514

Ion	Ratio	Lower	Upper
126	100		
91	307.9	0.0	623.8

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#75
1,3,5-Trimethylbenzene
Concen: 17.122 ug/l
RT: 11.077 min Scan# 3044
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:105 Resp: 1018594

Ion	Ratio	Lower	Upper
105	100		
120	46.7	37.3	55.9





#76
4-Chlorotoluene
Concen: 16.404 ug/l
RT: 11.087 min Scan# 30
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:126 Resp: 284996

Ion	Ratio	Lower	Upper
126	100		
91	351.8	0.0	703.6



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#77
tert-Butylbenzene
Concen: 16.693 ug/l
RT: 11.409 min Scan# 3147
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:119 Resp: 1004388

Ion	Ratio	Lower	Upper
119	100		
91	59.1	29.4	88.3
134	21.7	17.6	26.4





#78
1,2,4-Trimethylbenzene
Concen: 17.598 ug/l
RT: 11.457 min Scan# 31
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:105 Resp: 1038794

Ion	Ratio	Lower	Upper
105	100		
120	43.4	21.9	65.7

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#79
sec-Butylbenzene
Concen: 17.036 ug/l
RT: 11.630 min Scan# 3216
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:105 Resp: 1305019

Ion	Ratio	Lower	Upper
105	100		
134	18.9	15.1	22.7





#80
Nitrobenzene
Concen: 75.007 ug/l
RT: 13.200 min Scan# 37
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 77 Resp: 39246

Ion	Ratio	Lower	Upper
77	100		
123	42.5	18.9	67.1
65	12.5	11.9	15.1

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#81
p-Isopropyltoluene
Concen: 17.605 ug/l
RT: 11.782 min Scan# 3263
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:119 Resp: 1064335

Ion	Ratio	Lower	Upper
119	100		
134	25.7	20.3	30.5
91	24.1	19.4	29.2





#82
1,3-Dichlorobenzene
Concen: 15.863 ug/l
RT: 11.733 min Scan# 32
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:146 Resp: 532282

Ion	Ratio	Lower	Upper
146	100		
111	41.1	32.8	49.2
148	64.5	51.1	76.7

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#83
1,4-Dichlorobenzene
Concen: 16.378 ug/l
RT: 11.827 min Scan# 3277
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:146 Resp: 537516

Ion	Ratio	Lower	Upper
146	100		
111	39.8	32.1	48.1
148	63.5	50.2	75.4





#84
n-Butylbenzene
Concen: 18.246 ug/l
RT: 12.196 min Scan# 33
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion: 91 Resp: 994690

Ion	Ratio	Lower	Upper
91	100		
92	52.1	41.8	62.8
134	23.4	18.6	28.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#85
1,2-Dichlorobenzene
Concen: 16.323 ug/l
RT: 12.200 min Scan# 3393
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:146 Resp: 526346

Ion	Ratio	Lower	Upper
146	100		
111	43.4	21.9	65.7
148	64.1	32.3	96.9





#86
1,2-Dibromo-3-Chloropropane
Concen: 16.565 ug/l
RT: 12.987 min Scan# 36
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Tgt Ion: 75 Resp: 39960

Ion	Ratio	Lower	Upper
75	100		
155	81.3	63.5	95.3
157	103.4	81.8	122.6

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#87
1,2,4-Trichlorobenzene
Concen: 18.320 ug/l
RT: 13.830 min Scan# 3900
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:180 Resp: 288112

Ion	Ratio	Lower	Upper
180	100		
182	96.1	76.6	115.0
145	30.9	25.4	38.2





#88
Hexachlorobutadiene
Concen: 15.648 ug/l
RT: 14.010 min Scan# 39
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:225 Resp: 175875

Ion	Ratio	Lower	Upper
225	100		
223	62.6	49.5	74.3
227	64.4	51.0	76.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#89
Naphthalene
Concen: 15.328 ug/l
RT: 14.077 min Scan# 3977
Delta R.T. -0.000 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:128 Resp: 515515

Ion	Ratio	Lower	Upper
128	100		
127	13.2	10.6	16.0
129	10.5	8.6	13.0





#90
1,2,3-Trichlorobenzene
Concen: 18.007 ug/l
RT: 14.318 min Scan# 40
Delta R.T. -0.003 min
Lab File: VU063224.D
Acq: 10 Feb 2025 15:33

Tgt Ion:180 Resp: 277101

Ion	Ratio	Lower	Upper
180	100		
182	95.6	78.2	117.2
145	32.0	26.1	39.1



Instrument :
MSVOA_U
ClientSampleId :
VSTDICC015

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quantitation Report (QT Reviewed)

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063225.D
Acq On : 10 Feb 2025 16:45
Operator : MD/SY
Sample : VSTDICV010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 04:20:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 04:17:17 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Fluorobenzene	6.103	96	54746	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.624	95	23243	1.287	ug/l	0.00
Spiked Amount 1.000			Recovery = 129.000%			
68) 1,2-Dichlorobenzene-d4	12.184	152	20943	1.115	ug/l	0.00
Spiked Amount 1.000			Recovery = 112.000%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.377	85	122859	6.907	ug/l	99
3) Chloromethane	1.515	50	158995	7.762	ug/l	99
4) Vinyl Chloride	1.599	62	173502	8.561	ug/l	99
5) Bromomethane	1.821	94	55063	5.874	ug/l	99
6) Chloroethane	1.898	64	111524	8.736	ug/l	99
7) Trichlorofluoromethane	2.113	101	220887	9.197	ug/l	100
8) 1,1,2-Trichloro-1,2,2-...	2.560	101	142270	10.436	ug/l	99
9) 1,1-Dichloroethene	2.560	96	143602	10.339	ug/l	99
10) Iodomethane	2.702	142	204632	9.372	ug/l	100
11) Allyl Chloride	2.904	41	219588	11.005	ug/l	97
12) Acrylonitrile	3.300	53	166008	51.821	ug/l	98
13) Acetone	2.615	43	116503	47.295	ug/l	99
14) Carbon Disulfide	2.772	76	421701	8.683	ug/l	100
15) Methylene Chloride	3.026	84	165284	9.632	ug/l	99
16) trans-1,2-Dichloroethene	3.335	96	154618	9.754	ug/l	97
17) 1,1-Dichloroethane	3.849	63	301514	10.091	ug/l	100
18) 2-Butanone	4.689	43	197169	50.293	ug/l	98
19) Cyclohexane	5.370	56	248051m	10.331	ug/l	
20) Methylcyclohexane	6.750	83	263935	11.086	ug/l	100
21) 2,2-Dichloropropane	4.644	77	247364	10.612	ug/l	100
22) cis-1,2-Dichloroethene	4.647	96	185608	10.837	ug/l	99
23) Diethyl Ether	2.364	59	112591	9.447	ug/l	99
24) tert-Butyl Alcohol	3.187	59	66161	51.182	ug/l #	88
25) Methyl tert-Butyl Ether	3.348	73	369805	10.661	ug/l	99
26) Bromochloromethane	4.956	128	68843	9.196	ug/l	99
27) Chloroform	5.071	83	304221	10.089	ug/l	100
28) 1,1,1-Trichloroethane	5.300	97	250902	10.272	ug/l	99
29) 1,1-Dichloropropene	5.512	75	208602	9.535	ug/l	99
30) Carbon Tetrachloride	5.509	117	210911	10.069	ug/l	98
31) Isopropyl Ether	3.975	45	498428	11.684	ug/l #	1
34) Propionitrile	4.756	54	132730	112.251	ug/l #	98
35) Benzene	5.759	78	678147	10.081	ug/l	100
36) 1,2-Dichloroethane	5.779	62	180895	9.317	ug/l	100
37) Trichloroethene	6.531	130	151261	9.455	ug/l	99
38) 1,2-Dichloropropane	6.775	63	170018	9.657	ug/l	100
39) Methacrylonitrile	4.959	41	50755	11.518	ug/l	98
40) Methyl acrylate	4.833	55	86382	10.636	ug/l	98
41) Tetrahydrofuran	5.039	42	133021	51.903	ug/l	96
42) 1-Chlorobutane	5.441	56	317470	10.606	ug/l	99
43) Dibromomethane	6.907	93	87369	9.800	ug/l	97
44) Bromodichloromethane	7.094	83	225336	10.859	ug/l	98
45) 4-Methyl-2-Pentanone	7.779	43	519812	55.614	ug/l	100
46) t-1,4-Dichloro-2-butene	10.823	75	53637m	12.207	ug/l	

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063225.D
Acq On : 10 Feb 2025 16:45
Operator : MD/SY
Sample : VSTDICV010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025

Quant Time: Feb 11 04:20:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 04:17:17 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units		Dev(Min)
47) Methyl methacrylate	6.952	69	81105	10.801	ug/l		96
48) Ethyl methacrylate	8.325	69	166665	11.820	ug/l		99
49) Toluene	7.959	92	415448	10.739	ug/l		100
50) t-1,3-Dichloropropene	8.200	75	207350	10.914	ug/l		98
51) cis-1,3-Dichloropropene	7.595	75	242070	10.315	ug/l		97
52) 1,1,2-Trichloroethane	8.390	97	119029	9.902	ug/l		99
53) 1,3-Dichloropropane	8.563	76	207689	9.732	ug/l		99
54) 2-Hexanone	8.676	43	369589	57.943	ug/l		99
55) Dibromochloromethane	8.798	129	144391	10.441	ug/l		98
56) 1,2-Dibromoethane	8.914	107	111510	9.891	ug/l		99
58) Tetrachloroethene	8.544	164	135201	10.255	ug/l		98
59) Chlorobenzene	9.438	112	435888	10.678	ug/l		99
60) 1,1,1,2-Tetrachloroethane	9.525	131	145439	9.912	ug/l		99
61) Pentachloroethane	11.418	117	146273	11.159	ug/l		99
62) Hexachloroethane	12.467	117	135009	11.642	ug/l		99
63) Ethyl Benzene	9.560	91	794797	11.289	ug/l		100
64) m/p-Xylenes	9.682	106	610698	23.222	ug/l		99
65) o-Xylene	10.090	106	294864	11.453	ug/l		100
66) Styrene	10.103	104	499423	12.190	ug/l		100
67) Bromoform	10.280	173	87483	11.146	ug/l		99
69) Isopropylbenzene	10.473	105	777638	12.848	ug/l		100
70) 1,1,2,2-Tetrachloroethane	10.772	83	169972	10.491	ug/l		99
71) 1,2,3-Trichloropropane	10.814	75	120161m	9.872	ug/l		
72) Bromobenzene	10.772	156	183865	11.275	ug/l		99
73) n-propylbenzene	10.897	120	210467	12.145	ug/l		99
74) 2-Chlorotoluene	10.978	126	191413	11.986	ug/l		96
75) 1,3,5-Trimethylbenzene	11.078	105	684766	12.210	ug/l		100
76) 4-Chlorotoluene	11.087	126	194674	11.887	ug/l		100
77) tert-Butylbenzene	11.409	119	683840	12.057	ug/l		100
78) 1,2,4-Trimethylbenzene	11.457	105	696739	12.521	ug/l		100
79) sec-Butylbenzene	11.634	105	893150	12.368	ug/l		100
80) Nitrobenzene	13.212	77	5344	14.120	ug/l	#	81
81) p-Isopropyltoluene	11.785	119	740341	12.990	ug/l		100
82) 1,3-Dichlorobenzene	11.737	146	376778	11.912	ug/l		99
83) 1,4-Dichlorobenzene	11.827	146	376860	12.181	ug/l		99
84) n-Butylbenzene	12.200	91	707039	13.835	ug/l		99
85) 1,2-Dichlorobenzene	12.203	146	361247	11.884	ug/l		99
86) 1,2-Dibromo-3-Chloropr...	12.987	75	25391	11.165	ug/l		96
87) 1,2,4-Trichlorobenzene	13.830	180	207953	14.027	ug/l		99
88) Hexachlorobutadiene	14.010	225	123242	11.632	ug/l		99
89) Naphthalene	14.077	128	342725	10.999	ug/l		99
90) 1,2,3-Trichlorobenzene	14.322	180	175916	12.127	ug/l		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063225.D
Acq On : 10 Feb 2025 16:45
Operator : MD/SY
Sample : VSTDICV010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Feb 11 04:20:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 04:17:17 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.103 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 96 Resp: 5474

Ion	Ratio	Lower	Upper
96	100		
70	21.0	15.6	23.4
95	9.1	7.3	10.9
97	0.0	0.0	0.0



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#2
Dichlorodifluoromethane
Concen: 6.907 ug/l
RT: 1.377 min Scan# 27
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 85 Resp: 122859

Ion	Ratio	Lower	Upper
85	100		
87	32.8	16.0	48.0





#3
Chloromethane
Concen: 7.762 ug/l
RT: 1.515 min Scan# 7(
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 50 Resp: 15899

Ion	Ratio	Lower	Upper
50	100		
52	33.0	25.8	38.8



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#4
Vinyl Chloride
Concen: 8.561 ug/l
RT: 1.599 min Scan# 96
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 62 Resp: 173502

Ion	Ratio	Lower	Upper
62	100		
64	31.2	25.4	38.0





#5
Bromomethane
Concen: 5.874 ug/l
RT: 1.821 min Scan# 1
Delta R.T. -0.022 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 94 Resp: 5506

Ion	Ratio	Lower	Upper
94	100		
96	93.9	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#6
Chloroethane
Concen: 8.736 ug/l
RT: 1.898 min Scan# 189
Delta R.T. -0.026 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 64 Resp: 111524

Ion	Ratio	Lower	Upper
64	100		
66	32.0	25.8	38.8





#7
Trichlorofluoromethane
Concen: 9.197 ug/l
RT: 2.113 min Scan# 2
Delta R.T. -0.016 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:101 Resp: 22088

Ion	Ratio	Lower	Upper
101	100		
103	64.1	51.4	77.2

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 10.436 ug/l
RT: 2.560 min Scan# 395
Delta R.T. -0.010 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:101 Resp: 142270

Ion	Ratio	Lower	Upper
101	100		
85	44.3	35.4	53.0
151	71.9	58.5	87.7





#9
1,1-Dichloroethene
Concen: 10.339 ug/l
RT: 2.560 min Scan# 3
Delta R.T. -0.006 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 96 Resp: 14360

Ion	Ratio	Lower	Upper
96	100		
61	164.9	0.0	492.9
98	63.7	0.0	124.0

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#10
Iodomethane
Concen: 9.372 ug/l
RT: 2.702 min Scan# 439
Delta R.T. -0.006 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:142 Resp: 204632

Ion	Ratio	Lower	Upper
142	100		
127	43.0	34.5	51.7





#11
Allyl Chloride
Concen: 11.005 ug/l
RT: 2.904 min Scan# 5
Delta R.T. -0.006 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 41 Resp: 21958

Ion	Ratio	Lower	Upper
41	100		
39	69.8	57.9	86.9



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#12
Acrylonitrile
Concen: 51.821 ug/l
RT: 3.300 min Scan# 625
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 53 Resp: 166008

Ion	Ratio	Lower	Upper
53	100		
52	82.1	64.2	96.2
51	36.7	30.8	46.2





#13
Acetone
Concen: 47.295 ug/l
RT: 2.615 min Scan# 4
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 43 Resp: 11650

Ion	Ratio	Lower	Upper
43	100		
58	35.0	27.4	41.0



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#14
Carbon Disulfide
Concen: 8.683 ug/l
RT: 2.772 min Scan# 461
Delta R.T. -0.010 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 76 Resp: 421701

Ion	Ratio	Lower	Upper
76	100		
78	9.2	7.2	10.8





#15
Methylene Chloride
Concen: 9.632 ug/l
RT: 3.026 min Scan# 54
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 84 Resp: 16528

Ion	Ratio	Lower	Upper
84	100		
49	119.4	96.4	144.6
51	36.8	0.0	74.8
86	62.6	51.3	76.9



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#16
trans-1,2-Dichloroethene
Concen: 9.754 ug/l
RT: 3.335 min Scan# 636
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 96 Resp: 154618

Ion	Ratio	Lower	Upper
96	100		
61	138.0	113.4	170.2
98	62.9	51.2	76.8





#17
1,1-Dichloroethane
Concen: 10.091 ug/l
RT: 3.849 min Scan# 7
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 63 Resp: 30151

Ion	Ratio	Lower	Upper
63	100		
98	6.6	3.3	9.9
100	4.3	2.1	6.2



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#18
2-Butanone
Concen: 50.293 ug/l
RT: 4.689 min Scan# 1057
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 43 Resp: 197169

Ion	Ratio	Lower	Upper
43	100		
57	9.1	0.0	17.0





#19
Cyclohexane
Concen: 10.331 ug/l m
RT: 5.370 min Scan# 1
Delta R.T. -0.006 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 56 Resp: 24805

Ion	Ratio	Lower	Upper
56	100		
69	31.4	24.5	36.7
84	82.3	65.2	97.8

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#20
Methylcyclohexane
Concen: 11.086 ug/l
RT: 6.750 min Scan# 1698
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 83 Resp: 263935

Ion	Ratio	Lower	Upper
83	100		
55	78.4	63.1	94.7
98	44.0	35.2	52.8





#21
2,2-Dichloropropane
Concen: 10.612 ug/l
RT: 4.644 min Scan# 1
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 77 Resp: 24736

Ion	Ratio	Lower	Upper
77	100		
97	23.0	18.5	27.7



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#22
cis-1,2-Dichloroethene
Concen: 10.837 ug/l
RT: 4.647 min Scan# 1044
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 96 Resp: 185608

Ion	Ratio	Lower	Upper
96	100		
61	150.7	0.0	373.3
98	64.8	31.9	95.9





#23
Diethyl Ether
Concen: 9.447 ug/l
RT: 2.364 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 59 Resp: 11259

Ion	Ratio	Lower	Upper
59	100		
45	70.8	57.8	86.6
74	71.3	57.7	86.5



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#24
tert-Butyl Alcohol
Concen: 51.182 ug/l
RT: 3.187 min Scan# 590
Delta R.T. 0.029 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 59 Resp: 66161

Ion	Ratio	Lower	Upper
59	100		
57	5.2	7.5	11.3#





#25
Methyl tert-Butyl Ether
Concen: 10.661 ug/l
RT: 3.348 min Scan# 64
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 73 Resp: 36980

Ion	Ratio	Lower	Upper
73	100		
43	18.8	14.9	22.3

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#26
Bromochloromethane
Concen: 9.196 ug/l
RT: 4.956 min Scan# 1140
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:128 Resp: 68843

Ion	Ratio	Lower	Upper
128	100		
49	173.8	0.0	343.4
130	129.8	102.9	154.3





#27
Chloroform
Concen: 10.089 ug/l
RT: 5.071 min Scan# 11
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 83 Resp: 30422

Ion	Ratio	Lower	Upper
83	100		
85	64.5	0.0	129.8



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#28
1,1,1-Trichloroethane
Concen: 10.272 ug/l
RT: 5.300 min Scan# 1247
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 97 Resp: 250902

Ion	Ratio	Lower	Upper
97	100		
99	64.5	32.4	97.0
61	46.9	23.8	71.2





#29
1,1-Dichloropropene
Concen: 9.535 ug/l
RT: 5.512 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 75 Resp: 20860

Ion	Ratio	Lower	Upper
75	100		
110	34.6	17.2	51.5
77	31.6	24.6	37.0



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#30
Carbon Tetrachloride
Concen: 10.069 ug/l
RT: 5.509 min Scan# 1312
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:117 Resp: 210911

Ion	Ratio	Lower	Upper
117	100		
119	94.4	76.7	115.1
121	30.2	25.5	38.3





#31
Isopropyl Ether
Concen: 11.684 ug/l
RT: 3.975 min Scan# 8
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 45 Resp: 49842

Ion	Ratio	Lower	Upper
45	100		
43	260.1	25.7	77.1#



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#34
Propionitrile
Concen: 112.251 ug/l
RT: 4.756 min Scan# 1078
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 54 Resp: 132730

Ion	Ratio	Lower	Upper
54	100		
52	19.8	16.3	24.5
55	16.1	11.8	17.6
40	6.5	4.3	6.5#





#35
Benzene
Concen: 10.081 ug/l
RT: 5.759 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 78 Resp: 67814

Ion	Ratio	Lower	Upper
78	100		
77	23.9	19.0	28.4

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#36
1,2-Dichloroethane
Concen: 9.317 ug/l
RT: 5.779 min Scan# 1396
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 62 Resp: 180895

Ion	Ratio	Lower	Upper
62	100		
98	8.5	6.9	10.3





#37
Trichloroethene
Concen: 9.455 ug/l
RT: 6.531 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:130 Resp: 15126

Ion	Ratio	Lower	Upper
130	100		
95	103.3	83.2	124.8



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#38
1,2-Dichloropropane
Concen: 9.657 ug/l
RT: 6.775 min Scan# 1706
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 63 Resp: 170018

Ion	Ratio	Lower	Upper
63	100		
65	31.4	25.3	37.9





#39
Methacrylonitrile
Concen: 11.518 ug/l
RT: 4.959 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 41 Resp: 5075

Ion	Ratio	Lower	Upper
41	100		
67	83.5	64.5	96.7
39	52.5	43.0	64.6
52	32.3	25.3	37.9



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#40
Methyl acrylate
Concen: 10.636 ug/l
RT: 4.833 min Scan# 1102
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 55 Resp: 86382

Ion	Ratio	Lower	Upper
55	100		
85	16.1	13.3	19.9
58	8.2	7.3	10.9
42	8.1	6.9	10.3





#41
Tetrahydrofuran
Concen: 51.903 ug/l
RT: 5.039 min Scan# 11
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 42 Resp: 13302

Ion	Ratio	Lower	Upper
42	100		
72	47.8	41.5	62.3
71	44.8	37.2	55.8



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#42
1-Chlorobutane
Concen: 10.606 ug/l
RT: 5.441 min Scan# 1291
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 56 Resp: 317470

Ion	Ratio	Lower	Upper
56	100		
41	52.1	26.3	78.8





#43
Dibromomethane
Concen: 9.800 ug/l
RT: 6.907 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 93 Resp: 8736

Ion	Ratio	Lower	Upper
93	100		
95	81.9	67.2	100.8
174	91.4	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#44
Bromodichloromethane
Concen: 10.859 ug/l
RT: 7.094 min Scan# 1805
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 83 Resp: 225336

Ion	Ratio	Lower	Upper
83	100		
85	62.9	51.7	77.5
127	8.0	6.7	10.1





#45
4-Methyl-2-Pentanone
Concen: 55.614 ug/l
RT: 7.779 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 43 Resp: 51981

Ion	Ratio	Lower	Upper
43	100		
58	41.8	20.8	62.5
85	18.8	14.8	22.2



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#46
t-1,4-Dichloro-2-butene
Concen: 12.207 ug/l m
RT: 10.823 min Scan# 2965
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 75 Resp: 53637

Ion	Ratio	Lower	Upper
75	100		
53	63.5	64.5	96.7#
89	41.6	43.4	65.2#
88	31.5	31.2	46.8





#47
Methyl methacrylate
Concen: 10.801 ug/l
RT: 6.952 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 69 Resp: 8110

Ion	Ratio	Lower	Upper
69	100		
41	104.9	0.0	217.0
39	57.5	47.7	71.5
100	31.1	27.4	41.0



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#48
Ethyl methacrylate
Concen: 11.820 ug/l
RT: 8.325 min Scan# 2188
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 69 Resp: 166665

Ion	Ratio	Lower	Upper
69	100		
41	62.4	30.6	92.0





#49
Toluene
Concen: 10.739 ug/l
RT: 7.959 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 92 Resp: 41544

Ion	Ratio	Lower	Upper
92	100		
91	176.7	141.8	212.6

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#50
t-1,3-Dichloropropene
Concen: 10.914 ug/l
RT: 8.200 min Scan# 2149
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 75 Resp: 207350

Ion	Ratio	Lower	Upper
75	100		
77	31.5	25.9	38.9





#51
cis-1,3-Dichloropropene
Concen: 10.315 ug/l
RT: 7.595 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 75 Resp: 24207

Ion	Ratio	Lower	Upper
75	100		
77	32.2	25.3	37.9
39	44.2	33.5	50.3



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#52
1,1,2-Trichloroethane
Concen: 9.902 ug/l
RT: 8.390 min Scan# 2208
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 97 Resp: 119029

Ion	Ratio	Lower	Upper
97	100		
83	90.8	73.0	109.4
85	57.6	46.3	69.5
99	62.6	48.5	72.7





#53
1,3-Dichloropropane
Concen: 9.732 ug/l
RT: 8.563 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 76 Resp: 20768

Ion	Ratio	Lower	Upper
76	100		
78	32.4	26.3	39.5



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#54
2-Hexanone
Concen: 57.943 ug/l
RT: 8.676 min Scan# 2297
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 43 Resp: 369589

Ion	Ratio	Lower	Upper
43	100		
58	57.8	38.0	78.0
57	20.2	0.0	39.1





#55
Dibromochloromethane
Concen: 10.441 ug/l
RT: 8.798 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:129 Resp: 14439

Ion	Ratio	Lower	Upper
129	100		
127	77.3	60.4	90.6

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#56
1,2-Dibromoethane
Concen: 9.891 ug/l
RT: 8.914 min Scan# 2371
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:107 Resp: 111510

Ion	Ratio	Lower	Upper
107	100		
109	93.0	0.0	187.8





#57
4-Bromofluorobenzene
Concen: 1.287 ug/l
RT: 10.624 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 95 Resp: 2324

Ion	Ratio	Lower	Upper
95	100		
174	75.2	58.6	88.0
176	74.4	58.2	87.4



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#58
Tetrachloroethene
Concen: 10.255 ug/l
RT: 8.544 min Scan# 2256
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:164 Resp: 135201

Ion	Ratio	Lower	Upper
164	100		
166	129.1	101.4	152.0
129	99.4	77.0	115.4
131	95.8	76.3	114.5





#59
Chlorobenzene
Concen: 10.678 ug/l
RT: 9.438 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:112 Resp: 43588

Ion	Ratio	Lower	Upper
112	100		
114	31.6	25.7	38.5



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#60
1,1,1,2-Tetrachloroethane
Concen: 9.912 ug/l
RT: 9.525 min Scan# 2561
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:131 Resp: 145439

Ion	Ratio	Lower	Upper
131	100		
133	96.9	76.7	115.1
119	69.2	54.4	81.6





#61
Pentachloroethane
Concen: 11.159 ug/l
RT: 11.418 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:117 Resp: 14627

Ion	Ratio	Lower	Upper
117	100		
167	81.8	65.0	97.6



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#62
Hexachloroethane
Concen: 11.642 ug/l
RT: 12.467 min Scan# 3476
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:117 Resp: 135009

Ion	Ratio	Lower	Upper
117	100		
201	75.6	61.3	91.9





#63
Ethyl Benzene
Concen: 11.289 ug/l
RT: 9.560 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 91 Resp: 79479

Ion	Ratio	Lower	Upper
91	100		
106	30.3	24.2	36.2

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#64
m/p-Xylenes
Concen: 23.222 ug/l
RT: 9.682 min Scan# 2610
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:106 Resp: 610698

Ion	Ratio	Lower	Upper
106	100		
91	210.4	166.9	250.3





#65
o-Xylene
Concen: 11.453 ug/l
RT: 10.090 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:106 Resp: 29486

Ion	Ratio	Lower	Upper
106	100		
91	221.8	110.9	332.9

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#66
Styrene
Concen: 12.190 ug/l
RT: 10.103 min Scan# 2741
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:104 Resp: 499423

Ion	Ratio	Lower	Upper
104	100		
78	51.3	41.2	61.8
103	55.8	44.8	67.2





#67
Bromoform
Concen: 11.146 ug/l
RT: 10.280 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:173 Resp: 8748

Ion	Ratio	Lower	Upper
173	100		
175	48.2	39.0	58.4
254	9.7	7.7	11.5



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#68
1,2-Dichlorobenzene-d4
Concen: 1.115 ug/l
RT: 12.184 min Scan# 3388
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:152 Resp: 20943

Ion	Ratio	Lower	Upper
152	100		
115	139.6	0.0	275.2
150	331.3	0.0	658.4





#69
Isopropylbenzene
Concen: 12.848 ug/l
RT: 10.473 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:105 Resp: 77763

Ion	Ratio	Lower	Upper
105	100		
120	25.6	12.8	38.4

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#70
1,1,2,2-Tetrachloroethane
Concen: 10.491 ug/l
RT: 10.772 min Scan# 2949
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 83 Resp: 169972

Ion	Ratio	Lower	Upper
83	100		
131	8.9	7.4	11.0
85	64.0	51.8	77.8





#71
1,2,3-Trichloropropane
Concen: 9.872 ug/l m
RT: 10.814 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 75 Resp: 12016

Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0
110	34.7	0.0	77.0
97	18.5	0.0	42.2



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#72
Bromobenzene
Concen: 11.275 ug/l
RT: 10.772 min Scan# 2949
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:156 Resp: 183865

Ion	Ratio	Lower	Upper
156	100		
77	171.5	0.0	343.6
158	97.6	0.0	193.0





#73
n-propylbenzene
Concen: 12.145 ug/l
RT: 10.897 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:120 Resp: 21046

Ion	Ratio	Lower	Upper
120	100		
91	460.3	369.8	554.6



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#74
2-Chlorotoluene
Concen: 11.986 ug/l
RT: 10.978 min Scan# 3013
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:126 Resp: 191413

Ion	Ratio	Lower	Upper
126	100		
91	304.4	0.0	623.8





#75
1,3,5-Trimethylbenzene
Concen: 12.210 ug/l
RT: 11.078 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:105 Resp: 68476

Ion	Ratio	Lower	Upper
105	100		
120	46.6	37.3	55.9

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#76
4-Chlorotoluene
Concen: 11.887 ug/l
RT: 11.087 min Scan# 3047
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:126 Resp: 194674

Ion	Ratio	Lower	Upper
126	100		
91	351.2	0.0	703.6





#77
tert-Butylbenzene
Concen: 12.057 ug/l
RT: 11.409 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:119 Resp: 68384

Ion	Ratio	Lower	Upper
119	100		
91	59.0	29.4	88.3
134	21.5	17.6	26.4



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#78
1,2,4-Trimethylbenzene
Concen: 12.521 ug/l
RT: 11.457 min Scan# 3162
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:105 Resp: 696739

Ion	Ratio	Lower	Upper
105	100		
120	43.8	21.9	65.7





#79
sec-Butylbenzene
Concen: 12.368 ug/l
RT: 11.634 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:105 Resp: 89315

Ion	Ratio	Lower	Upper
105	100		
134	18.9	15.1	22.7

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025





#80
Nitrobenzene
Concen: 14.120 ug/l
RT: 13.212 min Scan# 3708
Delta R.T. 0.010 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 77 Resp: 5344

Ion	Ratio	Lower	Upper
77	100		
123	52.1	18.9	67.1
65	0.0	11.9	15.1#





#81
p-Isopropyltoluene
Concen: 12.990 ug/l
RT: 11.785 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:119 Resp: 74034

Ion	Ratio	Lower	Upper
119	100		
134	25.4	20.3	30.5
91	24.1	19.4	29.2



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#82
1,3-Dichlorobenzene
Concen: 11.912 ug/l
RT: 11.737 min Scan# 3249
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:146 Resp: 376778

Ion	Ratio	Lower	Upper
146	100		
111	41.2	32.8	49.2
148	63.4	51.1	76.7





#83
1,4-Dichlorobenzene
Concen: 12.181 ug/l
RT: 11.827 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:146 Resp: 37686

Ion	Ratio	Lower	Upper
146	100		
111	39.7	32.1	48.1
148	63.5	50.2	75.4



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#84
n-Butylbenzene
Concen: 13.835 ug/l
RT: 12.200 min Scan# 3393
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 91 Resp: 707039

Ion	Ratio	Lower	Upper
91	100		
92	51.6	41.8	62.8
134	23.4	18.6	28.0





#85
1,2-Dichlorobenzene
Concen: 11.884 ug/l
RT: 12.203 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:146 Resp: 36124

Ion	Ratio	Lower	Upper
146	100		
111	43.2	21.9	65.7
148	63.7	32.3	96.9



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#86
1,2-Dibromo-3-Chloropropane
Concen: 11.165 ug/l
RT: 12.987 min Scan# 3638
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion: 75 Resp: 25391

Ion	Ratio	Lower	Upper
75	100		
155	76.7	63.5	95.3
157	97.7	81.8	122.6





#87
1,2,4-Trichlorobenzene
Concen: 14.027 ug/l
RT: 13.830 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:180 Resp: 20795

Ion	Ratio	Lower	Upper
180	100		
182	95.2	76.6	115.0
145	30.4	25.4	38.2



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#88
Hexachlorobutadiene
Concen: 11.632 ug/l
RT: 14.010 min Scan# 3956
Delta R.T. -0.003 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:225 Resp: 123242

Ion	Ratio	Lower	Upper
225	100		
223	62.4	49.5	74.3
227	64.1	51.0	76.4





#89
Naphthalene
Concen: 10.999 ug/l
RT: 14.077 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:128 Resp: 34272

Ion	Ratio	Lower	Upper
128	100		
127	12.9	10.6	16.0
129	10.5	8.6	13.0



Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Manual Integrations
APPROVED

Reviewed By :Amit Patel 02/12/2025
Supervised By :Mahesh Dadoda 02/12/2025



#90
1,2,3-Trichlorobenzene
Concen: 12.127 ug/l
RT: 14.322 min Scan# 4053
Delta R.T. 0.000 min
Lab File: VU063225.D
Acq: 10 Feb 2025 16:45

Tgt Ion:180 Resp: 175916

Ion	Ratio	Lower	Upper
180	100		
182	96.8	78.2	117.2
145	32.9	26.1	39.1



Evaluate Continuing Calibration Report

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063225.D
Acq On : 10 Feb 2025 16:45
Operator : MD/SY
Sample : VSTDICV010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Quant Time: Feb 11 04:20:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 04:17:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 200%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	i	Fluorobenzene	1.000	1.000	0.0	95	0.00
2	T	Dichlorodifluoromethane	0.325	0.224	31.1#	70	0.00
3	t	Chloromethane	0.374	0.290	22.5	78	0.00
4	Rt	Vinyl Chloride	0.370	0.317	14.3	82	0.00
5	T	Bromomethane	0.181	0.101	44.2#	50	-0.02
6	T	Chloroethane	0.233	0.204	12.4	87	-0.03
7	T	Trichlorofluoromethane	0.439	0.403	8.2	90	-0.02
8		1,1,2-Trichloro-1,2,2-trifl	0.249	0.260	-4.4	107	0.00
9	Rt	1,1-Dichloroethene	0.254	0.262	-3.1	100	0.00
10	t	Iodomethane	0.399	0.374	6.3	87	0.00
11	t	Allyl Chloride	0.364	0.401	-10.2	104	0.00
12	t	Acrylonitrile	0.059	0.152	-157.6#	221#	0.00
13	T	Acetone	0.045	0.043	4.4	85	0.00
14	T	Carbon Disulfide	0.887	0.770	13.2	85	0.00
15	RT	Methylene Chloride	0.313	0.302	3.5	94	0.00
16	RT	trans-1,2-Dichloroethene	0.290	0.282	2.8	93	0.00
17	t	1,1-Dichloroethane	0.546	0.551	-0.9	97	0.00
18	T	2-Butanone	0.072	0.072	0.0	87	0.00
19		Cyclohexane	0.439	0.453	-3.2	95	0.00
20		Methylcyclohexane	0.435	0.482	-10.8	102	0.00
21	T	2,2-Dichloropropane	0.426	0.452	-6.1	102	0.00
22	RT	cis-1,2-Dichloroethene	0.313	0.339	-8.3	102	0.00
23	t	Diethyl Ether	0.218	0.206	5.5	90	0.00
24	t	tert-Butyl Alcohol	0.026	0.012	53.8#	48	0.03
25	t	Methyl tert-Butyl Ether	0.634	0.675	-6.5	95	0.00
26	t	Bromochloromethane	0.137	0.126	8.0	87	0.00
27	t	Chloroform	0.551	0.556	-0.9	96	0.00
28	RT	1,1,1-Trichloroethane	0.446	0.458	-2.7	98	0.00
29	T	1,1-Dichloropropene	0.400	0.381	4.8	89	0.00
30	RT	Carbon Tetrachloride	0.383	0.385	-0.5	97	0.00
31	t	Isopropyl Ether	0.779	0.910	-16.8	106	0.00
32		Ethyl-t-butyl ether	0.709	0.000	100.0#	0#	-4.48#
33		Tert-Amyl methyl ether	0.619	0.000	100.0#	0#	-5.93#
34	t	Propionitrile	0.022	0.048	-118.2#	185	0.00
35	RT	Benzene	1.229	1.239	-0.8	95	0.00
36	RT	1,2-Dichloroethane	0.355	0.330	7.0	88	0.00
37	RT	Trichloroethene	0.292	0.276	5.5	90	0.00
38	Rt	1,2-Dichloropropane	0.322	0.311	3.4	90	0.00
39	t	Methacrylonitrile	0.080	0.093	-16.2	90	0.00
40	t	Methyl acrylate	0.148	0.158	-6.8	90	0.00
41	t	Tetrahydrofuran	0.047	0.121	-157.4#	224#	0.00
42	t	1-Chlorobutane	0.547	0.580	-6.0	97	0.00
43	T	Dibromomethane	0.163	0.160	1.8	93	0.00
44	T	Bromodichloromethane	0.379	0.412	-8.7	100	0.00
45	T	4-Methyl-2-Pentanone	0.171	0.190	-11.1	91	0.00
46	t	t-1,4-Dichloro-2-butene	0.080	0.049	38.8#	55	0.00
47	t	Methyl methacrylate	0.137	0.074	46.0#	44	0.00
48	t	Ethyl methacrylate	0.258	0.304	-17.8	92	0.00
49	Rt	Toluene	0.707	0.759	-7.4	97	0.00

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063225.D
Acq On : 10 Feb 2025 16:45
Operator : MD/SY
Sample : VSTDICV010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Quant Time: Feb 11 04:20:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 04:17:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 200%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50	T	t-1,3-Dichloropropene	0.347	0.379	-9.2	94	0.00
51	T	cis-1,3-Dichloropropene	0.429	0.442	-3.0	91	0.00
52	RT	1,1,2-Trichloroethane	0.220	0.217	1.4	90	0.00
53	t	1,3-Dichloropropane	0.390	0.379	2.8	89	0.00
54	t	2-Hexanone	0.117	0.135	-15.4	94	0.00
55	t	Dibromochloromethane	0.253	0.264	-4.3	93	0.00
56	T	1,2-Dibromoethane	0.206	0.204	1.0	88	0.00
57	S	4-Bromofluorobenzene	0.330	0.425	-28.8	112	0.00
58	RT	Tetrachloroethene	0.241	0.247	-2.5	100	0.00
59	Rt	Chlorobenzene	0.746	0.796	-6.7	98	0.00
60	T	1,1,1,2-Tetrachloroethane	0.268	0.266	0.7	92	0.00
61	t	Pentachloroethane	0.239	0.267	-11.7	105	0.00
62	t	Hexachloroethane	0.212	0.247	-16.5	105	0.00
63	Rt	Ethyl Benzene	1.286	1.452	-12.9	98	0.00
64	RT	m/p-Xylenes	0.480	0.558	-16.3	99	0.00
65	RT	o-Xylene	0.470	0.539	-14.7	99	0.00
66	RT	Styrene	0.748	0.912	-21.9	101	0.00
67	t	Bromoform	0.143	0.160	-11.9	97	0.00
68	S	1,2-Dichlorobenzene-d4	0.343	0.383	-11.7	105	0.00
69	T	Isopropylbenzene	1.106	1.420	-28.4	111	0.00
70	T	1,1,2,2-Tetrachloroethane	0.296	0.310	-4.7	95	0.00
71	T	1,2,3-Trichloropropane	0.222	0.219	1.4	102	0.00
72	t	Bromobenzene	0.298	0.336	-12.8	101	0.00
73	t	n-propylbenzene	0.317	0.384	-21.1	103	0.00
74	t	2-Chlorotoluene	0.292	0.350	-19.9	105	0.00
75	t	1,3,5-Trimethylbenzene	1.024	1.251	-22.2	103	0.00
76	t	4-Chlorotoluene	0.299	0.356	-19.1	105	0.00
77	t	tert-Butylbenzene	1.036	1.249	-20.6	106	0.00
78	t	1,2,4-Trimethylbenzene	1.016	1.273	-25.3	104	0.00
79	t	sec-Butylbenzene	1.319	1.631	-23.7	107	0.00
80		Nitrobenzene	0.007	0.002	71.4#	21	0.00
81	t	p-Isopropyltoluene	1.041	1.352	-29.9	110	0.00
82	t	1,3-Dichlorobenzene	0.578	0.688	-19.0	109	0.00
83	Rt	1,4-Dichlorobenzene	0.565	0.688	-21.8	108	0.00
84	t	n-Butylbenzene	0.933	1.291	-38.4#	117	0.00
85	Rt	1,2-Dichlorobenzene	0.555	0.660	-18.9	109	0.00
86	t	1,2-Dibromo-3-Chloropropane	0.042	0.046	-9.5	90	0.00
87	Rt	1,2,4-Trichlorobenzene	0.271	0.380	-40.2#	117	0.00
88	t	Hexachlorobutadiene	0.194	0.225	-16.0	117	0.00
89	t	Naphthalene	0.436	0.626	-43.6#	107	0.00
90	t	1,2,3-Trichlorobenzene	0.265	0.321	-21.1	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Evaluate Continuing Calibration Report

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063225.D
Acq On : 10 Feb 2025 16:45
Operator : MD/SY
Sample : VSTDICV010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Quant Time: Feb 11 04:20:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 04:17:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 200%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	i	Fluorobenzene	1.000	1.000	0.0	95	0.00
2	T	Dichlorodifluoromethane	10.000	6.907	30.9#	70	0.00
3	t	Chloromethane	10.000	7.762	22.4	78	0.00
4	Rt	Vinyl Chloride	10.000	8.561	14.4	82	0.00
5	T	Bromomethane	10.000	5.874	41.3#	50	-0.02
6	T	Chloroethane	10.000	8.736	12.6	87	-0.03
7	T	Trichlorofluoromethane	10.000	9.197	8.0	90	-0.02
8		1,1,2-Trichloro-1,2,2-trifl	10.000	10.436	-4.4	107	0.00
9	Rt	1,1-Dichloroethene	10.000	10.339	-3.4	100	0.00
10	t	Iodomethane	10.000	9.372	6.3	87	0.00
11	t	Allyl Chloride	10.000	11.005	-10.1	104	0.00
12	t	Acrylonitrile	20.000	51.821	-159.1#	221	0.00
13	T	Acetone	50.000	47.295	5.4	85	0.00
14	T	Carbon Disulfide	10.000	8.683	13.2	85	0.00
15	RT	Methylene Chloride	10.000	9.632	3.7	94	0.00
16	RT	trans-1,2-Dichloroethene	10.000	9.754	2.5	93	0.00
17	t	1,1-Dichloroethane	10.000	10.091	-0.9	97	0.00
18	T	2-Butanone	50.000	50.293	-0.6	87	0.00
19		Cyclohexane	10.000	10.331	-3.3	95	0.00
20		Methylcyclohexane	10.000	11.086	-10.9	102	0.00
21	T	2,2-Dichloropropane	10.000	10.612	-6.1	102	0.00
22	RT	cis-1,2-Dichloroethene	10.000	10.837	-8.4	102	0.00
23	t	Diethyl Ether	10.000	9.447	5.5	90	0.00
24	t	tert-Butyl Alcohol	100.000	51.182	48.8#	48	0.03
25	t	Methyl tert-Butyl Ether	10.000	10.661	-6.6	95	0.00
26	t	Bromochloromethane	10.000	9.196	8.0	87	0.00
27	t	Chloroform	10.000	10.089	-0.9	96	0.00
28	RT	1,1,1-Trichloroethane	10.000	10.272	-2.7	98	0.00
29	T	1,1-Dichloropropene	10.000	9.535	4.6	89	0.00
30	RT	Carbon Tetrachloride	10.000	10.069	-0.7	97	0.00
31	t	Isopropyl Ether	10.000	11.684	-16.8	106	0.00
32		Ethyl-t-butyl ether	10.000	0.000	100.0#	0	-4.48#
33		Tert-Amyl methyl ether	10.000	0.000	100.0#	0	-5.93#
34	t	Propionitrile	50.000	112.251	-124.5#	185	0.00
35	RT	Benzene	10.000	10.081	-0.8	95	0.00
36	RT	1,2-Dichloroethane	10.000	9.317	6.8	88	0.00
37	RT	Trichloroethene	10.000	9.455	5.4	90	0.00
38	Rt	1,2-Dichloropropane	10.000	9.657	3.4	90	0.00
39	t	Methacrylonitrile	10.000	11.518	-15.2	90	0.00
40	t	Methyl acrylate	10.000	10.636	-6.4	90	0.00
41	t	Tetrahydrofuran	20.000	51.903	-159.5#	224	0.00
42	t	1-Chlorobutane	10.000	10.606	-6.1	97	0.00
43	T	Dibromomethane	10.000	9.800	2.0	93	0.00
44	T	Bromodichloromethane	10.000	10.859	-8.6	100	0.00
45	T	4-Methyl-2-Pentanone	50.000	55.614	-11.2	91	0.00
46	t	t-1,4-Dichloro-2-butene	20.000	12.207	39.0#	55	0.00
47	t	Methyl methacrylate	20.000	10.801	46.0#	44	0.00
48	t	Ethyl methacrylate	10.000	11.820	-18.2	92	0.00
49	Rt	Toluene	10.000	10.739	-7.4	97	0.00

ICAL Form

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063225.D
Acq On : 10 Feb 2025 16:45
Operator : MD/SY
Sample : VSTDICV010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
ICVVU021025

Quant Time: Feb 11 04:20:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 04:17:17 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 200%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
50	T	t-1,3-Dichloropropene	10.000	10.914	-9.1	94	0.00
51	T	cis-1,3-Dichloropropene	10.000	10.315	-3.1	91	0.00
52	RT	1,1,2-Trichloroethane	10.000	9.902	1.0	90	0.00
53	t	1,3-Dichloropropane	10.000	9.732	2.7	89	0.00
54	t	2-Hexanone	50.000	57.943	-15.9	94	0.00
55	t	Dibromochloromethane	10.000	10.441	-4.4	93	0.00
56	T	1,2-Dibromoethane	10.000	9.891	1.1	88	0.00
57	S	4-Bromofluorobenzene	1.000	1.287	-28.7	112	0.00
58	RT	Tetrachloroethene	10.000	10.255	-2.6	100	0.00
59	Rt	Chlorobenzene	10.000	10.678	-6.8	98	0.00
60	T	1,1,1,2-Tetrachloroethane	10.000	9.912	0.9	92	0.00
61	t	Pentachloroethane	10.000	11.159	-11.6	105	0.00
62	t	Hexachloroethane	10.000	11.642	-16.4	105	0.00
63	Rt	Ethyl Benzene	10.000	11.289	-12.9	98	0.00
64	RT	m/p-Xylenes	20.000	23.222	-16.1	99	0.00
65	RT	o-Xylene	10.000	11.453	-14.5	99	0.00
66	RT	Styrene	10.000	12.190	-21.9	101	0.00
67	t	Bromoform	10.000	11.146	-11.5	97	0.00
68	S	1,2-Dichlorobenzene-d4	1.000	1.115	-11.5	105	0.00
69	T	Isopropylbenzene	10.000	12.848	-28.5	111	0.00
70	T	1,1,2,2-Tetrachloroethane	10.000	10.491	-4.9	95	0.00
71	T	1,2,3-Trichloropropane	10.000	9.872	1.3	102	0.00
72	t	Bromobenzene	10.000	11.275	-12.8	101	0.00
73	t	n-propylbenzene	10.000	12.145	-21.4	103	0.00
74	t	2-Chlorotoluene	10.000	11.986	-19.9	105	0.00
75	t	1,3,5-Trimethylbenzene	10.000	12.210	-22.1	103	0.00
76	t	4-Chlorotoluene	10.000	11.887	-18.9	105	0.00
77	t	tert-Butylbenzene	10.000	12.057	-20.6	106	0.00
78	t	1,2,4-Trimethylbenzene	10.000	12.521	-25.2	104	0.00
79	t	sec-Butylbenzene	10.000	12.368	-23.7	107	0.00
80		Nitrobenzene	50.000	14.120	71.8#	21	0.00
81	t	p-Isopropyltoluene	10.000	12.990	-29.9	110	0.00
82	t	1,3-Dichlorobenzene	10.000	11.912	-19.1	109	0.00
83	Rt	1,4-Dichlorobenzene	10.000	12.181	-21.8	108	0.00
84	t	n-Butylbenzene	10.000	13.835	-38.4#	117	0.00
85	Rt	1,2-Dichlorobenzene	10.000	11.884	-18.8	109	0.00
86	t	1,2-Dibromo-3-Chloropropane	10.000	11.165	-11.6	90	0.00
87	Rt	1,2,4-Trichlorobenzene	10.000	14.027	-40.3#	117	0.00
88	t	Hexachlorobutadiene	10.000	11.632	-16.3	117	0.00
89	t	Naphthalene	10.000	10.999	-10.0	107	0.00
90	t	1,2,3-Trichlorobenzene	10.000	12.127	-21.3	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Raw Data: **VU063249.D**

Alliance TECHNICAL GROUP

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172 SDG No.: Q1172

Instrument ID: MSVOA_U Calibration Date/Time: 02/12/2025 10:48

Lab File ID: VU063249.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:59 15:33

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.325	0.315		-3.08	30
Chloromethane	0.374	0.351		-6.15	30
Vinyl Chloride	0.370	0.363		-1.89	30
Bromomethane	0.181	0.201		11.05	30
Chloroethane	0.233	0.219		-6.01	30
Tetrahydrofuran	0.047	0.047		0	30
Trichlorofluoromethane	0.439	0.437		-0.46	30
1,1,2-Trichloro-1,2,2-trifluoroethane	0.249	0.257		3.21	30
tert-Butyl Alcohol	0.026	0.019		-26.92	30
Diethyl Ether	0.218	0.210		-3.67	30
1,1-Dichloroethene	0.254	0.260		2.36	30
Acrylonitrile	0.059	0.060		1.7	30
Acetone	0.045	0.040		-11.11	30
Carbon Disulfide	0.887	0.865		-2.48	30
Methyl tert-Butyl Ether	0.634	0.653		3	30
Methyl acrylate	0.148	0.153		3.38	30
Methylene Chloride	0.313	0.307		-1.92	30
trans-1,2-Dichloroethene	0.290	0.296		2.07	30
1,1-Dichloroethane	0.546	0.551		0.92	30
Cyclohexane	0.439	0.473		7.74	30
2-Butanone	0.072	0.070		-2.78	30
Carbon Tetrachloride	0.383	0.394		2.87	30
2,2-Dichloropropane	0.426	0.438		2.82	30
cis-1,2-Dichloroethene	0.313	0.322		2.88	30
Bromochloromethane	0.137	0.138		0.73	30
Chloroform	0.551	0.553		0.36	30
1,1,1-Trichloroethane	0.446	0.454		1.79	30
Methylcyclohexane	0.435	0.494		13.56	30
1,1-Dichloropropene	0.400	0.428		7	30
Propionitrile	0.022	0.022		0	30
Benzene	1.229	1.264		2.85	30
1,2-Dichloroethane	0.355	0.352		-0.85	30
Trichloroethene	0.292	0.299		2.4	30
1,2-Dichloropropane	0.322	0.330		2.48	30
1-Chlorobutane	0.547	0.571		4.39	30
Dibromomethane	0.163	0.163		0	30
Bromodichloromethane	0.379	0.391		3.17	30
4-Methyl-2-Pentanone	0.171	0.180		5.26	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Raw Data: **VU063249.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172 SDG No.: Q1172

Instrument ID: MSVOA_U Calibration Date/Time: 02/12/2025 10:48

Lab File ID: VU063249.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:59 15:33

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Toluene	0.707	0.772		9.19	30
t-1,3-Dichloropropene	0.347	0.374		7.78	30
cis-1,3-Dichloropropene	0.429	0.448		4.43	30
1,1,2-Trichloroethane	0.220	0.230		4.55	30
1,3-Dichloropropane	0.390	0.401		2.82	30
2-Hexanone	0.117	0.123		5.13	30
Dibromochloromethane	0.253	0.262		3.56	30
1,2-Dibromoethane	0.206	0.210		1.94	30
Tetrachloroethene	0.241	0.261		8.3	30
Chlorobenzene	0.746	0.783		4.96	30
1,1,1,2-Tetrachloroethane	0.268	0.281		4.85	30
Hexachloroethane	0.212	0.221		4.24	30
Ethyl Benzene	1.286	1.442		12.13	30
m/p-Xylenes	0.480	0.553		15.21	30
o-Xylene	0.470	0.526		11.91	30
Styrene	0.748	0.871		16.44	30
Bromoform	0.143	0.150		4.89	30
Isopropylbenzene	1.106	1.255		13.47	30
1,1,2,2-Tetrachloroethane	0.296	0.293		-1.01	30
1,2,3-Trichloropropane	0.222	0.215		-3.15	30
Bromobenzene	0.298	0.319		7.05	30
n-propylbenzene	0.317	0.375		18.3	30
2-Chlorotoluene	0.292	0.325		11.3	30
1,3,5-Trimethylbenzene	1.024	1.181		15.33	30
4-Chlorotoluene	0.299	0.340		13.71	30
tert-Butylbenzene	1.036	1.142		10.23	30
1,2,4-Trimethylbenzene	1.016	1.181		16.24	30
sec-Butylbenzene	1.319	1.520		15.24	30
p-Isopropyltoluene	1.041	1.231		18.25	30
1,3-Dichlorobenzene	0.578	0.618		6.92	30
1,4-Dichlorobenzene	0.565	0.617		9.2	30
n-Butylbenzene	0.933	1.107		18.65	30
1,2-Dichlorobenzene	0.555	0.584		5.22	30
1,2-Dibromo-3-Chloropropane	0.042	0.044		4.76	30
1,2,4-Trichlorobenzene	0.271	0.329		21.4	30
Hexachlorobutadiene	0.194	0.215		10.82	30
Naphthalene	0.436	0.504		15.6	30
1,2,3-Trichlorobenzene	0.265	0.314		18.49	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172 SDG No.: Q1172

Instrument ID: MSVOA_U Calibration Date/Time: 02/12/2025 10:48

Lab File ID: VU063249.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:59 15:33

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
1,2-Dichlorobenzene-d4	0.343	0.341		-0.58	30
4-Bromofluorobenzene	0.330	0.438		32.73	30
Iodomethane	0.399	0.417		4.51	30
Allyl Chloride	0.364	0.364		0	30
Methacrylonitrile	0.080	0.088		10	30
Isopropyl Ether	0.779	0.810		3.98	30
Methyl methacrylate	0.137	0.151		10.22	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063249.D
Acq On : 12 Feb 2025 10:48
Operator : MD/SY
Sample : VSTDCCC010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Feb 13 03:06:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Fluorobenzene	6.107	96	54296	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.624	95	23785	1.328	ug/l	0.00
Spiked Amount 1.000			Recovery = 133.000%			
68) 1,2-Dichlorobenzene-d4	12.184	152	18535	0.995	ug/l	0.00
Spiked Amount 1.000			Recovery = 100.000%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.380	85	171269	9.709	ug/l	97
3) Chloromethane	1.515	50	190441	9.374	ug/l	99
4) Vinyl Chloride	1.596	62	196947	9.798	ug/l	99
5) Bromomethane	1.843	94	109059	11.104	ug/l	100
6) Chloroethane	1.920	64	118841	9.387	ug/l	100
7) Trichlorofluoromethane	2.129	101	237258	9.961	ug/l	99
8) 1,1,2-Trichloro-1,2,2-...	2.570	101	139748	10.336	ug/l	99
9) 1,1-Dichloroethene	2.570	96	141173	10.249	ug/l	97
10) Iodomethane	2.708	142	226361	10.453	ug/l	100
11) Allyl Chloride	2.911	41	197730	9.992	ug/l	98
12) Acrylonitrile	3.306	53	65351	20.569	ug/l	99
13) Acetone	2.618	43	109824	44.953	ug/l	99
14) Carbon Disulfide	2.782	76	469449	9.747	ug/l	100
15) Methylene Chloride	3.033	84	166812	9.802	ug/l	97
16) trans-1,2-Dichloroethene	3.338	96	160953	10.238	ug/l	99
17) 1,1-Dichloroethane	3.853	63	299422	10.104	ug/l	99
18) 2-Butanone	4.689	43	189558	48.752	ug/l	98
19) Cyclohexane	5.374	56	256943m	10.790	ug/l	
20) Methylcyclohexane	6.753	83	268292	11.362	ug/l	98
21) 2,2-Dichloropropane	4.647	77	237576	10.276	ug/l	100
22) cis-1,2-Dichloroethene	4.653	96	174819	10.291	ug/l	99
23) Diethyl Ether	2.367	59	114120	9.655	ug/l	98
24) tert-Butyl Alcohol	3.174	59	102081	83.500	ug/l	99
25) Methyl tert-Butyl Ether	3.348	73	354554	10.306	ug/l	99
26) Bromochloromethane	4.959	128	74753	10.068	ug/l	99
27) Chloroform	5.071	83	300401	10.045	ug/l	100
28) 1,1,1-Trichloroethane	5.303	97	246535	10.177	ug/l	100
29) 1,1-Dichloropropene	5.512	75	232266	10.705	ug/l	99
30) Carbon Tetrachloride	5.512	117	213796	10.291	ug/l	98
31) Isopropyl Ether	3.975	45	439672	10.392	ug/l	99
32) Ethyl-t-butyl ether	4.486	59	404830	10.522	ug/l	99
33) Tert-Amyl methyl ether	5.927	73	364333	10.838	ug/l	100
34) Propionitrile	4.759	54	60683	51.745	ug/l #	98
35) Benzene	5.759	78	686051	10.283	ug/l	100
36) 1,2-Dichloroethane	5.779	62	191321	9.936	ug/l	99
37) Trichloroethene	6.531	130	162142	10.219	ug/l	95
38) 1,2-Dichloropropane	6.779	63	179409	10.274	ug/l	98
39) Methacrylonitrile	4.959	41	47930	10.967	ug/l	94
40) Methyl acrylate	4.837	55	83334	10.346	ug/l #	98
41) Tetrahydrofuran	5.039	42	51395	20.220	ug/l	96
42) 1-Chlorobutane	5.444	56	309896	10.439	ug/l	100
43) Dibromomethane	6.907	93	88575	10.017	ug/l	98
44) Bromodichloromethane	7.094	83	212526	10.327	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063249.D
Acq On : 12 Feb 2025 10:48
Operator : MD/SY
Sample : VSTDCCC010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Feb 13 03:06:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
45) 4-Methyl-2-Pentanone	7.779	43	488624	52.711	ug/l	99
46) t-1,4-Dichloro-2-butene	10.820	75	95729m	21.967	ug/l	
47) Methyl methacrylate	6.949	69	164430	22.078	ug/l	99
48) Ethyl methacrylate	8.322	69	160597	11.484	ug/l	99
49) Toluene	7.959	92	419404	10.932	ug/l	99
50) t-1,3-Dichloropropene	8.200	75	203081	10.778	ug/l	99
51) cis-1,3-Dichloropropene	7.595	75	243151	10.447	ug/l	98
52) 1,1,2-Trichloroethane	8.386	97	125016	10.486	ug/l	97
53) 1,3-Dichloropropane	8.563	76	217877	10.294	ug/l	99
54) 2-Hexanone	8.672	43	333762	52.760	ug/l	98
55) Dibromochloromethane	8.798	129	142483	10.388	ug/l	99
56) 1,2-Dibromoethane	8.914	107	114002	10.196	ug/l	98
58) Tetrachloroethene	8.544	164	141699	10.837	ug/l	99
59) Chlorobenzene	9.438	112	424889	10.494	ug/l	99
60) 1,1,1,2-Tetrachloroethane	9.521	131	152380	10.471	ug/l	98
61) Pentachloroethane	11.415	117	128136	9.856	ug/l	98
62) Hexachloroethane	12.463	117	120228	10.453	ug/l	98
63) Ethyl Benzene	9.560	91	782784	11.211	ug/l	100
64) m/p-Xylenes	9.682	106	600186	23.011	ug/l	99
65) o-Xylene	10.090	106	285666	11.188	ug/l	98
66) Styrene	10.103	104	473092	11.643	ug/l	100
67) Bromoform	10.280	173	81654	10.490	ug/l	100
69) Isopropylbenzene	10.473	105	681530	11.354	ug/l	99
70) 1,1,2,2-Tetrachloroethane	10.769	83	159278	9.913	ug/l	100
71) 1,2,3-Trichloropropane	10.814	75	116559m	9.656	ug/l	
72) Bromobenzene	10.772	156	173177	10.707	ug/l	99
73) n-propylbenzene	10.894	120	203622	11.848	ug/l	99
74) 2-Chlorotoluene	10.975	126	176492	11.143	ug/l	99
75) 1,3,5-Trimethylbenzene	11.078	105	641052	11.526	ug/l	100
76) 4-Chlorotoluene	11.087	126	184692	11.371	ug/l	96
77) tert-Butylbenzene	11.409	119	619981	11.021	ug/l	100
78) 1,2,4-Trimethylbenzene	11.457	105	641240	11.619	ug/l	100
79) sec-Butylbenzene	11.634	105	825347	11.524	ug/l	100
80) Nitrobenzene	13.203	77	20299m	43.211	ug/l	
81) p-Isopropyltoluene	11.782	119	668338	11.824	ug/l	100
82) 1,3-Dichlorobenzene	11.737	146	335328	10.689	ug/l	100
83) 1,4-Dichlorobenzene	11.827	146	335268	10.926	ug/l	99
84) n-Butylbenzene	12.200	91	601201	11.862	ug/l	99
85) 1,2-Dichlorobenzene	12.203	146	316850	10.510	ug/l	99
86) 1,2-Dibromo-3-Chloropr...	12.984	75	23790	10.548	ug/l	99
87) 1,2,4-Trichlorobenzene	13.830	180	178545	12.143	ug/l	99
88) Hexachlorobutadiene	14.010	225	116891	11.124	ug/l	99
89) Naphthalene	14.077	128	273523	8.979	ug/l	100
90) 1,2,3-Trichlorobenzene	14.319	180	170303	11.837	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063249.D
Acq On : 12 Feb 2025 10:48
Operator : MD/SY
Sample : VSTDCCC010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Feb 13 03:06:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.107 min Scan# 14
Delta R.T. 0.004 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 96 Resp: 5429

Ion	Ratio	Lower	Upper
96	100		
70	21.1	15.6	23.4
95	9.1	7.3	10.9
97	0.0	0.0	0.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#2
Dichlorodifluoromethane
Concen: 9.709 ug/l
RT: 1.380 min Scan# 28
Delta R.T. 0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 85 Resp: 171269

Ion	Ratio	Lower	Upper
85	100		
87	33.8	16.0	48.0





#3
Chloromethane
Concen: 9.374 ug/l
RT: 1.515 min Scan# 7
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 50 Resp: 19044

Ion	Ratio	Lower	Upper
50	100		
52	32.9	25.8	38.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#4
Vinyl Chloride
Concen: 9.798 ug/l
RT: 1.596 min Scan# 95
Delta R.T. -0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 62 Resp: 196947

Ion	Ratio	Lower	Upper
62	100		
64	32.3	25.4	38.0





#5
Bromomethane
Concen: 11.104 ug/l
RT: 1.843 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 94 Resp: 10905

Ion	Ratio	Lower	Upper
94	100		
96	94.2	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#6
Chloroethane
Concen: 9.387 ug/l
RT: 1.920 min Scan# 196
Delta R.T. -0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 64 Resp: 118841

Ion	Ratio	Lower	Upper
64	100		
66	32.4	25.8	38.8





#7
Trichlorofluoromethane
Concen: 9.961 ug/l
RT: 2.129 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:101 Resp: 23725

Ion	Ratio	Lower	Upper
101	100		
103	65.2	51.4	77.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 10.336 ug/l
RT: 2.570 min Scan# 398
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:101 Resp: 139748

Ion	Ratio	Lower	Upper
101	100		
85	44.3	35.4	53.0
151	72.1	58.5	87.7





#9
1,1-Dichloroethene
Concen: 10.249 ug/l
RT: 2.570 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 96 Resp: 14117

Ion	Ratio	Lower	Upper
96	100		
61	158.7	0.0	492.9
98	61.1	0.0	124.0

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#10
Iodomethane
Concen: 10.453 ug/l
RT: 2.708 min Scan# 441
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:142 Resp: 226361

Ion	Ratio	Lower	Upper
142	100		
127	42.8	34.5	51.7





#11
Allyl Chloride
Concen: 9.992 ug/l
RT: 2.911 min Scan# 5
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 41 Resp: 19773

Ion	Ratio	Lower	Upper
41	100		
39	73.7	57.9	86.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#12
Acrylonitrile
Concen: 20.569 ug/l
RT: 3.306 min Scan# 627
Delta R.T. 0.006 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 53 Resp: 65351

Ion	Ratio	Lower	Upper
53	100		
52	79.6	64.2	96.2
51	38.8	30.8	46.2





#13
Acetone
Concen: 44.953 ug/l
RT: 2.618 min Scan# 41
Delta R.T. 0.006 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 43 Resp: 10982

Ion	Ratio	Lower	Upper
43	100		
58	33.5	27.4	41.0

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#14
Carbon Disulfide
Concen: 9.747 ug/l
RT: 2.782 min Scan# 464
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 76 Resp: 469449

Ion	Ratio	Lower	Upper
76	100		
78	9.0	7.2	10.8





#15
Methylene Chloride
Concen: 9.802 ug/l
RT: 3.033 min Scan# 54
Delta R.T. 0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 84 Resp: 16681

Ion	Ratio	Lower	Upper
84	100		
49	114.7	96.4	144.6
51	36.0	0.0	74.8
86	63.8	51.3	76.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#16
trans-1,2-Dichloroethene
Concen: 10.238 ug/l
RT: 3.338 min Scan# 637
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 96 Resp: 160953

Ion	Ratio	Lower	Upper
96	100		
61	142.6	113.4	170.2
98	62.0	51.2	76.8





#17
1,1-Dichloroethane
Concen: 10.104 ug/l
RT: 3.853 min Scan# 7
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 63 Resp: 29942

Ion	Ratio	Lower	Upper
63	100		
98	6.3	3.3	9.9
100	4.3	2.1	6.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#18
2-Butanone
Concen: 48.752 ug/l
RT: 4.689 min Scan# 1057
Delta R.T. 0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 43 Resp: 189558

Ion	Ratio	Lower	Upper
43	100		
57	7.8	0.0	17.0





#19
Cyclohexane
Concen: 10.790 ug/l m
RT: 5.374 min Scan# 1
Delta R.T. -0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 56 Resp: 25694

Ion	Ratio	Lower	Upper
56	100		
69	30.7	24.5	36.7
84	83.4	65.2	97.8

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#20
Methylcyclohexane
Concen: 11.362 ug/l
RT: 6.753 min Scan# 1699
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 83 Resp: 268292

Ion	Ratio	Lower	Upper
83	100		
55	77.2	63.1	94.7
98	44.5	35.2	52.8





#21
2,2-Dichloropropane
Concen: 10.276 ug/l
RT: 4.647 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 77 Resp: 23757

Ion	Ratio	Lower	Upper
77	100		
97	23.1	18.5	27.7



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#22
cis-1,2-Dichloroethene
Concen: 10.291 ug/l
RT: 4.653 min Scan# 1046
Delta R.T. 0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 96 Resp: 174819

Ion	Ratio	Lower	Upper
96	100		
61	149.6	0.0	373.3
98	65.4	31.9	95.9





#23
Diethyl Ether
Concen: 9.655 ug/l
RT: 2.367 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 59 Resp: 11412

Ion	Ratio	Lower	Upper
59	100		
45	69.1	57.8	86.6
74	71.6	57.7	86.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#24
tert-Butyl Alcohol
Concen: 83.500 ug/l
RT: 3.174 min Scan# 586
Delta R.T. 0.016 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 59 Resp: 102081

Ion	Ratio	Lower	Upper
59	100		
57	9.9	7.5	11.3





#25
Methyl tert-Butyl Ether
Concen: 10.306 ug/l
RT: 3.348 min Scan# 64
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 73 Resp: 354554

Ion	Ratio	Lower	Upper
73	100		
43	18.8	14.9	22.3

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#26
Bromochloromethane
Concen: 10.068 ug/l
RT: 4.959 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:128 Resp: 74753

Ion	Ratio	Lower	Upper
128	100		
49	169.2	0.0	343.4
130	127.7	102.9	154.3





#27
Chloroform
Concen: 10.045 ug/l
RT: 5.071 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 83 Resp: 30040

Ion	Ratio	Lower	Upper
83	100		
85	65.2	0.0	129.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#28
1,1,1-Trichloroethane
Concen: 10.177 ug/l
RT: 5.303 min Scan# 1248
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 97 Resp: 246535

Ion	Ratio	Lower	Upper
97	100		
99	64.4	32.4	97.0
61	47.2	23.8	71.2





#29
1,1-Dichloropropene
Concen: 10.705 ug/l
RT: 5.512 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 75 Resp: 23226

Ion	Ratio	Lower	Upper
75	100		
110	34.9	17.2	51.5
77	31.1	24.6	37.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#30
Carbon Tetrachloride
Concen: 10.291 ug/l
RT: 5.512 min Scan# 1313
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:117 Resp: 213796

Ion	Ratio	Lower	Upper
117	100		
119	94.2	76.7	115.1
121	29.7	25.5	38.3





#31
Isopropyl Ether
Concen: 10.392 ug/l
RT: 3.975 min Scan# 8
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 45 Resp: 43967

Ion	Ratio	Lower	Upper
45	100		
43	51.8	25.7	77.1



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#32
Ethyl-t-butyl ether
Concen: 10.522 ug/l
RT: 4.486 min Scan# 994
Delta R.T. 0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 59 Resp: 404830

Ion	Ratio	Lower	Upper
59	100		
87	41.4	32.6	49.0





#33
Tert-Amyl methyl ether
Concen: 10.838 ug/l
RT: 5.927 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 73 Resp: 36433

Ion	Ratio	Lower	Upper
73	100		
87	23.4	18.9	28.3
71	11.4	9.0	13.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#34
Propionitrile
Concen: 51.745 ug/l
RT: 4.759 min Scan# 1079
Delta R.T. 0.006 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 54 Resp: 60683

Ion	Ratio	Lower	Upper
54	100		
52	20.2	16.3	24.5
55	14.7	11.8	17.6
40	2.1	4.3	6.5#





#35
Benzene
Concen: 10.283 ug/l
RT: 5.759 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 78 Resp: 68605

Ion	Ratio	Lower	Upper
78	100		
77	23.9	19.0	28.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#36
1,2-Dichloroethane
Concen: 9.936 ug/l
RT: 5.779 min Scan# 1396
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 62 Resp: 191321

Ion	Ratio	Lower	Upper
62	100		
98	8.9	6.9	10.3





#37
Trichloroethene
Concen: 10.219 ug/l
RT: 6.531 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:130 Resp: 16214

Ion	Ratio	Lower	Upper
130	100		
95	109.0	83.2	124.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#38
1,2-Dichloropropane
Concen: 10.274 ug/l
RT: 6.779 min Scan# 1707
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 63 Resp: 179409

Ion	Ratio	Lower	Upper
63	100		
65	32.5	25.3	37.9





#39
Methacrylonitrile
Concen: 10.967 ug/l
RT: 4.959 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 41 Resp: 4793

Ion	Ratio	Lower	Upper
41	100		
67	86.5	64.5	96.7
39	56.0	43.0	64.6
52	36.1	25.3	37.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#40
Methyl acrylate
Concen: 10.346 ug/l
RT: 4.837 min Scan# 1103
Delta R.T. 0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 55 Resp: 83334

Ion	Ratio	Lower	Upper
55	100		
85	16.2	13.3	19.9
58	7.2	7.3	10.9#
42	8.2	6.9	10.3





#41
Tetrahydrofuran
Concen: 20.220 ug/l
RT: 5.039 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 42 Resp: 5139

Ion	Ratio	Lower	Upper
42	100		
72	50.3	41.5	62.3
71	43.0	37.2	55.8



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#42
1-Chlorobutane
Concen: 10.439 ug/l
RT: 5.444 min Scan# 1292
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 56 Resp: 309896

Ion	Ratio	Lower	Upper
56	100		
41	52.5	26.3	78.8





#43
Dibromomethane
Concen: 10.017 ug/l
RT: 6.907 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 93 Resp: 8857

Ion	Ratio	Lower	Upper
93	100		
95	83.1	67.2	100.8
174	91.3	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#44
Bromodichloromethane
Concen: 10.327 ug/l
RT: 7.094 min Scan# 1805
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 83 Resp: 212526

Ion	Ratio	Lower	Upper
83	100		
85	64.2	51.7	77.5
127	8.2	6.7	10.1





#45
4-Methyl-2-Pentanone
Concen: 52.711 ug/l
RT: 7.779 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 43 Resp: 48862

Ion	Ratio	Lower	Upper
43	100		
58	41.1	20.8	62.5
85	18.2	14.8	22.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#46
t-1,4-Dichloro-2-butene
Concen: 21.967 ug/l m
RT: 10.820 min Scan# 2964
Delta R.T. -0.006 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 75 Resp: 95729

Ion	Ratio	Lower	Upper
75	100		
53	70.5	64.5	96.7
89	46.8	43.4	65.2
88	34.7	31.2	46.8





#47
Methyl methacrylate
Concen: 22.078 ug/l
RT: 6.949 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 69 Resp: 16443

Ion	Ratio	Lower	Upper
69	100		
41	107.4	0.0	217.0
39	60.0	47.7	71.5
100	34.6	27.4	41.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#48
Ethyl methacrylate
Concen: 11.484 ug/l
RT: 8.322 min Scan# 2187
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 69 Resp: 160597

Ion	Ratio	Lower	Upper
69	100		
41	60.5	30.6	92.0





#49
Toluene
Concen: 10.932 ug/l
RT: 7.959 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 92 Resp: 41940

Ion	Ratio	Lower	Upper
92	100		
91	175.8	141.8	212.6



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#50
t-1,3-Dichloropropene
Concen: 10.778 ug/l
RT: 8.200 min Scan# 2149
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 75 Resp: 203081

Ion	Ratio	Lower	Upper
75	100		
77	31.7	25.9	38.9





#51
cis-1,3-Dichloropropene
Concen: 10.447 ug/l
RT: 7.595 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 75 Resp: 24315

Ion	Ratio	Lower	Upper
75	100		
77	32.2	25.3	37.9
39	43.2	33.5	50.3



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#52
1,1,2-Trichloroethane
Concen: 10.486 ug/l
RT: 8.386 min Scan# 2207
Delta R.T. -0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 97 Resp: 125016

Ion	Ratio	Lower	Upper
97	100		
83	86.7	73.0	109.4
85	57.3	46.3	69.5
99	59.8	48.5	72.7





#53
1,3-Dichloropropane
Concen: 10.294 ug/l
RT: 8.563 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 76 Resp: 21787

Ion	Ratio	Lower	Upper
76	100		
78	32.2	26.3	39.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#54
2-Hexanone
Concen: 52.760 ug/l
RT: 8.672 min Scan# 2296
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 43 Resp: 333762

Ion	Ratio	Lower	Upper
43	100		
58	57.0	38.0	78.0
57	19.9	0.0	39.1





#55
Dibromochloromethane
Concen: 10.388 ug/l
RT: 8.798 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:129 Resp: 14248

Ion	Ratio	Lower	Upper
129	100		
127	76.8	60.4	90.6



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#56
1,2-Dibromoethane
Concen: 10.196 ug/l
RT: 8.914 min Scan# 2371
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:107 Resp: 114002

Ion	Ratio	Lower	Upper
107	100		
109	95.7	0.0	187.8





#57
4-Bromofluorobenzene
Concen: 1.328 ug/l
RT: 10.624 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 95 Resp: 2378

Ion	Ratio	Lower	Upper
95	100		
174	75.4	58.6	88.0
176	74.1	58.2	87.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#58
Tetrachloroethene
Concen: 10.837 ug/l
RT: 8.544 min Scan# 2256
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:164 Resp: 141699

Ion	Ratio	Lower	Upper
164	100		
166	127.2	101.4	152.0
129	96.7	77.0	115.4
131	93.3	76.3	114.5





#59
Chlorobenzene
Concen: 10.494 ug/l
RT: 9.438 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:112 Resp: 42488

Ion	Ratio	Lower	Upper
112	100		
114	32.7	25.7	38.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#60
1,1,1,2-Tetrachloroethane
Concen: 10.471 ug/l
RT: 9.521 min Scan# 2560
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:131 Resp: 152380

Ion	Ratio	Lower	Upper
131	100		
133	94.5	76.7	115.1
119	66.3	54.4	81.6





#61
Pentachloroethane
Concen: 9.856 ug/l
RT: 11.415 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:117 Resp: 12813

Ion	Ratio	Lower	Upper
117	100		
167	79.3	65.0	97.6

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#62
Hexachloroethane
Concen: 10.453 ug/l
RT: 12.463 min Scan# 3475
Delta R.T. -0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:117 Resp: 120228

Ion	Ratio	Lower	Upper
117	100		
201	75.2	61.3	91.9





#63
Ethyl Benzene
Concen: 11.211 ug/l
RT: 9.560 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 91 Resp: 78278

Ion	Ratio	Lower	Upper
91	100		
106	29.9	24.2	36.2

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#64
m/p-Xylenes
Concen: 23.011 ug/l
RT: 9.682 min Scan# 2610
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:106 Resp: 600186

Ion	Ratio	Lower	Upper
106	100		
91	209.8	166.9	250.3





#65
o-Xylene
Concen: 11.188 ug/l
RT: 10.090 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:106 Resp: 28566

Ion	Ratio	Lower	Upper
106	100		
91	219.0	110.9	332.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#66
Styrene
Concen: 11.643 ug/l
RT: 10.103 min Scan# 2741
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:104 Resp: 473092

Ion	Ratio	Lower	Upper
104	100		
78	51.0	41.2	61.8
103	55.9	44.8	67.2





#67
Bromoform
Concen: 10.490 ug/l
RT: 10.280 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:173 Resp: 8165

Ion	Ratio	Lower	Upper
173	100		
175	48.8	39.0	58.4
254	9.2	7.7	11.5



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#68
1,2-Dichlorobenzene-d4
Concen: 0.995 ug/l
RT: 12.184 min Scan# 3388
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:152 Resp: 18535

Ion	Ratio	Lower	Upper
152	100		
115	145.2	0.0	275.2
150	333.4	0.0	658.4





#69
Isopropylbenzene
Concen: 11.354 ug/l
RT: 10.473 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:105 Resp: 68153

Ion	Ratio	Lower	Upper
105	100		
120	25.9	12.8	38.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#70
1,1,2,2-Tetrachloroethane
Concen: 9.913 ug/l
RT: 10.769 min Scan# 2948
Delta R.T. -0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 83 Resp: 159278

Ion	Ratio	Lower	Upper
83	100		
131	9.1	7.4	11.0
85	64.6	51.8	77.8





#71
1,2,3-Trichloropropane
Concen: 9.656 ug/l m
RT: 10.814 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 75 Resp: 11655

Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0
110	33.4	0.0	77.0
97	18.7	0.0	42.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#72
Bromobenzene
Concen: 10.707 ug/l
RT: 10.772 min Scan# 2949
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:156 Resp: 173177

Ion	Ratio	Lower	Upper
156	100		
77	169.7	0.0	343.6
158	96.7	0.0	193.0





#73
n-propylbenzene
Concen: 11.848 ug/l
RT: 10.894 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:120 Resp: 20362

Ion	Ratio	Lower	Upper
120	100		
91	459.9	369.8	554.6



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#74
2-Chlorotoluene
Concen: 11.143 ug/l
RT: 10.975 min Scan# 3012
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:126 Resp: 176492

Ion	Ratio	Lower	Upper
126	100		
91	309.4	0.0	623.8





#75
1,3,5-Trimethylbenzene
Concen: 11.526 ug/l
RT: 11.078 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:105 Resp: 64105

Ion	Ratio	Lower	Upper
105	100		
120	46.9	37.3	55.9

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#76
4-Chlorotoluene
Concen: 11.371 ug/l
RT: 11.087 min Scan# 3047
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:126 Resp: 184692

Ion	Ratio	Lower	Upper
126	100		
91	342.8	0.0	703.6





#77
tert-Butylbenzene
Concen: 11.021 ug/l
RT: 11.409 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:119 Resp: 61998

Ion	Ratio	Lower	Upper
119	100		
91	59.1	29.4	88.3
134	22.1	17.6	26.4

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#78
1,2,4-Trimethylbenzene
Concen: 11.619 ug/l
RT: 11.457 min Scan# 3162
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:105 Resp: 641240

Ion	Ratio	Lower	Upper
105	100		
120	43.6	21.9	65.7





#79
sec-Butylbenzene
Concen: 11.524 ug/l
RT: 11.634 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:105 Resp: 82534

Ion	Ratio	Lower	Upper
105	100		
134	19.0	15.1	22.7

Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#80
Nitrobenzene
Concen: 43.211 ug/l m
RT: 13.203 min Scan# 3705
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 77 Resp: 20299

Ion	Ratio	Lower	Upper
77	100		
123	36.8	18.9	67.1
65	10.6	11.9	15.1#





#81
p-Isopropyltoluene
Concen: 11.824 ug/l
RT: 11.782 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:119 Resp: 66833

Ion	Ratio	Lower	Upper
119	100		
134	25.7	20.3	30.5
91	24.2	19.4	29.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#82
1,3-Dichlorobenzene
Concen: 10.689 ug/l
RT: 11.737 min Scan# 3249
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:146 Resp: 335328

Ion	Ratio	Lower	Upper
146	100		
111	40.8	32.8	49.2
148	63.5	51.1	76.7





#83
1,4-Dichlorobenzene
Concen: 10.926 ug/l
RT: 11.827 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:146 Resp: 33526

Ion	Ratio	Lower	Upper
146	100		
111	39.2	32.1	48.1
148	63.4	50.2	75.4



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#84
n-Butylbenzene
Concen: 11.862 ug/l
RT: 12.200 min Scan# 3393
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 91 Resp: 601201

Ion	Ratio	Lower	Upper
91	100		
92	51.7	41.8	62.8
134	23.8	18.6	28.0





#85
1,2-Dichlorobenzene
Concen: 10.510 ug/l
RT: 12.203 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:146 Resp: 31685

Ion	Ratio	Lower	Upper
146	100		
111	42.9	21.9	65.7
148	64.9	32.3	96.9



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#86
1,2-Dibromo-3-Chloropropane
Concen: 10.548 ug/l
RT: 12.984 min Scan# 3637
Delta R.T. -0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion: 75 Resp: 23790

Ion	Ratio	Lower	Upper
75	100		
155	80.5	63.5	95.3
157	103.4	81.8	122.6





#87
1,2,4-Trichlorobenzene
Concen: 12.143 ug/l
RT: 13.830 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:180 Resp: 17854

Ion	Ratio	Lower	Upper
180	100		
182	95.3	76.6	115.0
145	30.6	25.4	38.2



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#88
Hexachlorobutadiene
Concen: 11.124 ug/l
RT: 14.010 min Scan# 3956
Delta R.T. -0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:225 Resp: 116891

Ion	Ratio	Lower	Upper
225	100		
223	62.3	49.5	74.3
227	64.3	51.0	76.4





#89
Naphthalene
Concen: 8.979 ug/l
RT: 14.077 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:128 Resp: 27352

Ion	Ratio	Lower	Upper
128	100		
127	13.2	10.6	16.0
129	10.6	8.6	13.0



Instrument :
MSVOA_U
ClientSampleId :
VSTDCCC010

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#90
1,2,3-Trichlorobenzene
Concen: 11.837 ug/l
RT: 14.319 min Scan# 4052
Delta R.T. -0.003 min
Lab File: VU063249.D
Acq: 12 Feb 2025 10:48

Tgt Ion:180 Resp: 170303

Ion	Ratio	Lower	Upper
180	100		
182	97.1	78.2	117.2
145	31.7	26.1	39.1



Sample Results: **VU063249.D**

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063249.D
Acq On : 12 Feb 2025 10:48
Operator : MD/SY
Sample : VSTDCCC010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC010

Quant Time: Feb 13 03:06:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 200%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	i	Fluorobenzene	1.000	1.000	0.0	94	0.00
2	T	Dichlorodifluoromethane	0.325	0.315	3.1	97	0.00
3	t	Chloromethane	0.374	0.351	6.1	93	0.00
4	Rt	Vinyl Chloride	0.370	0.363	1.9	93	0.00
5	T	Bromomethane	0.181	0.201	-11.0	99	0.00
6	T	Chloroethane	0.233	0.219	6.0	93	0.00
7	T	Trichlorofluoromethane	0.439	0.437	0.5	97	0.00
8		1,1,2-Trichloro-1,2,2-trifl	0.249	0.257	-3.2	105	0.00
9	Rt	1,1-Dichloroethene	0.254	0.260	-2.4	98	0.00
10	t	Iodomethane	0.399	0.417	-4.5	97	0.00
11	t	Allyl Chloride	0.364	0.364	0.0	94	0.00
12	t	Acrylonitrile	0.059	0.060	-1.7	87	0.00
13	T	Acetone	0.045	0.040	11.1	80	0.00
14	T	Carbon Disulfide	0.887	0.865	2.5	94	0.00
15	RT	Methylene Chloride	0.313	0.307	1.9	95	0.00
16	RT	trans-1,2-Dichloroethene	0.290	0.296	-2.1	97	0.00
17	t	1,1-Dichloroethane	0.546	0.551	-0.9	96	0.00
18	T	2-Butanone	0.072	0.070	2.8	83	0.00
19		Cyclohexane	0.439	0.473	-7.7	98	0.00
20		Methylcyclohexane	0.435	0.494	-13.6	104	0.00
21	T	2,2-Dichloropropane	0.426	0.438	-2.8	98	0.00
22	RT	cis-1,2-Dichloroethene	0.313	0.322	-2.9	96	0.00
23	t	Diethyl Ether	0.218	0.210	3.7	91	0.00
24	t	tert-Butyl Alcohol	0.026	0.019	26.9	75	0.02
25	t	Methyl tert-Butyl Ether	0.634	0.653	-3.0	91	0.00
26	t	Bromochloromethane	0.137	0.138	-0.7	95	0.00
27	t	Chloroform	0.551	0.553	-0.4	95	0.00
28	RT	1,1,1-Trichloroethane	0.446	0.454	-1.8	97	0.00
29	T	1,1-Dichloropropene	0.400	0.428	-7.0	99	0.00
30	RT	Carbon Tetrachloride	0.383	0.394	-2.9	98	0.00
31	t	Isopropyl Ether	0.779	0.810	-4.0	93	0.00
32		Ethyl-t-butyl ether	0.709	0.746	-5.2	93	0.00
33		Tert-Amyl methyl ether	0.619	0.671	-8.4	92	0.00
34	t	Propionitrile	0.022	0.022	0.0	85	0.00
35	RT	Benzene	1.229	1.264	-2.8	96	0.00
36	RT	1,2-Dichloroethane	0.355	0.352	0.8	93	0.00
37	RT	Trichloroethene	0.292	0.299	-2.4	96	0.00
38	Rt	1,2-Dichloropropane	0.322	0.330	-2.5	95	0.00
39	t	Methacrylonitrile	0.080	0.088	-10.0	85	0.00
40	t	Methyl acrylate	0.148	0.153	-3.4	87	0.00
41	t	Tetrahydrofuran	0.047	0.047	0.0	86	0.00
42	t	1-Chlorobutane	0.547	0.571	-4.4	95	0.00
43	T	Dibromomethane	0.163	0.163	0.0	94	0.00
44	T	Bromodichloromethane	0.379	0.391	-3.2	94	0.00
45	T	4-Methyl-2-Pentanone	0.171	0.180	-5.3	86	0.00
46	t	t-1,4-Dichloro-2-butene	0.080	0.088	-10.0	98	0.00
47	t	Methyl methacrylate	0.137	0.151	-10.2	88	0.00
48	t	Ethyl methacrylate	0.258	0.296	-14.7	88	0.00
49	Rt	Toluene	0.707	0.772	-9.2	97	0.00

Sample Results: **VU063249.D**

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063249.D
Acq On : 12 Feb 2025 10:48
Operator : MD/SY
Sample : VSTDCCC010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC010

Quant Time: Feb 13 03:06:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 200%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
50	T	t-1,3-Dichloropropene	0.347	0.374	-7.8	92	0.00
51	T	cis-1,3-Dichloropropene	0.429	0.448	-4.4	92	0.00
52	RT	1,1,2-Trichloroethane	0.220	0.230	-4.5	95	0.00
53	t	1,3-Dichloropropane	0.390	0.401	-2.8	93	0.00
54	t	2-Hexanone	0.117	0.123	-5.1	85	0.00
55	t	Dibromochloromethane	0.253	0.262	-3.6	92	0.00
56	T	1,2-Dibromoethane	0.206	0.210	-1.9	90	0.00
57	S	4-Bromofluorobenzene	0.330	0.438	-32.7#	115	0.00
58	RT	Tetrachloroethene	0.241	0.261	-8.3	105	0.00
59	Rt	Chlorobenzene	0.746	0.783	-5.0	96	0.00
60	T	1,1,1,2-Tetrachloroethane	0.268	0.281	-4.9	96	0.00
61	t	Pentachloroethane	0.239	0.236	1.3	92	0.00
62	t	Hexachloroethane	0.212	0.221	-4.2	94	0.00
63	Rt	Ethyl Benzene	1.286	1.442	-12.1	96	0.00
64	RT	m/p-Xylenes	0.480	0.553	-15.2	97	0.00
65	RT	o-Xylene	0.470	0.526	-11.9	96	0.00
66	RT	Styrene	0.748	0.871	-16.4	96	0.00
67	t	Bromoform	0.143	0.150	-4.9	91	0.00
68	S	1,2-Dichlorobenzene-d4	0.343	0.341	0.6	93	0.00
69	T	Isopropylbenzene	1.106	1.255	-13.5	97	0.00
70	T	1,1,2,2-Tetrachloroethane	0.296	0.293	1.0	89	0.00
71	T	1,2,3-Trichloropropane	0.222	0.215	3.2	99	0.00
72	t	Bromobenzene	0.298	0.319	-7.0	95	0.00
73	t	n-propylbenzene	0.317	0.375	-18.3	99	0.00
74	t	2-Chlorotoluene	0.292	0.325	-11.3	97	0.00
75	t	1,3,5-Trimethylbenzene	1.024	1.181	-15.3	97	0.00
76	t	4-Chlorotoluene	0.299	0.340	-13.7	99	0.00
77	t	tert-Butylbenzene	1.036	1.142	-10.2	96	0.00
78	t	1,2,4-Trimethylbenzene	1.016	1.181	-16.2	96	0.00
79	t	sec-Butylbenzene	1.319	1.520	-15.2	99	0.00
80		Nitrobenzene	0.007	0.007	0.0	79	0.00
81	t	p-Isopropyltoluene	1.041	1.231	-18.3	99	0.00
82	t	1,3-Dichlorobenzene	0.578	0.618	-6.9	97	0.00
83	Rt	1,4-Dichlorobenzene	0.565	0.617	-9.2	96	0.00
84	t	n-Butylbenzene	0.933	1.107	-18.6	100	0.00
85	Rt	1,2-Dichlorobenzene	0.555	0.584	-5.2	95	0.00
86	t	1,2-Dibromo-3-Chloropropane	0.042	0.044	-4.8	85	0.00
87	Rt	1,2,4-Trichlorobenzene	0.271	0.329	-21.4	100	0.00
88	t	Hexachlorobutadiene	0.194	0.215	-10.8	111	0.00
89	t	Naphthalene	0.436	0.504	-15.6	85	0.00
90	t	1,2,3-Trichlorobenzene	0.265	0.314	-18.5	97	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Sample Results: **VU063249.D**

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063249.D
Acq On : 12 Feb 2025 10:48
Operator : MD/SY
Sample : VSTDCCC010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC010

Quant Time: Feb 13 03:06:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 200%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1	i	Fluorobenzene	1.000	1.000	0.0	94	0.00
2	T	Dichlorodifluoromethane	10.000	9.709	2.9	97	0.00
3	t	Chloromethane	10.000	9.374	6.3	93	0.00
4	Rt	Vinyl Chloride	10.000	9.798	2.0	93	0.00
5	T	Bromomethane	10.000	11.104	-11.0	99	0.00
6	T	Chloroethane	10.000	9.387	6.1	93	0.00
7	T	Trichlorofluoromethane	10.000	9.961	0.4	97	0.00
8		1,1,2-Trichloro-1,2,2-trifl	10.000	10.336	-3.4	105	0.00
9	Rt	1,1-Dichloroethene	10.000	10.249	-2.5	98	0.00
10	t	Iodomethane	10.000	10.453	-4.5	97	0.00
11	t	Allyl Chloride	10.000	9.992	0.1	94	0.00
12	t	Acrylonitrile	20.000	20.569	-2.8	87	0.00
13	T	Acetone	50.000	44.953	10.1	80	0.00
14	T	Carbon Disulfide	10.000	9.747	2.5	94	0.00
15	RT	Methylene Chloride	10.000	9.802	2.0	95	0.00
16	RT	trans-1,2-Dichloroethene	10.000	10.238	-2.4	97	0.00
17	t	1,1-Dichloroethane	10.000	10.104	-1.0	96	0.00
18	T	2-Butanone	50.000	48.752	2.5	83	0.00
19		Cyclohexane	10.000	10.790	-7.9	98	0.00
20		Methylcyclohexane	10.000	11.362	-13.6	104	0.00
21	T	2,2-Dichloropropane	10.000	10.276	-2.8	98	0.00
22	RT	cis-1,2-Dichloroethene	10.000	10.291	-2.9	96	0.00
23	t	Diethyl Ether	10.000	9.655	3.5	91	0.00
24	t	tert-Butyl Alcohol	100.000	83.500	16.5	75	0.02
25	t	Methyl tert-Butyl Ether	10.000	10.306	-3.1	91	0.00
26	t	Bromochloromethane	10.000	10.068	-0.7	95	0.00
27	t	Chloroform	10.000	10.045	-0.4	95	0.00
28	RT	1,1,1-Trichloroethane	10.000	10.177	-1.8	97	0.00
29	T	1,1-Dichloropropene	10.000	10.705	-7.1	99	0.00
30	RT	Carbon Tetrachloride	10.000	10.291	-2.9	98	0.00
31	t	Isopropyl Ether	10.000	10.392	-3.9	93	0.00
32		Ethyl-t-butyl ether	10.000	10.522	-5.2	93	0.00
33		Tert-Amyl methyl ether	10.000	10.838	-8.4	92	0.00
34	t	Propionitrile	50.000	51.745	-3.5	85	0.00
35	RT	Benzene	10.000	10.283	-2.8	96	0.00
36	RT	1,2-Dichloroethane	10.000	9.936	0.6	93	0.00
37	RT	Trichloroethene	10.000	10.219	-2.2	96	0.00
38	Rt	1,2-Dichloropropane	10.000	10.274	-2.7	95	0.00
39	t	Methacrylonitrile	10.000	10.967	-9.7	85	0.00
40	t	Methyl acrylate	10.000	10.346	-3.5	87	0.00
41	t	Tetrahydrofuran	20.000	20.220	-1.1	86	0.00
42	t	1-Chlorobutane	10.000	10.439	-4.4	95	0.00
43	T	Dibromomethane	10.000	10.017	-0.2	94	0.00
44	T	Bromodichloromethane	10.000	10.327	-3.3	94	0.00
45	T	4-Methyl-2-Pentanone	50.000	52.711	-5.4	86	0.00
46	t	t-1,4-Dichloro-2-butene	20.000	21.967	-9.8	98	0.00
47	t	Methyl methacrylate	20.000	22.078	-10.4	88	0.00
48	t	Ethyl methacrylate	10.000	11.484	-14.8	88	0.00
49	Rt	Toluene	10.000	10.932	-9.3	97	0.00

Sample Results: **VU063249.D**

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063249.D
Acq On : 12 Feb 2025 10:48
Operator : MD/SY
Sample : VSTDCCC010
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_U
LabSampleId :
VSTDCCC010

Quant Time: Feb 13 03:06:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 30% Max. Rel. Area : 200%

		Compound	Amount	Calc.	%Dev	Area%	Dev(min)
50	T	t-1,3-Dichloropropene	10.000	10.778	-7.8	92	0.00
51	T	cis-1,3-Dichloropropene	10.000	10.447	-4.5	92	0.00
52	RT	1,1,2-Trichloroethane	10.000	10.486	-4.9	95	0.00
53	t	1,3-Dichloropropane	10.000	10.294	-2.9	93	0.00
54	t	2-Hexanone	50.000	52.760	-5.5	85	0.00
55	t	Dibromochloromethane	10.000	10.388	-3.9	92	0.00
56	T	1,2-Dibromoethane	10.000	10.196	-2.0	90	0.00
57	S	4-Bromofluorobenzene	1.000	1.328	-32.8#	115	0.00
58	RT	Tetrachloroethene	10.000	10.837	-8.4	105	0.00
59	Rt	Chlorobenzene	10.000	10.494	-4.9	96	0.00
60	T	1,1,1,2-Tetrachloroethane	10.000	10.471	-4.7	96	0.00
61	t	Pentachloroethane	10.000	9.856	1.4	92	0.00
62	t	Hexachloroethane	10.000	10.453	-4.5	94	0.00
63	Rt	Ethyl Benzene	10.000	11.211	-12.1	96	0.00
64	RT	m/p-Xylenes	20.000	23.011	-15.1	97	0.00
65	RT	o-Xylene	10.000	11.188	-11.9	96	0.00
66	RT	Styrene	10.000	11.643	-16.4	96	0.00
67	t	Bromoform	10.000	10.490	-4.9	91	0.00
68	S	1,2-Dichlorobenzene-d4	1.000	0.995	0.5	93	0.00
69	T	Isopropylbenzene	10.000	11.354	-13.5	97	0.00
70	T	1,1,2,2-Tetrachloroethane	10.000	9.913	0.9	89	0.00
71	T	1,2,3-Trichloropropane	10.000	9.656	3.4	99	0.00
72	t	Bromobenzene	10.000	10.707	-7.1	95	0.00
73	t	n-propylbenzene	10.000	11.848	-18.5	99	0.00
74	t	2-Chlorotoluene	10.000	11.143	-11.4	97	0.00
75	t	1,3,5-Trimethylbenzene	10.000	11.526	-15.3	97	0.00
76	t	4-Chlorotoluene	10.000	11.371	-13.7	99	0.00
77	t	tert-Butylbenzene	10.000	11.021	-10.2	96	0.00
78	t	1,2,4-Trimethylbenzene	10.000	11.619	-16.2	96	0.00
79	t	sec-Butylbenzene	10.000	11.524	-15.2	99	0.00
80		Nitrobenzene	50.000	43.211	13.6	79	0.00
81	t	p-Isopropyltoluene	10.000	11.824	-18.2	99	0.00
82	t	1,3-Dichlorobenzene	10.000	10.689	-6.9	97	0.00
83	Rt	1,4-Dichlorobenzene	10.000	10.926	-9.3	96	0.00
84	t	n-Butylbenzene	10.000	11.862	-18.6	100	0.00
85	Rt	1,2-Dichlorobenzene	10.000	10.510	-5.1	95	0.00
86	t	1,2-Dibromo-3-Chloropropane	10.000	10.548	-5.5	85	0.00
87	Rt	1,2,4-Trichlorobenzene	10.000	12.143	-21.4	100	0.00
88	t	Hexachlorobutadiene	10.000	11.124	-11.2	111	0.00
89	t	Naphthalene	10.000	8.979	10.2	85	0.00
90	t	1,2,3-Trichlorobenzene	10.000	11.837	-18.4	97	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0

QC SAMPLE DATA

Sample Results: **VU063218.D**

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021025\
Data File : VU063218.D
Acq On : 10 Feb 2025 09:06
Operator : MD/SY
Sample : BFB
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Title : METHOD 524.2 VOLATILES DRINKING WATER
Last Update : Tue Feb 11 08:42:19 2025



AutoFind: Scans 2902, 2903, 2904; Background Corrected with Scan 2890

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	7489	PASS
75	95	30	60	52.7	20232	PASS
95	95	100	100	100.0	38421	PASS
96	95	5	9	6.5	2505	PASS
173	174	0.00	2	0.5	139	PASS
174	95	50	100	79.1	30381	PASS
175	174	5	9	7.3	2224	PASS
176	174	95	101	98.1	29792	PASS
177	176	5	9	6.2	1844	PASS

Sample Results: **VU063248.D**

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063248.D
Acq On : 12 Feb 2025 10:13
Operator : MD/SY
Sample : BFB
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
BFB

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Title : METHOD 524.2 VOLATILES DRINKING WATER
Last Update : Tue Feb 11 08:42:19 2025





AutoFind: Scans 2902, 2903, 2904; Background Corrected with Scan 2889

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.3	24832	PASS
75	95	30	60	51.4	66152	PASS
95	95	100	100	100.0	128739	PASS
96	95	5	9	6.7	8666	PASS
173	174	0.00	2	1.2	1241	PASS
174	95	50	100	78.7	101285	PASS
175	174	5	9	7.5	7621	PASS
176	174	95	101	96.2	97483	PASS
177	176	5	9	6.6	6463	PASS

Raw Data: **VU063250.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	
Project:	NJ Drinking Water PT	Date Received:	
Client Sample ID:	VU0212WBL01	SDG No.:	Q1172
Lab Sample ID:	VU0212WBL01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063250.D	1		02/12/25 12:17	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.14	U	0.14	0.50	ug/L
74-87-3	Chloromethane	0.13	U	0.13	0.50	ug/L
75-01-4	Vinyl Chloride	0.13	U	0.13	0.50	ug/L
74-83-9	Bromomethane	0.18	U	0.18	0.50	ug/L
75-00-3	Chloroethane	0.14	U	0.14	0.50	ug/L
109-99-9	Tetrahydrofuran	0.44	U	0.44	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.21	U	0.21	0.50	ug/L
76-13-1	1,1,2-Trichloro-1,2,2-trifluoroethane	0.14	U	0.14	0.50	ug/L
75-65-0	tert-Butyl Alcohol	8.60	U	8.60	10.0	ug/L
60-29-7	Diethyl Ether	0.13	U	0.13	0.50	ug/L
75-35-4	1,1-Dichloroethene	0.12	U	0.12	0.50	ug/L
107-13-1	Acrylonitrile	0.44	U	0.44	1.00	ug/L
67-64-1	Acetone	1.10	U	1.10	2.50	ug/L
75-15-0	Carbon Disulfide	0.13	U	0.13	0.50	ug/L
1634-04-4	Methyl tert-Butyl Ether	0.12	U	0.12	0.50	ug/L
96-33-3	Methyl acrylate	0.28	U	0.28	0.50	ug/L
75-09-2	Methylene Chloride	0.47	U	0.47	0.50	ug/L
156-60-5	trans-1,2-Dichloroethene	0.14	U	0.14	0.50	ug/L
75-34-3	1,1-Dichloroethane	0.13	U	0.13	0.50	ug/L
110-82-7	Cyclohexane	0.14	U	0.14	0.50	ug/L
78-93-3	2-Butanone	0.68	U	0.68	2.50	ug/L
56-23-5	Carbon Tetrachloride	0.14	U	0.14	0.50	ug/L
594-20-7	2,2-Dichloropropane	0.14	U	0.14	0.50	ug/L
156-59-2	cis-1,2-Dichloroethene	0.13	U	0.13	0.50	ug/L
74-97-5	Bromochloromethane	0.16	U	0.16	0.50	ug/L
67-66-3	Chloroform	0.13	U	0.13	0.50	ug/L
71-55-6	1,1,1-Trichloroethane	0.12	U	0.12	0.50	ug/L
108-87-2	Methylcyclohexane	0.12	U	0.12	0.50	ug/L
563-58-6	1,1-Dichloropropene	0.11	U	0.11	0.50	ug/L
107-12-0	Propionitrile	1.00	U	1.00	2.50	ug/L

Raw Data: **VU063250.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	
Project:	NJ Drinking Water PT	Date Received:	
Client Sample ID:	VU0212WBL01	SDG No.:	Q1172
Lab Sample ID:	VU0212WBL01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063250.D	1		02/12/25 12:17	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
71-43-2	Benzene	0.11	U	0.11	0.50	ug/L
107-06-2	1,2-Dichloroethane	0.16	U	0.16	0.50	ug/L
79-01-6	Trichloroethene	0.13	U	0.13	0.50	ug/L
78-87-5	1,2-Dichloropropane	0.13	U	0.13	0.50	ug/L
109-69-3	1-Chlorobutane	0.12	U	0.12	0.50	ug/L
74-95-3	Dibromomethane	0.14	U	0.14	0.50	ug/L
75-27-4	Bromodichloromethane	0.12	U	0.12	0.50	ug/L
108-10-1	4-Methyl-2-Pentanone	0.60	U	0.60	2.50	ug/L
108-88-3	Toluene	0.11	U	0.11	0.50	ug/L
10061-02-6	t-1,3-Dichloropropene	0.11	U	0.11	0.50	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.11	U	0.11	0.50	ug/L
79-00-5	1,1,2-Trichloroethane	0.13	U	0.13	0.50	ug/L
142-28-9	1,3-Dichloropropane	0.13	U	0.13	0.50	ug/L
591-78-6	2-Hexanone	0.57	U	0.57	2.50	ug/L
124-48-1	Dibromochloromethane	0.13	U	0.13	0.50	ug/L
106-93-4	1,2-Dibromoethane	0.13	U	0.13	0.50	ug/L
127-18-4	Tetrachloroethene	0.14	U	0.14	0.50	ug/L
108-90-7	Chlorobenzene	0.11	U	0.11	0.50	ug/L
630-20-6	1,1,1,2-Tetrachloroethane	0.13	U	0.13	0.50	ug/L
67-72-1	Hexachloroethane	0.12	U	0.12	0.50	ug/L
100-41-4	Ethyl Benzene	0.12	U	0.12	0.50	ug/L
179601-23-1	m/p-Xylenes	0.23	U	0.23	1.00	ug/L
1330-20-7	Total Xylenes	0.35	U	0.35	1.50	ug/L
95-47-6	o-Xylene	0.12	U	0.12	0.50	ug/L
100-42-5	Styrene	0.13	U	0.13	0.50	ug/L
75-25-2	Bromoform	0.14	U	0.14	0.50	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	0.50	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.12	U	0.12	0.50	ug/L
96-18-4	1,2,3-Trichloropropane	0.21	U	0.21	0.50	ug/L
108-86-1	Bromobenzene	0.13	U	0.13	0.50	ug/L
103-65-1	n-propylbenzene	0.16	U	0.16	0.50	ug/L

Raw Data: **VU063250.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	
Project:	NJ Drinking Water PT	Date Received:	
Client Sample ID:	VU0212WBL01	SDG No.:	Q1172
Lab Sample ID:	VU0212WBL01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063250.D	1		02/12/25 12:17	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
95-49-8	2-Chlorotoluene	0.14	U	0.14	0.50	ug/L
108-67-8	1,3,5-Trimethylbenzene	0.13	U	0.13	0.50	ug/L
106-43-4	4-Chlorotoluene	0.14	U	0.14	0.50	ug/L
98-06-6	tert-Butylbenzene	0.11	U	0.11	0.50	ug/L
95-63-6	1,2,4-Trimethylbenzene	0.13	U	0.13	0.50	ug/L
135-98-8	sec-Butylbenzene	0.13	U	0.13	0.50	ug/L
99-87-6	p-Isopropyltoluene	0.16	U	0.16	0.50	ug/L
541-73-1	1,3-Dichlorobenzene	0.13	U	0.13	0.50	ug/L
106-46-7	1,4-Dichlorobenzene	0.14	U	0.14	0.50	ug/L
104-51-8	n-Butylbenzene	0.28	U	0.28	0.50	ug/L
95-50-1	1,2-Dichlorobenzene	0.14	U	0.14	0.50	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.23	U	0.23	0.50	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.21	U	0.21	0.50	ug/L
87-68-3	Hexachlorobutadiene	0.14	U	0.14	0.50	ug/L
91-20-3	Naphthalene	0.31	U	0.31	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.25	U	0.25	0.50	ug/L
74-88-4	Iodomethane	0.16	U	0.16	1.00	ug/L
107-05-1	Allyl Chloride	0.11	U	0.11	0.50	ug/L
126-98-7	Methacrylonitrile	0.19	U	0.19	0.50	ug/L
108-20-3	Isopropyl Ether	0.12	U	0.12	0.50	ug/L
80-62-6	Methyl methacrylate	0.24	U	0.24	1.00	ug/L
SURROGATES						
2199-69-1	1,2-Dichlorobenzene-d4	0.78		70 - 130	78%	SPK: 1
460-00-4	4-Bromofluorobenzene	0.82		70 - 130	82%	SPK: 1
INTERNAL STANDARDS						
462-06-6	Fluorobenzene	48000	6.107			
	Tert-Amyl methyl ether					

Raw Data: **VU063250.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	
Project:	NJ Drinking Water PT	Date Received:	
Client Sample ID:	VU0212WBL01	SDG No.:	Q1172
Lab Sample ID:	VU0212WBL01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063250.D	1		02/12/25 12:17	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063250.D
Acq On : 12 Feb 2025 12:17
Operator : MD/SY
Sample : VU0212WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBL01

Quant Time: Feb 13 03:10:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Fluorobenzene	6.107	96	47954	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.627	95	12922	0.817	ug/l	0.00
Spiked Amount 1.000			Recovery =	82.000%		
68) 1,2-Dichlorobenzene-d4	12.190	152	12781	0.777	ug/l	0.00
Spiked Amount 1.000			Recovery =	78.000%		
Target Compounds						Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063250.D
Acq On : 12 Feb 2025 12:17
Operator : MD/SY
Sample : VU0212WBL01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBL01

Quant Time: Feb 13 03:10:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.107 min Scan# 14
Delta R.T. 0.004 min
Lab File: VU063250.D
Acq: 12 Feb 2025 12:17

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBL01

Tgt Ion: 96 Resp: 47954

Ion	Ratio	Lower	Upper
96	100		
70	20.6	15.6	23.4
95	9.6	7.3	10.9
97	6.5	0.0	0.0#





#57
4-Bromofluorobenzene
Concen: 0.817 ug/l
RT: 10.627 min Scan# 2904
Delta R.T. 0.003 min
Lab File: VU063250.D
Acq: 12 Feb 2025 12:17

Tgt Ion: 95 Resp: 12922

Ion	Ratio	Lower	Upper
95	100		
174	82.6	58.6	88.0
176	72.7	58.2	87.4





#68
1,2-Dichlorobenzene-d4
Concen: 0.777 ug/l
RT: 12.190 min Scan# 3
Delta R.T. 0.006 min
Lab File: VU063250.D
Acq: 12 Feb 2025 12:17

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBL01

Tgt Ion:152 Resp: 12781

Ion	Ratio	Lower	Upper
152	100		
115	60.8	0.0	275.2
150	158.5	0.0	658.4



Raw Data: **VU063251.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	
Project:	NJ Drinking Water PT	Date Received:	
Client Sample ID:	VU0212WBS01	SDG No.:	Q1172
Lab Sample ID:	VU0212WBS01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063251.D	1		02/12/25 12:56	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	2.00		0.14	0.50	ug/L
74-87-3	Chloromethane	2.00		0.13	0.50	ug/L
75-01-4	Vinyl Chloride	2.00		0.13	0.50	ug/L
74-83-9	Bromomethane	2.20		0.18	0.50	ug/L
75-00-3	Chloroethane	2.00		0.14	0.50	ug/L
109-99-9	Tetrahydrofuran	3.80		0.44	1.00	ug/L
75-69-4	Trichlorofluoromethane	2.00		0.21	0.50	ug/L
76-13-1	1,1,2-Trichloro-1,2,2-trifluoroethane	2.00		0.14	0.50	ug/L
75-65-0	tert-Butyl Alcohol	21.2		8.60	10.0	ug/L
60-29-7	Diethyl Ether	1.90		0.13	0.50	ug/L
75-35-4	1,1-Dichloroethene	2.00		0.12	0.50	ug/L
107-13-1	Acrylonitrile	4.10		0.44	1.00	ug/L
67-64-1	Acetone	10.3		1.10	2.50	ug/L
75-15-0	Carbon Disulfide	1.90		0.13	0.50	ug/L
1634-04-4	Methyl tert-Butyl Ether	1.90		0.12	0.50	ug/L
96-33-3	Methyl acrylate	2.00		0.28	0.50	ug/L
75-09-2	Methylene Chloride	2.00		0.47	0.50	ug/L
156-60-5	trans-1,2-Dichloroethene	2.00		0.14	0.50	ug/L
75-34-3	1,1-Dichloroethane	2.00		0.13	0.50	ug/L
110-82-7	Cyclohexane	1.90		0.14	0.50	ug/L
78-93-3	2-Butanone	8.90		0.68	2.50	ug/L
56-23-5	Carbon Tetrachloride	1.90		0.14	0.50	ug/L
594-20-7	2,2-Dichloropropane	2.00		0.14	0.50	ug/L
156-59-2	cis-1,2-Dichloroethene	2.00		0.13	0.50	ug/L
74-97-5	Bromochloromethane	2.00		0.16	0.50	ug/L
67-66-3	Chloroform	2.00		0.13	0.50	ug/L
71-55-6	1,1,1-Trichloroethane	2.00		0.12	0.50	ug/L
108-87-2	Methylcyclohexane	2.00		0.12	0.50	ug/L
563-58-6	1,1-Dichloropropene	2.00		0.11	0.50	ug/L
107-12-0	Propionitrile	10.4		1.00	2.50	ug/L

Raw Data: **VU063251.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	
Project:	NJ Drinking Water PT	Date Received:	
Client Sample ID:	VU0212WBS01	SDG No.:	Q1172
Lab Sample ID:	VU0212WBS01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063251.D	1		02/12/25 12:56	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
71-43-2	Benzene	2.00		0.11	0.50	ug/L
107-06-2	1,2-Dichloroethane	2.00		0.16	0.50	ug/L
79-01-6	Trichloroethene	2.00		0.13	0.50	ug/L
78-87-5	1,2-Dichloropropane	2.00		0.13	0.50	ug/L
109-69-3	1-Chlorobutane	1.90		0.12	0.50	ug/L
74-95-3	Dibromomethane	2.00		0.14	0.50	ug/L
75-27-4	Bromodichloromethane	2.00		0.12	0.50	ug/L
108-10-1	4-Methyl-2-Pentanone	9.50		0.60	2.50	ug/L
108-88-3	Toluene	2.00		0.11	0.50	ug/L
10061-02-6	t-1,3-Dichloropropene	2.00		0.11	0.50	ug/L
10061-01-5	cis-1,3-Dichloropropene	2.00		0.11	0.50	ug/L
79-00-5	1,1,2-Trichloroethane	2.00		0.13	0.50	ug/L
142-28-9	1,3-Dichloropropane	2.00		0.13	0.50	ug/L
591-78-6	2-Hexanone	9.00		0.57	2.50	ug/L
124-48-1	Dibromochloromethane	2.00		0.13	0.50	ug/L
106-93-4	1,2-Dibromoethane	1.90		0.13	0.50	ug/L
127-18-4	Tetrachloroethene	2.10		0.14	0.50	ug/L
108-90-7	Chlorobenzene	2.00		0.11	0.50	ug/L
630-20-6	1,1,1,2-Tetrachloroethane	2.00		0.13	0.50	ug/L
67-72-1	Hexachloroethane	1.90		0.12	0.50	ug/L
100-41-4	Ethyl Benzene	1.90		0.12	0.50	ug/L
179601-23-1	m/p-Xylenes	3.90		0.23	1.00	ug/L
1330-20-7	Total Xylenes	5.90		0.35	1.50	ug/L
95-47-6	o-Xylene	2.00		0.12	0.50	ug/L
100-42-5	Styrene	1.90		0.13	0.50	ug/L
75-25-2	Bromoform	1.90		0.14	0.50	ug/L
98-82-8	Isopropylbenzene	2.00		0.13	0.50	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	2.00		0.12	0.50	ug/L
96-18-4	1,2,3-Trichloropropane	2.00		0.21	0.50	ug/L
108-86-1	Bromobenzene	2.00		0.13	0.50	ug/L
103-65-1	n-propylbenzene	1.90		0.16	0.50	ug/L

Raw Data: **VU063251.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	
Project:	NJ Drinking Water PT	Date Received:	
Client Sample ID:	VU0212WBS01	SDG No.:	Q1172
Lab Sample ID:	VU0212WBS01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063251.D	1		02/12/25 12:56	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
95-49-8	2-Chlorotoluene	2.00		0.14	0.50	ug/L
108-67-8	1,3,5-Trimethylbenzene	2.00		0.13	0.50	ug/L
106-43-4	4-Chlorotoluene	2.00		0.14	0.50	ug/L
98-06-6	tert-Butylbenzene	2.00		0.11	0.50	ug/L
95-63-6	1,2,4-Trimethylbenzene	1.90		0.13	0.50	ug/L
135-98-8	sec-Butylbenzene	2.00		0.13	0.50	ug/L
99-87-6	p-Isopropyltoluene	2.00		0.16	0.50	ug/L
541-73-1	1,3-Dichlorobenzene	2.00		0.13	0.50	ug/L
106-46-7	1,4-Dichlorobenzene	2.00		0.14	0.50	ug/L
104-51-8	n-Butylbenzene	1.90		0.28	0.50	ug/L
95-50-1	1,2-Dichlorobenzene	2.00		0.14	0.50	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	1.70		0.23	0.50	ug/L
120-82-1	1,2,4-Trichlorobenzene	1.80		0.21	0.50	ug/L
87-68-3	Hexachlorobutadiene	2.10		0.14	0.50	ug/L
91-20-3	Naphthalene	1.70		0.31	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	1.60		0.25	0.50	ug/L
74-88-4	Iodomethane	2.00		0.16	1.00	ug/L
107-05-1	Allyl Chloride	2.00		0.11	0.50	ug/L
126-98-7	Methacrylonitrile	2.00		0.19	0.50	ug/L
108-20-3	Isopropyl Ether	1.90		0.12	0.50	ug/L
80-62-6	Methyl methacrylate	3.90		0.24	1.00	ug/L
SURROGATES						
2199-69-1	1,2-Dichlorobenzene-d4	0.98		70 - 130	98%	SPK: 1
460-00-4	4-Bromofluorobenzene	1.00		70 - 130	103%	SPK: 1
INTERNAL STANDARDS						
462-06-6	Fluorobenzene	51100	6.103			
	Tert-Amyl methyl ether					

Raw Data: **VU063251.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	
Project:	NJ Drinking Water PT	Date Received:	
Client Sample ID:	VU0212WBS01	SDG No.:	Q1172
Lab Sample ID:	VU0212WBS01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063251.D	1		02/12/25 12:56	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063251.D
Acq On : 12 Feb 2025 12:56
Operator : MD/SY
Sample : VU0212WBS01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Feb 13 03:10:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Fluorobenzene	6.103	96	51131	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.627	95	17337	1.028	ug/l	0.00
Spiked Amount 1.000			Recovery = 103.000%			
68) 1,2-Dichlorobenzene-d4	12.187	152	17238	0.983	ug/l	0.00
Spiked Amount 1.000			Recovery = 98.000%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.377	85	32874	1.979	ug/l	95
3) Chloromethane	1.515	50	37771	1.974	ug/l	97
4) Vinyl Chloride	1.599	62	36923	1.951	ug/l	98
5) Bromomethane	1.824	94	20801	2.249	ug/l	98
6) Chloroethane	1.911	64	23326	1.956	ug/l	97
7) Trichlorofluoromethane	2.123	101	44667	1.991	ug/l	95
8) 1,1,2-Trichloro-1,2,2-...	2.563	101	25640	2.014	ug/l	98
9) 1,1-Dichloroethene	2.567	96	25934	1.999	ug/l	98
10) Iodomethane	2.708	142	41607	2.040	ug/l	100
11) Allyl Chloride	2.907	41	36781	1.974	ug/l	99
12) Acrylonitrile	3.316	53	12135	4.056	ug/l	93
13) Acetone	2.615	43	23779	10.336	ug/l	95
14) Carbon Disulfide	2.779	76	85777	1.891	ug/l	97
15) Methylene Chloride	3.030	84	31815	1.985	ug/l	95
16) trans-1,2-Dichloroethene	3.338	96	29107	1.966	ug/l	94
17) 1,1-Dichloroethane	3.850	63	55417	1.986	ug/l	99
18) 2-Butanone	4.702	43	32440	8.860	ug/l	97
19) Cyclohexane	5.374	56	42779m	1.908	ug/l	
20) Methylcyclohexane	6.750	83	44086	1.983	ug/l	98
21) 2,2-Dichloropropane	4.647	77	43300	1.989	ug/l	99
22) cis-1,2-Dichloroethene	4.653	96	31577	1.974	ug/l	99
23) Diethyl Ether	2.364	59	21191	1.904	ug/l	99
24) tert-Butyl Alcohol	3.174	59	29904	21.171	ug/l #	93
25) Methyl tert-Butyl Ether	3.348	73	63104	1.948	ug/l	97
26) Bromochloromethane	4.959	128	14144	2.023	ug/l	98
27) Chloroform	5.075	83	57173	2.030	ug/l	99
28) 1,1,1-Trichloroethane	5.300	97	45409	1.990	ug/l	98
29) 1,1-Dichloropropene	5.515	75	40742	1.994	ug/l	99
30) Carbon Tetrachloride	5.509	117	37714	1.928	ug/l	99
31) Isopropyl Ether	3.975	45	76353	1.916	ug/l	96
32) Ethyl-t-butyl ether	4.483	59	68833	1.900	ug/l	100
33) Tert-Amyl methyl ether	5.927	73	61563	1.945	ug/l	99
34) Propionitrile	4.772	54	11521m	10.432	ug/l	
35) Benzene	5.763	78	125217	1.993	ug/l	99
36) 1,2-Dichloroethane	5.782	62	36496	2.013	ug/l	100
37) Trichloroethene	6.534	130	30224	2.023	ug/l	93
38) 1,2-Dichloropropane	6.779	63	32315	1.965	ug/l	98
39) Methacrylonitrile	4.972	41	8096	1.967	ug/l	93
40) Methyl acrylate	4.853	55	15111m	1.992	ug/l	
41) Tetrahydrofuran	5.049	42	9062	3.786	ug/l	95
42) 1-Chlorobutane	5.444	56	54007	1.932	ug/l	99
43) Dibromomethane	6.907	93	16622	1.996	ug/l	96
44) Bromodichloromethane	7.094	83	39273	2.026	ug/l	94

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063251.D
Acq On : 12 Feb 2025 12:56
Operator : MD/SY
Sample : VU0212WBS01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Feb 13 03:10:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
45) 4-Methyl-2-Pentanone	7.782	43	83073	9.516	ug/l	98
46) t-1,4-Dichloro-2-butene	10.820	75	16859m	4.108	ug/l	
47) Methyl methacrylate	6.956	69	27267	3.888	ug/l	98
48) Ethyl methacrylate	8.325	69	25396	1.928	ug/l	98
49) Toluene	7.959	92	71512	1.979	ug/l	99
50) t-1,3-Dichloropropene	8.203	75	34712	1.956	ug/l	99
51) cis-1,3-Dichloropropene	7.599	75	42816	1.953	ug/l	98
52) 1,1,2-Trichloroethane	8.390	97	22745	2.026	ug/l	98
53) 1,3-Dichloropropane	8.566	76	39631	1.988	ug/l	99
54) 2-Hexanone	8.679	43	53643	9.005	ug/l	99
55) Dibromochloromethane	8.798	129	25441	1.970	ug/l	97
56) 1,2-Dibromoethane	8.914	107	20372	1.935	ug/l	97
58) Tetrachloroethene	8.544	164	25526	2.073	ug/l	98
59) Chlorobenzene	9.438	112	75420	1.978	ug/l	99
60) 1,1,1,2-Tetrachloroethane	9.525	131	27395	1.999	ug/l	99
61) Pentachloroethane	11.415	117	23808	1.945	ug/l	93
62) Hexachloroethane	12.463	117	20151	1.861	ug/l	100
63) Ethyl Benzene	9.560	91	124557	1.894	ug/l	98
64) m/p-Xylenes	9.685	106	96092	3.912	ug/l	99
65) o-Xylene	10.090	106	48728	2.027	ug/l	95
66) Styrene	10.107	104	73539	1.922	ug/l	100
67) Bromoform	10.283	173	14014	1.912	ug/l	96
69) Isopropylbenzene	10.473	105	110230	1.950	ug/l	99
70) 1,1,2,2-Tetrachloroethane	10.772	83	29840	1.972	ug/l	99
71) 1,2,3-Trichloropropane	10.814	75	23230m	2.043	ug/l	
72) Bromobenzene	10.775	156	29771	1.955	ug/l	98
73) n-propylbenzene	10.897	120	31267	1.932	ug/l	99
74) 2-Chlorotoluene	10.978	126	30285	2.030	ug/l	98
75) 1,3,5-Trimethylbenzene	11.078	105	102401	1.955	ug/l	99
76) 4-Chlorotoluene	11.090	126	30015	1.962	ug/l	99
77) tert-Butylbenzene	11.409	119	103577	1.955	ug/l	99
78) 1,2,4-Trimethylbenzene	11.460	105	100810	1.940	ug/l	100
79) sec-Butylbenzene	11.634	105	133023	1.972	ug/l	100
80) Nitrobenzene	13.229	77	2781m	9.568	ug/l	
81) p-Isopropyltoluene	11.785	119	104011	1.954	ug/l	100
82) 1,3-Dichlorobenzene	11.737	146	57876	1.959	ug/l	99
83) 1,4-Dichlorobenzene	11.830	146	57047	1.974	ug/l	99
84) n-Butylbenzene	12.200	91	88658	1.858	ug/l	99
85) 1,2-Dichlorobenzene	12.203	146	56355	1.985	ug/l	99
86) 1,2-Dibromo-3-Chloropr...	12.991	75	3704	1.744	ug/l	91
87) 1,2,4-Trichlorobenzene	13.836	180	24294	1.755	ug/l	99
88) Hexachlorobutadiene	14.010	225	20391	2.061	ug/l	98
89) Naphthalene	14.084	128	31946	1.689	ug/l	98
90) 1,2,3-Trichlorobenzene	14.325	180	21647	1.598	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063251.D
Acq On : 12 Feb 2025 12:56
Operator : MD/SY
Sample : VU0212WBS01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Feb 13 03:10:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.103 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 96 Resp: 5113

Ion	Ratio	Lower	Upper
96	100		
70	20.1	15.6	23.4
95	9.2	7.3	10.9
97	6.9	0.0	0.0#



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#2
Dichlorodifluoromethane
Concen: 1.979 ug/l
RT: 1.377 min Scan# 27
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 85 Resp: 32874

Ion	Ratio	Lower	Upper
85	100		
87	34.8	16.0	48.0





#3
Chloromethane
Concen: 1.974 ug/l
RT: 1.515 min Scan# 7
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 50 Resp: 3777

Ion	Ratio	Lower	Upper
50	100		
52	34.1	25.8	38.8



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#4
Vinyl Chloride
Concen: 1.951 ug/l
RT: 1.599 min Scan# 96
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 62 Resp: 36923

Ion	Ratio	Lower	Upper
62	100		
64	32.7	25.4	38.0





#5
Bromomethane
Concen: 2.249 ug/l
RT: 1.824 min Scan# 1
Delta R.T. -0.019 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 94 Resp: 2080

Ion	Ratio	Lower	Upper
94	100		
96	92.4	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#6
Chloroethane
Concen: 1.956 ug/l
RT: 1.911 min Scan# 193
Delta R.T. -0.013 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 64 Resp: 23326

Ion	Ratio	Lower	Upper
64	100		
66	34.1	25.8	38.8





#7
Trichlorofluoromethane
Concen: 1.991 ug/l
RT: 2.123 min Scan# 2
Delta R.T. -0.006 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:101 Resp: 4466

Ion	Ratio	Lower	Upper
101	100		
103	67.9	51.4	77.2

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 2.014 ug/l
RT: 2.563 min Scan# 396
Delta R.T. -0.006 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:101 Resp: 25640

Ion	Ratio	Lower	Upper
101	100		
85	44.6	35.4	53.0
151	74.9	58.5	87.7





#9
1,1-Dichloroethene
Concen: 1.999 ug/l
RT: 2.567 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 96 Resp: 2593

Ion	Ratio	Lower	Upper
96	100		
61	161.7	0.0	492.9
98	64.4	0.0	124.0

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#10
Iodomethane
Concen: 2.040 ug/l
RT: 2.708 min Scan# 441
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:142 Resp: 41607

Ion	Ratio	Lower	Upper
142	100		
127	43.2	34.5	51.7





#11
Allyl Chloride
Concen: 1.974 ug/l
RT: 2.907 min Scan# 5
Delta R.T. -0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 41 Resp: 3678

Ion	Ratio	Lower	Upper
41	100		
39	73.4	57.9	86.9



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#12
Acrylonitrile
Concen: 4.056 ug/l
RT: 3.316 min Scan# 630
Delta R.T. 0.016 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 53 Resp: 12135

Ion	Ratio	Lower	Upper
53	100		
52	71.4	64.2	96.2
51	38.3	30.8	46.2





#13
Acetone
Concen: 10.336 ug/l
RT: 2.615 min Scan# 4
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 43 Resp: 2377

Ion	Ratio	Lower	Upper
43	100		
58	37.2	27.4	41.0

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#14
Carbon Disulfide
Concen: 1.891 ug/l
RT: 2.779 min Scan# 463
Delta R.T. -0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 76 Resp: 85777

Ion	Ratio	Lower	Upper
76	100		
78	10.0	7.2	10.8





#15
Methylene Chloride
Concen: 1.985 ug/l
RT: 3.030 min Scan# 54
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 84 Resp: 3181

Ion	Ratio	Lower	Upper
84	100		
49	113.8	96.4	144.6
51	35.6	0.0	74.8
86	61.1	51.3	76.9



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#16
trans-1,2-Dichloroethene
Concen: 1.966 ug/l
RT: 3.338 min Scan# 637
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 96 Resp: 29107

Ion	Ratio	Lower	Upper
96	100		
61	147.5	113.4	170.2
98	71.0	51.2	76.8





#17
1,1-Dichloroethane
Concen: 1.986 ug/l
RT: 3.850 min Scan# 7
Delta R.T. -0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 63 Resp: 5541

Ion	Ratio	Lower	Upper
63	100		
98	6.0	3.3	9.9
100	4.1	2.1	6.2



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#18
2-Butanone
Concen: 8.860 ug/l
RT: 4.702 min Scan# 1061
Delta R.T. 0.016 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 43 Resp: 32440

Ion	Ratio	Lower	Upper
43	100		
57	9.7	0.0	17.0





#19
Cyclohexane
Concen: 1.908 ug/l m
RT: 5.374 min Scan# 1
Delta R.T. -0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 56 Resp: 4277

Ion	Ratio	Lower	Upper
56	100		
69	30.7	24.5	36.7
84	82.8	65.2	97.8

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#20
Methylcyclohexane
Concen: 1.983 ug/l
RT: 6.750 min Scan# 1698
Delta R.T. -0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 83 Resp: 44086

Ion	Ratio	Lower	Upper
83	100		
55	80.7	63.1	94.7
98	45.3	35.2	52.8





#21
2,2-Dichloropropane
Concen: 1.989 ug/l
RT: 4.647 min Scan# 1
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 77 Resp: 4330

Ion	Ratio	Lower	Upper
77	100		
97	23.4	18.5	27.7



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#22
cis-1,2-Dichloroethene
Concen: 1.974 ug/l
RT: 4.653 min Scan# 1046
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 96 Resp: 31577

Ion	Ratio	Lower	Upper
96	100		
61	150.5	0.0	373.3
98	63.2	31.9	95.9





#23
Diethyl Ether
Concen: 1.904 ug/l
RT: 2.364 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 59 Resp: 2119

Ion	Ratio	Lower	Upper
59	100		
45	73.1	57.8	86.6
74	73.3	57.7	86.5

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#24
tert-Butyl Alcohol
Concen: 21.171 ug/l
RT: 3.174 min Scan# 586
Delta R.T. 0.016 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 59 Resp: 29904

Ion	Ratio	Lower	Upper
59	100		
57	6.7	7.5	11.3#





#25
Methyl tert-Butyl Ether
Concen: 1.948 ug/l
RT: 3.348 min Scan# 64
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 73 Resp: 6310

Ion	Ratio	Lower	Upper
73	100		
43	19.9	14.9	22.3

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#26
Bromochloromethane
Concen: 2.023 ug/l
RT: 4.959 min Scan# 1141
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:128 Resp: 14144

Ion	Ratio	Lower	Upper
128	100		
49	167.7	0.0	343.4
130	126.7	102.9	154.3





#27
Chloroform
Concen: 2.030 ug/l
RT: 5.075 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 83 Resp: 5717

Ion	Ratio	Lower	Upper
83	100		
85	64.5	0.0	129.8

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#28
1,1,1-Trichloroethane
Concen: 1.990 ug/l
RT: 5.300 min Scan# 1247
Delta R.T. -0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 97 Resp: 45409

Ion	Ratio	Lower	Upper
97	100		
99	63.6	32.4	97.0
61	48.7	23.8	71.2





#29
1,1-Dichloropropene
Concen: 1.994 ug/l
RT: 5.515 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 75 Resp: 4074

Ion	Ratio	Lower	Upper
75	100		
110	33.7	17.2	51.5
77	31.2	24.6	37.0

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#30
Carbon Tetrachloride
Concen: 1.928 ug/l
RT: 5.509 min Scan# 1312
Delta R.T. -0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:117 Resp: 37714

Ion	Ratio	Lower	Upper
117	100		
119	95.6	76.7	115.1
121	30.2	25.5	38.3





#31
Isopropyl Ether
Concen: 1.916 ug/l
RT: 3.975 min Scan# 8
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 45 Resp: 7635

Ion	Ratio	Lower	Upper
45	100		
43	54.0	25.7	77.1



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#32
Ethyl-t-butyl ether
Concen: 1.900 ug/l
RT: 4.483 min Scan# 993
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 59 Resp: 68833

Ion	Ratio	Lower	Upper
59	100		
87	41.1	32.6	49.0





#33
Tert-Amyl methyl ether
Concen: 1.945 ug/l
RT: 5.927 min Scan# 14
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 73 Resp: 6156

Ion	Ratio	Lower	Upper
73	100		
87	23.4	18.9	28.3
71	12.0	9.0	13.4

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#34
Propionitrile
Concen: 10.432 ug/l m
RT: 4.772 min Scan# 1083
Delta R.T. 0.019 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 54 Resp: 11521

Ion	Ratio	Lower	Upper
54	100		
52	19.3	16.3	24.5
55	12.8	11.8	17.6
40	2.8	4.3	6.5#





#35
Benzene
Concen: 1.993 ug/l
RT: 5.763 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 78 Resp: 12521

Ion	Ratio	Lower	Upper
78	100		
77	24.3	19.0	28.4

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#36
1,2-Dichloroethane
Concen: 2.013 ug/l
RT: 5.782 min Scan# 1397
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 62 Resp: 36496

Ion	Ratio	Lower	Upper
62	100		
98	8.4	6.9	10.3





#37
Trichloroethene
Concen: 2.023 ug/l
RT: 6.534 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:130 Resp: 3022

Ion	Ratio	Lower	Upper
130	100		
95	110.9	83.2	124.8



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#38
1,2-Dichloropropane
Concen: 1.965 ug/l
RT: 6.779 min Scan# 1707
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 63 Resp: 32315

Ion	Ratio	Lower	Upper
63	100		
65	32.7	25.3	37.9





#39
Methacrylonitrile
Concen: 1.967 ug/l
RT: 4.972 min Scan# 1
Delta R.T. 0.013 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 41 Resp: 809

Ion	Ratio	Lower	Upper
41	100		
67	87.2	64.5	96.7
39	47.9	43.0	64.6
52	28.8	25.3	37.9



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#40
Methyl acrylate
Concen: 1.992 ug/l m
RT: 4.853 min Scan# 1108
Delta R.T. 0.019 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 55 Resp: 15111

Ion	Ratio	Lower	Upper
55	100		
85	0.0	13.3	19.9#
58	6.9	7.3	10.9#
42	8.8	6.9	10.3





#41
Tetrahydrofuran
Concen: 3.786 ug/l
RT: 5.049 min Scan# 1
Delta R.T. 0.010 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 42 Resp: 906

Ion	Ratio	Lower	Upper
42	100		
72	48.1	41.5	62.3
71	43.5	37.2	55.8



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#42
1-Chlorobutane
Concen: 1.932 ug/l
RT: 5.444 min Scan# 1292
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 56 Resp: 54007

Ion	Ratio	Lower	Upper
56	100		
41	52.1	26.3	78.8





#43
Dibromomethane
Concen: 1.996 ug/l
RT: 6.907 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 93 Resp: 1662

Ion	Ratio	Lower	Upper
93	100		
95	80.7	67.2	100.8
174	90.0	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#44
Bromodichloromethane
Concen: 2.026 ug/l
RT: 7.094 min Scan# 1805
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 83 Resp: 39273

Ion	Ratio	Lower	Upper
83	100		
85	59.8	51.7	77.5
127	7.8	6.7	10.1





#45
4-Methyl-2-Pentanone
Concen: 9.516 ug/l
RT: 7.782 min Scan# 2
Delta R.T. 0.006 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 43 Resp: 8307

Ion	Ratio	Lower	Upper
43	100		
58	40.9	20.8	62.5
85	17.4	14.8	22.2



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#46
t-1,4-Dichloro-2-butene
Concen: 4.108 ug/l m
RT: 10.820 min Scan# 2964
Delta R.T. -0.006 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 75 Resp: 16859

Ion	Ratio	Lower	Upper
75	100		
53	66.5	64.5	96.7
89	40.2	43.4	65.2#
88	32.4	31.2	46.8





#47
Methyl methacrylate
Concen: 3.888 ug/l
RT: 6.956 min Scan# 1
Delta R.T. 0.006 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 69 Resp: 2726

Ion	Ratio	Lower	Upper
69	100		
41	107.9	0.0	217.0
39	63.1	47.7	71.5
100	32.0	27.4	41.0



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#48
Ethyl methacrylate
Concen: 1.928 ug/l
RT: 8.325 min Scan# 2188
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 69 Resp: 25396

Ion	Ratio	Lower	Upper
69	100		
41	59.5	30.6	92.0





#49
Toluene
Concen: 1.979 ug/l
RT: 7.959 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 92 Resp: 7151

Ion	Ratio	Lower	Upper
92	100		
91	178.7	141.8	212.6

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#50
t-1,3-Dichloropropene
Concen: 1.956 ug/l
RT: 8.203 min Scan# 2150
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 75 Resp: 34712

Ion	Ratio	Lower	Upper
75	100		
77	31.6	25.9	38.9





#51
cis-1,3-Dichloropropene
Concen: 1.953 ug/l
RT: 7.599 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 75 Resp: 4281

Ion	Ratio	Lower	Upper
75	100		
77	30.5	25.3	37.9
39	43.1	33.5	50.3



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#52
1,1,2-Trichloroethane
Concen: 2.026 ug/l
RT: 8.390 min Scan# 2208
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 97 Resp: 22745

Ion	Ratio	Lower	Upper
97	100		
83	89.5	73.0	109.4
85	56.9	46.3	69.5
99	61.8	48.5	72.7





#53
1,3-Dichloropropane
Concen: 1.988 ug/l
RT: 8.566 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 76 Resp: 3963

Ion	Ratio	Lower	Upper
76	100		
78	32.6	26.3	39.5



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#54
2-Hexanone
Concen: 9.005 ug/l
RT: 8.679 min Scan# 2298
Delta R.T. 0.006 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 43 Resp: 53643

Ion	Ratio	Lower	Upper
43	100		
58	57.0	38.0	78.0
57	19.5	0.0	39.1





#55
Dibromochloromethane
Concen: 1.970 ug/l
RT: 8.798 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:129 Resp: 2544

Ion	Ratio	Lower	Upper
129	100		
127	77.8	60.4	90.6



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#56
1,2-Dibromoethane
Concen: 1.935 ug/l
RT: 8.914 min Scan# 2371
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:107 Resp: 20372

Ion	Ratio	Lower	Upper
107	100		
109	96.6	0.0	187.8





#57
4-Bromofluorobenzene
Concen: 1.028 ug/l
RT: 10.627 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 95 Resp: 1733

Ion	Ratio	Lower	Upper
95	100		
174	78.4	58.6	88.0
176	76.4	58.2	87.4



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#58
Tetrachloroethene
Concen: 2.073 ug/l
RT: 8.544 min Scan# 2256
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:164 Resp: 25526

Ion	Ratio	Lower	Upper
164	100		
166	125.4	101.4	152.0
129	99.5	77.0	115.4
131	98.0	76.3	114.5





#59
Chlorobenzene
Concen: 1.978 ug/l
RT: 9.438 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:112 Resp: 7542

Ion	Ratio	Lower	Upper
112	100		
114	32.6	25.7	38.5



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#60
1,1,1,2-Tetrachloroethane
Concen: 1.999 ug/l
RT: 9.525 min Scan# 2561
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:131 Resp: 27395

Ion	Ratio	Lower	Upper
131	100		
133	95.1	76.7	115.1
119	68.2	54.4	81.6





#61
Pentachloroethane
Concen: 1.945 ug/l
RT: 11.415 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:117 Resp: 2380

Ion	Ratio	Lower	Upper
117	100		
167	75.0	65.0	97.6

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#62
Hexachloroethane
Concen: 1.861 ug/l
RT: 12.463 min Scan# 3475
Delta R.T. -0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:117 Resp: 20151

Ion	Ratio	Lower	Upper
117	100		
201	76.3	61.3	91.9





#63
Ethyl Benzene
Concen: 1.894 ug/l
RT: 9.560 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 91 Resp: 12455

Ion	Ratio	Lower	Upper
91	100		
106	31.2	24.2	36.2

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#64
m/p-Xylenes
Concen: 3.912 ug/l
RT: 9.685 min Scan# 2611
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:106 Resp: 96092

Ion	Ratio	Lower	Upper
106	100		
91	209.4	166.9	250.3





#65
o-Xylene
Concen: 2.027 ug/l
RT: 10.090 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:106 Resp: 4872

Ion	Ratio	Lower	Upper
106	100		
91	213.4	110.9	332.9



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#66
Styrene
Concen: 1.922 ug/l
RT: 10.107 min Scan# 2742
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:104 Resp: 73539

Ion	Ratio	Lower	Upper
104	100		
78	51.8	41.2	61.8
103	55.7	44.8	67.2





#67
Bromoform
Concen: 1.912 ug/l
RT: 10.283 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:173 Resp: 1401

Ion	Ratio	Lower	Upper
173	100		
175	51.7	39.0	58.4
254	9.8	7.7	11.5

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#68
1,2-Dichlorobenzene-d4
Concen: 0.983 ug/l
RT: 12.187 min Scan# 3389
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:152 Resp: 17238

Ion	Ratio	Lower	Upper
152	100		
115	85.5	0.0	275.2
150	189.7	0.0	658.4





#69
Isopropylbenzene
Concen: 1.950 ug/l
RT: 10.473 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:105 Resp: 11023

Ion	Ratio	Lower	Upper
105	100		
120	26.1	12.8	38.4

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#70
1,1,2,2-Tetrachloroethane
Concen: 1.972 ug/l
RT: 10.772 min Scan# 2949
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 83 Resp: 29840

Ion	Ratio	Lower	Upper
83	100		
131	9.2	7.4	11.0
85	64.3	51.8	77.8





#71
1,2,3-Trichloropropane
Concen: 2.043 ug/l m
RT: 10.814 min Scan# 2
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 75 Resp: 2323

Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0
110	33.3	0.0	77.0
97	18.1	0.0	42.2



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#72
Bromobenzene
Concen: 1.955 ug/l
RT: 10.775 min Scan# 2950
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:156 Resp: 29771

Ion	Ratio	Lower	Upper
156	100		
77	173.0	0.0	343.6
158	99.7	0.0	193.0





#73
n-propylbenzene
Concen: 1.932 ug/l
RT: 10.897 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:120 Resp: 3126

Ion	Ratio	Lower	Upper
120	100		
91	463.8	369.8	554.6



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#74
2-Chlorotoluene
Concen: 2.030 ug/l
RT: 10.978 min Scan# 3013
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:126 Resp: 30285

Ion	Ratio	Lower	Upper
126	100		
91	307.8	0.0	623.8





#75
1,3,5-Trimethylbenzene
Concen: 1.955 ug/l
RT: 11.078 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:105 Resp: 10240

Ion	Ratio	Lower	Upper
105	100		
120	47.4	37.3	55.9

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#76
4-Chlorotoluene
Concen: 1.962 ug/l
RT: 11.090 min Scan# 3048
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:126 Resp: 30015

Ion	Ratio	Lower	Upper
126	100		
91	349.1	0.0	703.6





#77
tert-Butylbenzene
Concen: 1.955 ug/l
RT: 11.409 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:119 Resp: 10357

Ion	Ratio	Lower	Upper
119	100		
91	58.3	29.4	88.3
134	21.5	17.6	26.4



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#78
1,2,4-Trimethylbenzene
Concen: 1.940 ug/l
RT: 11.460 min Scan# 3163
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:105 Resp: 100810

Ion	Ratio	Lower	Upper
105	100		
120	44.0	21.9	65.7





#79
sec-Butylbenzene
Concen: 1.972 ug/l
RT: 11.634 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:105 Resp: 13302

Ion	Ratio	Lower	Upper
105	100		
134	19.0	15.1	22.7

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#80
Nitrobenzene
Concen: 9.568 ug/l m
RT: 13.229 min Scan# 3713
Delta R.T. 0.026 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 77 Resp: 2781

Ion	Ratio	Lower	Upper
77	100		
123	22.5	18.9	67.1
65	3.1	11.9	15.1#





#81
p-Isopropyltoluene
Concen: 1.954 ug/l
RT: 11.785 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:119 Resp: 10401

Ion	Ratio	Lower	Upper
119	100		
134	25.4	20.3	30.5
91	24.3	19.4	29.2



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#82
1,3-Dichlorobenzene
Concen: 1.959 ug/l
RT: 11.737 min Scan# 3249
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:146 Resp: 57876

Ion	Ratio	Lower	Upper
146	100		
111	42.2	32.8	49.2
148	64.2	51.1	76.7





#83
1,4-Dichlorobenzene
Concen: 1.974 ug/l
RT: 11.830 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:146 Resp: 5704

Ion	Ratio	Lower	Upper
146	100		
111	40.6	32.1	48.1
148	63.9	50.2	75.4

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#84
n-Butylbenzene
Concen: 1.858 ug/l
RT: 12.200 min Scan# 3393
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 91 Resp: 88658

Ion	Ratio	Lower	Upper
91	100		
92	51.7	41.8	62.8
134	22.9	18.6	28.0





#85
1,2-Dichlorobenzene
Concen: 1.985 ug/l
RT: 12.203 min Scan# 3
Delta R.T. 0.000 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:146 Resp: 5635

Ion	Ratio	Lower	Upper
146	100		
111	43.7	21.9	65.7
148	63.4	32.3	96.9



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#86
1,2-Dibromo-3-Chloropropane
Concen: 1.744 ug/l
RT: 12.991 min Scan# 3639
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion: 75 Resp: 3704

Ion	Ratio	Lower	Upper
75	100		
155	87.3	63.5	95.3
157	110.6	81.8	122.6





#87
1,2,4-Trichlorobenzene
Concen: 1.755 ug/l
RT: 13.836 min Scan# 3
Delta R.T. 0.006 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:180 Resp: 2429

Ion	Ratio	Lower	Upper
180	100		
182	96.5	76.6	115.0
145	31.3	25.4	38.2



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#88
Hexachlorobutadiene
Concen: 2.061 ug/l
RT: 14.010 min Scan# 3956
Delta R.T. -0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:225 Resp: 20391

Ion	Ratio	Lower	Upper
225	100		
223	63.8	49.5	74.3
227	64.3	51.0	76.4





#89
Naphthalene
Concen: 1.689 ug/l
RT: 14.084 min Scan# 3
Delta R.T. 0.006 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:128 Resp: 3194

Ion	Ratio	Lower	Upper
128	100		
127	14.0	10.6	16.0
129	10.0	8.6	13.0



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBS01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#90
1,2,3-Trichlorobenzene
Concen: 1.598 ug/l
RT: 14.325 min Scan# 4054
Delta R.T. 0.003 min
Lab File: VU063251.D
Acq: 12 Feb 2025 12:56

Tgt Ion:180 Resp: 21647

Ion	Ratio	Lower	Upper
180	100		
182	100.5	78.2	117.2
145	31.9	26.1	39.1



Raw Data: **VU063252.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	
Project:	NJ Drinking Water PT	Date Received:	
Client Sample ID:	VU0212WBSD01	SDG No.:	Q1172
Lab Sample ID:	VU0212WBSD01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063252.D	1		02/12/25 13:20	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	2.10		0.14	0.50	ug/L
74-87-3	Chloromethane	2.10		0.13	0.50	ug/L
75-01-4	Vinyl Chloride	2.10		0.13	0.50	ug/L
74-83-9	Bromomethane	2.30		0.18	0.50	ug/L
75-00-3	Chloroethane	2.10		0.14	0.50	ug/L
109-99-9	Tetrahydrofuran	3.80		0.44	1.00	ug/L
75-69-4	Trichlorofluoromethane	2.10		0.21	0.50	ug/L
76-13-1	1,1,2-Trichloro-1,2,2-trifluoroethane	2.10		0.14	0.50	ug/L
75-65-0	tert-Butyl Alcohol	15.6		8.60	10.0	ug/L
60-29-7	Diethyl Ether	2.10		0.13	0.50	ug/L
75-35-4	1,1-Dichloroethene	2.10		0.12	0.50	ug/L
107-13-1	Acrylonitrile	4.30		0.44	1.00	ug/L
67-64-1	Acetone	10.0		1.10	2.50	ug/L
75-15-0	Carbon Disulfide	2.00		0.13	0.50	ug/L
1634-04-4	Methyl tert-Butyl Ether	2.00		0.12	0.50	ug/L
96-33-3	Methyl acrylate	2.00		0.28	0.50	ug/L
75-09-2	Methylene Chloride	2.10		0.47	0.50	ug/L
156-60-5	trans-1,2-Dichloroethene	2.00		0.14	0.50	ug/L
75-34-3	1,1-Dichloroethane	2.10		0.13	0.50	ug/L
110-82-7	Cyclohexane	2.00		0.14	0.50	ug/L
78-93-3	2-Butanone	9.20		0.68	2.50	ug/L
56-23-5	Carbon Tetrachloride	2.10		0.14	0.50	ug/L
594-20-7	2,2-Dichloropropane	2.10		0.14	0.50	ug/L
156-59-2	cis-1,2-Dichloroethene	2.10		0.13	0.50	ug/L
74-97-5	Bromochloromethane	2.10		0.16	0.50	ug/L
67-66-3	Chloroform	2.10		0.13	0.50	ug/L
71-55-6	1,1,1-Trichloroethane	2.10		0.12	0.50	ug/L
108-87-2	Methylcyclohexane	2.00		0.12	0.50	ug/L
563-58-6	1,1-Dichloropropene	2.10		0.11	0.50	ug/L
107-12-0	Propionitrile	9.80		1.00	2.50	ug/L

Raw Data: **VU063252.D**

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	
Project:	NJ Drinking Water PT	Date Received:	
Client Sample ID:	VU0212WBSD01	SDG No.:	Q1172
Lab Sample ID:	VU0212WBSD01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063252.D	1		02/12/25 13:20	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
71-43-2	Benzene	2.10		0.11	0.50	ug/L
107-06-2	1,2-Dichloroethane	2.10		0.16	0.50	ug/L
79-01-6	Trichloroethene	2.10		0.13	0.50	ug/L
78-87-5	1,2-Dichloropropane	2.10		0.13	0.50	ug/L
109-69-3	1-Chlorobutane	2.00		0.12	0.50	ug/L
74-95-3	Dibromomethane	2.10		0.14	0.50	ug/L
75-27-4	Bromodichloromethane	2.10		0.12	0.50	ug/L
108-10-1	4-Methyl-2-Pentanone	9.70		0.60	2.50	ug/L
108-88-3	Toluene	2.10		0.11	0.50	ug/L
10061-02-6	t-1,3-Dichloropropene	2.10		0.11	0.50	ug/L
10061-01-5	cis-1,3-Dichloropropene	2.00		0.11	0.50	ug/L
79-00-5	1,1,2-Trichloroethane	2.10		0.13	0.50	ug/L
142-28-9	1,3-Dichloropropane	2.10		0.13	0.50	ug/L
591-78-6	2-Hexanone	9.10		0.57	2.50	ug/L
124-48-1	Dibromochloromethane	2.10		0.13	0.50	ug/L
106-93-4	1,2-Dibromoethane	2.10		0.13	0.50	ug/L
127-18-4	Tetrachloroethene	2.10		0.14	0.50	ug/L
108-90-7	Chlorobenzene	2.10		0.11	0.50	ug/L
630-20-6	1,1,1,2-Tetrachloroethane	2.10		0.13	0.50	ug/L
67-72-1	Hexachloroethane	2.00		0.12	0.50	ug/L
100-41-4	Ethyl Benzene	2.00		0.12	0.50	ug/L
179601-23-1	m/p-Xylenes	4.10		0.23	1.00	ug/L
1330-20-7	Total Xylenes	6.20		0.35	1.50	ug/L
95-47-6	o-Xylene	2.10		0.12	0.50	ug/L
100-42-5	Styrene	2.00		0.13	0.50	ug/L
75-25-2	Bromoform	2.00		0.14	0.50	ug/L
98-82-8	Isopropylbenzene	2.00		0.13	0.50	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	2.00		0.12	0.50	ug/L
96-18-4	1,2,3-Trichloropropane	1.80		0.21	0.50	ug/L
108-86-1	Bromobenzene	2.10		0.13	0.50	ug/L
103-65-1	n-propylbenzene	1.90		0.16	0.50	ug/L

Raw Data: **VU063252.D**

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	
Project:	NJ Drinking Water PT	Date Received:	
Client Sample ID:	VU0212WBSD01	SDG No.:	Q1172
Lab Sample ID:	VU0212WBSD01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063252.D	1		02/12/25 13:20	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
95-49-8	2-Chlorotoluene	2.00		0.14	0.50	ug/L
108-67-8	1,3,5-Trimethylbenzene	2.00		0.13	0.50	ug/L
106-43-4	4-Chlorotoluene	2.10		0.14	0.50	ug/L
98-06-6	tert-Butylbenzene	2.00		0.11	0.50	ug/L
95-63-6	1,2,4-Trimethylbenzene	2.00		0.13	0.50	ug/L
135-98-8	sec-Butylbenzene	2.10		0.13	0.50	ug/L
99-87-6	p-Isopropyltoluene	2.00		0.16	0.50	ug/L
541-73-1	1,3-Dichlorobenzene	2.10		0.13	0.50	ug/L
106-46-7	1,4-Dichlorobenzene	2.10		0.14	0.50	ug/L
104-51-8	n-Butylbenzene	1.90		0.28	0.50	ug/L
95-50-1	1,2-Dichlorobenzene	2.00		0.14	0.50	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	2.00		0.23	0.50	ug/L
120-82-1	1,2,4-Trichlorobenzene	1.90		0.21	0.50	ug/L
87-68-3	Hexachlorobutadiene	2.10		0.14	0.50	ug/L
91-20-3	Naphthalene	2.00		0.31	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	1.90		0.25	0.50	ug/L
74-88-4	Iodomethane	2.20		0.16	1.00	ug/L
107-05-1	Allyl Chloride	2.00		0.11	0.50	ug/L
126-98-7	Methacrylonitrile	1.90		0.19	0.50	ug/L
108-20-3	Isopropyl Ether	2.00		0.12	0.50	ug/L
80-62-6	Methyl methacrylate	4.10		0.24	1.00	ug/L
SURROGATES						
2199-69-1	1,2-Dichlorobenzene-d4	0.96		70 - 130	96%	SPK: 1
460-00-4	4-Bromofluorobenzene	1.00		70 - 130	103%	SPK: 1
INTERNAL STANDARDS						
462-06-6	Fluorobenzene	51200	6.107			
	Tert-Amyl methyl ether					

Alliance TECHNICAL GROUP
284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	
Project:	NJ Drinking Water PT	Date Received:	
Client Sample ID:	VU0212WBSD01	SDG No.:	Q1172
Lab Sample ID:	VU0212WBSD01	Matrix:	Water
Analytical Method:	E524.2	% Solid:	0
Sample Wt/Vol:	25 Units: mL	Final Vol:	25000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group3
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VU063252.D	1		02/12/25 13:20	VU021225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063252.D
Acq On : 12 Feb 2025 13:20
Operator : MD/SY
Sample : VU0212WBSD01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Feb 13 03:11:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Fluorobenzene	6.107	96	51180	1.000	ug/l	# 0.00
System Monitoring Compounds						
57) 4-Bromofluorobenzene	10.627	95	17310	1.025	ug/l	0.00
Spiked Amount 1.000			Recovery = 103.000%			
68) 1,2-Dichlorobenzene-d4	12.187	152	16808	0.957	ug/l	0.00
Spiked Amount 1.000			Recovery = 96.000%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.380	85	34139	2.053	ug/l	100
3) Chloromethane	1.515	50	39500	2.063	ug/l	99
4) Vinyl Chloride	1.599	62	39806	2.101	ug/l	96
5) Bromomethane	1.846	94	21678	2.342	ug/l	95
6) Chloroethane	1.923	64	24495	2.053	ug/l	100
7) Trichlorofluoromethane	2.129	101	47061	2.096	ug/l	98
8) 1,1,2-Trichloro-1,2,2-...	2.570	101	26155	2.052	ug/l	97
9) 1,1-Dichloroethene	2.570	96	26932	2.074	ug/l	96
10) Iodomethane	2.711	142	44058	2.158	ug/l	99
11) Allyl Chloride	2.914	41	37130	1.991	ug/l	98
12) Acrylonitrile	3.313	53	12746m	4.256	ug/l	
13) Acetone	2.615	43	23082	10.023	ug/l	98
14) Carbon Disulfide	2.782	76	92105	2.029	ug/l	99
15) Methylene Chloride	3.033	84	33386	2.081	ug/l	100
16) trans-1,2-Dichloroethene	3.341	96	30224	2.039	ug/l	98
17) 1,1-Dichloroethane	3.853	63	57716	2.066	ug/l	99
18) 2-Butanone	4.698	43	33717	9.200	ug/l	98
19) Cyclohexane	5.373	56	44849m	1.998	ug/l	
20) Methylcyclohexane	6.750	83	45015	2.022	ug/l	98
21) 2,2-Dichloropropane	4.650	77	45151	2.072	ug/l	99
22) cis-1,2-Dichloroethene	4.660	96	34321	2.143	ug/l	98
23) Diethyl Ether	2.367	59	23235	2.085	ug/l	97
24) tert-Butyl Alcohol	3.165	59	24034	15.625	ug/l #	94
25) Methyl tert-Butyl Ether	3.348	73	65478	2.019	ug/l	98
26) Bromochloromethane	4.959	128	14577	2.083	ug/l	100
27) Chloroform	5.071	83	60016	2.129	ug/l	98
28) 1,1,1-Trichloroethane	5.300	97	47367	2.074	ug/l	98
29) 1,1-Dichloropropene	5.515	75	42057	2.056	ug/l	99
30) Carbon Tetrachloride	5.512	117	41248	2.106	ug/l	98
31) Isopropyl Ether	3.978	45	80738	2.025	ug/l	96
32) Ethyl-t-butyl ether	4.486	59	74455	2.053	ug/l	99
33) Tert-Amyl methyl ether	5.927	73	65827	2.077	ug/l	99
34) Propionitrile	4.775	54	10831	9.798	ug/l #	74
35) Benzene	5.763	78	129848	2.065	ug/l	99
36) 1,2-Dichloroethane	5.782	62	38095	2.099	ug/l	99
37) Trichloroethene	6.534	130	31471	2.104	ug/l	98
38) 1,2-Dichloropropane	6.779	63	34356	2.087	ug/l	99
39) Methacrylonitrile	4.968	41	7822	1.899	ug/l	90
40) Methyl acrylate	4.853	55	15431m	2.032	ug/l	
41) Tetrahydrofuran	5.052	42	9203	3.841	ug/l	99
42) 1-Chlorobutane	5.447	56	56138	2.006	ug/l	96
43) Dibromomethane	6.910	93	17301	2.076	ug/l	96
44) Bromodichloromethane	7.094	83	40765	2.101	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063252.D
Acq On : 12 Feb 2025 13:20
Operator : MD/SY
Sample : VU0212WBSD01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Feb 13 03:11:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
45) 4-Methyl-2-Pentanone	7.779	43	84886	9.715	ug/l	99
46) t-1,4-Dichloro-2-butene	10.820	75	16629m	4.048	ug/l	
47) Methyl methacrylate	6.952	69	28606	4.075	ug/l	96
48) Ethyl methacrylate	8.325	69	25786	1.956	ug/l	95
49) Toluene	7.959	92	74526	2.061	ug/l	99
50) t-1,3-Dichloropropene	8.203	75	36777	2.071	ug/l	98
51) cis-1,3-Dichloropropene	7.598	75	43806	1.997	ug/l	95
52) 1,1,2-Trichloroethane	8.389	97	23471	2.089	ug/l	96
53) 1,3-Dichloropropane	8.563	76	40985	2.054	ug/l	100
54) 2-Hexanone	8.676	43	54296	9.106	ug/l	97
55) Dibromochloromethane	8.798	129	26971	2.086	ug/l	97
56) 1,2-Dibromoethane	8.917	107	21945	2.082	ug/l	99
58) Tetrachloroethene	8.544	164	26364	2.139	ug/l	97
59) Chlorobenzene	9.438	112	80285	2.104	ug/l	100
60) 1,1,1,2-Tetrachloroethane	9.521	131	28939	2.110	ug/l	99
61) Pentachloroethane	11.415	117	24586	2.006	ug/l	96
62) Hexachloroethane	12.463	117	21415	1.975	ug/l	99
63) Ethyl Benzene	9.560	91	131789	2.002	ug/l	99
64) m/p-Xylenes	9.685	106	99562	4.050	ug/l	98
65) o-Xylene	10.090	106	50162	2.084	ug/l	99
66) Styrene	10.106	104	76780	2.005	ug/l	98
67) Bromoform	10.280	173	15038	2.049	ug/l	98
69) Isopropylbenzene	10.473	105	115836	2.047	ug/l	98
70) 1,1,2,2-Tetrachloroethane	10.772	83	30615	2.021	ug/l	96
71) 1,2,3-Trichloropropane	10.814	75	20581m	1.809	ug/l	
72) Bromobenzene	10.775	156	31609	2.073	ug/l	86
73) n-propylbenzene	10.894	120	31375	1.937	ug/l	97
74) 2-Chlorotoluene	10.975	126	29907	2.003	ug/l	96
75) 1,3,5-Trimethylbenzene	11.077	105	105820	2.018	ug/l	99
76) 4-Chlorotoluene	11.090	126	31442	2.054	ug/l	95
77) tert-Butylbenzene	11.409	119	108063	2.038	ug/l	98
78) 1,2,4-Trimethylbenzene	11.457	105	101607	1.953	ug/l	98
79) sec-Butylbenzene	11.634	105	138797	2.056	ug/l	99
80) Nitrobenzene	13.222	77	3316m	10.663	ug/l	
81) p-Isopropyltoluene	11.782	119	104541	1.962	ug/l	98
82) 1,3-Dichlorobenzene	11.737	146	60800	2.056	ug/l	100
83) 1,4-Dichlorobenzene	11.827	146	60154	2.080	ug/l	99
84) n-Butylbenzene	12.200	91	88457	1.852	ug/l	99
85) 1,2-Dichlorobenzene	12.206	146	57516	2.024	ug/l	98
86) 1,2-Dibromo-3-Chloropr...	12.991	75	4329	2.036	ug/l	94
87) 1,2,4-Trichlorobenzene	13.833	180	26285	1.896	ug/l	94
88) Hexachlorobutadiene	14.010	225	20626	2.082	ug/l	99
89) Naphthalene	14.084	128	41502	1.996	ug/l	99
90) 1,2,3-Trichlorobenzene	14.322	180	26436	1.949	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_U\Data\VU021225\
Data File : VU063252.D
Acq On : 12 Feb 2025 13:20
Operator : MD/SY
Sample : VU0212WBSD01
Misc : 25.0mL/MSVOA_U/WATER
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Feb 13 03:11:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_U\Method\524U021025DW.M
Quant Title : METHOD 524.2 VOLATILES DRINKING WATER
QLast Update : Tue Feb 11 08:42:19 2025
Response via : Initial Calibration

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#1
Fluorobenzene
Concen: 1.000 ug/l
RT: 6.107 min Scan# 14
Delta R.T. 0.004 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 96 Resp: 5118

Ion	Ratio	Lower	Upper
96	100		
70	20.2	15.6	23.4
95	8.7	7.3	10.9
97	6.2	0.0	0.0#



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#2
Dichlorodifluoromethane
Concen: 2.053 ug/l
RT: 1.380 min Scan# 28
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 85 Resp: 34139

Ion	Ratio	Lower	Upper
85	100		
87	32.0	16.0	48.0





#3
Chloromethane
Concen: 2.063 ug/l
RT: 1.515 min Scan# 7(
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 50 Resp: 3950(

Ion	Ratio	Lower	Upper
50	100		
52	31.7	25.8	38.8



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#4
Vinyl Chloride
Concen: 2.101 ug/l
RT: 1.599 min Scan# 96
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 62 Resp: 39806

Ion	Ratio	Lower	Upper
62	100		
64	33.8	25.4	38.0





#5
Bromomethane
Concen: 2.342 ug/l
RT: 1.846 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 94 Resp: 2167

Ion	Ratio	Lower	Upper
94	100		
96	90.2	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#6
Chloroethane
Concen: 2.053 ug/l
RT: 1.923 min Scan# 197
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 64 Resp: 24495

Ion	Ratio	Lower	Upper
64	100		
66	32.2	25.8	38.8





#7
Trichlorofluoromethane
Concen: 2.096 ug/l
RT: 2.129 min Scan# 2
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:101 Resp: 4706

Ion	Ratio	Lower	Upper
101	100		
103	66.1	51.4	77.2

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#8
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 2.052 ug/l
RT: 2.570 min Scan# 398
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:101 Resp: 26155

Ion	Ratio	Lower	Upper
101	100		
85	46.1	35.4	53.0
151	75.4	58.5	87.7





#9
1,1-Dichloroethene
Concen: 2.074 ug/l
RT: 2.570 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 96 Resp: 2693

Ion	Ratio	Lower	Upper
96	100		
61	169.5	0.0	492.9
98	65.5	0.0	124.0

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#10
Iodomethane
Concen: 2.158 ug/l
RT: 2.711 min Scan# 442
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:142 Resp: 44058

Ion	Ratio	Lower	Upper
142	100		
127	42.7	34.5	51.7





#11
Allyl Chloride
Concen: 1.991 ug/l
RT: 2.914 min Scan# 5
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 41 Resp: 3713

Ion	Ratio	Lower	Upper
41	100		
39	73.8	57.9	86.9

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#12
Acrylonitrile
Concen: 4.256 ug/l m
RT: 3.313 min Scan# 629
Delta R.T. 0.013 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 53 Resp: 12746

Ion	Ratio	Lower	Upper
53	100		
52	71.5	64.2	96.2
51	38.2	30.8	46.2





#13
Acetone
Concen: 10.023 ug/l
RT: 2.615 min Scan# 4
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 43 Resp: 2308

Ion	Ratio	Lower	Upper
43	100		
58	33.0	27.4	41.0



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#14
Carbon Disulfide
Concen: 2.029 ug/l
RT: 2.782 min Scan# 464
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 76 Resp: 92105

Ion	Ratio	Lower	Upper
76	100		
78	9.4	7.2	10.8





#15
Methylene Chloride
Concen: 2.081 ug/l
RT: 3.033 min Scan# 54
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 84 Resp: 3338

Ion	Ratio	Lower	Upper
84	100		
49	120.5	96.4	144.6
51	37.1	0.0	74.8
86	64.8	51.3	76.9



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#16
trans-1,2-Dichloroethene
Concen: 2.039 ug/l
RT: 3.341 min Scan# 638
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 96 Resp: 30224

Ion	Ratio	Lower	Upper
96	100		
61	143.3	113.4	170.2
98	65.9	51.2	76.8





#17
1,1-Dichloroethane
Concen: 2.066 ug/l
RT: 3.853 min Scan# 7
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 63 Resp: 5771

Ion	Ratio	Lower	Upper
63	100		
98	6.9	3.3	9.9
100	3.7	2.1	6.2



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#18
2-Butanone
Concen: 9.200 ug/l
RT: 4.698 min Scan# 1060
Delta R.T. 0.013 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 43 Resp: 33717

Ion	Ratio	Lower	Upper
43	100		
57	9.4	0.0	17.0





#19
Cyclohexane
Concen: 1.998 ug/l m
RT: 5.373 min Scan# 1
Delta R.T. -0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 56 Resp: 4484

Ion	Ratio	Lower	Upper
56	100		
69	30.9	24.5	36.7
84	86.6	65.2	97.8



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#20
Methylcyclohexane
Concen: 2.022 ug/l
RT: 6.750 min Scan# 1698
Delta R.T. -0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 83 Resp: 45015

Ion	Ratio	Lower	Upper
83	100		
55	80.7	63.1	94.7
98	44.5	35.2	52.8





#21
2,2-Dichloropropane
Concen: 2.072 ug/l
RT: 4.650 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 77 Resp: 4515

Ion	Ratio	Lower	Upper
77	100		
97	23.4	18.5	27.7



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#22
cis-1,2-Dichloroethene
Concen: 2.143 ug/l
RT: 4.660 min Scan# 1048
Delta R.T. 0.010 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 96 Resp: 34321

Ion	Ratio	Lower	Upper
96	100		
61	145.5	0.0	373.3
98	63.1	31.9	95.9





#23
Diethyl Ether
Concen: 2.085 ug/l
RT: 2.367 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 59 Resp: 2323

Ion	Ratio	Lower	Upper
59	100		
45	69.5	57.8	86.6
74	70.1	57.7	86.5

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#24
tert-Butyl Alcohol
Concen: 15.625 ug/l
RT: 3.165 min Scan# 583
Delta R.T. 0.006 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 59 Resp: 24034

Ion	Ratio	Lower	Upper
59	100		
57	7.2	7.5	11.3#





#25
Methyl tert-Butyl Ether
Concen: 2.019 ug/l
RT: 3.348 min Scan# 64
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 73 Resp: 6547

Ion	Ratio	Lower	Upper
73	100		
43	19.7	14.9	22.3

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#26
Bromochloromethane
Concen: 2.083 ug/l
RT: 4.959 min Scan# 1141
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:128 Resp: 14577

Ion	Ratio	Lower	Upper
128	100		
49	171.6	0.0	343.4
130	129.1	102.9	154.3





#27
Chloroform
Concen: 2.129 ug/l
RT: 5.071 min Scan# 1
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 83 Resp: 6001

Ion	Ratio	Lower	Upper
83	100		
85	63.4	0.0	129.8

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#28
1,1,1-Trichloroethane
Concen: 2.074 ug/l
RT: 5.300 min Scan# 1247
Delta R.T. -0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 97 Resp: 47367

Ion	Ratio	Lower	Upper
97	100		
99	65.6	32.4	97.0
61	49.8	23.8	71.2





#29
1,1-Dichloropropene
Concen: 2.056 ug/l
RT: 5.515 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 75 Resp: 4205

Ion	Ratio	Lower	Upper
75	100		
110	34.9	17.2	51.5
77	31.8	24.6	37.0



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#30
Carbon Tetrachloride
Concen: 2.106 ug/l
RT: 5.512 min Scan# 1313
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:117 Resp: 41248

Ion	Ratio	Lower	Upper
117	100		
119	97.0	76.7	115.1
121	29.9	25.5	38.3





#31
Isopropyl Ether
Concen: 2.025 ug/l
RT: 3.978 min Scan# 8
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 45 Resp: 8073

Ion	Ratio	Lower	Upper
45	100		
43	54.5	25.7	77.1



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#32
Ethyl-t-butyl ether
Concen: 2.053 ug/l
RT: 4.486 min Scan# 994
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 59 Resp: 74455

Ion	Ratio	Lower	Upper
59	100		
87	41.3	32.6	49.0





#33
Tert-Amyl methyl ether
Concen: 2.077 ug/l
RT: 5.927 min Scan# 14
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 73 Resp: 6582

Ion	Ratio	Lower	Upper
73	100		
87	23.6	18.9	28.3
71	12.1	9.0	13.4

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#34
Propionitrile
Concen: 9.798 ug/l
RT: 4.775 min Scan# 1084
Delta R.T. 0.022 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 54 Resp: 10831

Ion	Ratio	Lower	Upper
54	100		
52	10.5	16.3	24.5#
55	0.0	11.8	17.6#
40	0.0	4.3	6.5#





#35
Benzene
Concen: 2.065 ug/l
RT: 5.763 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 78 Resp: 12984

Ion	Ratio	Lower	Upper
78	100		
77	23.4	19.0	28.4

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#36
1,2-Dichloroethane
Concen: 2.099 ug/l
RT: 5.782 min Scan# 1397
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 62 Resp: 38095

Ion	Ratio	Lower	Upper
62	100		
98	8.1	6.9	10.3





#37
Trichloroethene
Concen: 2.104 ug/l
RT: 6.534 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:130 Resp: 3147

Ion	Ratio	Lower	Upper
130	100		
95	102.3	83.2	124.8



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#38
1,2-Dichloropropane
Concen: 2.087 ug/l
RT: 6.779 min Scan# 1707
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 63 Resp: 34356

Ion	Ratio	Lower	Upper
63	100		
65	32.1	25.3	37.9





#39
Methacrylonitrile
Concen: 1.899 ug/l
RT: 4.968 min Scan# 1
Delta R.T. 0.010 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 41 Resp: 782

Ion	Ratio	Lower	Upper
41	100		
67	92.7	64.5	96.7
39	56.4	43.0	64.6
52	37.1	25.3	37.9



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#40
Methyl acrylate
Concen: 2.032 ug/l m
RT: 4.853 min Scan# 1108
Delta R.T. 0.019 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 55 Resp: 15431

Ion	Ratio	Lower	Upper
55	100		
85	0.0	13.3	19.9#
58	4.2	7.3	10.9#
42	7.3	6.9	10.3





#41
Tetrahydrofuran
Concen: 3.841 ug/l
RT: 5.052 min Scan# 1
Delta R.T. 0.013 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 42 Resp: 920

Ion	Ratio	Lower	Upper
42	100		
72	51.0	41.5	62.3
71	45.5	37.2	55.8



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#42
1-Chlorobutane
Concen: 2.006 ug/l
RT: 5.447 min Scan# 1293
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 56 Resp: 56138

Ion	Ratio	Lower	Upper
56	100		
41	49.5	26.3	78.8





#43
Dibromomethane
Concen: 2.076 ug/l
RT: 6.910 min Scan# 1
Delta R.T. 0.006 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 93 Resp: 1730

Ion	Ratio	Lower	Upper
93	100		
95	79.8	67.2	100.8
174	91.2	75.7	113.5



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#44
Bromodichloromethane
Concen: 2.101 ug/l
RT: 7.094 min Scan# 1805
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 83 Resp: 40765

Ion	Ratio	Lower	Upper
83	100		
85	67.9	51.7	77.5
127	7.1	6.7	10.1





#45
4-Methyl-2-Pentanone
Concen: 9.715 ug/l
RT: 7.779 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 43 Resp: 8488

Ion	Ratio	Lower	Upper
43	100		
58	41.3	20.8	62.5
85	18.8	14.8	22.2

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#46
t-1,4-Dichloro-2-butene
Concen: 4.048 ug/l m
RT: 10.820 min Scan# 2964
Delta R.T. -0.006 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 75 Resp: 16629

Ion	Ratio	Lower	Upper
75	100		
53	64.9	64.5	96.7
89	40.4	43.4	65.2#
88	30.7	31.2	46.8#





#47
Methyl methacrylate
Concen: 4.075 ug/l
RT: 6.952 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 69 Resp: 2860

Ion	Ratio	Lower	Upper
69	100		
41	102.5	0.0	217.0
39	61.5	47.7	71.5
100	32.6	27.4	41.0



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#48
Ethyl methacrylate
Concen: 1.956 ug/l
RT: 8.325 min Scan# 2188
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 69 Resp: 25786

Ion	Ratio	Lower	Upper
69	100		
41	65.2	30.6	92.0





#49
Toluene
Concen: 2.061 ug/l
RT: 7.959 min Scan# 2
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 92 Resp: 7452

Ion	Ratio	Lower	Upper
92	100		
91	179.2	141.8	212.6

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#50
t-1,3-Dichloropropene
Concen: 2.071 ug/l
RT: 8.203 min Scan# 2150
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 75 Resp: 36777

Ion	Ratio	Lower	Upper
75	100		
77	31.5	25.9	38.9





#51
cis-1,3-Dichloropropene
Concen: 1.997 ug/l
RT: 7.598 min Scan# 1
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 75 Resp: 4380

Ion	Ratio	Lower	Upper
75	100		
77	31.1	25.3	37.9
39	47.3	33.5	50.3



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#52
1,1,2-Trichloroethane
Concen: 2.089 ug/l
RT: 8.389 min Scan# 2208
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 97 Resp: 23471

Ion	Ratio	Lower	Upper
97	100		
83	86.2	73.0	109.4
85	57.1	46.3	69.5
99	64.4	48.5	72.7





#53
1,3-Dichloropropane
Concen: 2.054 ug/l
RT: 8.563 min Scan# 2
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 76 Resp: 4098

Ion	Ratio	Lower	Upper
76	100		
78	33.1	26.3	39.5

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#54
2-Hexanone
Concen: 9.106 ug/l
RT: 8.676 min Scan# 2297
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 43 Resp: 54296

Ion	Ratio	Lower	Upper
43	100		
58	55.5	38.0	78.0
57	17.8	0.0	39.1





#55
Dibromochloromethane
Concen: 2.086 ug/l
RT: 8.798 min Scan# 2
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:129 Resp: 2697

Ion	Ratio	Lower	Upper
129	100		
127	78.1	60.4	90.6



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#56
1,2-Dibromoethane
Concen: 2.082 ug/l
RT: 8.917 min Scan# 2372
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:107 Resp: 21945

Ion	Ratio	Lower	Upper
107	100		
109	92.8	0.0	187.8





#57
4-Bromofluorobenzene
Concen: 1.025 ug/l
RT: 10.627 min Scan# 2
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 95 Resp: 1731

Ion	Ratio	Lower	Upper
95	100		
174	78.2	58.6	88.0
176	74.8	58.2	87.4



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#58
Tetrachloroethene
Concen: 2.139 ug/l
RT: 8.544 min Scan# 2256
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:164 Resp: 26364

Ion	Ratio	Lower	Upper
164	100		
166	133.0	101.4	152.0
129	97.9	77.0	115.4
131	94.9	76.3	114.5





#59
Chlorobenzene
Concen: 2.104 ug/l
RT: 9.438 min Scan# 2
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:112 Resp: 8028

Ion	Ratio	Lower	Upper
112	100		
114	31.9	25.7	38.5



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#60
1,1,1,2-Tetrachloroethane
Concen: 2.110 ug/l
RT: 9.521 min Scan# 2560
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:131 Resp: 28939

Ion	Ratio	Lower	Upper
131	100		
133	93.9	76.7	115.1
119	68.4	54.4	81.6





#61
Pentachloroethane
Concen: 2.006 ug/l
RT: 11.415 min Scan# 3
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:117 Resp: 2458

Ion	Ratio	Lower	Upper
117	100		
167	78.1	65.0	97.6

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#62
Hexachloroethane
Concen: 1.975 ug/l
RT: 12.463 min Scan# 3475
Delta R.T. -0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:117 Resp: 21415

Ion	Ratio	Lower	Upper
117	100		
201	75.7	61.3	91.9





#63
Ethyl Benzene
Concen: 2.002 ug/l
RT: 9.560 min Scan# 2
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 91 Resp: 13178

Ion	Ratio	Lower	Upper
91	100		
106	30.6	24.2	36.2

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#64
m/p-Xylenes
Concen: 4.050 ug/l
RT: 9.685 min Scan# 2611
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:106 Resp: 99562

Ion	Ratio	Lower	Upper
106	100		
91	211.5	166.9	250.3





#65
o-Xylene
Concen: 2.084 ug/l
RT: 10.090 min Scan# 2
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:106 Resp: 5016

Ion	Ratio	Lower	Upper
106	100		
91	219.6	110.9	332.9

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#66
Styrene
Concen: 2.005 ug/l
RT: 10.106 min Scan# 2742
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:104 Resp: 76780

Ion	Ratio	Lower	Upper
104	100		
78	51.2	41.2	61.8
103	57.9	44.8	67.2





#67
Bromoform
Concen: 2.049 ug/l
RT: 10.280 min Scan# 2
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:173 Resp: 1503

Ion	Ratio	Lower	Upper
173	100		
175	47.5	39.0	58.4
254	9.2	7.7	11.5

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#68
1,2-Dichlorobenzene-d4
Concen: 0.957 ug/l
RT: 12.187 min Scan# 3389
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:152 Resp: 16808

Ion	Ratio	Lower	Upper
152	100		
115	78.9	0.0	275.2
150	189.9	0.0	658.4





#69
Isopropylbenzene
Concen: 2.047 ug/l
RT: 10.473 min Scan# 2
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:105 Resp: 11583

Ion	Ratio	Lower	Upper
105	100		
120	24.8	12.8	38.4

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#70
1,1,2,2-Tetrachloroethane
Concen: 2.021 ug/l
RT: 10.772 min Scan# 2949
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 83 Resp: 30615

Ion	Ratio	Lower	Upper
83	100		
131	10.2	7.4	11.0
85	67.8	51.8	77.8





#71
1,2,3-Trichloropropane
Concen: 1.809 ug/l m
RT: 10.814 min Scan# 2
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 75 Resp: 2058

Ion	Ratio	Lower	Upper
75	100		
77	0.0	0.0	0.0
110	38.6	0.0	77.0
97	20.0	0.0	42.2



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#72
Bromobenzene
Concen: 2.073 ug/l
RT: 10.775 min Scan# 2950
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:156 Resp: 31609

Ion	Ratio	Lower	Upper
156	100		
77	200.9	0.0	343.6
158	98.0	0.0	193.0





#73
n-propylbenzene
Concen: 1.937 ug/l
RT: 10.894 min Scan# 2
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:120 Resp: 3137

Ion	Ratio	Lower	Upper
120	100		
91	468.7	369.8	554.6

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#74
2-Chlorotoluene
Concen: 2.003 ug/l
RT: 10.975 min Scan# 3012
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:126 Resp: 29907

Ion	Ratio	Lower	Upper
126	100		
91	320.6	0.0	623.8





#75
1,3,5-Trimethylbenzene
Concen: 2.018 ug/l
RT: 11.077 min Scan# 3
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:105 Resp: 10582

Ion	Ratio	Lower	Upper
105	100		
120	47.2	37.3	55.9

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#76
4-Chlorotoluene
Concen: 2.054 ug/l
RT: 11.090 min Scan# 3048
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:126 Resp: 31442

Ion	Ratio	Lower	Upper
126	100		
91	340.1	0.0	703.6





#77
tert-Butylbenzene
Concen: 2.038 ug/l
RT: 11.409 min Scan# 3
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:119 Resp: 10806

Ion	Ratio	Lower	Upper
119	100		
91	56.9	29.4	88.3
134	22.0	17.6	26.4

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#78
1,2,4-Trimethylbenzene
Concen: 1.953 ug/l
RT: 11.457 min Scan# 3162
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:105 Resp: 101607

Ion	Ratio	Lower	Upper
105	100		
120	44.8	21.9	65.7





#79
sec-Butylbenzene
Concen: 2.056 ug/l
RT: 11.634 min Scan# 3
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:105 Resp: 13879

Ion	Ratio	Lower	Upper
105	100		
134	19.3	15.1	22.7

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#80
Nitrobenzene
Concen: 10.663 ug/l m
RT: 13.222 min Scan# 3711
Delta R.T. 0.019 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 77 Resp: 3316

Ion	Ratio	Lower	Upper
77	100		
123	32.6	18.9	67.1
65	3.7	11.9	15.1#





#81
p-Isopropyltoluene
Concen: 1.962 ug/l
RT: 11.782 min Scan# 3
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:119 Resp: 10454

Ion	Ratio	Lower	Upper
119	100		
134	26.0	20.3	30.5
91	25.3	19.4	29.2



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#82
1,3-Dichlorobenzene
Concen: 2.056 ug/l
RT: 11.737 min Scan# 3249
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:146 Resp: 60800

Ion	Ratio	Lower	Upper
146	100		
111	40.9	32.8	49.2
148	63.3	51.1	76.7





#83
1,4-Dichlorobenzene
Concen: 2.080 ug/l
RT: 11.827 min Scan# 3
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:146 Resp: 6015

Ion	Ratio	Lower	Upper
146	100		
111	38.9	32.1	48.1
148	62.2	50.2	75.4

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#84
n-Butylbenzene
Concen: 1.852 ug/l
RT: 12.200 min Scan# 3393
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 91 Resp: 88457

Ion	Ratio	Lower	Upper
91	100		
92	53.0	41.8	62.8
134	23.3	18.6	28.0





#85
1,2-Dichlorobenzene
Concen: 2.024 ug/l
RT: 12.206 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Tgt Ion:146 Resp: 5751

Ion	Ratio	Lower	Upper
146	100		
111	41.6	21.9	65.7
148	63.8	32.3	96.9

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025





#86
1,2-Dibromo-3-Chloropropane
Concen: 2.036 ug/l
RT: 12.991 min Scan# 3639
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion: 75 Resp: 4329

Ion	Ratio	Lower	Upper
75	100		
155	73.4	63.5	95.3
157	97.4	81.8	122.6





#87
1,2,4-Trichlorobenzene
Concen: 1.896 ug/l
RT: 13.833 min Scan# 3
Delta R.T. 0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:180 Resp: 2628

Ion	Ratio	Lower	Upper
180	100		
182	103.2	76.6	115.0
145	31.9	25.4	38.2



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#88
Hexachlorobutadiene
Concen: 2.082 ug/l
RT: 14.010 min Scan# 3956
Delta R.T. -0.003 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:225 Resp: 20626

Ion	Ratio	Lower	Upper
225	100		
223	61.4	49.5	74.3
227	62.0	51.0	76.4





#89
Naphthalene
Concen: 1.996 ug/l
RT: 14.084 min Scan# 3
Delta R.T. 0.006 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:128 Resp: 4150

Ion	Ratio	Lower	Upper
128	100		
127	13.1	10.6	16.0
129	10.0	8.6	13.0



Instrument :
MSVOA_U
ClientSampleId :
VU0212WBSD01

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/17/2025
Supervised By :Mahesh Dadoda 02/17/2025



#90
1,2,3-Trichlorobenzene
Concen: 1.949 ug/l
RT: 14.322 min Scan# 4053
Delta R.T. -0.000 min
Lab File: VU063252.D
Acq: 12 Feb 2025 13:20

Tgt Ion:180 Resp: 26436

Ion	Ratio	Lower	Upper
180	100		
182	98.2	78.2	117.2
145	30.8	26.1	39.1



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VU021025	Instrument	MSVOA_u

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC0.5	VU063219.D	1,2,3-Trichloropropane	Amit	2/12/2025 4:22:31 PM	MMDadoda	2/12/2025 4:22:33 PM	Peak Integrated by Software
VSTDICC0.5	VU063219.D	2-Butanone	Amit	2/12/2025 4:22:31 PM	MMDadoda	2/12/2025 4:22:33 PM	Peak Integrated by Software
VSTDICC0.5	VU063219.D	2-Hexanone	Amit	2/12/2025 4:22:31 PM	MMDadoda	2/12/2025 4:22:33 PM	Peak Integrated by Software
VSTDICC0.5	VU063219.D	Acrylonitrile	Amit	2/12/2025 4:22:31 PM	MMDadoda	2/12/2025 4:22:33 PM	Peak Integrated by Software
VSTDICC0.5	VU063219.D	Cyclohexane	Amit	2/12/2025 4:22:31 PM	MMDadoda	2/12/2025 4:22:33 PM	Peak Integrated by Software
VSTDICC0.5	VU063219.D	Methacrylonitrile	Amit	2/12/2025 4:22:31 PM	MMDadoda	2/12/2025 4:22:33 PM	Peak Integrated by Software
VSTDICC0.5	VU063219.D	Methyl acrylate	Amit	2/12/2025 4:22:31 PM	MMDadoda	2/12/2025 4:22:33 PM	Peak Integrated by Software
VSTDICC0.5	VU063219.D	Methyl methacrylate	Amit	2/12/2025 4:22:31 PM	MMDadoda	2/12/2025 4:22:33 PM	Peak Integrated by Software
VSTDICC0.5	VU063219.D	Propionitrile	Amit	2/12/2025 4:22:31 PM	MMDadoda	2/12/2025 4:22:33 PM	Peak Integrated by Software
VSTDICC0.5	VU063219.D	t-1,4-Dichloro-2-butene	Amit	2/12/2025 4:22:31 PM	MMDadoda	2/12/2025 4:22:33 PM	Peak Integrated by Software
VSTDICC001	VU063220.D	1,2,3-Trichloropropane	Amit	2/12/2025 4:22:52 PM	MMDadoda	2/12/2025 4:23:07 PM	Peak Integrated by Software
VSTDICC001	VU063220.D	2-Butanone	Amit	2/12/2025 4:22:52 PM	MMDadoda	2/12/2025 4:23:07 PM	Peak Integrated by Software
VSTDICC001	VU063220.D	2-Hexanone	Amit	2/12/2025 4:22:52 PM	MMDadoda	2/12/2025 4:23:07 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VU021025	Instrument	MSVOA_u

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC001	VU063220.D	Cyclohexane	Amit	2/12/2025 4:22:52 PM	MMDadoda	2/12/2025 4:23:07 PM	Peak Integrated by Software
VSTDICC001	VU063220.D	Methyl acrylate	Amit	2/12/2025 4:22:52 PM	MMDadoda	2/12/2025 4:23:07 PM	Peak Integrated by Software
VSTDICC001	VU063220.D	Propionitrile	Amit	2/12/2025 4:22:52 PM	MMDadoda	2/12/2025 4:23:07 PM	Peak Integrated by Software
VSTDICC001	VU063220.D	t-1,4-Dichloro-2-butene	Amit	2/12/2025 4:22:52 PM	MMDadoda	2/12/2025 4:23:07 PM	Peak Integrated by Software
VSTDICC002	VU063221.D	1,2,3-Trichloropropane	Amit	2/12/2025 4:23:13 PM	MMDadoda	2/12/2025 4:23:18 PM	Peak Integrated by Software
VSTDICC002	VU063221.D	Cyclohexane	Amit	2/12/2025 4:23:13 PM	MMDadoda	2/12/2025 4:23:18 PM	Peak Integrated by Software
VSTDICC002	VU063221.D	Methyl acrylate	Amit	2/12/2025 4:23:13 PM	MMDadoda	2/12/2025 4:23:18 PM	Peak Integrated by Software
VSTDICC002	VU063221.D	Nitrobenzene	Amit	2/12/2025 4:23:13 PM	MMDadoda	2/12/2025 4:23:18 PM	Peak Integrated by Software
VSTDICC002	VU063221.D	t-1,4-Dichloro-2-butene	Amit	2/12/2025 4:23:13 PM	MMDadoda	2/12/2025 4:23:18 PM	Peak Integrated by Software
VSTDICC005	VU063222.D	1,2,3-Trichloropropane	Amit	2/12/2025 4:23:29 PM	MMDadoda	2/12/2025 4:23:42 PM	Peak Integrated by Software
VSTDICC005	VU063222.D	Cyclohexane	Amit	2/12/2025 4:23:29 PM	MMDadoda	2/12/2025 4:23:42 PM	Peak Integrated by Software
VSTDICC005	VU063222.D	Methyl acrylate	Amit	2/12/2025 4:23:29 PM	MMDadoda	2/12/2025 4:23:42 PM	Peak Integrated by Software
VSTDICC005	VU063222.D	Nitrobenzene	Amit	2/12/2025 4:23:29 PM	MMDadoda	2/12/2025 4:23:42 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VU021025	Instrument	MSVOA_u

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC005	VU063222.D	t-1,4-Dichloro-2-butene	Amit	2/12/2025 4:23:29 PM	MMDadoda	2/12/2025 4:23:42 PM	Peak Integrated by Software
VSTDICCC010	VU063223.D	1,2,3-Trichloropropane	Amit	2/12/2025 4:23:38 PM	MMDadoda	2/12/2025 4:23:53 PM	Peak Integrated by Software
VSTDICCC010	VU063223.D	Cyclohexane	Amit	2/12/2025 4:23:38 PM	MMDadoda	2/12/2025 4:23:53 PM	Peak Integrated by Software
VSTDICCC010	VU063223.D	t-1,4-Dichloro-2-butene	Amit	2/12/2025 4:23:38 PM	MMDadoda	2/12/2025 4:23:53 PM	Peak Integrated by Software
VSTDICC015	VU063224.D	1,2,3-Trichloropropane	Amit	2/12/2025 4:23:42 PM	MMDadoda	2/12/2025 4:23:57 PM	Peak Integrated by Software
VSTDICC015	VU063224.D	Cyclohexane	Amit	2/12/2025 4:23:42 PM	MMDadoda	2/12/2025 4:23:57 PM	Peak Integrated by Software
VSTDICC015	VU063224.D	t-1,4-Dichloro-2-butene	Amit	2/12/2025 4:23:42 PM	MMDadoda	2/12/2025 4:23:57 PM	Peak Integrated by Software
VSTDICV010	VU063225.D	1,2,3-Trichloropropane	Amit	2/12/2025 4:23:46 PM	MMDadoda	2/12/2025 4:24:01 PM	Peak Integrated by Software
VSTDICV010	VU063225.D	Cyclohexane	Amit	2/12/2025 4:23:46 PM	MMDadoda	2/12/2025 4:24:01 PM	Peak Integrated by Software
VSTDICV010	VU063225.D	t-1,4-Dichloro-2-butene	Amit	2/12/2025 4:23:46 PM	MMDadoda	2/12/2025 4:24:01 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VU021225	Instrument	MSVOA_u

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC010	VU063249.D	1,2,3-Trichloropropane	SAM	2/17/2025 1:09:01 PM	MMDadoda	2/17/2025 1:09:52 PM	Peak Integrated by Software
VSTDCCC010	VU063249.D	Cyclohexane	SAM	2/17/2025 1:09:01 PM	MMDadoda	2/17/2025 1:09:52 PM	Peak Integrated by Software
VSTDCCC010	VU063249.D	Nitrobenzene	SAM	2/17/2025 1:09:01 PM	MMDadoda	2/17/2025 1:09:52 PM	Peak Integrated by Software
VSTDCCC010	VU063249.D	t-1,4-Dichloro-2-butene	SAM	2/17/2025 1:09:01 PM	MMDadoda	2/17/2025 1:09:52 PM	Peak Integrated by Software
VU0212WBS01	VU063251.D	1,2,3-Trichloropropane	SAM	2/17/2025 1:09:02 PM	MMDadoda	2/17/2025 1:09:54 PM	Peak Integrated by Software
VU0212WBS01	VU063251.D	Cyclohexane	SAM	2/17/2025 1:09:02 PM	MMDadoda	2/17/2025 1:09:54 PM	Peak Integrated by Software
VU0212WBS01	VU063251.D	Methyl acrylate	SAM	2/17/2025 1:09:02 PM	MMDadoda	2/17/2025 1:09:54 PM	Peak Integrated by Software
VU0212WBS01	VU063251.D	Nitrobenzene	SAM	2/17/2025 1:09:02 PM	MMDadoda	2/17/2025 1:09:54 PM	Peak Integrated by Software
VU0212WBS01	VU063251.D	Propionitrile	SAM	2/17/2025 1:09:02 PM	MMDadoda	2/17/2025 1:09:54 PM	Peak Integrated by Software
VU0212WBS01	VU063251.D	t-1,4-Dichloro-2-butene	SAM	2/17/2025 1:09:02 PM	MMDadoda	2/17/2025 1:09:54 PM	Peak Integrated by Software
VU0212WBSD01	VU063252.D	1,2,3-Trichloropropane	SAM	2/17/2025 1:09:04 PM	MMDadoda	2/17/2025 1:09:55 PM	Peak Integrated by Software
VU0212WBSD01	VU063252.D	Acrylonitrile	SAM	2/17/2025 1:09:04 PM	MMDadoda	2/17/2025 1:09:55 PM	Peak Integrated by Software
VU0212WBSD01	VU063252.D	Cyclohexane	SAM	2/17/2025 1:09:04 PM	MMDadoda	2/17/2025 1:09:55 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	VU021225	Instrument	MSVOA_u

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VU0212WBSD01	VU063252.D	Methyl acrylate	SAM	2/17/2025 1:09:04 PM	MMDadoda	2/17/2025 1:09:55 PM	Peak Integrated by Software
VU0212WBSD01	VU063252.D	Nitrobenzene	SAM	2/17/2025 1:09:04 PM	MMDadoda	2/17/2025 1:09:55 PM	Peak Integrated by Software
VU0212WBSD01	VU063252.D	t-1,4-Dichloro-2-butene	SAM	2/17/2025 1:09:04 PM	MMDadoda	2/17/2025 1:09:55 PM	Peak Integrated by Software
VSTDCCC010	VU063260.D	1,2,3-Trichloropropane	SAM	2/17/2025 1:09:15 PM	MMDadoda	2/17/2025 1:10:07 PM	Peak Integrated by Software
VSTDCCC010	VU063260.D	Cyclohexane	SAM	2/17/2025 1:09:15 PM	MMDadoda	2/17/2025 1:10:07 PM	Peak Integrated by Software
VSTDCCC010	VU063260.D	Nitrobenzene	SAM	2/17/2025 1:09:15 PM	MMDadoda	2/17/2025 1:10:07 PM	Peak Integrated by Software
VSTDCCC010	VU063260.D	t-1,4-Dichloro-2-butene	SAM	2/17/2025 1:09:15 PM	MMDadoda	2/17/2025 1:10:07 PM	Peak Integrated by Software

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_U

Daily Analysis Runlog For Sequence/QCBatch ID # VU021025

Review By	Mahesh Dadoda	Review On	2/13/2025 12:08:39 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/13/2025 12:10:26 PM		
SubDirectory	VU021025	HP Acquire Method	MSVOA_U	HP Processing Method	524u021025dw.m

STD. NAME	STD REF.#
Tune/Reschk	VP132982
Initial Calibration Stds	VP132975,VP132976,VP132977,VP132978,VP132979,VP132980
CCC	
Internal Standard/PEM	VP132884
ICV/I.BLK	VP132981
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VU063218.D	10 Feb 2025 09:06	MD/SY	Ok
2	VSTDICC0.5	VU063219.D	10 Feb 2025 12:59	MD/SY	Ok,M
3	VSTDICC001	VU063220.D	10 Feb 2025 13:23	MD/SY	Ok,M
4	VSTDICC002	VU063221.D	10 Feb 2025 13:58	MD/SY	Ok,M
5	VSTDICC005	VU063222.D	10 Feb 2025 14:23	MD/SY	Ok,M
6	VSTDICCC010	VU063223.D	10 Feb 2025 15:06	MD/SY	Ok,M
7	VSTDICC015	VU063224.D	10 Feb 2025 15:33	MD/SY	Ok,M
8	VSTDICV010	VU063225.D	10 Feb 2025 16:45	MD/SY	ReRun

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_U

Daily Analysis Runlog For Sequence/QCBatch ID # VU021225

Review By	Semsettin Yesilyurt	Review On	2/17/2025 1:09:26 PM				
Supervise By	Mahesh Dadoda	Supervise On	2/17/2025 1:10:15 PM				
SubDirectory	VU021225	HP Acquire Method	MSVOA_U		HP Processing Method	524u021025dw.m	

STD. NAME	STD REF.#
Tune/Reschk	VP133014
Initial Calibration Stds	
CCC	VP133004,VP133005,MDL VP133006,VP133007,VP133008,LOQ VP133009,VP133013
Internal Standard/PEM	VP132884
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VU063248.D	12 Feb 2025 10:13	MD/SY	Ok
2	VSTDCCC010	VU063249.D	12 Feb 2025 10:48	MD/SY	Ok,M
3	VU0212WBL01	VU063250.D	12 Feb 2025 12:17	MD/SY	Ok
4	VU0212WBS01	VU063251.D	12 Feb 2025 12:56	MD/SY	Ok,M
5	VU0212WBSD01	VU063252.D	12 Feb 2025 13:20	MD/SY	Ok,M
6	Q1172-10	VU063253.D	12 Feb 2025 14:20	MD/SY	Dilution
7	Q1172-10DL	VU063254.D	12 Feb 2025 14:58	MD/SY	Ok
8	Q1168-09	VU063255.D	12 Feb 2025 15:30	MD/SY	Ok,M
9	Q1168-09	VU063256.D	12 Feb 2025 15:55	MD/SY	Ok,M
10	Q1168-09	VU063257.D	12 Feb 2025 16:20	MD/SY	Ok,M
11	Q1168-08	VU063258.D	12 Feb 2025 16:44	MD/SY	Ok,M
12	Q1168-08	VU063259.D	12 Feb 2025 17:09	MD/SY	Ok,M
13	VSTDCCC010	VU063260.D	12 Feb 2025 18:23	MD/SY	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_U

Daily Analysis Runlog For Sequence/QCBatch ID # VU021025

Review By	Mahesh Dadoda	Review On	2/13/2025 12:08:39 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	2/13/2025 12:10:26 PM		
SubDirectory	VU021025	HP Acquire Method	MSVOA_U	HP Processing Method	524u021025dw.m

STD. NAME	STD REF.#
Tune/Reschk	VP132982
Initial Calibration Stds	VP132975,VP132976,VP132977,VP132978,VP132979,VP132980
CCC	
Internal Standard/PEM	VP132884
ICV/I.BLK	VP132981
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VU063218.D	10 Feb 2025 09:06		MD/SY	Ok
2	VSTDICC0.5	VSTDICC0.5	VU063219.D	10 Feb 2025 12:59		MD/SY	Ok,M
3	VSTDICC001	VSTDICC001	VU063220.D	10 Feb 2025 13:23	524.2	MD/SY	Ok,M
4	VSTDICC002	VSTDICC002	VU063221.D	10 Feb 2025 13:58	Drinking water method	MD/SY	Ok,M
5	VSTDICC005	VSTDICC005	VU063222.D	10 Feb 2025 14:23		MD/SY	Ok,M
6	VSTDICCC010	VSTDICCC010	VU063223.D	10 Feb 2025 15:06		MD/SY	Ok,M
7	VSTDICC015	VSTDICC015	VU063224.D	10 Feb 2025 15:33		MD/SY	Ok,M
8	VSTDICV010	ICVVU021025	VU063225.D	10 Feb 2025 16:45	Fail ,rerun	MD/SY	ReRun

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Instrument ID: MSVOA_U

Daily Analysis Runlog For Sequence/QCBatch ID # VU021225

Review By	Semsettin Yesilyurt	Review On	2/17/2025 1:09:26 PM		
Supervise By	Mahesh Dadoda	Supervise On	2/17/2025 1:10:15 PM		
SubDirectory	VU021225	HP Acquire Method	MSVOA_U	HP Processing Method	524u021025dw.m

STD. NAME	STD REF.#
Tune/Reschk	VP133014
Initial Calibration Stds	
CCC	VP133004,VP133005,MDL VP133006,VP133007,VP133008,LOQ VP133009,VP133013
Internal Standard/PEM	VP132884
ICV/I.BLK	
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VU063248.D	12 Feb 2025 10:13		MD/SY	Ok
2	VSTDCCC010	VSTDCCC010	VU063249.D	12 Feb 2025 10:48		MD/SY	Ok,M
3	VU0212WBL01	VU0212WBL01	VU063250.D	12 Feb 2025 12:17		MD/SY	Ok
4	VU0212WBS01	VU0212WBS01	VU063251.D	12 Feb 2025 12:56		MD/SY	Ok,M
5	VU0212WBSD01	VU0212WBSD01	VU063252.D	12 Feb 2025 13:20		MD/SY	Ok,M
6	Q1172-10	PT-ADD-WS	VU063253.D	12 Feb 2025 14:20	PT-ADD ,Need 5X	MD/SY	Dilution
7	Q1172-10DL	PT-ADD-WSDL	VU063254.D	12 Feb 2025 14:58		MD/SY	Ok
8	Q1168-09	MDL-WATER-03-QT1-2	VU063255.D	12 Feb 2025 15:30	mdl-0.25 ppb	MD/SY	Ok,M
9	Q1168-09	MDL-WATER-03-QT1-2	VU063256.D	12 Feb 2025 15:55	mdl-0.4 ppb	MD/SY	Ok,M
10	Q1168-09	MDL-WATER-03-QT1-2	VU063257.D	12 Feb 2025 16:20	mdl-0.8 ppb	MD/SY	Ok,M
11	Q1168-08	LOQ-WATER-02-QT1-2	VU063258.D	12 Feb 2025 16:44	RL-check-0.5ppb	MD/SY	Ok,M
12	Q1168-08	LOQ-WATER-02-QT1-2	VU063259.D	12 Feb 2025 17:09	RL-check-1.0ppb	MD/SY	Ok,M
13	VSTDCCC010	VSTDCCC010EC	VU063260.D	12 Feb 2025 18:23		MD/SY	Ok,M

M : Manual Integration

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Prep Standard - Chemical Standard Summary

Order ID : Q1172

Test : VOCMS Group3

Prepbatch ID :

Sequence ID/Qc Batch ID: VU021225,

Standard ID :
VP131767,VP132883,VP132884,VP133004,VP133005,VP133007,VP133008,VP133013,VP133014,

Chemical ID :
LOQ VP133009,MDL VP133006,V13391,V13879,V14134,V14154,V14419,V14624,V14724,V14756,V14837,W3112,

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
218	BFB, 25PPM	VP131767	11/22/2024	05/18/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 11/27/2024

FROM 0.50000ml of V13391 + 49.50000ml of V14154 = Final Quantity: 50.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
553	524 Calibration CC Mix Working STD, 25 PPM	VP132883	02/05/2025	04/07/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 02/14/2025

FROM 0.12500ml of V13879 + 0.12500ml of V14419 + 0.12500ml of V14756 + 0.12500ml of V14837 + 0.25000ml of V14724 + 9.24600ml of V14624 = Final Quantity: 10.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
552	524 Internal STD and Surrogate Mix, 5 PPM	VP132884	02/05/2025	07/13/2025	Semsettin Yesilyurt	None	None	Mahesh Dadoda 02/14/2025

FROM 0.02500ml of V14134 + 9.97500ml of V14624 = Final Quantity: 10.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1131	10 PPB CCC, 524.2	VP133004	02/12/2025	02/13/2025	Amit Patel	None	None	Mahesh Dadoda 02/14/2025

FROM 39.97600ml of W3112 + 0.00800ml of VP132884 + 0.01600ml of VP132883 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1131	10 PPB CCC, 524.2	VP133005	02/12/2025	02/13/2025	Amit Patel	None	None	Mahesh Dadoda 02/14/2025

FROM 39.97600ml of W3112 + 0.00800ml of VP132884 + 0.01600ml of VP132883 = Final Quantity: 40.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
2411	0.4 PPB 524 MDL	VP133007	02/12/2025	02/13/2025	Amit Patel	None	None	Mahesh Dadoda 02/14/2025

FROM 39.99000ml of W3112 + 0.00060ml of VP132883 + 0.00800ml of VP132884 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
3919	524.2 MDL 0.8 PPB	VP133008	02/12/2025	02/13/2025	Amit Patel	None	None	Mahesh Dadoda 02/14/2025

FROM 39.99000ml of W3112 + 0.00130ml of VP132883 + 0.00800ml of VP132884 = Final Quantity: 40.000 ml

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1898	524 LOD LOQ, 1PPB	VP133013	02/12/2025	02/13/2025	Amit Patel	None	None	Mahesh Dadoda 02/14/2025

FROM 39.99000ml of W3112 + 0.00160ml of VP132883 + 0.00800ml of VP132884 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOC STANDARD PREPARATION LOG

Recipe ID	NAME	NO.	Prep Date	Expiration Date	Prepared By	ScaleID	PipetteID	Supervised By
1580	BFB TUNE CHECK-524.2	VP133014	02/12/2025	02/13/2025	Romaben Patel	None	None	Mahesh Dadoda 02/14/2025

FROM 39.99000ml of W3112 + 0.00160ml of VP131767 = Final Quantity: 40.000 ml

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30067 / BFB tuneing solution	A0191805	11/22/2025	11/22/2024 / SAM	01/13/2023 / SAM	V13391

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	564323 / Custom Oxygenates Standard	A0199211	04/17/2025	10/17/2024 / SAM	06/30/2023 / SAM	V13879

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30201 / VOA Mix,500 series method, 524 Internal Std., 2000ug/mL. P&TM, 1mL/ampul	A0168982	02/05/2026	02/05/2025 / SAM	01/18/2024 / SAM	V14134

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	22L0562016	05/18/2025	11/18/2024 / pedro	02/06/2024 / SAM	V14154

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30601 / VOA Mega Mix, Drinking Water VOA Mega Mix, 524.2 Rev 4.1, 1mL, 2000ug/mL P&TM	A0204639	10/17/2025	10/17/2024 / SAM	06/04/2024 / SAM	V14419

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA9077-02 / Methanol, Purge/Trap (cs=6x1L)	23I0762004	07/13/2025	01/13/2025 / SAM	11/26/2024 / SAM	V14624

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30006 / VOA Mix, CLP method Calibration Std #1 ketones 5000uq/ml, PTM, 1ml	A02110618	07/10/2025	01/10/2025 / SAM	12/17/2024 / SAM	V14724

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	30042 / VOA Mix,500 series method 502.2 Calibration Std #1 gases, 2000uq/ml, PTM, 1ml	A0216826	07/10/2025	01/10/2025 / SAM	12/17/2024 / SAM	V14756

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	560065 / Custom Standard, 524 Std w/ COC [CS 8005]	A0220861	07/20/2025	01/20/2025 / SAM	01/16/2025 / SAM	V14837

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

avantor™ | J.T.Baker®

Material No.: 9077-02
Batch No.: 2310762004
Manufactured Date: 2023-08-11
Expiration Date: 2026-08-10
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.5 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.2
Titrable Base (µeq/g)	≤ 0.10	0.01
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis – Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory,Research,or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Ken Koehnlein
Sr. Manager, Quality Assurance

RESTEK®
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30201 **Lot No.:** A0168982

Description : 524 Internal Std / Surrogate Mix

524 Internal Std/Surrogate Mix 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : February 29, 2028 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound		Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Fluorobenzene **CAS #** 462-06-6 **Purity** 99%	(Lot BCBK8171V)	2,008.0 µg/mL	+/- 11.7841 µg/mL +/- 112.5980 µg/mL +/- 115.2321 µg/mL		Gravimetric Unstressed Stressed
2	1-Bromo-4-fluorobenzene (BFB) **CAS #** 460-00-4 **Purity** 99%	(Lot 20401KO)	2,010.0 µg/mL	+/- 11.7958 µg/mL +/- 112.7101 µg/mL +/- 115.3469 µg/mL		Gravimetric Unstressed Stressed
3	1,2-Dichlorobenzene-d4 **CAS #** 2199-69-1 **Purity** 99%	(Lot M-2097)	2,015.5 µg/mL	+/- 11.8281 µg/mL +/- 113.0185 µg/mL +/- 115.6625 µg/mL		Gravimetric Unstressed Stressed

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler - Operations Tech I

Date Mixed: 11-Feb-2021 **Balance:** 1128360905

Alexis Shelow - Operations Tech I

Date Passed: 12-Feb-2021

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35: The certified combined stressed uncertainty value (includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ stressed} = k\sqrt{U^2_{gravimetric} + U^2_{homogeneity} + U^2_{storage\ stability} + U^2_{shipping\ stability}}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- It is important to note that the shipping stability uncertainty was obtained under temperature extremes for specific time intervals; therefore, the certified combined stressed uncertainty value should only be applied to the product if it was stored at non-standard temperature conditions up to and including 7 days. Contact Restek Technical Service at www.restek.com/Contact-Us for use recommendations if your shipment was in-transit for more than 7 days at non-standard temperature conditions.
- Apply the certified combined unstressed uncertainty value if the product was received under standard shipping conditions. Apply the certified combined stressed uncertainty value if the product was received under non-standard conditions as specified below.

Label Conditions	Standard Conditions	Non-Standard Conditions
25°C Nominal (Room Temperature)	< 60°C	≥ 60°C up to 7 days
10°C or colder (Refrigerate)	< 40°C	≥ 40°C up to 7 days
0°C or colder (Freezer) -20°C or colder (Deep Freezer)	< 25°C	≥ 25°C up to 7 days

- Separate (not combined) uncertainty values for gravimetric uncertainty are also displayed on the certificate, if needed, separate homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty values are available by contacting Restek Technical Service at www.restek.com/Contact-Us.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampules are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampules. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30067 **Lot No.:** A0191805

Description : 4-Bromofluorobenzene Standard

4-Bromofluorobenzene Standard 2,500µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1-Bromo-4-fluorobenzene (BFB)	460-00-4	184975	99%	2,483.9 µg/mL	+/- 139.5488

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Alicia Leathers - Operation Technician I **Date Mixed:** 17-Nov-2022 **Balance Serial #** B251644995

Jennifer Pollino - Operations Tech III - ARM QC **Date Passed:** 21-Nov-2022

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 564323 **Lot No.:** A0199211

Description : Custom Oxygenates Standard

Custom Oxygenates Standard 2,000-10,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : June 30, 2028 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	tert-Butanol (TBA)	75-65-0	101619K21F-1	99%	10,093.2 µg/mL	+/- 125.6116
2	Diisopropyl ether (DIPE)	108-20-3	STBK3450	99%	2,011.0 µg/mL	+/- 25.0950
3	Ethyl-tert-butyl ether (ETBE)	637-92-3	MKCP5997	99%	2,009.8 µg/mL	+/- 25.0800
4	tert-Amyl methyl ether (TAME)	994-05-8	HMBJ0825	99%	2,009.2 µg/mL	+/- 25.0726
5	tert-Amyl ethyl ether (TAEE)	919-94-8	IKVYB	97%	2,010.4 µg/mL	+/- 25.0878

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Bryan Snyder - Operations Tech I

Date Mixed: 22-Jun-2023 **Balance Serial #** 1128342314

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 23-Jun-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus

ilac-MRA A2LA ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA A2LA ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30601 **Lot No.:** A0204639

Description : Drinking Water VOA MegaMix™, 524.2 Rev 4.1

Drinking Water VOA Mega Mix 524.2 Rev 4.1, 2000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2026 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Diethyl ether (ethyl ether)	60-29-7	SHBQ1495	99%	2,016.9 µg/mL	+/- 70.1908
2	1,1-dichloroethene	75-35-4	SHBG8609V	99%	2,009.6 µg/mL	+/- 69.9229
3	Iodomethane (methyl iodide)	74-88-4	MKCN8012	99%	2,016.5 µg/mL	+/- 70.1787
4	Allyl chloride (3-chloropropene)	107-05-1	RD221118RSR	99%	2,017.0 µg/mL	+/- 69.7168
5	Methylene chloride (dichloromethane)	75-09-2	231383	99%	2,013.2 µg/mL	+/- 70.0499
6	Carbon disulfide	75-15-0	N28F701	99%	2,017.0 µg/mL	+/- 70.1961
7	Acrylonitrile	107-13-1	102466R02E	99%	2,017.1 µg/mL	+/- 70.1995
8	Methyl-tert-butyl ether (MTBE)	1634-04-4	SHBP0179	99%	2,017.0 µg/mL	+/- 69.7168
9	trans-1,2-Dichloroethene	156-60-5	MKCP9516	99%	2,011.9 µg/mL	+/- 70.0038
10	1,1-Dichloroethane	75-34-3	852900	99%	2,010.5 µg/mL	+/- 69.9560
11	Propionitrile	107-12-0	BCCH7430	99%	2,017.0 µg/mL	+/- 70.1943
12	2,2-Dichloropropane	594-20-7	RD230426	99%	2,013.2 µg/mL	+/- 70.0652
13	cis-1,2-Dichloroethene	156-59-2	MKCP7830	99%	2,014.0 µg/mL	+/- 70.0903
14	Methacrylonitrile	126-98-7	1012014	99%	2,015.7 µg/mL	+/- 70.1491
15	Methyl acrylate	96-33-3	SHBG6616V	99%	2,019.0 µg/mL	+/- 70.2639
16	chloroform	67-66-3	SHBN8469	99%	2,009.7 µg/mL	+/- 69.9273

17	Bromochloromethane	74-97-5	230810JLM	99%	2,016.0	µg/mL	+/- 70.1613
18	Tetrahydrofuran	109-99-9	SHBQ0910	99%	2,019.6	µg/mL	+/- 70.2865
19	1,1,1-trichloroethane	71-55-6	RD230728RSR	99%	2,011.1	µg/mL	+/- 69.9769
20	1-Chlorobutane (Butyl chloride)	109-69-3	SHBC2651V	99%	2,015.0	µg/mL	+/- 69.6476
21	1,1-Dichloropropene	563-58-6	230825JLM	99%	2,018.9	µg/mL	+/- 70.2629
22	carbon tetrachloride	56-23-5	SHBP4875	99%	2,011.5	µg/mL	+/- 69.9890
23	1,2-Dichloroethane	107-06-2	SHBQ0693	99%	2,008.7	µg/mL	+/- 69.8916
24	Benzene	71-43-2	MKCS3357	99%	2,017.4	µg/mL	+/- 70.2100
25	Trichloroethene	79-01-6	SHBN3720	99%	2,008.3	µg/mL	+/- 69.8786
26	1,2-Dichloropropane	78-87-5	BCBR0882V	99%	2,012.1	µg/mL	+/- 70.0117
27	Methyl methacrylate	80-62-6	MKCQ2756	99%	2,017.7	µg/mL	+/- 70.2204
28	Chloroacetonitrile	107-14-2	MKBG6249V	99%	2,006.0	µg/mL	+/- 69.3366
29	bromodichloromethane	75-27-4	MKCF8470	99%	2,012.6	µg/mL	+/- 70.0273
30	Dibromomethane	74-95-3	10233302	99%	2,014.7	µg/mL	+/- 70.1153
31	2-Nitropropane	79-46-9	BCCB9352	97%	2,015.9	µg/mL	+/- 70.1562
32	cis-1,3-Dichloropropene	10061-01-5	RD230406RSR	99%	2,005.0	µg/mL	+/- 69.7655
33	Toluene	108-88-3	MKCS9989	99%	2,019.0	µg/mL	+/- 70.2643
34	Ethyl methacrylate	97-63-2	MKCN6206	97%	2,015.4	µg/mL	+/- 70.1393
35	trans-1,3-Dichloropropene	10061-02-6	RD230727RSR	99%	2,011.3	µg/mL	+/- 69.9838
36	1,1,2-Trichloroethane	79-00-5	FGB01	99%	2,013.2	µg/mL	+/- 70.0491
37	1,3-Dichloropropane	142-28-9	BCCH5357	99%	2,017.1	µg/mL	+/- 70.2002
38	Tetrachloroethene	127-18-4	SHBQ0051	99%	2,011.5	µg/mL	+/- 69.9908
39	dibromochloromethane	124-48-1	MKCQ4517	99%	2,006.6	µg/mL	+/- 69.8185
40	1,2-Dibromoethane (EDB)	106-93-4	BCCH7113	99%	2,009.0	µg/mL	+/- 69.9176
41	Chlorobenzene	108-90-7	SHBN6640	99%	2,009.8	µg/mL	+/- 69.9299
42	1,1,1,2-Tetrachloroethane	630-20-6	GC01	99%	2,013.8	µg/mL	+/- 70.0833
43	Ethylbenzene	100-41-4	094632L21G	99%	2,006.8	µg/mL	+/- 69.8411
44	m-Xylene	108-38-3	SHBN6673	99%	2,018.7	µg/mL	+/- 70.2559
45	p-Xylene	106-42-3	SHBP5191	99%	2,008.0	µg/mL	+/- 69.8828
46	o-Xylene	95-47-6	SHBN5105	99%	2,016.3	µg/mL	+/- 70.1724
47	Styrene	100-42-5	MKCQ3390	99%	2,014.8	µg/mL	+/- 70.1209
48	Isopropylbenzene (cumene)	98-82-8	Z20D022	99%	2,011.4	µg/mL	+/- 70.0026
49	bromoform	75-25-2	050494L04R	99%	2,009.6	µg/mL	+/- 69.9255
50	1,1,2,2-Tetrachloroethane	79-34-5	OXACF	99%	2,011.7	µg/mL	+/- 69.9986
51	1,2,3-Trichloropropane	96-18-4	Q91-34	98%	2,013.8	µg/mL	+/- 70.0841
52	trans-1,4-dichloro-2-butene	110-57-6	RP231113CTH	94%	2,017.2	µg/mL	+/- 69.7251

53	n-Propylbenzene	103-65-1	095067T18C	99%	2,018.4	µg/mL	+/- 70.2434
54	Bromobenzene	108-86-1	MKCQ7174	99%	2,016.9	µg/mL	+/- 70.1919
55	1,3,5-Trimethylbenzene	108-67-8	BCCF4166	99%	2,017.0	µg/mL	+/- 70.1961
56	2-Chlorotoluene	95-49-8	235783M23T	99%	2,017.8	µg/mL	+/- 70.2253
57	4-Chlorotoluene	106-43-4	BCCG9286	99%	2,014.1	µg/mL	+/- 70.0958
58	tert-Butylbenzene	98-06-6	STBJ1937	99%	2,005.2	µg/mL	+/- 69.7868
59	1,2,4-Trimethylbenzene	95-63-6	MKCS3775	99%	2,015.9	µg/mL	+/- 70.1571
60	Pentachloroethane	76-01-7	13550700	97%	2,012.8	µg/mL	+/- 69.5699
61	sec-Butylbenzene	135-98-8	MKCP2266	99%	2,011.0	µg/mL	+/- 69.9872
62	p-Isopropyltoluene (p-Cymene)	99-87-6	MKCR6143	99%	2,014.6	µg/mL	+/- 70.1111
63	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	2,003.2	µg/mL	+/- 69.7020
64	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	2,015.0	µg/mL	+/- 70.1108
65	n-Butylbenzene	104-51-8	09418JJ	99%	2,005.3	µg/mL	+/- 69.7882
66	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	2,009.0	µg/mL	+/- 69.9020
67	Hexachloroethane	67-72-1	QTORH	99%	2,016.0	µg/mL	+/- 69.6822
68	1,2-Dibromo-3-chloropropane	96-12-8	HBMVB	97%	2,005.1	µg/mL	+/- 69.7821
69	Nitrobenzene	98-95-3	10224044	99%	2,017.9	µg/mL	+/- 70.2256
70	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	2,015.0	µg/mL	+/- 70.1251
71	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	2,001.7	µg/mL	+/- 69.6639
72	Naphthalene	91-20-3	STBL1057	99%	2,008.9	µg/mL	+/- 69.9149
73	1,2,3-Trichlorobenzene	87-61-6	MKBX7627V	99%	2,012.3	µg/mL	+/- 70.0318

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:
helium-constant pressure 30 psi

Temp. Program:
40°C (hold 6 min.) to 240°C
@ 6°C/min. (hold 10 min.)

Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
MSD

Split Vent:
20.0 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

John Friedline - Operations Technician I

Date Mixed: 20-Nov-2023 **Balance Serial #** 1128342314

Dillan Murphy - Operations Technician I

Date Passed: 29-Nov-2023

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 V.4

Certificate of Analysis

chromatographic plus

V14697-to-14726

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A0210618

Description : VOA Calibration Mix #1

VOA Calibration Mix #1 5,000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2027 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Acetone	67-64-1	SHBQ8504	99%	5,014.8 µg/mL	+/- 173.2883
2	2-Butanone (MEK)	78-93-3	SHBQ4704	99%	5,012.4 µg/mL	+/- 173.2054
3	4-Methyl-2-pentanone (MIBK)	108-10-1	SHBP9200	99%	5,011.6 µg/mL	+/- 173.1777
4	2-Hexanone	591-78-6	MKCQ6663	99%	5,013.0 µg/mL	+/- 173.2261

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Quality Confirmation Test

Column:
105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:
hydrogen-constant pressure 11.0 psi.

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
FID

Split Vent:
40 ml/min

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Dakota Parson - Operations Technician I

Date Mixed: 22-Apr-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 12/17/24

CERTIFIED REFERENCE MATERIAL

30 Vial

Certificate of Analysis

chromatographic plus

V14727 to
V14756

ACCREDITED
ISO 17034 Accredited
Reference Material Producer
Certificate #3222.01

ACCREDITED
ISO/IEC 17025 Accredited
Testing Laboratory
Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30042 **Lot No.:** A0216826

Description : 502.2 Calibration Mix #1

502.2 Calibration Mix #1 2,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2031 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Dichlorodifluoromethane (CFC-12)	75-71-8	00022922	99%	2,000.9 µg/mL	+/- 112.4144
2	Chloromethane (methyl chloride)	74-87-3	00022694	99%	2,000.7 µg/mL	+/- 112.3998
3	Vinyl chloride	75-01-4	00015559	99%	2,000.3 µg/mL	+/- 112.3779
4	Bromomethane (methyl bromide)	74-83-9	00017022	99%	2,001.8 µg/mL	+/- 112.4650
5	Chloroethane (ethyl chloride)	75-00-3	107-401039114-1	99%	2,000.1 µg/mL	+/- 112.3700
6	Trichlorofluoromethane (CFC-11)	75-69-4	MKCJ8658	99%	2,000.7 µg/mL	+/- 112.3992

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:
60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:
helium-constant flow 2.0 mL/min.

Temp. Program:
40°C (hold 6 min.) to 100°C
@ 6°C/min.

Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
MSD

Split Vent:
Split ratio 10:1

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Tom Suckar - Mix Technician

Date Mixed: 23-Sep-2024 **Balance Serial #** B707717271

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 04-Oct-2024

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

RESTEK
110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

Rec 01/16/25

CERTIFIED REFERENCE MATERIAL

5 vial

Certificate of Analysis

chromatographic

V14837 to V14841

ACCREDITED ISO 17034 Accredited Reference Material Producer Certificate #3222.01

ACCREDITED ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 560065 **Lot No.:** A0220861

Description : Custom 524 Standard

Custom 524 Standard 2,000-10,000µg/mL, P&T Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2026 **Storage:** 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,1,2-Trichlorotrifluoroethane (CFC-113)	76-13-1	00022779	99%	2,009.0 µg/mL	+/- 69.4402
2	tert-Butanol (TBA)	75-65-0	SHBR5545	99%	10,036.0 µg/mL	+/- 346.8674
3	Acrylonitrile	107-13-1	102466R02E	99%	2,015.0 µg/mL	+/- 69.6476
4	Propionitrile	107-12-0	BCCL0691	99%	8,074.0 µg/mL	+/- 279.0744
5	Tetrahydrofuran	109-99-9	SHBR7392	99%	2,009.0 µg/mL	+/- 69.4402
6	Cyclohexane	110-82-7	SHBS0091	99%	2,014.0 µg/mL	+/- 69.6131
7	Methylcyclohexane	108-87-2	SHBR3777	99%	2,015.0 µg/mL	+/- 69.6476
8	Methyl methacrylate	80-62-6	MKCQ2756	99%	2,011.0 µg/mL	+/- 69.5094
9	trans-1,4-dichloro-2-butene	110-57-6	RD240719ECSB	97%	2,013.7 µg/mL	+/- 69.6034
10	Nitrobenzene	98-95-3	10224044	99%	8,026.0 µg/mL	+/- 277.4153

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: P&T Methanol

CAS # 67-56-1

Purity 99%

Quality Confirmation Test

Column:
60m x 0.25mm x 1.4µm
Rtx-502.2 (cat.#10916)

Carrier Gas:
helium-constant pressure 30 psi

Temp. Program:
40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:
200°C

Det. Temp:
250°C

Det. Type:
MSD

Split Vent:
25.0 ml/min.

Inj. Vol
1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 07-Jan-2025 **Balance Serial #** 1128342314

Dillan Murphy - Operations Technician I

Date Passed: 10-Jan-2025

Manufactured under Restek's ISO 9001:2015 Registered Quality System Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k\sqrt{u^2_{gravimetric} + u^2_{homogeneity} + u^2_{storage\ stability} + u^2_{shipping\ stability}}$$

  k is a coverage factor of 2, which gives a level of confidence of approximately 95%.
- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

avantor™ | J.T.Baker®

Material No.: 9077-02
Batch No.: 22L0562016
Manufactured Date: 2022-10-26
Expiration Date: 2025-10-25
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	≥ 99.9 %	100.0 %
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titrable Acid (µeq/g)	≤ 0.3	0.2
Titrable Base (µeq/g)	≤ 0.10	0.03
Water (by KF, coulometric)	≤ 0.08 %	< 0.01 %
Volatile Organic Trace Analysis - Below EPA 8260B CRQL	Conforms	Conforms

For Laboratory,Research,or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

James T Ethier
Jamie Ethier
Vice President Global Quality

SHIPPING DOCUMENTS

phenova® | A Phenomenex® Company
Certified Reference Materials

Packing List

Date	Order #
01/13/2025	333291

6390 Joyce Dr., #100
Golden, CO 80403

Tel: +1-303-940-0033
Fax: +1-303-940-0043
info@phenova.com
www.phenova.com

For terms and conditions of your order, please visit:
www.phenova.com/home/termsofsale

Ship To

Alliance Tech Group - Newark
ATTN: Sohil Jodhani
284 Sheffield St., #1
Mountainside, NJ 07092
USA

Received by: MA
01/15/2025 10:00

Customer PO #	Terms	PT Acct #	Customer #	Ship Via	F.O.B.
PO2-974	Net 30	ZCM-100	1500470	FedEx 2nd Day	Golden, CO

Qty Ordered	Qty Shipped	Qty Backorder	Part Number	Part Description	Study Number	Lot Number
			PT-TMSET-WS	WS Trace Metals Set : (TM and HG)		
1	1	0	PT-TM-WS	WS Trace Metals 1	WS0125	9099-04B
1	1	0	PT-HG-WS	WS Trace Metals Mercury	WS0125	9099-05
1	1	0	PT-MIN-WS	WS Minerals Only	WS0125	9099-51
1	1	0	PT-TURB-WS	WS Turbidity	WS0125	9099-13
1	1	0	PT-SIO2-WS	WS Silica	WS0125	9099-17
1	1	0	PT-RVOA-WS	WS Regulated Volatiles	WS0125	9099-21
1	1	0	PT-UNRVOA-WS	WS Unregulated Volatiles	WS0125	9099-22
1	1	0	PT-THM-WS	WS Trihalomethanes	WS0125	9099-23
1	1	0	PT-EDBCP-WS	WS EDB/DBCP/TCP	WS0125	9099-27
1	1	0	PT-ADD-WS	WS Gasoline Additives	WS0125	9099-36

CHEMTECH 284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488