

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1172

Client: Chemtech Consulting Group

Analytical Method: SW524.2

Datafile : VU063231.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VU0211WBSD01	Chloroform	2	1.90	ug/L	95	5		70	130	20
	Bromodichloromethane	2	2.00	ug/L	100	0		70	130	20
	Dibromochloromethane	2	1.90	ug/L	95	5		70	130	20
	Bromoform	2	1.80	ug/L	90	5		70	130	20