

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172 SDG NO.: Q1172
 Lab File ID: VU063228.D Date Analyzed: 02/11/2025
 Instrument ID: MSVOA_U Time Analyzed: 10:01
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	58984	6.11	0	0.00	0	0.00
UPPER LIMIT	76679.2	6.607	0		0	
LOWER LIMIT	41288.8	5.607	0		0	
EPA SAMPLE NO.						
PT-THM-WS	50312	6.11	0	0.00	0	0.00
PT-THM-WSDL	47942	6.10	0	0.00	0	0.00
VU0211WBL01	52216	6.11	0	0.00	0	0.00
VU0211WBS01	55425	6.11	0	0.00	0	0.00
VU0211WBSD01	54780	6.10	0	0.00	0	0.00

IS1 = Fluorobenzene

IS2 =

IS3 =

AREA UPPER LIMIT = +30% of internal standard area

AREA LOWER LIMIT = -30% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.