

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1172 SAS No.: Q1172 SDG No.: Q1172

Instrument ID: MSVOA_U Calibration Date/Time: 02/11/2025 10:01

Lab File ID: VU063228.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:59 15:33

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
1,2-Dibromoethane	0.206	0.190		-7.77	30
1,2,3-Trichloropropane	0.222	0.211		-4.95	30
1,2-Dibromo-3-Chloropropane	0.042	0.039		-7.14	30
1,2-Dichlorobenzene-d4	0.343	0.340		-0.88	30
4-Bromofluorobenzene	0.330	0.353		6.97	30

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.