

LAB CHRONICLE

OrderID:	Q1173	OrderDate:	1/23/2025 3:29:00 PM
Client:	Tetra Tech NUS, Inc.	Project:	CTO WE13
Contact:	Ernie Wu	Location:	N41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1173-02	BP-VPB-192-GW-60-6 2	Water			01/21/25			01/23/25
			SVOC-SIMGroup1	8270-Modified		01/24/25	01/29/25	
Q1173-03	BP-VPB-192-GW-100- 102	Water			01/21/25			01/23/25
			SVOC-SIMGroup1	8270-Modified		01/24/25	01/29/25	
Q1173-04	BP-VPB-192-GW-140- 142	Water			01/22/25			01/23/25
			SVOC-SIMGroup1	8270-Modified		01/24/25	01/29/25	
Q1173-05	BP-VPB-192-GW-200- 202	Water			01/22/25			01/23/25
			SVOC-SIMGroup1	8270-Modified		01/24/25	01/30/25	
Q1173-06	VPB192-HYD-202501 22	Water			01/22/25			01/23/25
			SVOC-SIMGroup1	8270-Modified		01/24/25	01/30/25	
Q1173-06RE	VPB192-HYD-202501 22RE	Water			01/22/25			01/23/25
			SVOC-SIMGroup1	8270-Modified		01/24/25	02/05/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q1173
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : BP-VPB-192-GW-140-142								
Q1173-04	BP-VPB-192-GW-140-14 WATER	1,4-Dioxane	0.160	J	0.1	0.28	0.28	ug/L
		Total Svoc :			0.16			
		Total Concentration:			0.16			
Client ID : BP-VPB-192-GW-200-202								
Q1173-05	BP-VPB-192-GW-200-20 WATER	1,4-Dioxane	0.250		0.08	0.24	0.24	ug/L
		Total Svoc :			0.25			
		Total Concentration:			0.25			



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/21/25
Project:	CTO WE13		Date Received:	01/23/25
Client Sample ID:	BP-VPB-192-GW-60-62		SDG No.:	Q1173
Lab Sample ID:	Q1173-02		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	590	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036119.D	1	01/24/25 11:00	01/29/25 22:19	PB166237

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.34	U	0.12	0.34	0.34	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.38		30 - 150		94%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		102%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.39		55 - 111		98%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.33		53 - 106		83%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.52		58 - 132		129%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2590	7.803				
1146-65-2	Naphthalene-d8	6320	10.6				
15067-26-2	Acenaphthene-d10	3480	14.442				
1517-22-2	Phenanthrene-d10	7080	17.186				
1719-03-5	Chrysene-d12	5870	21.367				
1520-96-3	Perylene-d12	5930	23.672				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/21/25
Project:	CTO WE13		Date Received:	01/23/25
Client Sample ID:	BP-VPB-192-GW-100-102		SDG No.:	Q1173
Lab Sample ID:	Q1173-03		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	450	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036120.D	1	01/24/25 11:00	01/29/25 22:55	PB166237

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.44	U	0.15	0.44	0.44	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.40		30 - 150		99%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.43		30 - 150		108%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.42		55 - 111		104%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		90%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.59	*	58 - 132		147%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2450		7.803			
1146-65-2	Naphthalene-d8	5670		10.59			
15067-26-2	Acenaphthene-d10	3080		14.442			
1517-22-2	Phenanthrene-d10	6240		17.186			
1719-03-5	Chrysene-d12	5340		21.367			
1520-96-3	Perylene-d12	5630		23.672			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	01/22/25
Project:	CTO WE13	Date Received:	01/23/25
Client Sample ID:	BP-VPB-192-GW-140-142	SDG No.:	Q1173
Lab Sample ID:	Q1173-04	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	710 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036121.D	1	01/24/25 11:00	01/29/25 23:31	PB166237

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.16	J	0.10	0.28	0.28	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		74%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 - 150		92%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		77%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.28		53 - 106		69%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.41		58 - 132		102%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2190	7.803				
1146-65-2	Naphthalene-d8	4950	10.6				
15067-26-2	Acenaphthene-d10	2530	14.441				
1517-22-2	Phenanthrene-d10	4820	17.186				
1719-03-5	Chrysene-d12	4380	21.367				
1520-96-3	Perylene-d12	4970	23.672				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	01/22/25
Project:	CTO WE13	Date Received:	01/23/25
Client Sample ID:	BP-VPB-192-GW-200-202	SDG No.:	Q1173
Lab Sample ID:	Q1173-05	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	830 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036122.D	1	01/24/25 11:00	01/30/25 00:07	PB166237

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.25		0.080	0.24	0.24	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.40		30 - 150		100%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.47		30 - 150		118%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.42		55 - 111		104%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.34		53 - 106		84%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.60	*	58 - 132		149%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2210		7.803			
1146-65-2	Naphthalene-d8	5160		10.59			
15067-26-2	Acenaphthene-d10	3020		14.437			
1517-22-2	Phenanthrene-d10	6550		17.181			
1719-03-5	Chrysene-d12	5370		21.367			
1520-96-3	Perylene-d12	4990		23.671			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	01/22/25
Project:	CTO WE13	Date Received:	01/23/25
Client Sample ID:	VPB192-HYD-20250122	SDG No.:	Q1173
Lab Sample ID:	Q1173-06	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	890 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036123.D	1	01/24/25 11:00	01/30/25 00:43	PB166237

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.22	U	0.080	0.22	0.22	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.035	*	30 - 150		9%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.36		30 - 150		90%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.40		55 - 111		100%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.35		53 - 106		88%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.56	*	58 - 132		140%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2520	7.803				
1146-65-2	Naphthalene-d8	5600	10.6				
15067-26-2	Acenaphthene-d10	3000	14.441				
1517-22-2	Phenanthrene-d10	5700	17.186				
1719-03-5	Chrysene-d12	4770	21.367				
1520-96-3	Perylene-d12	4940	23.672				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	01/22/25
Project:	CTO WE13	Date Received:	01/23/25
Client Sample ID:	VPB192-HYD-20250122RE	SDG No.:	Q1173
Lab Sample ID:	Q1173-06RE	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	890 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036306.D	1	01/24/25 11:00	02/05/25 20:35	PB166237

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.22	U	0.080	0.22	0.22	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.032	*	30 - 150		8%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.35		30 - 150		88%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		55 - 111		94%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.33		53 - 106		82%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.54	*	58 - 132		135%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2110	7.775				
1146-65-2	Naphthalene-d8	4570	10.562				
15067-26-2	Acenaphthene-d10	2300	14.409				
1517-22-2	Phenanthrene-d10	4460	17.161				
1719-03-5	Chrysene-d12	3800	21.34				
1520-96-3	Perylene-d12	3360	23.628				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q1173

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166237BL	PB166237BL	2-Methylnaphthalene-d10	0.4	0.40	100		30	150
		Fluoranthene-d10	0.4	0.44	110		30	150
		Nitrobenzene-d5	0.4	0.43	107		55	111
		2-Fluorobiphenyl	0.4	0.38	95		53	106
		Terphenyl-d14	0.4	0.46	115		58	132
PB166237BS	PB166237BS	2-Methylnaphthalene-d10	0.4	0.58	144		30	150
		Fluoranthene-d10	0.4	0.41	101		30	150
		Nitrobenzene-d5	0.4	0.43	108		55	111
		2-Fluorobiphenyl	0.4	0.40	99		53	106
		Terphenyl-d14	0.4	0.51	126		58	132
PB166237BSD	PB166237BSD	2-Methylnaphthalene-d10	0.4	0.60	150		30	150
		Fluoranthene-d10	0.4	0.41	103		30	150
		Nitrobenzene-d5	0.4	0.45	112	*	55	111
		2-Fluorobiphenyl	0.4	0.40	100		53	106
		Terphenyl-d14	0.4	0.52	131		58	132
Q1173-02	BP-VPB-192-GW-60-62	2-Methylnaphthalene-d10	0.4	0.38	94		30	150
		Fluoranthene-d10	0.4	0.41	102		30	150
		Nitrobenzene-d5	0.4	0.39	98		55	111
		2-Fluorobiphenyl	0.4	0.33	83		53	106
		Terphenyl-d14	0.4	0.52	129		58	132
Q1173-03	BP-VPB-192-GW-100-102	2-Methylnaphthalene-d10	0.4	0.40	99		30	150
		Fluoranthene-d10	0.4	0.43	108		30	150
		Nitrobenzene-d5	0.4	0.42	104		55	111
		2-Fluorobiphenyl	0.4	0.36	90		53	106
		Terphenyl-d14	0.4	0.59	147	*	58	132
Q1173-04	BP-VPB-192-GW-140-142	2-Methylnaphthalene-d10	0.4	0.30	74		30	150
		Fluoranthene-d10	0.4	0.37	92		30	150
		Nitrobenzene-d5	0.4	0.31	77		55	111
		2-Fluorobiphenyl	0.4	0.28	69		53	106
		Terphenyl-d14	0.4	0.41	102		58	132
Q1173-05	BP-VPB-192-GW-200-202	2-Methylnaphthalene-d10	0.4	0.40	100		30	150
		Fluoranthene-d10	0.4	0.47	118		30	150
		Nitrobenzene-d5	0.4	0.42	104		55	111
		2-Fluorobiphenyl	0.4	0.34	84		53	106
		Terphenyl-d14	0.4	0.60	149	*	58	132
Q1173-06	VPB192-HYD-20250122	2-Methylnaphthalene-d10	0.4	0.035	9	*	30	150
		Fluoranthene-d10	0.4	0.36	90		30	150
		Nitrobenzene-d5	0.4	0.40	100		55	111
		2-Fluorobiphenyl	0.4	0.35	88		53	106
		Terphenyl-d14	0.4	0.56	140	*	58	132
Q1173-06RE	VPB192-HYD-20250122RE	2-Methylnaphthalene-d10	0.4	0.032	8	*	30	150
		Fluoranthene-d10	0.4	0.35	88		30	150
		Nitrobenzene-d5	0.4	0.37	94		55	111
		2-Fluorobiphenyl	0.4	0.33	82		53	106
		Terphenyl-d14	0.4	0.54	135	*	58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1173

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified **DataFile:** BN036124.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB166237BS	1,4-Dioxane	0.4	0.34	ug/L	85				70	130	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1173

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified **DataFile:** BN036125.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB166237BSD	1,4-Dioxane	0.4	0.35	ug/L	88	3			70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB166237BL

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q1173

SAS No.: Q1173 SDG NO.: Q1173

Lab File ID: BN036329.D

Lab Sample ID: PB166237BL

Instrument ID: BNA_N

Date Extracted: 01/24/2025

Matrix: (soil/water) Water

Date Analyzed: 02/06/2025

Level: (low/med) LOW

Time Analyzed: 11:03

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB166237BS	PB166237BS	BN036124.D	01/30/2025
PB166237BSD	PB166237BSD	BN036125.D	01/30/2025
BP-VPB-192-GW-60-62	Q1173-02	BN036119.D	01/29/2025
BP-VPB-192-GW-100-102	Q1173-03	BN036120.D	01/29/2025
BP-VPB-192-GW-140-142	Q1173-04	BN036121.D	01/29/2025
BP-VPB-192-GW-200-202	Q1173-05	BN036122.D	01/30/2025
VPB192-HYD-20250122	Q1173-06	BN036123.D	01/30/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1173

SDG NO.: Q1173

Lab File ID: BN036009.D

DFTPP Injection Date: 01/22/2025

Instrument ID: BNA_N

DFTPP Injection Time: 09:44

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	48.9
68	Less than 2.0% of mass 69	0.5 (1.1) 1
69	Mass 69 relative abundance	45.7
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	47.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
275	10.0 - 60.0% of mass 198	24.1
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	9.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	11.5 (20.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036010.D	01/22/2025	11:02
SSTDICC0.2	SSTDICC0.2	BN036011.D	01/22/2025	11:38
SSTDICCC0.4	SSTDICCC0.4	BN036012.D	01/22/2025	12:13
SSTDICC0.8	SSTDICC0.8	BN036013.D	01/22/2025	12:49
SSTDICC1.6	SSTDICC1.6	BN036014.D	01/22/2025	13:25
SSTDICC3.2	SSTDICC3.2	BN036015.D	01/22/2025	14:01
SSTDICC5.0	SSTDICC5.0	BN036016.D	01/22/2025	14:36

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1173

SDG NO.: Q1173

Lab File ID: BN036111.D

DFTPP Injection Date: 01/29/2025

Instrument ID: BNA_N

DFTPP Injection Time: 17:27

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	42.5
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	42.8
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	47.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.2
275	10.0 - 60.0% of mass 198	26.8
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	11.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	12.7 (18.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036112.D	01/29/2025	18:06
BP-VPB-192-GW-60-62	Q1173-02	BN036119.D	01/29/2025	22:19
BP-VPB-192-GW-100-102	Q1173-03	BN036120.D	01/29/2025	22:55
BP-VPB-192-GW-140-142	Q1173-04	BN036121.D	01/29/2025	23:31
BP-VPB-192-GW-200-202	Q1173-05	BN036122.D	01/30/2025	00:07
VPB192-HYD-20250122	Q1173-06	BN036123.D	01/30/2025	00:43
PB166237BS	PB166237BS	BN036124.D	01/30/2025	01:19
PB166237BSD	PB166237BSD	BN036125.D	01/30/2025	01:55
SSTDCCC0.4EC	SSTDCCC0.4	BN036128.D	01/30/2025	03:43

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1173 SDG NO.: Q1173

Lab File ID: BN036302.D

DFTPP Injection Date: 02/05/2025

Instrument ID: BNA_N

DFTPP Injection Time: 18:08

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	49.1
68	Less than 2.0% of mass 69	0.1 (0.3) 1
69	Mass 69 relative abundance	45.4
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	47.2
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	26.2
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	9.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	11.5 (18.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036303.D	02/05/2025	18:47
VPB192-HYD-20250122RE	Q1173-06RE	BN036306.D	02/05/2025	20:35
SSTDCCC0.4EC	SSTDCCC0.4	BN036318.D	02/06/2025	03:45

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1173

SDG NO.: Q1173

Lab File ID: BN036319.D

DFTPP Injection Date: 02/06/2025

Instrument ID: BNA_N

DFTPP Injection Time: 05:01

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	50.8
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	47
70	Less than 2.0% of mass 69	0.2 (0.4) 1
127	10.0 - 80.0% of mass 198	48
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	8.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	9.5 (18.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036320.D	02/06/2025	05:40
PB166237BL	PB166237BL	BN036329.D	02/06/2025	11:03
SSTDCCC0.4EC	SSTDCCC0.4	BN036332.D	02/06/2025	12:54

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG NO.: Q1173

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 01/29/2025

Lab File ID: BN036112.D Time Analyzed: 18:06

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2094	7.803	4943	10.60	2715	14.44
UPPER LIMIT	4188	8.303	9886	11.1	5430	14.937
LOWER LIMIT	1047	7.303	2471.5	10.1	1357.5	13.937
EPA SAMPLE NO.						
01 PB166237BS	2273	7.80	5072	10.60	2643	14.44
02 PB166237BSD	2190	7.80	4796	10.59	2534	14.44
03 BP-VPB-192-GW-60-62	2585	7.80	6318	10.60	3478	14.44
04 BP-VPB-192-GW-100-102	2447	7.80	5665	10.59	3082	14.44
05 BP-VPB-192-GW-140-142	2194	7.80	4945	10.60	2527	14.44
06 BP-VPB-192-GW-200-202	2210	7.80	5163	10.59	3019	14.44
07 VPB192-HYD-20250122	2520	7.80	5602	10.60	2998	14.44

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG NO.: Q1173

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 01/29/2025

Lab File ID: BN036112.D Time Analyzed: 18:06

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	5698	17.194	4562	21.376	5021	23.677
UPPER LIMIT	11396	17.694	9124	21.876	10042	24.177
LOWER LIMIT	2849	16.694	2281	20.876	2510.5	23.177
EPA SAMPLE NO.						
01 PB166237BS	4937	17.19	3907	21.38	4551	23.67
02 PB166237BSD	4885	17.18	3758	21.38	4193	23.67
03 BP-VPB-192-GW-60-62	7077	17.19	5874	21.37	5930	23.67
04 BP-VPB-192-GW-100-102	6244	17.19	5342	21.37	5628	23.67
05 BP-VPB-192-GW-140-142	4819	17.19	4376	21.37	4969	23.67
06 BP-VPB-192-GW-200-202	6554	17.18	5368	21.37	4990	23.67
07 VPB192-HYD-20250122	5695	17.19	4766	21.37	4937	23.67

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG NO.: Q1173

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 02/05/2025

Lab File ID: BN036303.D Time Analyzed: 18:47

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2214	7.775	4864	10.56	2378	14.41
UPPER LIMIT	4428	8.275	9728	11.062	4756	14.909
LOWER LIMIT	1107	7.275	2432	10.062	1189	13.909
EPA SAMPLE NO.						
01 VPB192-HYD-20250122RE	2111	7.78	4569	10.56	2296	14.41

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG NO.: Q1173

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 02/05/2025

Lab File ID: BN036303.D Time Analyzed: 18:47

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	4362	17.161	3774	21.34	4195	23.628
UPPER LIMIT	8724	17.661	7548	21.84	8390	24.128
LOWER LIMIT	2181	16.661	1887	20.84	2097.5	23.128
EPA SAMPLE NO.						
01 VPB192-HYD-20250122RE	4463	17.16	3799	21.34	3362	23.63

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG NO.: Q1173

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 02/06/2025

Lab File ID: BN036320.D Time Analyzed: 05:40

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2017	7.775	4328	10.56	2158	14.41
UPPER LIMIT	4034	8.275	8656	11.062	4316	14.909
LOWER LIMIT	1008.5	7.275	2164	10.062	1079	13.909
EPA SAMPLE NO.						
01 PB166237BL	2019	7.78	4128	10.57	2036	14.42

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG NO.: Q1173

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 02/06/2025

Lab File ID: BN036320.D Time Analyzed: 05:40

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	4298	17.149	3803	21.34	4157	23.627
UPPER LIMIT	8596	17.649	7606	21.84	8314	24.127
LOWER LIMIT	2149	16.649	1901.5	20.84	2078.5	23.127
EPA SAMPLE NO.						
01 PB166237BL	3826	17.17	3617	21.35	3788	23.63

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	CTO WE13		Date Received:	
Client Sample ID:	PB166237BL		SDG No.:	Q1173
Lab Sample ID:	PB166237BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036329.D	1	01/24/25 11:00	02/06/25 11:03	PB166237

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.40		30 - 150		100%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.44		30 - 150		110%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.43		55 - 111		107%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.38		53 - 106		95%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.46		58 - 132		115%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2020	7.775				
1146-65-2	Naphthalene-d8	4130	10.573				
15067-26-2	Acenaphthene-d10	2040	14.42				
1517-22-2	Phenanthrene-d10	3830	17.173				
1719-03-5	Chrysene-d12	3620	21.349				
1520-96-3	Perylene-d12	3790	23.627				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	CTO WE13		Date Received:	
Client Sample ID:	PB166237BS		SDG No.:	Q1173
Lab Sample ID:	PB166237BS		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036124.D	1	01/24/25 11:00	01/30/25 01:19	PB166237

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.34		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.58		30 - 150		144%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		101%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.43		55 - 111		108%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.40		53 - 106		99%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.51		58 - 132		126%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2270	7.802				
1146-65-2	Naphthalene-d8	5070	10.6				
15067-26-2	Acenaphthene-d10	2640	14.441				
1517-22-2	Phenanthrene-d10	4940	17.186				
1719-03-5	Chrysene-d12	3910	21.375				
1520-96-3	Perylene-d12	4550	23.671				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	CTO WE13		Date Received:	
Client Sample ID:	PB166237BSD		SDG No.:	Q1173
Lab Sample ID:	PB166237BSD		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036125.D	1	01/24/25 11:00	01/30/25 01:55	PB166237

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.35		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.60		30 - 150		150%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		103%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.45	*	55 - 111		112%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.40		53 - 106		100%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.52		58 - 132		131%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2190	7.803				
1146-65-2	Naphthalene-d8	4800	10.59				
15067-26-2	Acenaphthene-d10	2530	14.436				
1517-22-2	Phenanthrene-d10	4890	17.181				
1719-03-5	Chrysene-d12	3760	21.375				
1520-96-3	Perylene-d12	4190	23.671				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN012225.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Jan 23 00:34:56 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN036010.D 0.2 =BN036011.D 0.4 =BN036012.D 0.8 =BN036013.D 1.6 =BN036014.D 3.2 =BN036015.D 5.0 =BN036016.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.452	0.460	0.469	0.477	0.447	0.408	0.417	0.447	5.81
3)	n-Nitrosodimet...	0.798	0.749	0.877	0.883	0.829	0.781	0.759	0.811	6.65
4) S	2-Fluorophenol	1.032	1.012	1.092	1.099	1.042	0.997	1.010	1.040	3.88
5) S	Phenol-d6	1.284	1.195	1.270	1.155	1.230	1.210	1.209	1.222	3.61
6)	bis(2-Chloroet...	1.024	0.979	1.056	0.929	0.993	0.952	0.952	0.984	4.53
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.377	0.356	0.399	0.333	0.397	0.388	0.394	0.378	6.55
9)	Naphthalene	1.149	1.141	1.250	1.137	1.184	1.141	1.131	1.162	3.68
10)	Hexachlorobuta...	0.383	0.369	0.404	0.371	0.388	0.359	0.353	0.375	4.74
11) SURR	2-Methylnaphth...	0.522	0.527	0.578	0.528	0.556	0.550	0.545	0.544	3.66
12)	2-Methylnaphth...	0.702	0.688	0.760	0.700	0.741	0.735	0.721	0.721	3.58
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.240	0.238	0.256	0.238	0.268	0.275	0.282	0.257	7.32
15) S	2-Fluorobiphenyl	1.806	1.736	1.934	1.787	1.819	1.693	1.724	1.786	4.47
16)	Acenaphthylene	1.835	1.826	2.011	1.840	1.940	1.889	1.936	1.897	3.65
17)	Acenaphthene	1.248	1.236	1.365	1.266	1.338	1.310	1.327	1.299	3.78
18)	Fluorene	1.583	1.482	1.633	1.550	1.739	1.703	1.700	1.627	5.76
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.071	0.081	0.095	0.089	0.101	0.107	0.108	0.093	14.73
21)	4-Bromophenyl-...	0.285	0.269	0.307	0.287	0.293	0.273	0.281	0.285	4.42
22)	Hexachlorobenzene	0.391	0.358	0.407	0.374	0.380	0.355	0.361	0.375	5.08
23)	Atrazine	0.185	0.194	0.218	0.204	0.216	0.209	0.215	0.206	6.05
24)	Pentachlorophenol	0.131	0.131	0.164	0.155	0.179	0.185	0.192	0.162	15.18
25)	Phenanthrene	1.154	1.158	1.302	1.172	1.226	1.182	1.219	1.202	4.33
26)	Anthracene	1.019	1.016	1.151	1.064	1.128	1.123	1.151	1.093	5.42
27) SURR	Fluoranthene-d10	1.005	1.006	1.111	0.994	0.959	1.078	1.101	1.036	5.75
28)	Fluoranthene	1.312	1.350	1.507	1.357	1.317	1.506	1.533	1.412	6.99
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.657	1.588	1.693	1.636	1.646	1.552	1.575	1.621	3.12
31) S	Terphenyl-d14	0.821	0.807	0.871	0.831	0.860	0.804	0.822	0.831	3.09
32)	Benzo(a)anthra...	1.445	1.403	1.503	1.411	1.513	1.448	1.433	1.451	2.93
33)	Chrysene	1.501	1.476	1.545	1.448	1.515	1.435	1.463	1.483	2.63
34)	Bis(2-ethylhex...	0.919	0.793	0.798	0.748	0.791	0.748	0.768	0.795	7.36
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN012225.M

36)	Indeno(1,2,3-c...	1.525	1.477	1.621	1.585	1.669	1.668	1.692	1.605	5.03
37)	Benzo(b)fluora...	1.443	1.380	1.497	1.429	1.475	1.444	1.510	1.454	3.03
38)	Benzo(k)fluora...	1.427	1.378	1.486	1.427	1.519	1.496	1.524	1.465	3.76
39) C	Benzo(a)pyrene	1.237	1.164	1.263	1.203	1.264	1.265	1.296	1.242	3.61
40)	Dibenzo(a,h)an...	1.187	1.169	1.290	1.279	1.337	1.338	1.356	1.279	5.86
41)	Benzo(g,h,i)pe...	1.338	1.308	1.426	1.387	1.438	1.428	1.436	1.394	3.75

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG No.: Q1173

Instrument ID: BNA_N Calibration Date/Time: 01/29/2025 18:06

Lab File ID: BN036112.D Init. Calib. Date(s): 01/22/2025 01/22/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:02 14:36

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.594		9.2	20.0
Fluoranthene-d10	1.036	1.092		5.4	20.0
2-Fluorophenol	1.040	1.195		14.9	20.0
Phenol-d6	1.222	1.488		21.8	20.0
Nitrobenzene-d5	0.378	0.398		5.3	20.0
2-Fluorobiphenyl	1.786	1.681		-5.9	20.0
2,4,6-Tribromophenol	0.257	0.246		-4.3	20.0
Terphenyl-d14	0.831	0.973		17.1	20.0
1,4-Dioxane	0.447	0.405		-9.4	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG No.: Q1173

Instrument ID: BNA_N Calibration Date/Time: 01/30/2025 03:43

Lab File ID: BN036128.D Init. Calib. Date(s): 01/22/2025 01/22/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:02 14:36

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.601		10.5	50.0
Fluoranthene-d10	1.036	1.159		11.9	50.0
2-Fluorophenol	1.040	1.189		14.3	50.0
Phenol-d6	1.222	1.499		22.7	50.0
Nitrobenzene-d5	0.378	0.413		9.3	50.0
2-Fluorobiphenyl	1.786	1.729		-3.2	50.0
2,4,6-Tribromophenol	0.257	0.251		-2.3	50.0
Terphenyl-d14	0.831	1.004		20.8	50.0
1,4-Dioxane	0.447	0.408		-8.7	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG No.: Q1173

Instrument ID: BNA_N Calibration Date/Time: 02/05/2025 18:47

Lab File ID: BN036303.D Init. Calib. Date(s): 01/22/2025 01/22/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:02 14:36

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.565		3.9	20.0
Fluoranthene-d10	1.036	1.052		1.5	20.0
2-Fluorophenol	1.040	1.145		10.1	20.0
Phenol-d6	1.222	1.324		8.3	20.0
Nitrobenzene-d5	0.378	0.409		8.2	20.0
2-Fluorobiphenyl	1.786	1.693		-5.2	20.0
2,4,6-Tribromophenol	0.257	0.198		-23.0	20.0
Terphenyl-d14	0.831	0.860		3.5	20.0
1,4-Dioxane	0.447	0.491		9.8	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG No.: Q1173

Instrument ID: BNA_N Calibration Date/Time: 02/06/2025 03:45

Lab File ID: BN036318.D Init. Calib. Date(s): 01/22/2025 01/22/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:02 14:36

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.576		5.9	50.0
Fluoranthene-d10	1.036	1.061		2.4	50.0
2-Fluorophenol	1.040	1.175		13.0	50.0
Phenol-d6	1.222	1.328		8.7	50.0
Nitrobenzene-d5	0.378	0.397		5.0	50.0
2-Fluorobiphenyl	1.786	1.502		-15.9	50.0
2,4,6-Tribromophenol	0.257	0.194		-24.5	50.0
Terphenyl-d14	0.831	0.907		9.1	50.0
1,4-Dioxane	0.447	0.478		6.9	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG No.: Q1173

Instrument ID: BNA_N Calibration Date/Time: 02/06/2025 05:40

Lab File ID: BN036320.D Init. Calib. Date(s): 01/22/2025 01/22/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:02 14:36

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.575		5.7	20.0
Fluoranthene-d10	1.036	1.072		3.5	20.0
2-Fluorophenol	1.040	1.170		12.5	20.0
Phenol-d6	1.222	1.367		11.9	20.0
Nitrobenzene-d5	0.378	0.417		10.3	20.0
2-Fluorobiphenyl	1.786	1.611		-9.8	20.0
2,4,6-Tribromophenol	0.257	0.200		-22.2	20.0
Terphenyl-d14	0.831	0.874		5.2	20.0
1,4-Dioxane	0.447	0.518		15.9	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG No.: Q1173

Instrument ID: BNA_N Calibration Date/Time: 02/06/2025 12:54

Lab File ID: BN036332.D Init. Calib. Date(s): 01/22/2025 01/22/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:02 14:36

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.544	0.570		4.8	50.0
Fluoranthene-d10	1.036	1.077		4.0	50.0
2-Fluorophenol	1.040	1.162		11.7	50.0
Phenol-d6	1.222	1.293		5.8	50.0
Nitrobenzene-d5	0.378	0.409		8.2	50.0
2-Fluorobiphenyl	1.786	1.549		-13.3	50.0
2,4,6-Tribromophenol	0.257	0.188		-26.8	50.0
Terphenyl-d14	0.831	0.868		4.5	50.0
1,4-Dioxane	0.447	0.486		8.7	50.0

All other compounds must meet a minimum RRF of 0.010.