



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
 Fax : 908 789 8922

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH  
 Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG NO.: Q1173  
 EPA Sample No.: SSTDCCC0.4 Date Analyzed: 01/29/2025  
 Lab File ID: BN036112.D Time Analyzed: 18:06  
 Instrument ID: BNA\_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #	
12 HOUR STD	2094	7.803	4943	10.60	2715	14.44	
UPPER LIMIT	4188	8.303	9886	11.1	5430	14.937	
LOWER LIMIT	1047	7.303	2471.5	10.1	1357.5	13.937	
EPA SAMPLE NO.							
01	PB166237BS	2273	7.80	5072	10.60	2643	14.44
02	PB166237BSD	2190	7.80	4796	10.59	2534	14.44
03	BP-VPB-192-GW-60-62	2585	7.80	6318	10.60	3478	14.44
04	BP-VPB-192-GW-100-102	2447	7.80	5665	10.59	3082	14.44
05	BP-VPB-192-GW-140-142	2194	7.80	4945	10.60	2527	14.44
06	BP-VPB-192-GW-200-202	2210	7.80	5163	10.59	3019	14.44
07	VPB192-HYD-20250122	2520	7.80	5602	10.60	2998	14.44

IS1 (DCB) = 1,4-Dichlorobenzene-d4  
 IS2 (NPT) = Naphthalene-d8  
 IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.