

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1173 SDG NO.: Q1173

Lab File ID: BN036319.D

DFTPP Injection Date: 02/06/2025

Instrument ID: BNA_N

DFTPP Injection Time: 05:01

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	50.8
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	47
70	Less than 2.0% of mass 69	0.2 (0.4) 1
127	10.0 - 80.0% of mass 198	48
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	23.8
365	Greater than 1% of mass 198	3.6
441	Present, but less than mass 443	8.7
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	9.5 (18.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036320.D	02/06/2025	05:40
PB166237BL	PB166237BL	BN036329.D	02/06/2025	11:03
SSTDCCC0.4EC	SSTDCCC0.4	BN036332.D	02/06/2025	12:54