

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG No.: Q1173

Instrument ID: MSVOA\_X Calibration Date/Time: 01/27/2025 10:08

Lab File ID: VX044711.D Init. Calib. Date(s): 01/15/2025 01/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:31 13:37

GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND                       | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|--------------------------------|-------|--------|------------|-------|-------|
| Chloromethane                  | 0.728 | 0.776  | 0.1        | 6.59  | 20    |
| Vinyl Chloride                 | 0.719 | 0.750  |            | 4.31  | 20    |
| Bromomethane                   | 0.172 | 0.181  |            | 5.23  | 20    |
| Chloroethane                   | 0.132 | 0.178  |            | 34.85 | 20    |
| Trichlorofluoromethane         | 0.795 | 0.780  |            | -1.89 | 20    |
| 1,1,2-Trichlorotrifluoroethane | 0.597 | 0.612  |            | 2.51  | 20    |
| 1,1-Dichloroethene             | 0.585 | 0.612  |            | 4.61  | 20    |
| Acetone                        | 0.299 | 0.288  |            | -3.68 | 20    |
| Carbon Disulfide               | 1.598 | 1.689  |            | 5.76  | 20    |
| Methyl tert-butyl Ether        | 2.121 | 2.198  |            | 3.63  | 20    |
| Methylene Chloride             | 0.695 | 0.734  |            | 5.61  | 20    |
| trans-1,2-Dichloroethene       | 0.611 | 0.618  |            | 1.15  | 20    |
| 1,1-Dichloroethane             | 1.175 | 1.237  | 0.1        | 5.28  | 20    |
| 2-Butanone                     | 0.444 | 0.470  |            | 5.86  | 20    |
| Carbon Tetrachloride           | 0.497 | 0.453  |            | -8.85 | 20    |
| cis-1,2-Dichloroethene         | 0.769 | 0.785  |            | 2.08  | 20    |
| Chloroform                     | 1.186 | 1.204  |            | 1.52  | 20    |
| 1,1,1-Trichloroethane          | 1.090 | 0.992  |            | -8.99 | 20    |
| Methylcyclohexane              | 0.587 | 0.601  |            | 2.38  | 20    |
| Benzene                        | 1.370 | 1.421  |            | 3.72  | 20    |
| 1,2-Dichloroethane             | 0.500 | 0.487  |            | -2.6  | 20    |
| Trichloroethene                | 0.329 | 0.332  |            | 0.91  | 20    |
| 1,2-Dichloropropane            | 0.345 | 0.367  |            | 6.38  | 20    |
| Bromodichloromethane           | 0.538 | 0.535  |            | -0.56 | 20    |
| 4-Methyl-2-Pentanone           | 0.479 | 0.506  |            | 5.64  | 20    |
| Toluene                        | 0.839 | 0.864  |            | 2.98  | 20    |
| t-1,3-Dichloropropene          | 0.556 | 0.582  |            | 4.68  | 20    |
| cis-1,3-Dichloropropene        | 0.603 | 0.618  |            | 2.49  | 20    |
| 1,1,2-Trichloroethane          | 0.339 | 0.351  |            | 3.54  | 20    |
| 2-Hexanone                     | 0.352 | 0.364  |            | 3.41  | 20    |
| Dibromochloromethane           | 0.398 | 0.411  |            | 3.27  | 20    |
| Tetrachloroethene              | 0.314 | 0.305  |            | -2.87 | 20    |
| Chlorobenzene                  | 1.062 | 1.085  | 0.3        | 2.17  | 20    |
| Ethyl Benzene                  | 1.853 | 1.874  |            | 1.13  | 20    |
| m/p-Xylenes                    | 0.688 | 0.702  |            | 2.04  | 20    |
| o-Xylene                       | 0.695 | 0.695  |            | 0     | 20    |
| Styrene                        | 1.123 | 1.192  |            | 6.14  | 20    |
| Bromoform                      | 0.314 | 0.313  | 0.1        | -0.32 | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG No.: Q1173

Instrument ID: MSVOA\_X Calibration Date/Time: 01/27/2025 10:08

Lab File ID: VX044711.D Init. Calib. Date(s): 01/15/2025 01/15/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:31 13:37

GC Column: DB-624UI ID: 0.18 (mm)

| COMPOUND                  | RRF   | RRF050 | MIN<br>RRF | %D    | MAX%D |
|---------------------------|-------|--------|------------|-------|-------|
| Isopropylbenzene          | 4.085 | 3.903  |            | -4.45 | 20    |
| 1,1,2,2-Tetrachloroethane | 1.411 | 1.377  | 0.3        | -2.41 | 20    |
| 1,3-Dichlorobenzene       | 1.656 | 1.722  |            | 3.99  | 20    |
| 1,4-Dichlorobenzene       | 1.674 | 1.716  |            | 2.51  | 20    |
| 1,2-Dichlorobenzene       | 1.700 | 1.665  |            | -2.06 | 20    |
| 1,2-Dichloroethane-d4     | 0.755 | 0.714  |            | -5.43 | 20    |
| Dibromofluoromethane      | 0.318 | 0.317  |            | -0.31 | 20    |
| Toluene-d8                | 1.108 | 1.118  |            | 0.9   | 20    |
| 4-Bromofluorobenzene      | 0.413 | 0.422  |            | 2.18  | 20    |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.