

Surrogate Summary

SDG No.: Q1173

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1173-01	BP-VPB-192-TB-20250121	1,2-Dichloroethane-d4	50	46.8	94	81	118
		Dibromofluoromethane	50	50.7	101	80	119
		Toluene-d8	50	55.2	110	89	112
		4-Bromofluorobenzene	50	54.2	108	85	114
Q1173-02	BP-VPB-192-GW-60-62	1,2-Dichloroethane-d4	50	49.0	98	81	118
		Dibromofluoromethane	50	51.4	103	80	119
		Toluene-d8	50	55.7	111	89	112
		4-Bromofluorobenzene	50	52.2	104	85	114
Q1173-03	BP-VPB-192-GW-100-102	1,2-Dichloroethane-d4	50	47.9	96	81	118
		Dibromofluoromethane	50	50.8	102	80	119
		Toluene-d8	50	55.2	110	89	112
		4-Bromofluorobenzene	50	52.4	105	85	114
Q1173-04	BP-VPB-192-GW-140-142	1,2-Dichloroethane-d4	50	48.7	97	81	118
		Dibromofluoromethane	50	51.2	102	80	119
		Toluene-d8	50	54.9	110	89	112
		4-Bromofluorobenzene	50	50.9	102	85	114
Q1173-05	BP-VPB-192-GW-200-202	1,2-Dichloroethane-d4	50	47.8	96	81	118
		Dibromofluoromethane	50	50.8	102	80	119
		Toluene-d8	50	54.5	109	89	112
		4-Bromofluorobenzene	50	53.6	107	85	114
Q1173-06	VPB192-HYD-20250122	1,2-Dichloroethane-d4	50	49.6	99	81	118
		Dibromofluoromethane	50	49.8	100	80	119
		Toluene-d8	50	54.4	109	89	112
		4-Bromofluorobenzene	50	53.6	107	85	114
VX0127WBL01	VX0127WBL01	1,2-Dichloroethane-d4	50	47.7	95	81	118
		Dibromofluoromethane	50	50.5	101	80	119
		Toluene-d8	50	54.6	109	89	112
		4-Bromofluorobenzene	50	56.1	112	85	114
VX0127WBS01	VX0127WBS01	1,2-Dichloroethane-d4	50	41.6	83	81	118
		Dibromofluoromethane	50	45.8	92	80	119
		Toluene-d8	50	48.4	97	89	112
		4-Bromofluorobenzene	50	46.8	94	85	114
VX0127WBSD0	VX0127WBSD01	1,2-Dichloroethane-d4	50	42.1	84	81	118
		Dibromofluoromethane	50	45.1	90	80	119
		Toluene-d8	50	48.1	96	89	112
		4-Bromofluorobenzene	50	46.9	94	85	114