

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1173 SAS No.: Q1173 SDG NO.: Q1173
Lab File ID: VX044711.D Date Analyzed: 01/27/2025
Instrument ID: MSVOA_X Time Analyzed: 10:08
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	123095	12.018				
UPPER LIMIT	246190	12.518				
LOWER LIMIT	61547.5	11.518				
EPA SAMPLE NO.						
BP-VPB-192-TB-20250121	115657	12.02				
BP-VPB-192-GW-60-62	111378	12.02				
BP-VPB-192-GW-100-102	115943	12.02				
BP-VPB-192-GW-140-142	114579	12.02				
BP-VPB-192-GW-200-202	112964	12.02				
VPB192-HYD-20250122	114051	12.02				
VX0127WBL01	124517	12.02				
VX0127WBS01	132232	12.02				
VX0127WBSD01	125692	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.