EPA SAMPLE NO.

ME2944S

FORM 5A - IN MATRIX SPIKE SAMPLE RECOVERY

Lab Name:	Chemtech Co	onsulting Group	Contract:	EPW14030			
Lab Code:	ACE	Case No.: 51900	MA No.:		SDG No.: ME2931		
Matrix :	Water		Analytical	Method:	Spectrophotometry		
% Solids:							
Concentrat	ion Units (µo	g/L, mg/L or mg/kg dry v	veight): ug/I	- -			

Analyte	Control Limit %R	Spiked Sample Result (SSR)	Q	Sample Result (SR)	Q	Spike Added (SA)	%R	Q
Cyanide	75 - 125	95		10.0	U	100	95	