EPA SAMPLE NO.

ME2944S

FORM 5A - IN MATRIX SPIKE SAMPLE RECOVERY

Lab Name:	Chemtech C	Consulting Group	Contract:	EPW14030			
Lab Code:	ACE	Case No.: 51900	MA No.:		SDG No.:	ME2931	
Matrix :	Water		Analytical	Method:	CVAA		
% Solids:							
Concentrat	ion Units (µ	ıg/L, mg/L or mg/kg dry v	reight): ug/l	L			

Analyte	Control Limit %R	Spiked Sample Result (SSR)	Q	Sample Result (SR)	Q	Spike Added (SA)	%R	Q
Mercury	75 - 125	1.0		0.2	U	1.0	101	