

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1180 SAS No.: Q1180 SDG NO.: Q1180

Continuing Calib Date: 01/28/2025 Initial Calibration Date(s): 01/16/2025 01/17/2025

Continuing Calib Time: 16:25 Initial Calibration Time(s): 16:46 02:19

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.85	4.85	4.78	4.92	0.00
Aroclor-1016-2	(2)	4.87	4.87	4.80	4.94	0.00
Aroclor-1016-3	(3)	4.93	4.93	4.86	5.00	0.00
Aroclor-1016-4	(4)	5.04	5.04	4.97	5.11	0.00
Aroclor-1016-5	(5)	5.35	5.35	5.28	5.42	0.00
Aroclor-1260-1	(1)	6.56	6.56	6.49	6.63	0.00
Aroclor-1260-2	(2)	6.85	6.85	6.78	6.92	0.00
Aroclor-1260-3	(3)	7.24	7.24	7.17	7.31	0.00
Aroclor-1260-4	(4)	7.48	7.48	7.41	7.55	0.00
Aroclor-1260-5	(5)	7.82	7.82	7.75	7.89	0.00
TCX		3.63	3.63	3.58	3.68	0.00
DCB		9.52	9.52	9.42	9.62	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1180 SAS No.: Q1180 SDG NO.: Q1180

Continuing Calib Date: 01/28/2025 Initial Calibration Date(s): 01/16/2025 01/17/2025

Continuing Calib Time: 16:25 Initial Calibration Time(s): 16:46 02:19

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.08	4.08	4.01	4.15	0.00
Aroclor-1016-2	(2)	4.09	4.10	4.03	4.17	0.01
Aroclor-1016-3	(3)	4.27	4.28	4.21	4.35	0.01
Aroclor-1016-4	(4)	4.32	4.33	4.26	4.40	0.01
Aroclor-1016-5	(5)	4.54	4.55	4.48	4.62	0.01
Aroclor-1260-1	(1)	5.64	5.64	5.57	5.71	0.00
Aroclor-1260-2	(2)	5.85	5.85	5.78	5.92	0.01
Aroclor-1260-3	(3)	6.00	6.01	5.94	6.08	0.01
Aroclor-1260-4	(4)	6.51	6.51	6.44	6.58	0.00
Aroclor-1260-5	(5)	6.78	6.78	6.71	6.85	0.00
TCX		2.96	2.96	2.91	3.01	0.00
DCB		8.23	8.23	8.13	8.33	0.00