

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM012825

Review By	yogesh	Review On	1/29/2025 7:58:19 AM
Supervise By	mohammad	Supervise On	2/5/2025 9:47:47 AM
SubDirectory	BM012825	HP Acquire Method	BNA_M
		HP Processing Method	bm012825

STD. NAME	STD REF.#
Tune/Reschk	SP6633
Initial Calibration Stds	SP6690,SP6691,SP6692,SP6693,SP6694,SP6695
CCC	SP6692
Internal Standard/PEM	S12648,10ul/1000ul sample
ICV/I.BLK	SP6697
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP066	DFTPP066	BM049461.D	28 Jan 2025 09:33		RC/JU	Ok
2	SSTDCCC020	SSTD020079	BM049462.D	28 Jan 2025 10:13	A Fresh Calibration is required.	RC/JU	Not Ok
3	SSTD00572	SSTD005013	BM049463.D	28 Jan 2025 10:52		RC/JU	Ok,M
4	SSTD01073	SSTD010014	BM049464.D	28 Jan 2025 11:31		RC/JU	Ok,M
5	SSTD02074	SSTD020015	BM049465.D	28 Jan 2025 12:10		RC/JU	Ok
6	SSTD04075	SSTD040016	BM049466.D	28 Jan 2025 12:50		RC/JU	Ok,M
7	SSTD08076	SSTD080017	BM049467.D	28 Jan 2025 13:29		RC/JU	Ok,M
8	SSTD16077	SSTD160018	BM049468.D	28 Jan 2025 14:08		RC/JU	Ok,M
9	SSTDICV020	SICV019	BM049469.D	28 Jan 2025 14:47		RC/JU	Ok
10	PB166286BL	SBLK286	BM049470.D	28 Jan 2025 15:27	Analysed for Contamination check	RC/JU	Not Ok
11	SSTDCCC020	SSTD020079	BM049471.D	28 Jan 2025 16:45		RC/JU	Ok,M
12	PB166311BL	SBLK311	BM049472.D	28 Jan 2025 17:24		RC/JU	Ok,M
13	PB166287BL	SBLK287	BM049473.D	28 Jan 2025 18:04		RC/JU	Ok,M
14	Q1188-01	GCQB8	BM049474.D	28 Jan 2025 18:43		RC/JU	Ok
15	Q1188-02	GCQB9	BM049475.D	28 Jan 2025 19:22		RC/JU	Ok,M
16	Q1188-03	GCQC0	BM049476.D	28 Jan 2025 20:01		RC/JU	Ok,M
17	Q1188-04	GCQC1	BM049477.D	28 Jan 2025 20:40		RC/JU	Ok,M

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM012825

Review By	yogesh	Review On	1/29/2025 7:58:19 AM		
Supervise By	mohammad	Supervise On	2/5/2025 9:47:47 AM		
SubDirectory	BM012825	HP Acquire Method	BNA_M	HP Processing Method	bm012825
STD. NAME	STD REF.#				
Tune/Reschk	SP6633				
Initial Calibration Stds	SP6690,SP6691,SP6692,SP6693,SP6694,SP6695				
CCC	SP6692				
Internal Standard/PEM	S12648,10ul/1000ul sample				
ICV/I.BLK	SP6697				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1188-05MS	GCQC1MS	BM049478.D	28 Jan 2025 21:19		RC/JU	Ok
19	Q1188-06MSD	GCQC1MSD	BM049479.D	28 Jan 2025 21:59		RC/JU	Ok,M
20	Q1184-09	E2959	BM049480.D	28 Jan 2025 22:38		RC/JU	Ok,M
21	Q1184-10MS	E2959MS	BM049481.D	28 Jan 2025 23:17		RC/JU	Ok,M
22	Q1184-11MSD	E2959MSD	BM049482.D	28 Jan 2025 23:56		RC/JU	Ok,M
23	PB166311BS	SLCS311	BM049483.D	29 Jan 2025 00:35		RC/JU	Ok,M
24	PB166287BS	SLCS287	BM049484.D	29 Jan 2025 01:15		RC/JU	Ok
25	PB166286BS	SLCS286	BM049485.D	29 Jan 2025 01:54		RC/JU	Ok,M
26	SSTDCCC020	SSTD020080	BM049486.D	29 Jan 2025 03:12		RC/JU	Ok
27	PB166304BL	SBLK304	BM049487.D	29 Jan 2025 03:52		RC/JU	Ok,M
28	PB166314BL	SBLK314	BM049488.D	29 Jan 2025 04:31		RC/JU	Ok
29	Q1184-01	E2948	BM049489.D	29 Jan 2025 05:10		RC/JU	Ok,M
30	Q1184-02	E2949	BM049490.D	29 Jan 2025 05:49		RC/JU	Ok,M
31	Q1184-03	E2955	BM049491.D	29 Jan 2025 06:28		RC/JU	Ok,M
32	Q1184-04	E2956	BM049492.D	29 Jan 2025 07:08		RC/JU	Ok,M
33	Q1184-05	E2957	BM049493.D	29 Jan 2025 07:47		RC/JU	Ok,M
34	Q1184-16	E2958	BM049494.D	29 Jan 2025 08:26		RC/JU	Ok,M
35	Q1184-17	E2971	BM049495.D	29 Jan 2025 09:05		RC/JU	Ok,M
36	Q1139-06MEDL	DCZH3MEDL	BM049496.D	29 Jan 2025 09:45		RC/JU	Ok

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM012825

Review By	yogesh	Review On	1/29/2025 7:58:19 AM		
Supervise By	mohammad	Supervise On	2/5/2025 9:47:47 AM		
SubDirectory	BM012825	HP Acquire Method	BNA_M	HP Processing Method	bm012825
STD. NAME	STD REF.#				
Tune/Reschk	SP6633				
Initial Calibration Stds	SP6690,SP6691,SP6692,SP6693,SP6694,SP6695				
CCC	SP6692				
Internal Standard/PEM	S12648,10ul/1000ul sample				
ICV/I.BLK	SP6697				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

37	Q1139-07MEDL	DCZH4MEDL	BM049497.D	29 Jan 2025 10:24		RC/JU	Ok,M
38	Q1139-12MEDL	DCZH9MEDL	BM049498.D	29 Jan 2025 11:03		RC/JU	Ok,M
39	SSTDCCC020	SSTD020081	BM049499.D	29 Jan 2025 11:42		RC/JU	Ok,M
40	PB166286BL	SBLK286	BM049500.D	29 Jan 2025 12:22		RC/JU	Ok,M
41	Q1188-07	GCQC2	BM049501.D	29 Jan 2025 13:01	PT	RC/JU	Ok,M
42	SSTDCCC020	SSTD020082	BM049502.D	29 Jan 2025 13:40		RC/JU	Ok
43	PB166307BL	SBLK307	BM049503.D	29 Jan 2025 14:20	Analysed for Contamination check	RC/JU	Not Ok
44	Q1184-15	E2966	BM049504.D	29 Jan 2025 14:59	Not Quantifiable and used for screening , Need 20X Dilution-EPA Communication for IS Failure	RC/JU	Not Ok
45	Q1184-14	E2965	BM049505.D	29 Jan 2025 15:39	Not Quantifiable and used for screening , Need 20X Dilution-EPA Communication for IS Failure	RC/JU	Not Ok
46	Q1184-12	E2963	BM049506.D	29 Jan 2025 16:18	Need 10X Dilution	RC/JU	Dilution
47	Q1184-15	E2966	BM049507.D	29 Jan 2025 16:57	Surrogate Fail For EPA Communication, Need Further 10X Dilution	RC/JU	Dilution
48	Q1184-13	E2967	BM049508.D	29 Jan 2025 17:36		RC/JU	Ok,M
49	Q1189-07	A44R0	BM049509.D	29 Jan 2025 18:16		RC/JU	Ok,M
50	Q1189-16	A44R7	BM049510.D	29 Jan 2025 18:55		RC/JU	Ok
51	Q1189-17	A44R8	BM049511.D	29 Jan 2025 19:34		RC/JU	Ok
52	Q1189-18	A44R9	BM049512.D	29 Jan 2025 20:13		RC/JU	Ok

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM012825

Review By	yogesh	Review On	1/29/2025 7:58:19 AM		
Supervise By	mohammad	Supervise On	2/5/2025 9:47:47 AM		
SubDirectory	BM012825	HP Acquire Method	BNA_M	HP Processing Method	bm012825
STD. NAME	STD REF.#				
Tune/Reschk	SP6633				
Initial Calibration Stds	SP6690,SP6691,SP6692,SP6693,SP6694,SP6695				
CCC	SP6692				
Internal Standard/PEM	S12648,10ul/1000ul sample				
ICV/I.BLK	SP6697				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

53	Q1184-07	E2961	BM049513.D	29 Jan 2025 20:53		RC/JU	Ok,M
54	Q1184-08	E2962	BM049514.D	29 Jan 2025 21:32		RC/JU	Ok,M
55	Q1184-06	E2960	BM049515.D	29 Jan 2025 22:11		RC/JU	Ok,M
56	PB166284BS	SLCS284	BM049516.D	29 Jan 2025 22:50		RC/JU	Ok,M
57	PB166307BS	SLCS307	BM049517.D	29 Jan 2025 23:30		RC/JU	Ok,M
58	SSTDCCC020EC	SSTD020083	BM049518.D	30 Jan 2025 00:48		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM013025

Review By	Jagrut	Review On	2/3/2025 3:25:33 PM
Supervise By	mohammad	Supervise On	2/4/2025 7:22:06 AM
SubDirectory	BM013025	HP Acquire Method	BNA_M
		HP Processing Method	bm012825

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	SP6633 SP6690,SP6691,SP6692,SP6693,SP6694,SP6695
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	SP6692 S12648,10ul/1000ul sample SP6697

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP067	DFTPP067	BM049519.D	30 Jan 2025 09:31		RC/JU	Ok
2	SSTDCCC020	SSTD020084	BM049520.D	30 Jan 2025 10:11		RC/JU	Ok,M
3	PB166284BL	SBLK284	BM049521.D	30 Jan 2025 10:50	Analyzed for Contamination check	RC/JU	Not Ok
4	Q1189-06	A44Q9	BM049522.D	30 Jan 2025 11:46		RC/JU	Ok
5	Q1189-21	A44T5	BM049523.D	30 Jan 2025 12:25		RC/JU	Ok
6	Q1189-15	A44R6	BM049524.D	30 Jan 2025 13:04	Internal Standard Fail	RC/JU	ReRun
7	SSTDCCC020	SSTD020085	BM049525.D	30 Jan 2025 13:44		RC/JU	Ok,M
8	PB166284BL	SBLK284	BM049526.D	30 Jan 2025 14:23		RC/JU	Ok
9	Q1189-13	A44R4	BM049527.D	30 Jan 2025 15:02	Run 10X dilution & then decide further dilution	RC/JU	Dilution
10	Q1189-13DL	A44R4DL	BM049528.D	30 Jan 2025 15:41	Need further 3X Dilution	RC/JU	Dilution
11	Q1189-13DL2	A44R4DL2	BM049529.D	30 Jan 2025 16:23		RC/JU	Ok
12	SSTDCCC020	SSTD020086	BM049530.D	30 Jan 2025 17:14		RC/JU	Ok,M
13	PB166307BL	SBLK307	BM049531.D	30 Jan 2025 17:53		RC/JU	Ok
14	Q1180-10	SVOC-GPC2-BLANK	BM049532.D	30 Jan 2025 18:32		RC/JU	Ok
15	Q1189-20	A44S1	BM049533.D	30 Jan 2025 19:11		RC/JU	Ok,M
16	Q1189-02	A44Q5	BM049534.D	30 Jan 2025 19:50	Need 2X Dilution	RC/JU	Dilution
17	Q1189-02DL	A44Q5DL	BM049535.D	30 Jan 2025 20:29		RC/JU	Ok,M

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM013025

Review By	Jagrut	Review On	2/3/2025 3:25:33 PM		
Supervise By	mohammad	Supervise On	2/4/2025 7:22:06 AM		
SubDirectory	BM013025	HP Acquire Method	BNA_M	HP Processing Method	bm012825
STD. NAME	STD REF.#				
Tune/Reschk	SP6633				
Initial Calibration Stds	SP6690,SP6691,SP6692,SP6693,SP6694,SP6695				
CCC	SP6692				
Internal Standard/PEM	S12648,10ul/1000ul sample				
ICV/I.BLK	SP6697				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q1189-15	A44R6	BM049536.D	30 Jan 2025 21:08		RC/JU	Ok,M
19	Q1189-14	A44R5	BM049537.D	30 Jan 2025 21:47		RC/JU	Ok,M
20	Q1189-01	A44Q3	BM049538.D	30 Jan 2025 22:26		RC/JU	Ok,M
21	Q1189-03	A44Q6	BM049539.D	30 Jan 2025 23:06		RC/JU	Ok,M
22	Q1189-04	A44Q7	BM049540.D	30 Jan 2025 23:45		RC/JU	Ok,M
23	Q1184-14	E2965	BM049541.D	31 Jan 2025 00:24	Surrogate Fail For EPA Communication, Need Further 5X Dilution	RC/JU	Dilution
24	Q1184-12DL	E2963DL	BM049542.D	31 Jan 2025 01:03		RC/JU	Ok,M
25	Q1184-15DL	E2966DL	BM049543.D	31 Jan 2025 01:42		RC/JU	Ok,M
26	PB166304BS	SLCS304	BM049544.D	31 Jan 2025 02:21		RC/JU	Ok,M
27	PB166314BS	SLCS314	BM049545.D	31 Jan 2025 03:00		RC/JU	Ok,M
28	SSTDCCC020	SSTD020087	BM049546.D	31 Jan 2025 04:18		RC/JU	Ok,M
29	PB166286BL	SBLK286	BM049547.D	31 Jan 2025 04:57		RC/JU	Ok
30	Q1189-05	A44Q8	BM049548.D	31 Jan 2025 05:37		RC/JU	Ok,M
31	Q1189-11	A44R2	BM049549.D	31 Jan 2025 06:16		RC/JU	Ok,M
32	Q1189-19	A44S0	BM049550.D	31 Jan 2025 06:55		RC/JU	Ok
33	SSTDCCC020	SSTD020088	BM049551.D	31 Jan 2025 10:02		RC/JU	Ok,M
34	PB166295BL	SBLK295	BM049552.D	31 Jan 2025 10:41		RC/JU	Ok,M
35	Q1184-14DL	E2965DL	BM049553.D	31 Jan 2025 11:20	Not matching which first run	RC/JU	Not Ok

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM013025

Review By	Jagrut	Review On	2/3/2025 3:25:33 PM		
Supervise By	mohammad	Supervise On	2/4/2025 7:22:06 AM		
SubDirectory	BM013025	HP Acquire Method	BNA_M	HP Processing Method	bm012825
STD. NAME	STD REF.#				
Tune/Reschk	SP6633				
Initial Calibration Stds	SP6690,SP6691,SP6692,SP6693,SP6694,SP6695				
CCC	SP6692				
Internal Standard/PEM	S12648,10ul/1000ul sample				
ICV/I.BLK	SP6697				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

36	Q1184-14DL	E2965DL	BM049554.D	31 Jan 2025 12:16		RC/JU	Ok,M
37	Q1190-15	A44T4	BM049555.D	31 Jan 2025 13:20		RC/JU	Ok,M
38	Q1190-03	A44S4	BM049556.D	31 Jan 2025 13:59	Need 5X Dilution	RC/JU	Dilution
39	Q1190-02	A44S3	BM049557.D	31 Jan 2025 14:38	EPA Communication for Analysing with 10X. Need 3X Further Dilution	RC/JU	Dilution
40	Q1190-08	A44S7	BM049558.D	31 Jan 2025 15:17	Need straight run	RC/JU	Not Ok
41	Q1190-03DL	A44S4DL	BM049559.D	31 Jan 2025 15:57		RC/JU	Ok,M
42	Q1190-08	A44S7	BM049560.D	31 Jan 2025 16:36		RC/JU	Ok,M
43	Q1190-02DL	A44S3DL	BM049561.D	31 Jan 2025 17:15		RC/JU	Ok,M
44	SSTDCCC020EC	SSTD020089	BM049562.D	31 Jan 2025 17:54		RC/JU	Ok,M

M : Manual Integration