

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1180 SAS No.: Q1180 SDG NO.: Q1180

EPA Sample No.: SSTD020086 Date Analyzed: 01/30/2025

Lab File ID: BM049530.D Time Analyzed: 17:14

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	203355	7.51	783720	10.28	549944	14.16
UPPER LIMIT	406710	8.01	1567440	10.775	1099890	14.657
LOWER LIMIT	101678	7.01	391860	9.775	274972	13.657
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	222879	7.51	830885	10.27	571254	14.15

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.