

Surrogate Summary

SDG No.: Q1180

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1180-01	Storage-Blank-SOIL-REF-6	1,2-Dichloroethane-d4	50	49.7	99	74	125
		Dibromofluoromethane	50	50.9	102	75	124
		Toluene-d8	50	55.2	110	86	113
		4-Bromofluorobenzene	50	51.1	102	77	121
Q1180-02	Storage-Blank-WATER-REF-5	1,2-Dichloroethane-d4	50	49.2	98	74	125
		Dibromofluoromethane	50	50.5	101	75	124
		Toluene-d8	50	55.2	110	86	113
		4-Bromofluorobenzene	50	52.9	106	77	121
Q1180-03	Storage-Blank-WATER-REF-4	1,2-Dichloroethane-d4	50	49.0	98	74	125
		Dibromofluoromethane	50	50.1	100	75	124
		Toluene-d8	50	55.0	110	86	113
		4-Bromofluorobenzene	50	52.3	105	77	121
Q1180-04	Storage-Blank-SAMPLE-MANAGEMENT	1,2-Dichloroethane-d4	50	53.7	107	74	125
		Dibromofluoromethane	50	52.8	106	75	124
		Toluene-d8	50	49.9	100	86	113
		4-Bromofluorobenzene	50	42.9	86	77	121
VN0128WBL01	VN0128WBL01	1,2-Dichloroethane-d4	50	53.7	107	74	125
		Dibromofluoromethane	50	52.2	104	75	124
		Toluene-d8	50	50.0	100	86	113
		4-Bromofluorobenzene	50	44.2	88	77	121
VN0128WBS01	VN0128WBS01	1,2-Dichloroethane-d4	50	51.1	102	74	125
		Dibromofluoromethane	50	53.2	106	75	124
		Toluene-d8	50	54.1	108	86	113
		4-Bromofluorobenzene	50	52.0	104	77	121
VN0128WBSD01	VN0128WBSD01	1,2-Dichloroethane-d4	50	50.3	101	74	125
		Dibromofluoromethane	50	50.2	100	75	124
		Toluene-d8	50	51.1	102	86	113
		4-Bromofluorobenzene	50	51.0	102	77	121
VX0127WBL01	VX0127WBL01	1,2-Dichloroethane-d4	50	47.7	95	74	125
		Dibromofluoromethane	50	50.5	101	75	124
		Toluene-d8	50	54.6	109	86	113
		4-Bromofluorobenzene	50	56.1	112	77	121
VX0127WBS01	VX0127WBS01	1,2-Dichloroethane-d4	50	41.6	83	74	125
		Dibromofluoromethane	50	45.8	92	75	124
		Toluene-d8	50	48.4	97	86	113
		4-Bromofluorobenzene	50	46.8	94	77	121