

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1180 SAS No.: Q1180 SDG NO.: Q1180
 Lab File ID: VN085526.D Date Analyzed: 01/28/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:53
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	203638	8.22	333806	9.09	302551	11.87
UPPER LIMIT	407276	8.718	667612	9.594	605102	12.365
LOWER LIMIT	101819	7.718	166903	8.594	151276	11.365
EPA SAMPLE NO.						
Storage-Blank-SAMPLE-MANAGEMENT	180283	8.22	325340	9.10	283619	11.87
VN0128WBL01	183717	8.22	332601	9.10	291066	11.87
VN0128WBS01	185562	8.22	307569	9.10	271043	11.87
VN0128WBSD01	182154	8.22	309854	9.10	273221	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.