

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1180 SAS No.: Q1180 SDG NO.: Q1180
 Lab File ID: VN085526.D Date Analyzed: 01/28/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:53
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	155713	13.788				
UPPER LIMIT	311426	14.288				
LOWER LIMIT	77856.5	13.288				
EPA SAMPLE NO.						
Storage-Blank-SAMPLE-MANAGEMENT	106780	13.79				
VN0128WBL01	116250	13.79				
VN0128WBS01	126721	13.79				
VN0128WBSD01	127977	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.