

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51974 MA No.: _____ SDG No.: MBHH97
Instrument ID: P4 Start Date: 01/28/2025
Analytical Method: ICP-AES Run Batch: LB134460
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	0.0	0	10	9.1	-9	1250	1200	-4
Barium	0	0.030	0	200	180	-10	2500	2400	-4
Cadmium	0	-0	0	5	5.0	0	625	630	1
Chromium	0	-0	0	10	11	10	3875	4100	6
Lead	0	-0	0	10	10	0	6250	6200	-1
Selenium	0	0.0	0	35	34	-3	1250	1200	-4
Silver	0	0.0	0	10	9.8	-2	312.5	300	-4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51974 MA No.: _____ SDG No.: MBHH97
Instrument ID: P4 Start Date: 01/28/2025
Analytical Method: ICP-AES Run Batch: LB134460
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	2500	2500	0	5000	5000	0	10000	10000	0
Barium	5000	4900	-2	10000	10000	0	20000	20000	0
Cadmium	1250	1300	4	2500	2600	4	5000	4900	-2
Chromium	7750	8100	5	15500	16000	3	31000	30000	-3
Lead	12500	13000	4	25000	26000	4	50000	49000	-2
Selenium	2500	2500	0	5000	5000	0	10000	10000	0
Silver	625	600	-4	1250	1300	4	2500	2500	0

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51974 MA No.: SDG No.: MBHH97
Instrument ID: P4 Start Date: 01/28/2025
Analytical Method: ICP-AES Run Batch: LB134460
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	0.0	0						
Barium	0	0.0	0						
Cadmium	0	0.0	0						
Chromium	0	0.0	0						
Lead	0	0.0	0						
Selenium	0	0.0	0						
Silver	0	0.0	0						