

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51974 MA No. : SDG No.: MBHH97
 Initial Calibration Verification Source : MP84213
 Continuing Calibration Verification Source : MP84216
 Run Batch: LB134460 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV004				ID: CCV025				ID: CCV026		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	1000	1000	102	0	5000	5100	102	0	5000	100	0
Barium	520	560	108	0	10000	11000	108	2	10000	102	1
Cadmium	510	490	96	0	2500	2600	105	0	2600	103	0
Chromium	520	550	105	0	15500	16000	106	0	16000	103	0
Lead	1000	980	98	0	25000	26000	102	0	25000	100	0
Selenium	1000	990	99	1	5000	5200	105	0	5000	100	0
Silver	250	230	92	0	1250	1400	109	0	1300	104	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51974 MA No. : SDG No.: MBHH97
 Initial Calibration Verification Source : MP84213
 Continuing Calibration Verification Source : MP84216
 Run Batch: LB134460 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV027				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					5000	4900	98	0			
Barium					10000	9900	99	1			
Cadmium					2500	2600	104	0			
Chromium					15500	16000	104	0			
Lead					25000	25000	100	0			
Selenium					5000	4900	98	1			
Silver					1250	1300	103	0			