

LABORATORY CONTROL SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51974 MA No.: _____ SDG No.: MBHH97
Matrix : Water Preparation Method: 200.7
Analytical Method: ICP-AES Preparation Batch: PB166272
Concentration Units ($\mu\text{g/L}$, mg/L , mg/kg dry weight, or μg): $\mu\text{g/L}$

Analyte	True	Found	%R
Arsenic	20	20	100
Barium	400	360	91
Cadmium	10	11	107
Chromium	20	23	114
Lead	20	19	97
Selenium	70	68	98
Silver	20	20	99