

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134460

Review By	Kareem Khairalla	Review On	1/29/2025 10:02:19 AM
Supervise By	Sarabjit Jaswal	Supervise On	1/30/2025 7:27:43 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84212,MP84210,MP84209,MP84208,MP84207,MP84206		
ICV Standard	MP84213		
CCV Standard	MP84216		
ICSA Standard	MP84214,MP84215		
CRI Standard			
LCS Standard	MP84063		
Chk Standard	MP84218,MP84219		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	01/28/25 17:57		Kareem	OK
2	S1	S01	CAL2	01/28/25 18:01		Kareem	OK
3	S2	S02	CAL3	01/28/25 18:05		Kareem	OK
4	S3	S03	CAL4	01/28/25 18:09		Kareem	OK
5	S4	S04	CAL5	01/28/25 18:14		Kareem	OK
6	S5	S05	CAL6	01/28/25 18:18		Kareem	OK
7	S6	S06	CAL7	01/28/25 18:22		Kareem	OK
8	ICV004	ICV004	ICV	01/28/25 19:08	Inst. paused for calibration review	Kareem	OK
9	ICB004	ICB004	ICB	01/28/25 19:12		Kareem	OK
10	ICSA004	ICSA004	ICSA	01/28/25 19:29	Inst. paused for review	Kareem	OK
11	ICSAB004	ICSAB004	ICSAB	01/28/25 19:33		Kareem	OK
12	CCV025	CCV025	CCV	01/28/25 19:37		Kareem	OK
13	CCB025	CCB025	CCB	01/28/25 19:41		Kareem	OK
14	Q1136-02	YE8F7	SAM	01/28/25 19:45		Kareem	OK
15	Q1136-03	YE8F7D	DUP	01/28/25 19:49		Kareem	OK
16	Q1136-02L	YE8F7L	SD	01/28/25 19:53		Kareem	OK
17	Q1136-04	YE8F7S	MS	01/28/25 19:57		Kareem	OK
18	Q1151-01	MC0AZ0	SAM	01/28/25 20:01	Cr,Co,Ni high	Kareem	Dilution

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134460

Review By	Kareem Khairalla	Review On	1/29/2025 10:02:19 AM
Supervise By	Sarabjit Jaswal	Supervise On	1/30/2025 7:27:43 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84212,MP84210,MP84209,MP84208,MP84207,MP84206		
ICV Standard	MP84213		
CCV Standard	MP84216		
ICSA Standard	MP84214,MP84215		
CRI Standard			
LCS Standard	MP84063		
Chk Standard	MP84218,MP84219		

19	Q1151-01DL	MC0AZ0	SAM	01/28/25 20:05	10x for Cr,Co,Ni, Still Cr,Co,Ni high	Kareem	Dilution
20	Q1151-02DL	MC0AZ1	SAM	01/28/25 20:10	2x for Co,Ni high, Still Co,Ni high	Kareem	Dilution
21	Q1151-03DL	MC0AZ2	SAM	01/28/25 20:13	2x for Co,Ni	Kareem	Confirms
22	Q1151-05DL	MC0AZ4	SAM	01/28/25 20:17	2x for Co	Kareem	Confirms
23	Q1151-08A	MC0AZ7A	PS	01/28/25 20:21	PS for As,Be,Cu,Mn,Se,Zn(0.0152 ML MP84220,0.035 ML MP84252,0.0048 ML M5715,0.0076 ML M6128,0.007 ML MP84254,0.01 ML M5819)	Kareem	OK
24	Q1163-01DL	ME2946	SAM	01/28/25 20:25	5x for Fe	Kareem	Confirms
25	Q1163-02DL	ME2947	SAM	01/28/25 20:29	2x for Ca	Kareem	Confirms
26	Q1163-03DL	ME2947D	DUP	01/28/25 20:33	2x for Ca	Kareem	Confirms
27	Q1163-02LDL	ME2947L	SD	01/28/25 20:37	10x for Ca	Kareem	Confirms
28	Q1163-04DL	ME2947S	MS	01/28/25 20:41	not required	Kareem	Not Ok
29	PB166236TB	LEB236	MB	01/28/25 20:45		Kareem	OK
30	Q1164-01	ME2946	SAM	01/28/25 20:50		Kareem	OK
31	Q1164-02	ME2947	SAM	01/28/25 20:54		Kareem	OK
32	Q1164-03	ME2947D	DUP	01/28/25 20:58		Kareem	OK
33	Q1164-02L	ME2947L	SD	01/28/25 21:02		Kareem	OK
34	CCV026	CCV026	CCV	01/28/25 21:11		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134460

Review By	Kareem Khairalla	Review On	1/29/2025 10:02:19 AM
Supervise By	Sarabjit Jaswal	Supervise On	1/30/2025 7:27:43 AM
STD. NAME	STD REF.#		
ICAL Standard	MP84204,MP84212,MP84210,MP84209,MP84208,MP84207,MP84206		
ICV Standard	MP84213		
CCV Standard	MP84216		
ICSA Standard	MP84214,MP84215		
CRI Standard			
LCS Standard	MP84063		
Chk Standard	MP84218,MP84219		

35	CCB026	CCB026	CCB	01/28/25 21:15		Kareem	OK
36	Q1164-04	ME2947S	MS	01/28/25 21:19		Kareem	OK
37	Q1182-01	MBHH96	SAM	01/28/25 21:24		Kareem	OK
38	Q1182-02	MBHH97	SAM	01/28/25 21:28		Kareem	OK
39	Q1182-03	MBHH98	SAM	01/28/25 21:32		Kareem	OK
40	Q1182-04	MBHH99	SAM	01/28/25 21:36		Kareem	OK
41	Q1182-05	MBHH99D	DUP	01/28/25 21:40		Kareem	OK
42	Q1182-04L	MBHH99L	SD	01/28/25 21:44		Kareem	OK
43	Q1182-06	MBHH99S	MS	01/28/25 21:49		Kareem	OK
44	PB166272BL	PBW272	MB	01/28/25 21:53		Kareem	OK
45	PB166272BS	LCS272	LCS	01/28/25 21:57		Kareem	OK
46	PB166267TB	LEB267	MB	01/28/25 22:01		Kareem	OK
47	Q1160-01	YE8C9	SAM	01/28/25 22:05		Kareem	OK
48	Q1160-02	YE8C9D	DUP	01/28/25 22:09		Kareem	OK
49	Q1160-01L	YE8C9L	SD	01/28/25 22:13		Kareem	OK
50	Q1160-03	YE8C9S	MS	01/28/25 22:17		Kareem	OK
51	Q1160-04	YE8D0	SAM	01/28/25 22:21		Kareem	OK
52	Q1160-05	YE8D3	SAM	01/28/25 22:25		Kareem	OK
53	PB166267BL	PBW267	MB	01/28/25 22:29		Kareem	OK
54	PB166267BS	LCS267	LCS	01/28/25 22:33		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134460

Review By	Kareem Khairalla	Review On	1/29/2025 10:02:19 AM
Supervise By	Sarabjit Jaswal	Supervise On	1/30/2025 7:27:43 AM

STD. NAME	STD REF.#
ICAL Standard	MP84204,MP84212,MP84210,MP84209,MP84208,MP84207,MP84206
ICV Standard	MP84213
CCV Standard	MP84216
ICSA Standard	MP84214,MP84215
CRI Standard	
LCS Standard	MP84063
Chk Standard	MP84218,MP84219

55	CCV027	CCV027	CCV	01/28/25 22:42		Kareem	OK
56	CCB027	CCB027	CCB	01/28/25 22:46		Kareem	OK