

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51900 MA No. : SDG No.: ME2948

Initial Calibration Verification Source : MP84371

Continuing Calibration Verification Source : MP84373

Run Batch: LB134618 Analytical Method: CVAA

Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV090				ID: CCV037				ID: CCV038		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	4.0	100		5.0	5.1	101		4.9	98	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51900 MA No. : SDG No.: ME2948
Initial Calibration Verification Source : MP84371
Continuing Calibration Verification Source : MP84373
Run Batch: LB134618 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV039				ID: CCV040		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	4.9	99		5.0	99	