

FORM 6 - IN
DUPLICATES

Lab Name: Chemtech Consulting Group Contract: EPW14030
 Lab Code: ACE Case No.: 51900 MA No. : SDG No.: ME2948
 Matrix : Water Analytical Method: ICP-MS
 % Solids:
 Concentration Units ($\mu\text{g/L}$, mg/L or mg/kg dry weight): ug/L

Analyte	Control Limit	Sample (S)	Q	Duplicate (D)	Q	RPD	Q
Aluminum		20.0	U	20.0	U		
Antimony		2.0	U	2.0	U		
Arsenic		0.13	J	0.13	J		
Barium	10.0	16		16		2	
Beryllium		1.0	U	1.0	U		
Cadmium		1.0	U	1.0	U		
Calcium		420000		430000		3	
Chromium	2.0	2.0	U	0.13	J		
Cobalt		0.06	J	0.05	J		
Copper	2.0	0.43	J	3.5		156	*
Iron		1400		1500		9	
Lead		1.0	U	1.0	U		
Magnesium		150000		150000		0	
Manganese		51		51		1	
Nickel		0.24	J	0.27	J		
Potassium		4700		4700		0	
Selenium		5.0	U	5.0	U		
Silver		1.0	U	1.0	U		
Sodium		190000		190000		1	
Thallium		1.0	U	1.0	U		
Vanadium		5.0	U	5.0	U		
Zinc	5.0	5.0	U	0.93	J		