

Instrument ID: TOC

Daily Analysis Runlog For Sequence/QC Batch ID # LB134451

Review By	Niha	Review On	1/31/2025 9:19:01 AM
Supervise By	Iwona	Supervise On	1/31/2025 11:12:44 AM
SubDirectory	LB134451	Test	TOC
STD. NAME	STD REF.#		
ICAL Standard	WP111441,WP111442,WP111443,WP111444,WP111445,WP111446		
ICV Standard	WP111448		
CCV Standard	WP111580		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP111581		
Chk Standard	WP111453,WP111454,WP109953,WP111676,WP111677		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	0.0PPM	0.0PPM	CAL1	01/15/25 12:17		NF IZ	OK
2	0.5PPM	0.5PPM	CAL2	01/15/25 12:42		NF IZ	OK
3	1.0PPM	1.0PPM	CAL3	01/15/25 13:07		NF IZ	OK
4	2.0PPM	2.0PPM	CAL4	01/15/25 13:33		NF IZ	OK
5	5.0PPM	5.0PPM	CAL5	01/15/25 13:59		NF IZ	OK
6	10.0PPM	10.0PPM	CAL6	01/15/25 14:26		NF IZ	OK
7	20.0PPM	20.0PPM	CAL7	01/15/25 14:53		NF IZ	OK
8	ICV1	ICV1	ICV	01/15/25 15:19		NF IZ	OK
9	ICB1	ICB1	ICB	01/15/25 15:43		NF IZ	OK
10	IC-20	IC-20	SAM	01/15/25 16:07		NF IZ	OK
11	IC-R	IC-R	SAM	01/15/25 16:30		NF IZ	OK
12	CCV1	CCV1	CCV	01/28/25 10:32		NF IZ	OK
13	CCB1	CCB1	CCB	01/28/25 10:56		NF IZ	OK
14	LB134451BLW	LB134451BLW	MB	01/28/25 11:20		NF IZ	OK
15	LB134451BSW	LB134451BSW	LCS	01/28/25 11:45		NF IZ	OK
16	Q1168-11	LOQ-WATER-05-QT1	LOQ	01/28/25 12:33		NF IZ	OK
17	Q1168-10	LOD-MDL-WATER-04	SAM	01/28/25 12:57		NF IZ	OK
18	Q1193-01	TAPIAL3-MW04S-012	SAM	01/28/25 13:24		NF IZ	OK

Instrument ID: TOC

Daily Analysis Runlog For Sequence/QC Batch ID # LB134451

Review By	Niha	Review On	1/31/2025 9:19:01 AM
Supervise By	Iwona	Supervise On	1/31/2025 11:12:44 AM
SubDirectory	LB134451	Test	TOC
STD. NAME	STD REF.#		
ICAL Standard	WP111441,WP111442,WP111443,WP111444,WP111445,WP111446		
ICV Standard	WP111448		
CCV Standard	WP111580		
ICSA Standard	N/A		
CRI Standard	N/A		
LCS Standard	WP111581		
Chk Standard	WP111453,WP111454,WP109953,WP111676,WP111677		

19	Q1193-02	TAPIAL2-MW01-0124	SAM	01/28/25 14:14		NF IZ	OK
20	Q1193-02MS	TAPIAL2-MW01-0124	MS	01/28/25 14:51	2.0ml WP111578 +38.0ml Sample	NF IZ	OK
21	Q1193-02MSD	TAPIAL2-MW01-0124	MSD	01/28/25 15:17	2.0ml WP111578 +38.0ml Sample	NF IZ	OK
22	Q1201-01	TAPHHA-MW12-0127	SAM	01/28/25 15:43		NF IZ	OK
23	CCV2	CCV2	CCV	01/28/25 16:09		NF IZ	OK
24	CCB2	CCB2	CCB	01/28/25 16:33		NF IZ	OK