

Matrix Spike Summary

Client:	Weston Solutions	SDG No.:	Q1193
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID:	Q1193-02
Client ID:	TAPIAL2-MW01-012425-00-T2MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/L	75-125	12.0		2.40		10	1	96		01/28/2025

Matrix Spike Summary

Client:	Weston Solutions	SDG No.:	Q1193
Project:	Ft Meade Tipton Airfield Parcel RI - PO 0111169	Sample ID:	Q1193-02
Client ID:	TAPIAL2-MW01-012425-00-T2MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
TOC	mg/L	75-125	12.1		2.40		10	1	97		01/28/2025