

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Weston Solutions</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q1193</u>
Contract:	<u>WEST04</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1193</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q1193-01</u>	Client ID:	<u>TAPIAL3-MW04S-012425-00-T3DUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q1193-01DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1193
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1193 **SAS No.:** Q1193
Matrix: Water **Sample ID:** Q1193-01MS **Client ID:** TAPIAL3-MW04S-012425-00-T3MSD
Percent Solids for Sample: NA **Duplicate ID** Q1193-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	4.52		4.53		0		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1193
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1193 **SAS No.:** Q1193
Matrix: Water **Sample ID:** Q1193-02 **Client ID:** TAPIAL2-MW01-012425-00-T2DUP
Percent Solids for Sample: NA **Duplicate ID** Q1193-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	172		87.6		65	*	P
Antimony	ug/L	20	0.13	J	2.00	U	200.0		P
Arsenic	ug/L	20	3.67		3.54		4		P
Barium	ug/L	20	41.0		40.6		1		P
Beryllium	ug/L	20	1.00	U	1.00	U			P
Cadmium	ug/L	20	1.00	U	1.00	U			P
Calcium	ug/L	20	33800		33900		0		P
Chromium	ug/L	20	1.08	J	0.99	J	9		P
Cobalt	ug/L	20	0.49	J	0.49	J	0		P
Copper	ug/L	20	0.47	J	2.00	U	200.0		P
Iron	ug/L	20	18900		18400		3		P
Lead	ug/L	20	0.31	J	0.25	J	21		P
Magnesium	ug/L	20	2330		2290		2		P
Manganese	ug/L	20	159		155		3		P
Nickel	ug/L	20	0.95	J	0.93	J	2		P
Potassium	ug/L	20	1790		1740		3		P
Selenium	ug/L	20	5.00	U	5.00	U			P
Silver	ug/L	20	1.00	U	1.00	U			P
Sodium	ug/L	20	56200		55400		1		P
Thallium	ug/L	20	1.00	U	1.00	U			P
Vanadium	ug/L	20	1.09	J	0.94	J	15		P
Zinc	ug/L	20	1.02	J	1.36	J	29		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Weston Solutions **Level:** LOW **SDG No.:** Q1193
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1193 **SAS No.:** Q1193
Matrix: Water **Sample ID:** Q1193-02MS **Client ID:** TAPIAL2-MW01-012425-00-T2MSD
Percent Solids for Sample: NA **Duplicate ID** Q1193-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	8610		8470		2		P
Antimony	ug/L	20	513		491		4		P
Arsenic	ug/L	20	3.71		3.61		3		P
Barium	ug/L	20	2320		2220		4		P
Beryllium	ug/L	20	487		472		3		P
Cadmium	ug/L	20	513		483		6		P
Calcium	ug/L	20	79000		75600		4		P
Chromium	ug/L	20	481		466		3		P
Cobalt	ug/L	20	487		478		2		P
Copper	ug/L	20	4750		4660		2		P
Iron	ug/L	20	38700		38700		0		P
Lead	ug/L	20	2490		2380		5		P
Magnesium	ug/L	20	48000		47700		1		P
Manganese	ug/L	20	4840		4800		1		P
Nickel	ug/L	20	459		451		2		P
Potassium	ug/L	20	22800		22200		3		P
Selenium	ug/L	20	493		474		4		P
Silver	ug/L	20	88.7		84.8		4		P
Sodium	ug/L	20	97600		98100		1		P
Thallium	ug/L	20	496		473		5		P
Vanadium	ug/L	20	489		484		1		P
Zinc	ug/L	20	4730		4650		2		P