



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

Portal Partners Tri-Venture

Project Name: Amtrak Sawtooth Bridges 2025

Project # N/A

Chemtech Project # Q1194

Test Name: Metals ICP-TAL,Mercury

A. Number of Samples and Date of Receipt:

4 Solid samples were received on 01/27/2025.

3 Water samples were received on 01/27/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: EPH, EPH, Hexavalent Chromium, Mercury, Metals ICP-TAL, METALS-TAL, PCB, SVOC-TCL BNA -20, Trivalent Chromium, VOC-TCLVOA-10 and VOC-TCLVOA-10. This data package contains results for Metals ICP-TAL,Mercury.

C. Analytical Techniques:

The analysis of Metals ICP-TAL was based on method 6010D, digestion based on method 3050 (soils) and 3010 (waters).The analysis and digestion of Mercury was based on method 7470A. The analysis and digestion of Mercury was based on method 7471B.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

Sample B-110-SB02 was diluted due to high concentrations for Mercury.

The Blank Spike met requirements for all samples.

The Duplicate (B-113-SB02DUP) analysis met criteria for all samples except for Cadmium due to sample matrix interference.

The Matrix Spike (B-113-SB02MS) analysis met criteria for all samples except for Beryllium, Potassium and Selenium due to Chemical Interference during Digestion process. The Matrix Spike (EBMS) analysis met criteria for all samples except for Manganese due to Chemical Interference during Digestion Process.

The Matrix Spike Duplicate (B-113-SB02MSD) analysis met criteria for all samples except for Barium, Beryllium, Chromium, Copper, Potassium, Selenium and Vanadium due to Chemical Interference during Digestion Process. The Matrix Spike Duplicate (EBMSD) analysis met criteria for all samples except for Manganese due to Chemical Interference during Digestion Process.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution (B-113-SB02L) met criteria for all samples except for Calcium, Chromium, Iron, Magnesium and Manganese due to sample matrix interference.



E. Additional Comments:

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____